

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP092923\  
 Data File : BP017443.D  
 Acq On : 29 Sep 2023 13:22  
 Operator : MA/JU  
 Sample : 04497-06  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 BBBP1

Quant Time: Sep 29 22:51:18 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP092023.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Fri Sep 29 22:31:03 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.957  | 152  | 182535   | 20.000 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.793 | 136  | 743649   | 20.000 | ng/ul | -0.01    |
| 38) Acenaphthene-d10          | 14.645 | 164  | 464234   | 20.000 | ng/ul | -0.01    |
| 64) Phenanthrene-d10          | 17.433 | 188  | 1061038  | 20.000 | ng/ul | -0.01    |
| 79) Chrysene-d12              | 21.569 | 240  | 981774   | 20.000 | ng/ul | 0.00     |
| 88) Perylene-d12              | 24.145 | 264  | 1107284  | 20.000 | ng/ul | -0.01    |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8             | 3.299  | 96   | 18387    | 4.137  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.740  | 84   | 129446   | 10.419 | ng/ul | 0.00     |
| 7) Phenol-d5                  | 7.110  | 99   | 92198    | 6.150  | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.293  | 67   | 254681   | 28.150 | ng/ul | -0.01    |
| 11) 2-Chlorophenol-d4         | 7.475  | 132  | 277388   | 23.329 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.681  | 113  | 167298   | 14.354 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 9.163  | 128  | 177560   | 33.027 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.881  | 143  | 188119   | 32.226 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.416 | 165  | 330651   | 29.981 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.963 | 131  | 363236   | 22.736 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 14.063 | 166  | 1081083  | 32.508 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 14.345 | 160  | 1253441  | 32.771 | ng/ul | -0.01    |
| 54) 4-Nitrophenol-d4          | 14.892 | 143  | 26465    | 4.568  | ng/ul | 0.00     |
| 60) Fluorene-d10              | 15.651 | 176  | 956772   | 33.418 | ng/ul | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.792 | 200  | 118485   | 19.813 | ng/ul | -0.01    |
| 73) Anthracene-d10            | 17.539 | 188  | 1540984  | 32.747 | ng/ul | 0.00     |
| 81) Pyrene-d10                | 19.786 | 212  | 1836113  | 34.584 | ng/ul | -0.01    |
| 92) Benzo(a)pyrene-d12        | 23.974 | 264  | 1684140  | 31.590 | ng/ul | -0.01    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP092923\  
Data File : BP017443.D  
Acq On : 29 Sep 2023 13:22  
Operator : MA/JU  
Sample : 04497-06  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
BBBP1

Quant Time: Sep 29 22:51:18 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP092023.MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Fri Sep 29 22:31:03 2023  
Response via : Initial Calibration

