

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093022\
 Data File : BP011844.D
 Acq On : 30 Sep 2022 17:03
 Operator : CG/JU
 Sample : SSTD00556
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Quant Time: Oct 01 00:02:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 23:58:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	257919	20.000	ng/ul	0.00
20) Naphthalene-d8	10.705	136	1173088	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.540	164	802655	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.298	188	1714319	20.000	ng/ul	0.00
79) Chrysene-d12	21.380	240	1792943	20.000	ng/ul	0.00
88) Perylene-d12	23.816	264	1972095	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	10877	1.861	ng/uL	0.00
4) Pyridine-d5	3.740	84	68942	4.243	ng/ul	0.00
7) Phenol-d5	7.058	99	89112	4.175	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	54611	4.283	ng/ul	0.00
11) 2-Chlorophenol-d4	7.422	132	75063	4.605	ng/ul	0.00
15) 4-Methylphenol-d8	8.605	113	73698	4.503	ng/ul	0.00
21) Nitrobenzene-d5	9.058	128	37335	4.883	ng/ul	0.00
24) 2-Nitrophenol-d4	9.787	143	37098	5.721	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	73131	4.405	ng/ul	0.00
31) 4-Chloroaniline-d4	10.846	131	117342	4.630	ng/ul	0.00
46) Dimethylphthalate-d6	13.951	166	259628	4.806	ng/ul	0.00
49) Acenaphthylene-d8	14.234	160	284081	4.403	ng/ul	0.00
54) 4-Nitrophenol-d4	14.740	143	39871	4.192	ng/ul	0.00
60) Fluorene-d10	15.534	176	229417	4.784	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	32908	5.186	ng/ul	0.00
73) Anthracene-d10	17.398	188	356026	4.688	ng/ul	0.00
81) Pyrene-d10	19.628	212	406119	4.433	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.657	264	436017	4.551	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.346	88	12490	2.026	ng/uL#	91
5) Pyridine	3.758	79	66857	4.166	ng/ul	94
6) Benzaldehyde	7.034	77	47502	4.879	ng/ul	95
8) Phenol	7.081	94	95059	4.362	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.316	93	81780	4.755	ng/ul	98
12) 2-Chlorophenol	7.458	128	80816	4.682	ng/ul	100
13) 2-Methylphenol	8.340	108	73282	4.417	ng/ul	94
14) 2,2'-oxybis(1-Chloropr...	8.428	45	81536	3.685	ng/ul	98
16) Acetophenone	8.722	105	125365	4.779	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.705	70	57820	4.733	ng/ul	99
18) 4-Methylphenol	8.669	108	79876	4.496	ng/ul	100
19) Hexachloroethane	8.981	117	31279	4.584	ng/ul	94
22) Nitrobenzene	9.105	77	83807	4.643	ng/ul	99
23) Isophorone	9.634	82	170802	4.670	ng/ul	99
25) 2-Nitrophenol	9.816	139	43400	5.547	ng/ul	94
26) 2,4-Dimethylphenol	9.875	107	89334	4.529	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.116	93	114063	4.813	ng/ul	100
29) 2,4-Dichlorophenol	10.352	162	77032	4.537	ng/ul	98
30) Naphthalene	10.758	128	291405	4.828	ng/ul	99
32) 4-Chloroaniline	10.869	127	121554	4.694	ng/ul	96
33) Hexachlorobutadiene	11.040	225	50230	4.442	ng/ul	98
34) Caprolactam	11.646	113	24874	4.985	ng/ul	97
35) 4-Chloro-3-methylphenol	11.987	107	83076	4.922	ng/ul	99
36) 2-Methylnaphthalene	12.363	142	204633	5.060	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093022\
 Data File : BP011844.D
 Acq On : 30 Sep 2022 17:03
 Operator : CG/JU
 Sample : SST00556
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005617

Quant Time: Oct 01 00:02:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 23:58:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	204176	5.037	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.734	216	102624	4.200	ng/ul	97
40) Hexachlorocyclopentadiene	12.710	237	52194	3.593	ng/ul	97
41) 2,4,6-Trichlorophenol	12.975	196	64095	4.639	ng/ul	97
42) 2,4,5-Trichlorophenol	13.046	196	70791	4.664	ng/ul	99
43) 1,1'-Biphenyl	13.375	154	271254	4.484	ng/ul	100
44) 2-Chloronaphthalene	13.416	162	211548	4.566	ng/ul	99
45) 2-Nitroaniline	13.622	65	43923	4.577	ng/ul	100
47) Dimethylphthalate	13.999	163	277090	5.011	ng/ul	100
48) 2,6-Dinitrotoluene	14.116	165	44587	5.031	ng/ul	90
50) Acenaphthylene	14.263	152	321659	4.367	ng/ul	100
51) 3-Nitroaniline	14.451	138	48744	4.598	ng/ul#	98
52) Acenaphthene	14.604	153	235066	4.880	ng/ul	99
53) 2,4-Dinitrophenol	14.651	184	20838	5.371	ng/ul	98
55) 4-Nitrophenol	14.751	109	26580	3.676	ng/ul	92
56) Dibenzofuran	14.940	168	323540	4.735	ng/ul	99
57) 2,4-Dinitrotoluene	14.904	165	67015	5.431	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.169	232	62413	5.461	ng/ul	99
59) Diethylphthalate	15.363	149	270504	5.023	ng/ul	99
61) Fluorene	15.593	166	271525	4.916	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.587	204	133016	4.752	ng/ul	100
63) 4-Nitroaniline	15.610	138	51048	5.102	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.663	198	36980	5.381	ng/ul#	91
67) N-Nitrosodiphenylamine	15.798	169	231112	4.716	ng/ul	99
68) 4-Bromophenyl-phenylether	16.487	248	78585	4.444	ng/ul	98
69) Hexachlorobenzene	16.598	284	90522	4.315	ng/ul	98
70) Atrazine	16.757	200	78605	4.420	ng/ul	98
71) Pentachlorophenol	16.945	266	47332	4.222	ng/ul	99
72) Phenanthrene	17.340	178	437068	4.880	ng/ul	99
74) Anthracene	17.434	178	431480	4.812	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.340	216	105423	4.003	ng/uL	99
76) Pentachlorobenzene	14.857	250	113925	4.581	ng/uL	100
77) Carbazole	17.704	167	391706	4.957	ng/ul	99
78) Di-n-butylphthalate	18.251	149	463235	5.165	ng/ul	99
80) Fluoranthene	19.304	202	501019	4.553	ng/ul	99
82) Pyrene	19.657	202	535750	4.719	ng/ul	97
83) Butylbenzylphthalate	20.528	149	208498	5.665	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.298	252	180755	5.795	ng/ul	98
85) Benzo(a)anthracene	21.363	228	544097	5.014	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.286	149	342520	5.834	ng/ul	100
87) Chrysene	21.422	228	513103	4.896	ng/ul	98
89) Di-n-octyl phthalate	22.222	149	554505	5.117	ng/ul	100
90) Benzo(b)fluoranthene	23.069	252	551809	4.478	ng/ul	100
91) Benzo(k)fluoranthene	23.122	252	559072	4.740	ng/ul	98
93) Benzo(a)pyrene	23.704	252	499996	4.186	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.351	276	640611	4.647	ng/ul	100
95) Dibenzo(a,h)anthracene	26.374	278	574276	4.832	ng/ul	98
96) Benzo(g,h,i)perylene	27.133	276	564363	4.908	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA P

BNA P

ClientSampleId :

SSTD005617


Quant Time: Oct 01 00:02:28 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA P\Methods\SFAM-EPA-BP093022.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Sep 30 23:58:42 2022

Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.026 ng/uL
RT: 3.346 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 88 Resp: 12490
Ion Ratio Lower Upper
88 100
43 28.9 15.6 23.4#
58 64.8 48.6 72.8



#5
Pyridine
Concen: 4.166 ng/uL
RT: 3.758 min Scan# 131
Delta R.T. 0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion: 79 Resp: 66857
Ion Ratio Lower Upper
79 100
52 66.7 49.7 74.5
51 30.2 21.6 32.4









#6
Benzaldehyde
Concen: 4.879 ng/ul
RT: 7.034 min Scan# 688
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion: 77 Resp: 47502
Ion Ratio Lower Upper
77 100
105 104.5 87.0 130.6
106 99.2 83.8 125.6



#8
Phenol
Concen: 4.362 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion: 94 Resp: 95059
Ion Ratio Lower Upper
94 100
65 23.8 19.5 29.3
66 33.0 25.0 37.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 4.755 ng/ul
 RT: 7.316 min Scan# 736
 Delta R.T. 0.000 min
 Lab File: BP011844.D
 Acq: 30 Sep 2022 17:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617


Tgt Ion: 93 Resp: 81780
 Ion Ratio Lower Upper
 93 100
 63 68.8 56.2 84.4
 95 33.6 25.9 38.9



#12
 2-Chlorophenol
 Concen: 4.682 ng/ul
 RT: 7.458 min Scan# 760
 Delta R.T. 0.000 min
 Lab File: BP011844.D
 Acq: 30 Sep 2022 17:03

Tgt Ion: 128 Resp: 80816
 Ion Ratio Lower Upper
 128 100
 64 38.9 31.1 46.7
 130 31.8 25.5 38.3









#13
2-Methylphenol
Concen: 4.417 ng/ul
RT: 8.340 min Scan# 910
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:108 Resp: 73282
Ion Ratio Lower Upper
108 100
107 93.2 70.2 105.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 3.685 ng/ul
RT: 8.428 min Scan# 925
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion: 45 Resp: 81536
Ion Ratio Lower Upper
45 100
77 21.1 16.0 24.0
79 15.0 12.5 18.7









#16
Acetophenone
Concen: 4.779 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:105 Resp: 125365
Ion Ratio Lower Upper
105 100
77 70.5 55.4 83.0
51 21.0 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 4.733 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion: 70 Resp: 57820
Ion Ratio Lower Upper
70 100
42 41.4 32.7 49.1
101 9.8 8.4 12.6
130 25.1 19.4 29.2









#18
4-Methylphenol
Concen: 4.496 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:108 Resp: 79876
Ion Ratio Lower Upper
108 100
107 110.2 88.0 132.0



#19
Hexachloroethane
Concen: 4.584 ng/ul
RT: 8.981 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:117 Resp: 31279
Ion Ratio Lower Upper
117 100
201 109.4 82.4 123.6
199 69.2 51.6 77.4





#22

Nitrobenzene

Concen: 4.643 ng/ul

RT: 9.105 min Scan# 1104

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion: 77 Resp: 83807

Ion Ratio Lower Upper

77 100

123 51.2 41.6 62.4

65 11.8 9.7 14.5



#23

Isophorone

Concen: 4.670 ng/ul

RT: 9.634 min Scan# 1130

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03




Tgt Ion: 82 Resp: 170802


Ion Ratio Lower Upper

82 100

95 7.0 5.4 8.2

138 19.5 15.8 23.8









#25
2-Nitrophenol
Concen: 5.547 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:	139	Resp:	43400
Ion	Ratio	Lower	Upper
139	100		
65	43.4	32.2	48.4
109	27.8	19.6	29.4



#26
2,4-Dimethylphenol
Concen: 4.529 ng/ul
RT: 9.875 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:	107	Resp:	89334
Ion	Ratio	Lower	Upper
107	100		
121	55.0	43.0	64.6
122	93.0	74.9	112.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.813 ng/ul
 RT: 10.116 min Scan# 1212
 Delta R.T. 0.000 min
 Lab File: BP011844.D
 Acq: 30 Sep 2022 17:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617


Tgt Ion: 93 Resp: 114063
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.3 39.5
 123 11.9 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 4.537 ng/ul
 RT: 10.352 min Scan# 1252
 Delta R.T. 0.000 min
 Lab File: BP011844.D
 Acq: 30 Sep 2022 17:03

Tgt Ion: 162 Resp: 77032
 Ion Ratio Lower Upper
 162 100
 164 67.1 52.2 78.4
 98 34.8 27.0 40.4









#30
Naphthalene
Concen: 4.828 ng/ul
RT: 10.758 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:128 Resp: 291405
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 4.694 ng/ul
RT: 10.869 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:127 Resp: 121554
Ion Ratio Lower Upper
127 100
129 30.7 26.6 39.8









#33
Hexachlorobutadiene
Concen: 4.442 ng/ul
RT: 11.040 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:225 Resp: 50230
Ion Ratio Lower Upper
225 100
223 64.4 52.1 78.1
227 68.3 52.3 78.5



#34
Caprolactam
Concen: 4.985 ng/ul
RT: 11.646 min Scan# 1472
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:113 Resp: 24874
Ion Ratio Lower Upper
113 100
55 137.8 106.0 159.0
56 96.6 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 4.922 ng/ul
RT: 11.987 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:107 Resp: 83076
Ion Ratio Lower Upper
107 100
144 28.4 22.5 33.7
142 88.8 70.0 105.0



#36
2-Methylnaphthalene
Concen: 5.060 ng/ul
RT: 12.363 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:142 Resp: 204633
Ion Ratio Lower Upper
142 100
141 89.5 70.6 106.0









#37
1-Methylnaphthalene
Concen: 5.037 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:142 Resp: 204176
Ion Ratio Lower Upper
142 100
141 90.6 74.8 112.2
116 3.4 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.200 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:216 Resp: 102624
Ion Ratio Lower Upper
216 100
214 75.1 62.6 94.0
179 20.2 16.6 24.8
108 15.7 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 3.593 ng/ul

RT: 12.710 min Scan# 1653

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:237 Resp: 52194

Ion Ratio Lower Upper

237 100

235 60.9 51.0 76.6

272 13.9 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 4.639 ng/ul

RT: 12.975 min Scan# 1698

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03




Tgt Ion:196 Resp: 64095


Ion Ratio Lower Upper

196 100

198 95.0 78.7 118.1

200 31.0 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 4.664 ng/ul
RT: 13.046 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:196 Resp: 70791
Ion Ratio Lower Upper
196 100
198 94.8 76.6 114.8
200 30.5 25.6 38.4



#43
1,1'-Biphenyl
Concen: 4.484 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:154 Resp: 271254
Ion Ratio Lower Upper
154 100
153 40.0 31.8 47.6
76 11.9 9.5 14.3









#44
2-Chloronaphthalene
Concen: 4.566 ng/ul
RT: 13.416 min Scan# 1773
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:162 Resp: 211548
Ion Ratio Lower Upper
162 100
164 33.2 26.5 39.7
127 38.0 29.7 44.5



#45
2-Nitroaniline
Concen: 4.577 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion: 65 Resp: 43923
Ion Ratio Lower Upper
65 100
92 76.2 61.2 91.8
138 128.7 103.0 154.4









#47
Dimethylphthalate
Concen: 5.011 ng/ul
RT: 13.999 min Scan# 1872
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:163 Resp: 277090
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.5 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 5.031 ng/ul
RT: 14.116 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:165 Resp: 44587
Ion Ratio Lower Upper
165 100
89 56.7 39.1 58.7
121 23.7 16.7 25.1









#50
Acenaphthylene
Concen: 4.367 ng/ul
RT: 14.263 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:152 Resp: 321659
Ion Ratio Lower Upper
152 100
151 20.1 16.0 24.0
153 13.3 10.8 16.2



#51
3-Nitroaniline
Concen: 4.598 ng/ul
RT: 14.451 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:138 Resp: 48744
Ion Ratio Lower Upper
138 100
108 12.6 8.2 12.4#
92 105.4 85.6 128.4









#52
Acenaphthene
Concen: 4.880 ng/ul
RT: 14.604 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:153 Resp: 235066
Ion Ratio Lower Upper
153 100
152 47.3 38.4 57.6
154 89.1 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 5.371 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:184 Resp: 20838
Ion Ratio Lower Upper
184 100
63 53.4 44.2 66.4
154 58.0 45.8 68.6









#55
4-Nitrophenol
Concen: 3.676 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:109 Resp: 26580
Ion Ratio Lower Upper
109 100
139 168.1 125.0 187.6
65 135.6 101.9 152.9



#56
Dibenzofuran
Concen: 4.735 ng/ul
RT: 14.940 min Scan# 2032
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:168 Resp: 323540
Ion Ratio Lower Upper
168 100
139 36.2 29.4 44.2









#57
2,4-Dinitrotoluene
Concen: 5.431 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:165 Resp: 67015
Ion Ratio Lower Upper
165 100
63 36.3 28.2 42.2
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 5.461 ng/ul
RT: 15.169 min Scan# 2071
Delta R.T. 0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:232 Resp: 62413
Ion Ratio Lower Upper
232 100
131 44.3 36.2 54.2
130 2.1 1.7 2.5
166 31.5 25.0 37.4









#59
Diethylphthalate
Concen: 5.023 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:149 Resp: 270504
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 12.5 9.9 14.9



#61
Fluorene
Concen: 4.916 ng/ul
RT: 15.593 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:166 Resp: 271525
Ion Ratio Lower Upper
166 100
165 97.2 77.2 115.8
167 12.8 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 4.752 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:204 Resp: 133016
Ion Ratio Lower Upper
204 100
206 33.5 26.8 40.2
141 55.2 44.5 66.7



#63
4-Nitroaniline
Concen: 5.102 ng/ul
RT: 15.610 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:138 Resp: 51048
Ion Ratio Lower Upper
138 100
92 48.1 35.9 53.9
108 65.4 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 5.381 ng/ul
RT: 15.663 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:198 Resp: 36980
Ion Ratio Lower Upper
198 100
182 8.6 3.7 5.5#
77 24.5 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 4.716 ng/ul
RT: 15.798 min Scan# 2178
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:169 Resp: 231112
Ion Ratio Lower Upper
169 100
168 67.7 53.7 80.5
167 36.2 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 4.444 ng/ul
RT: 16.487 min Scan# 2295
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:248 Resp: 78585
Ion Ratio Lower Upper
248 100
250 96.6 78.6 118.0
141 67.4 56.0 84.0



#69
Hexachlorobenzene
Concen: 4.315 ng/ul
RT: 16.598 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:284 Resp: 90522
Ion Ratio Lower Upper
284 100
142 33.3 27.3 40.9
249 30.3 25.2 37.8





#70

Atrazine

Concen: 4.420 ng/ul

RT: 16.757 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:200 Resp: 78605

Ion Ratio Lower Upper

200 100

173 27.1 21.6 32.4

215 46.9 38.9 58.3



#71

Pentachlorophenol

Concen: 4.222 ng/ul

RT: 16.945 min Scan# 2373

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03




Tgt Ion:266 Resp: 47332


Ion Ratio Lower Upper

266 100

264 62.9 49.8 74.8

268 64.0 51.9 77.9





#72

Phenanthrene

Concen: 4.880 ng/ul

RT: 17.340 min Scan# 2440

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:178 Resp: 437068

Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 18.9 15.4 23.2



#74

Anthracene

Concen: 4.812 ng/ul

RT: 17.434 min Scan# 2456

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03




Tgt Ion:178 Resp: 431480


Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 18.8 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.003 ng/uL
RT: 13.340 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:216 Resp: 105423
Ion Ratio Lower Upper
216 100
214 78.1 62.6 93.8
218 48.9 37.4 56.0



#76
Pentachlorobenzene
Concen: 4.581 ng/uL
RT: 14.857 min Scan# 1818
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:250 Resp: 113925
Ion Ratio Lower Upper
250 100
248 64.3 51.4 77.0
252 63.8 51.5 77.3









#77
 Carbazole
 Concen: 4.957 ng/ul
 RT: 17.704 min Scan# 2502
 Delta R.T. 0.000 min
 Lab File: BP011844.D
 Acq: 30 Sep 2022 17:03

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005617

Tgt Ion:167 Resp: 391706




Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.1	25.7
139	12.4	10.1	15.1




#78
 Di-n-butylphthalate
 Concen: 5.165 ng/ul
 RT: 18.251 min Scan# 2595
 Delta R.T. 0.000 min
 Lab File: BP011844.D
 Acq: 30 Sep 2022 17:03

Tgt Ion:149 Resp: 463235

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.5	11.3
104	4.4	3.6	5.4





#80

Fluoranthene

Concen: 4.553 ng/ul

RT: 19.304 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:202 Resp: 501019

Ion Ratio Lower Upper

202 100

101 11.7 9.8 14.6

100 9.1 7.4 11.0



#82

Pyrene

Concen: 4.719 ng/ul

RT: 19.657 min Scan# 2834

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03




Tgt Ion:202 Resp: 535750


Ion Ratio Lower Upper

202 100

101 13.6 11.6 17.4

100 10.4 9.4 14.0





#83

Butylbenzylphthalate

Concen: 5.665 ng/ul

RT: 20.528 min Scan# 2982

Delta R.T. -0.006 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:149 Resp: 208498

Ion Ratio Lower Upper

149 100

91 67.8 50.2 75.4

206 20.3 16.9 25.3



#84

3,3'-Dichlorobenzidine

Concen: 5.795 ng/ul

RT: 21.298 min Scan# 3113

Delta R.T. -0.006 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03




Tgt Ion:252 Resp: 180755


Ion Ratio Lower Upper

252 100

254 62.9 51.4 77.0

126 12.7 8.9 13.3









#85
Benzo(a)anthracene
Concen: 5.014 ng/ul
RT: 21.363 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:228 Resp: 544097
Ion Ratio Lower Upper
228 100
229 19.0 15.8 23.6
226 26.0 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 5.834 ng/ul
RT: 21.286 min Scan# 3111
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:149 Resp: 342520
Ion Ratio Lower Upper
149 100
167 26.3 21.2 31.8
279 3.8 2.9 4.3





#87

Chrysene

Concen: 4.896 ng/ul

RT: 21.422 min Scan# 3134

Delta R.T. 0.000 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Instrument :

BNA_P

ClientSampleId :

SSTD005617





Tgt Ion:228 Resp: 513103

Ion Ratio Lower Upper

228 100

226 29.8 25.3 37.9

229 19.8 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 5.117 ng/ul




RT: 22.222 min Scan# 3270


Delta R.T. -0.006 min

Lab File: BP011844.D

Acq: 30 Sep 2022 17:03

Tgt Ion:149 Resp: 554505









#90
Benzo(b)fluoranthene
Concen: 4.478 ng/ul
RT: 23.069 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:252 Resp: 551809
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 11.6 9.2 13.8



#91
Benzo(k)fluoranthene
Concen: 4.740 ng/ul
RT: 23.122 min Scan# 3423
Delta R.T. 0.000 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:252 Resp: 559072
Ion Ratio Lower Upper
252 100
253 21.6 18.0 27.0
125 11.0 9.2 13.8









#93
Benzo(a)pyrene
Concen: 4.186 ng/ul
RT: 23.704 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03




Instrument :
BNA_P
ClientSampleId :
SSTD005617


Tgt Ion:252 Resp: 499996
Ion Ratio Lower Upper
252 100
253 22.7 17.7 26.5
125 13.2 9.8 14.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.647 ng/ul
RT: 26.351 min Scan# 3972
Delta R.T. -0.012 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:276 Resp: 640611
Ion Ratio Lower Upper
276 100
138 26.8 21.4 32.0
277 24.8 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 4.832 ng/ul
RT: 26.374 min Scan# 3977
Delta R.T. -0.006 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Instrument :
BNA_P
ClientSampleId :
SSTD005617

Tgt Ion:278 Resp: 574276
Ion Ratio Lower Upper
278 100
139 17.4 15.0 22.6
279 23.3 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 4.908 ng/ul
RT: 27.133 min Scan# 4105
Delta R.T. -0.012 min
Lab File: BP011844.D
Acq: 30 Sep 2022 17:03

Tgt Ion:276 Resp: 564363
Ion Ratio Lower Upper
276 100
138 24.4 19.8 29.6
277 23.6 18.6 27.8

