

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093022\
 Data File : BP011849.D
 Acq On : 30 Sep 2022 19:55
 Operator : CG/JU
 Sample : SSTD16061
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622

Quant Time: Oct 01 00:05:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 23:58:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.898	152	235268	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	1094076	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	714000	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.304	188	1578401	20.000	ng/u1	0.00
79) Chrysene-d12	21.398	240	1519344	20.000	ng/u1	0.01
88) Perylene-d12	23.833	264	1992554	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	321686	60.344	ng/uL	0.00
4) Pyridine-d5	3.728	84	2431902	164.077	ng/u1	0.00
7) Phenol-d5	7.069	99	3239874	166.393	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.234	67	1641871	141.160	ng/u1	0.00
11) 2-Chlorophenol-d4	7.434	132	2607976	175.399	ng/u1	0.01
15) 4-Methylphenol-d8	8.622	113	2614004	175.097	ng/u1	0.02
21) Nitrobenzene-d5	9.075	128	1352478	189.679	ng/u1	0.01
24) 2-Nitrophenol-d4	9.792	143	1537160	254.148	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	2637962	170.374	ng/u1	0.01
31) 4-Chloroaniline-d4	10.857	131	3710157	156.952	ng/u1	0.01
46) Dimethylphthalate-d6	13.969	166	7353107	153.000	ng/u1	0.02
49) Acenaphthylene-d8	14.245	160	8140161	141.816	ng/u1	0.01
54) 4-Nitrophenol-d4	14.775	143	1670825	197.503	ng/u1	0.04
60) Fluorene-d10	15.545	176	5994147	140.506	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.680	200	1572550	269.181	ng/u1	0.03
73) Anthracene-d10	17.304	188	1578401	22.576	ng/u1	-0.09
81) Pyrene-d10	19.639	212	10484985	135.053	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.692	264	13810104	142.672	ng/u1	0.03
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	337130	59.945	ng/uL	95
5) Pyridine	3.746	79	2341718	159.958	ng/u1	98
6) Benzaldehyde	7.034	77	852930	96.041	ng/u1	98
8) Phenol	7.098	94	3270457	164.514	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.328	93	2456705	156.586	ng/u1	97
12) 2-Chlorophenol	7.463	128	2727187	173.193	ng/u1	98
13) 2-Methylphenol	8.345	108	2562894	169.348	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.434	45	2343362	116.112	ng/u1	100
16) Acetophenone	8.740	105	3530519	147.545	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.745	70	1578318	141.645	ng/u1	95
18) 4-Methylphenol	8.698	108	2790928	172.211	ng/u1	97
19) Hexachloroethane	8.981	117	979981	157.437	ng/u1	97
22) Nitrobenzene	9.116	77	2696660	160.177	ng/u1	97
23) Isophorone	9.657	82	5402007	158.370	ng/u1	98
25) 2-Nitrophenol	9.828	139	1644392	225.337	ng/u1	97
26) 2,4-Dimethylphenol	9.892	107	2829053	153.791	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.128	93	3298477	149.219	ng/u1	100
29) 2,4-Dichlorophenol	10.363	162	2641764	166.827	ng/u1	99
30) Naphthalene	10.763	128	8071566	143.376	ng/u1	99
32) 4-Chloroaniline	10.881	127	3707261	153.502	ng/u1	99
33) Hexachlorobutadiene	11.045	225	1533767	145.437	ng/u1	99
34) Caprolactam	11.722	113	543811	116.851	ng/u1	100
35) 4-Chloro-3-methylphenol	12.016	107	2761958	175.441	ng/u1	99
36) 2-Methylnaphthalene	12.369	142	5606533	148.636	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093022\
 Data File : BP011849.D
 Acq On : 30 Sep 2022 19:55
 Operator : CG/JU
 Sample : SSTD16061
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622

Quant Time: Oct 01 00:05:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 23:58:42 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.592	142	5576405	147.496	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.739	216	3043180	140.015	ng/ul	98
40) Hexachlorocyclopentadiene	12.716	237	1838539	142.270	ng/ul	99
41) 2,4,6-Trichlorophenol	12.980	196	2237291	182.016	ng/ul	98
42) 2,4,5-Trichlorophenol	13.057	196	2423583	179.519	ng/ul	98
43) 1,1'-Biphenyl	13.386	154	6898958	128.190	ng/ul	97
44) 2-Chloronaphthalene	13.428	162	5776181	140.158	ng/ul	99
45) 2-Nitroaniline	13.639	65	1672780	195.945	ng/ul	94
47) Dimethylphthalate	14.016	163	7432142	151.087	ng/ul	100
48) 2,6-Dinitrotoluene	14.139	165	1782671	226.144	ng/ul	96
50) Acenaphthylene	14.275	152	8449417	128.948	ng/ul	98
51) 3-Nitroaniline	14.469	138	1861447	197.391	ng/ul	94
52) Acenaphthene	14.616	153	5913548	138.006	ng/ul	98
53) 2,4-Dinitrophenol	14.669	184	1319621	382.353	ng/ul	95
55) 4-Nitrophenol	14.792	109	1127614	175.295	ng/ul	94
56) Dibenzofuran	14.951	168	8131633	133.782	ng/ul	96
57) 2,4-Dinitrotoluene	14.927	165	2499602	227.711	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.180	232	2107618	207.317	ng/ul	94
59) Diethylphthalate	15.380	149	7566953	157.966	ng/ul	97
61) Fluorene	15.604	166	5735154	116.732	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.592	204	3089718	124.082	ng/ul	100
63) 4-Nitroaniline	15.651	138	1526655	171.534	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.698	198	1609560	254.369	ng/ul#	92
67) N-Nitrosodiphenylamine	15.810	169	5820622	128.990	ng/ul	99
68) 4-Bromophenyl-phenylether	16.492	248	2384986	146.494	ng/ul	96
69) Hexachlorobenzene	16.604	284	2678727	138.683	ng/ul	98
70) Atrazine	16.780	200	2452763	149.785	ng/ul	99
71) Pentachlorophenol	16.951	266	1934170	187.381	ng/ul	98
72) Phenanthrene	17.351	178	10770672	130.617	ng/ul	97
74) Anthracene	17.445	178	10202114	123.572	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.345	216	3168085	130.657	ng/uL	98
76) Pentachlorobenzene	14.869	250	3234648	141.264	ng/uL	99
77) Carbazole	17.716	167	10332414	142.018	ng/ul	96
78) Di-n-butylphthalate	18.263	149	12029774	145.670	ng/ul	99
80) Fluoranthene	19.315	202	12299249	131.895	ng/ul#	92
82) Pyrene	19.674	202	11660716	121.218	ng/ul#	90
83) Butylbenzylphthalate	20.539	149	6120535	196.239	ng/ul#	91
84) 3,3'-Dichlorobenzidine	21.315	252	4673546	176.823	ng/ul	97
85) Benzo(a)anthracene	21.380	228	12610964	137.133	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.292	149	7494348	150.626	ng/ul#	92
87) Chrysene	21.439	228	11338256	127.678	ng/ul	96
89) Di-n-octyl phthalate	22.233	149	14550757	132.909	ng/ul	100
90) Benzo(b)fluoranthene	23.098	252	16438326	132.038	ng/ul#	96
91) Benzo(k)fluoranthene	23.156	252	14262012	119.671	ng/ul#	96
93) Benzo(a)pyrene	23.751	252	14266861	118.218	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.415	276	20376376	146.293	ng/ul	95
95) Dibenzo(a,h)anthracene	26.456	278	15897990	132.388	ng/ul#	94
96) Benzo(g,h,i)perylene	27.221	276	19423947	167.174	ng/ul	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP093022\
 Data File : BP011849.D
 Acq On : 30 Sep 2022 19:55
 Operator : CG/JU
 Sample : SSTD16061
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622

Quant Time: Oct 01 00:05:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 30 23:58:42 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 59.945 ng/uL
 RT: 3.340 min Scan# 60
 Delta R.T. -0.000 min
 Lab File: BP011849.D
 Acq: 30 Sep 2022 19:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD160622


Tgt Ion: 88 Resp: 337130
 Ion Ratio Lower Upper
 88 100
 43 19.2 15.6 23.4
 58 65.5 48.6 72.8



#5
 Pyridine
 Concen: 159.958 ng/uL
 RT: 3.746 min Scan# 129
 Delta R.T. -0.006 min
 Lab File: BP011849.D
 Acq: 30 Sep 2022 19:55

Tgt Ion: 79 Resp: 2341718
 Ion Ratio Lower Upper
 79 100
 52 61.2 49.7 74.5
 51 25.8 21.6 32.4









#6
Benzaldehyde
Concen: 96.041 ng/ul
RT: 7.034 min Scan# 688
Delta R.T. -0.000 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 77 Resp: 852930
Ion Ratio Lower Upper
77 100
105 107.0 87.0 130.6
106 102.3 83.8 125.6



#8
Phenol
Concen: 164.514 ng/ul
RT: 7.098 min Scan# 699
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion: 94 Resp: 3270457
Ion Ratio Lower Upper
94 100
65 24.0 19.5 29.3
66 31.9 25.0 37.4









#10
Bis(2-Chloroethyl)ether
Concen: 156.586 ng/ul
RT: 7.328 min Scan# 71
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 93 Resp: 2456705
Ion Ratio Lower Upper
93 100
63 67.6 56.2 84.4
95 33.1 25.9 38.9



#12
2-Chlorophenol
Concen: 173.193 ng/ul
RT: 7.463 min Scan# 761
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:128 Resp: 2727187
Ion Ratio Lower Upper
128 100
64 37.2 31.1 46.7
130 32.4 25.5 38.3





#13

2-Methylphenol

Concen: 169.348 ng/ul

RT: 8.345 min Scan# 911

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :





SSTD160622

Tgt Ion:108 Resp: 2562894

Ion Ratio Lower Upper

108 100

107 88.0 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 116.112 ng/ul

RT: 8.434 min Scan# 926

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion: 45 Resp: 2343362


Ion Ratio Lower Upper

45 100

77 20.4 16.0 24.0

79 15.6 12.5 18.7









#16
Acetophenone
Concen: 147.545 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:105 Resp: 3530519
Ion Ratio Lower Upper
105 100
77 68.7 55.4 83.0
51 17.4 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 141.645 ng/ul
RT: 8.745 min Scan# 979
Delta R.T. 0.035 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion: 70 Resp: 1578318
Ion Ratio Lower Upper
70 100
42 37.8 32.7 49.1
101 11.6 8.4 12.6
130 27.2 19.4 29.2









#18
4-Methylphenol
Concen: 172.211 ng/ul
RT: 8.698 min Scan# 91
Delta R.T. 0.029 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:108 Resp: 2790928
Ion Ratio Lower Upper
108 100
107 107.0 88.0 132.0



#19
Hexachloroethane
Concen: 157.437 ng/ul
RT: 8.981 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:117 Resp: 979981
Ion Ratio Lower Upper
117 100
201 106.6 82.4 123.6
199 67.1 51.6 77.4









#22
Nitrobenzene
Concen: 160.177 ng/ul
RT: 9.116 min Scan# 1040
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion: 77 Resp: 2696660
Ion Ratio Lower Upper
77 100
123 54.0 41.6 62.4
65 12.6 9.7 14.5



#23
Isophorone
Concen: 158.370 ng/ul
RT: 9.657 min Scan# 1134
Delta R.T. 0.023 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion: 82 Resp: 5402007
Ion Ratio Lower Upper
82 100
95 7.3 5.4 8.2
138 20.9 15.8 23.8









#25
2-Nitrophenol
Concen: 225.337 ng/ul
RT: 9.828 min Scan# 1163
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:139 Resp: 1644392
Ion Ratio Lower Upper
139 100
65 38.6 32.2 48.4
109 25.8 19.6 29.4



#26
2,4-Dimethylphenol
Concen: 153.791 ng/ul
RT: 9.892 min Scan# 1174
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:107 Resp: 2829053
Ion Ratio Lower Upper
107 100
121 52.3 43.0 64.6
122 90.8 74.9 112.3





#27

Bis(2-Chloroethoxy)methane

Concen: 149.219 ng/ul

RT: 10.128 min Scan# 11

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion: 93 Resp: 3298477

Ion Ratio Lower Upper

93 100

95 33.0 26.3 39.5

123 12.2 9.5 14.3



#29

2,4-Dichlorophenol

Concen: 166.827 ng/ul

RT: 10.363 min Scan# 1254

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:162 Resp: 2641764


Ion Ratio Lower Upper

162 100

164 65.6 52.2 78.4

98 32.0 27.0 40.4









#30
Naphthalene
Concen: 143.376 ng/ul
RT: 10.763 min Scan# 1321
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:128 Resp: 8071566
Ion Ratio Lower Upper
128 100
129 11.5 8.9 13.3
127 13.7 10.5 15.7



#32
4-Chloroaniline
Concen: 153.502 ng/ul
RT: 10.881 min Scan# 1342
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:127 Resp: 3707261
Ion Ratio Lower Upper
127 100
129 32.8 26.6 39.8





#33

Hexachlorobutadiene

Concen: 145.437 ng/ul

RT: 11.045 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:225 Resp: 1533767

Ion Ratio Lower Upper

225 100

223 63.5 52.1 78.1

227 65.0 52.3 78.5



#34

Caprolactam

Concen: 116.851 ng/ul

RT: 11.722 min Scan# 1485

Delta R.T. 0.070 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:113 Resp: 543811


Ion Ratio Lower Upper

113 100

55 133.0 106.0 159.0

56 98.5 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 175.441 ng/ul
RT: 12.016 min Scan# 1535
Delta R.T. 0.029 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:107 Resp: 2761958
Ion Ratio Lower Upper
107 100
144 29.4 22.5 33.7
142 88.2 70.0 105.0



#36
2-Methylnaphthalene
Concen: 148.636 ng/ul
RT: 12.369 min Scan# 1595
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:142 Resp: 5606533
Ion Ratio Lower Upper
142 100
141 89.1 70.6 106.0





#37

1-Methylnaphthalene

Concen: 147.496 ng/ul

RT: 12.592 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:142 Resp: 5576405

Ion Ratio Lower Upper

142 100

141 92.2 74.8 112.2

116 3.7 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 140.015 ng/ul

RT: 12.739 min Scan# 1658

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Tgt Ion:216 Resp: 3043180




Ion Ratio Lower Upper


216 100

214 79.7 62.6 94.0

179 21.2 16.6 24.8

108 14.8 12.6 19.0









#40
Hexachlorocyclopentadiene
Concen: 142.270 ng/ul
RT: 12.716 min Scan# 1653
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:237 Resp: 1838539
Ion Ratio Lower Upper
237 100
235 62.8 51.0 76.6
272 13.1 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 182.016 ng/ul
RT: 12.980 min Scan# 1699
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:196 Resp: 2237291
Ion Ratio Lower Upper
196 100
198 95.8 78.7 118.1
200 31.1 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 179.519 ng/ul
RT: 13.057 min Scan# 1710
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:196 Resp: 2423583
Ion Ratio Lower Upper
196 100
198 93.9 76.6 114.8
200 31.2 25.6 38.4



#43
1,1'-Biphenyl
Concen: 128.190 ng/ul
RT: 13.386 min Scan# 1768
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:154 Resp: 6898958
Ion Ratio Lower Upper
154 100
153 41.4 31.8 47.6
76 10.4 9.5 14.3





#44

2-Chloronaphthalene

Concen: 140.158 ng/ul

RT: 13.428 min Scan# 11

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:162 Resp: 5776181

Ion Ratio Lower Upper

162 100

164 33.5 26.5 39.7

127 38.2 29.7 44.5



#45

2-Nitroaniline

Concen: 195.945 ng/ul

RT: 13.639 min Scan# 1811

Delta R.T. 0.017 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion: 65 Resp: 1672780


Ion Ratio Lower Upper

65 100

92 78.6 61.2 91.8

138 138.1 103.0 154.4









#47
Dimethylphthalate
Concen: 151.087 ng/ul
RT: 14.016 min Scan# 1875
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:163 Resp: 7432142
Ion Ratio Lower Upper
163 100
194 4.4 3.3 4.9
164 10.7 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 226.144 ng/ul
RT: 14.139 min Scan# 1896
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:165 Resp: 1782671
Ion Ratio Lower Upper
165 100
89 46.3 39.1 58.7
121 19.5 16.7 25.1





#50

Acenaphthylene

Concen: 128.948 ng/ul

RT: 14.275 min Scan# 1919

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:152 Resp: 8449417

Ion Ratio Lower Upper

152 100

151 21.4 16.0 24.0

153 14.0 10.8 16.2



#51

3-Nitroaniline

Concen: 197.391 ng/ul

RT: 14.469 min Scan# 1952

Delta R.T. 0.017 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:138 Resp: 1861447


Ion Ratio Lower Upper

138 100

108 9.2 8.2 12.4

92 100.2 85.6 128.4





#52

Acenaphthene

Concen: 138.006 ng/ul

RT: 14.616 min Scan# 1977

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:153 Resp: 5913548

Ion Ratio Lower Upper

153 100

152 49.0 38.4 57.6

154 89.6 70.6 105.8



#53

2,4-Dinitrophenol

Concen: 382.353 ng/ul

RT: 14.669 min Scan# 1986

Delta R.T. 0.017 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:184 Resp: 1319621


Ion Ratio Lower Upper

184 100

63 48.5 44.2 66.4

154 57.8 45.8 68.6









#55
4-Nitrophenol
Concen: 175.295 ng/ul
RT: 14.792 min Scan# 2007
Delta R.T. 0.041 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:109 Resp: 1127614
Ion Ratio Lower Upper
109 100
139 154.9 125.0 187.6
65 112.7 101.9 152.9



#56
Dibenzofuran
Concen: 133.782 ng/ul
RT: 14.951 min Scan# 2034
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:168 Resp: 8131633
Ion Ratio Lower Upper
168 100
139 39.0 29.4 44.2









#57
2,4-Dinitrotoluene
Concen: 227.711 ng/ul
RT: 14.927 min Scan# 2026
Delta R.T. 0.023 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:165 Resp: 2499602
Ion Ratio Lower Upper
165 100
63 31.0 28.2 42.2
182 2.9 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 207.317 ng/ul
RT: 15.180 min Scan# 2073
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:232 Resp: 2107618
Ion Ratio Lower Upper
232 100
131 39.9 36.2 54.2
130 2.0 1.7 2.5
166 29.7 25.0 37.4









#59
Diethylphthalate
Concen: 157.966 ng/ul
RT: 15.380 min Scan# 2107
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:149 Resp: 7566953
Ion Ratio Lower Upper
149 100
177 23.9 17.8 26.8
150 13.2 9.9 14.9



#61
Fluorene
Concen: 116.732 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:166 Resp: 5735154
Ion Ratio Lower Upper
166 100
165 98.0 77.2 115.8
167 13.9 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 124.082 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:204 Resp: 3089718
Ion Ratio Lower Upper
204 100
206 33.3 26.8 40.2
141 56.0 44.5 66.7



#63
4-Nitroaniline
Concen: 171.534 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. 0.035 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:138 Resp: 1526655
Ion Ratio Lower Upper
138 100
92 44.7 35.9 53.9
108 69.0 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 254.369 ng/ul
RT: 15.698 min Scan# 2156
Delta R.T. 0.029 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:198 Resp: 1609560
Ion Ratio Lower Upper
198 100
182 3.1 3.7 5.5#
77 16.9 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 128.990 ng/ul
RT: 15.810 min Scan# 2180
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:169 Resp: 5820622
Ion Ratio Lower Upper
169 100
168 67.7 53.7 80.5
167 37.7 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 146.494 ng/ul
RT: 16.492 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:248 Resp: 2384986
Ion Ratio Lower Upper
248 100
250 95.8 78.6 118.0
141 65.6 56.0 84.0



#69
Hexachlorobenzene
Concen: 138.683 ng/ul
RT: 16.604 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:284 Resp: 2678727
Ion Ratio Lower Upper
284 100
142 33.4 27.3 40.9
249 33.0 25.2 37.8









#70
Atrazine
Concen: 149.785 ng/ul
RT: 16.780 min Scan# 21
Delta R.T. 0.023 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:200 Resp: 2452763
Ion Ratio Lower Upper
200 100
173 26.1 21.6 32.4
215 49.4 38.9 58.3



#71
Pentachlorophenol
Concen: 187.381 ng/ul
RT: 16.951 min Scan# 2374
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:266 Resp: 1934170
Ion Ratio Lower Upper
266 100
264 64.2 49.8 74.8
268 63.3 51.9 77.9





#72

Phenanthrene

Concen: 130.617 ng/ul

RT: 17.351 min Scan# 2440

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:178 Resp:10770672

Ion Ratio Lower Upper

178 100

179 16.5 12.5 18.7

176 21.4 15.4 23.2



#74

Anthracene

Concen: 123.572 ng/ul

RT: 17.445 min Scan# 2458

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:178 Resp:10202114


Ion Ratio Lower Upper

178 100

179 16.9 12.3 18.5

176 21.0 15.4 23.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 130.657 ng/uL

RT: 13.345 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:216 Resp: 3168085

Ion Ratio Lower Upper

216 100

214 79.5 62.6 93.8

218 49.0 37.4 56.0



#76

Pentachlorobenzene

Concen: 141.264 ng/uL

RT: 14.869 min Scan# 2020

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:250 Resp: 3234648


Ion Ratio Lower Upper

250 100

248 65.1 51.4 77.0

252 64.1 51.5 77.3





#77

Carbazole

Concen: 142.018 ng/ul

RT: 17.716 min Scan# 2504

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:167 Resp:10332414

Ion Ratio Lower Upper

167 100

166 23.1 17.1 25.7

139 14.3 10.1 15.1



#78

Di-n-butylphthalate

Concen: 145.670 ng/ul

RT: 18.263 min Scan# 2597

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:149 Resp:12029774


Ion Ratio Lower Upper

149 100

150 10.0 7.5 11.3

104 4.8 3.6 5.4





#80

Fluoranthene

Concen: 131.895 ng/ul

RT: 19.315 min Scan# 21

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:202 Resp:12299249

Ion Ratio Lower Upper

202 100

101 8.8 9.8 14.6#

100 6.6 7.4 11.0#



#82

Pyrene

Concen: 121.218 ng/ul

RT: 19.674 min Scan# 2837

Delta R.T. 0.017 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:202 Resp:11660716


Ion Ratio Lower Upper

202 100

101 10.1 11.6 17.4#

100 8.0 9.4 14.0#





#83

Butylbenzylphthalate

Concen: 196.239 ng/ul

RT: 20.539 min Scan# 2983

Delta R.T. 0.006 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:149 Resp: 6120535

Ion Ratio Lower Upper

149 100

91 55.4 50.2 75.4

206 25.3 16.9 25.3#



#84

3,3'-Dichlorobenzidine

Concen: 176.823 ng/ul

RT: 21.315 min Scan# 3116

Delta R.T. 0.011 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:252 Resp: 4673546


Ion Ratio Lower Upper

252 100

254 65.9 51.4 77.0

126 9.4 8.9 13.3









#85
Benzo(a)anthracene
Concen: 137.133 ng/ul
RT: 21.380 min Scan# 3125
Delta R.T. 0.011 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:228 Resp:12610964
Ion Ratio Lower Upper
228 100
229 21.6 15.8 23.6
226 31.7 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 150.626 ng/ul
RT: 21.292 min Scan# 3112
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:149 Resp: 7494348
Ion Ratio Lower Upper
149 100
167 30.9 21.2 31.8
279 6.3 2.9 4.3









#87
Chrysene
Concen: 127.678 ng/ul
RT: 21.439 min Scan# 3134
Delta R.T. 0.017 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:228 Resp:11338256
Ion Ratio Lower Upper
228 100
226 34.3 25.3 37.9
229 21.4 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 132.909 ng/ul
RT: 22.233 min Scan# 3272
Delta R.T. 0.006 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:149 Resp:14550757









#90
Benzo(b)fluoranthene
Concen: 132.038 ng/ul
RT: 23.098 min Scan# 3415
Delta R.T. 0.023 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55




Instrument :
BNA_P
ClientSampleId :
SSTD160622


Tgt Ion:252 Resp:16438326
Ion Ratio Lower Upper
252 100
253 23.5 17.8 26.8
125 8.9 9.2 13.8#



#91
Benzo(k)fluoranthene
Concen: 119.671 ng/ul
RT: 23.156 min Scan# 3429
Delta R.T. 0.035 min
Lab File: BP011849.D
Acq: 30 Sep 2022 19:55

Tgt Ion:252 Resp:14262012
Ion Ratio Lower Upper
252 100
253 23.4 18.0 27.0
125 8.1 9.2 13.8#





#93

Benzo(a)pyrene

Concen: 118.218 ng/ul

RT: 23.751 min Scan# 3

Delta R.T. 0.035 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P

ClientSampleId :

SSTD160622





Tgt Ion:252 Resp:14266861

Ion Ratio Lower Upper

252 100

253 23.4 17.7 26.5

125 9.4 9.8 14.8#



#94

Indeno(1,2,3-cd)pyrene

Concen: 146.293 ng/ul

RT: 26.415 min Scan# 3983

Delta R.T. 0.053 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55




Tgt Ion:276 Resp:20376376


Ion Ratio Lower Upper

276 100

138 21.9 21.4 32.0

277 25.5 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 132.388 ng/ul

RT: 26.456 min Scan# 3977

Delta R.T. 0.076 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Instrument :

BNA_P





ClientSampleId :

SSTD160622

Tgt Ion:278 Resp:15897990

Ion Ratio Lower Upper

Ion	Ratio	Lower	Upper
278	100		
139	14.6	15.0	22.6#
279	25.0	19.0	28.6



#96

Benzo(g,h,i)perylene

Concen: 167.174 ng/ul

RT: 27.221 min Scan# 4120

Delta R.T. 0.076 min

Lab File: BP011849.D

Acq: 30 Sep 2022 19:55

Tgt Ion:276 Resp:19423947

Ion Ratio Lower Upper

Ion	Ratio	Lower	Upper
276	100		
138	21.8	19.8	29.6
277	24.9	18.6	27.8

