

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100121\
 Data File : BP007285.D
 Acq On : 01 Oct 2021 11:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 01 12:52:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:49:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	76	0.00
2	1,4-Dioxane	40.000	41.809	-4.5	83	0.00
3	Pyridine	40.000	40.995	-2.5	79	0.00
4	n-Nitrosodimethylamine	40.000	44.690	-11.7	86	0.00
5 S	2-Fluorophenol	80.000	80.291	-0.4	79	0.00
6	Aniline	40.000	38.360	4.1	73	0.00
7 S	Phenol-d6	80.000	77.736	2.8	74	0.00
8	2-Chlorophenol	40.000	39.059	2.4	76	0.00
9	Benzaldehyde	40.000	36.548	8.6	69	0.00
10 C	Phenol	40.000	38.346	4.1	74	0.00
11	bis(2-Chloroethyl)ether	40.000	37.689	5.8	73	0.00
12	1,3-Dichlorobenzene	40.000	38.616	3.5	76	0.00
13 C	1,4-Dichlorobenzene	40.000	38.473	3.8	76	0.00
14	1,2-Dichlorobenzene	40.000	38.248	4.4	75	0.00
15	Benzyl Alcohol	40.000	39.321	1.7	75	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.390	6.5	73	0.00
17	2-Methylphenol	40.000	38.008	5.0	73	0.00
18	Hexachloroethane	40.000	38.949	2.6	76	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.653	8.4	70	0.00
20	3+4-Methylphenols	40.000	37.840	5.4	72	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	71	0.00
22	Acetophenone	40.000	39.964	0.1	73	0.00
23 S	Nitrobenzene-d5	80.000	82.011	-2.5	74	0.00
24	Nitrobenzene	40.000	40.880	-2.2	73	0.00
25	Isophorone	40.000	39.639	0.9	71	0.00
26 C	2-Nitrophenol	40.000	42.103	-5.3	74	0.00
27	2,4-Dimethylphenol	40.000	39.585	1.0	71	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.128	4.7	69	0.00
29 C	2,4-Dichlorophenol	40.000	39.711	0.7	71	0.00
30	1,2,4-Trichlorobenzene	40.000	39.182	2.0	72	0.00
31	Naphthalene	40.000	39.092	2.3	71	0.00
32	Benzoic acid	40.000	41.185	-3.0	70	0.00
33	4-Chloroaniline	40.000	39.145	2.1	70	0.00
34 C	Hexachlorobutadiene	40.000	39.956	0.1	73	0.00
35	Caprolactam	40.000	40.588	-1.5	70	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.018	2.5	70	0.00
37	2-Methylnaphthalene	40.000	38.228	4.4	70	0.00
38	1-Methylnaphthalene	40.000	37.813	5.5	69	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	68	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.803	0.5	69	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.131	-2.8	70	0.00
42 S	2,4,6-Tribromophenol	80.000	78.950	1.3	68	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.336	-0.8	69	0.00
44	2,4,5-Trichlorophenol	40.000	39.939	0.2	68	0.00
45 S	2-Fluorobiphenyl	80.000	79.463	0.7	70	0.00
46	1,1'-Biphenyl	40.000	39.202	2.0	69	0.00
47	2-Chloronaphthalene	40.000	39.374	1.6	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100121\
 Data File : BP007285.D
 Acq On : 01 Oct 2021 11:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 01 12:52:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 12:49:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.016	-10.0	74	0.00
49	Acenaphthylene	40.000	39.713	0.7	68	0.00
50	Dimethylphthalate	40.000	39.060	2.3	68	0.00
51	2,6-Dinitrotoluene	40.000	40.381	-1.0	68	0.00
52 C	Acenaphthene	40.000	38.875	2.8	68	0.00
53	3-Nitroaniline	40.000	41.442	-3.6	69	0.00
54 P	2,4-Dinitrophenol	40.000	38.603	3.5	62	0.00
55	Dibenzofuran	40.000	38.931	2.7	68	0.00
56 P	4-Nitrophenol	40.000	39.000	2.5	63	0.00
57	2,4-Dinitrotoluene	40.000	41.492	-3.7	69	0.00
58	Fluorene	40.000	39.448	1.4	69	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.572	3.6	67	0.00
60	Diethylphthalate	40.000	39.387	1.5	69	0.00
61	4-Chlorophenyl-phenylether	40.000	38.906	2.7	68	0.00
62	4-Nitroaniline	40.000	42.582	-6.5	70	0.00
63	Azobenzene	40.000	40.626	-1.6	70	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	67	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.864	-2.2	65	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.807	0.5	68	0.00
67	4-Bromophenyl-phenylether	40.000	39.065	2.3	67	0.00
68	Hexachlorobenzene	40.000	39.032	2.4	68	0.00
69	Atrazine	40.000	39.279	1.8	66	0.00
70 C	Pentachlorophenol	40.000	34.270	14.3	57	0.00
71	Phenanthrene	40.000	39.477	1.3	68	0.00
72	Anthracene	40.000	39.800	0.5	68	0.00
73	Carbazole	40.000	40.213	-0.5	68	0.00
74	Di-n-butylphthalate	40.000	40.298	-0.7	68	0.00
75 C	Fluoranthene	40.000	39.816	0.5	68	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	69	0.00
77	Benzydine	40.000	34.857	12.9	60	0.00
78	Pyrene	40.000	38.717	3.2	68	0.00
79 S	Terphenyl-d14	80.000	78.523	1.8	69	0.00
80	Butylbenzylphthalate	40.000	40.439	-1.1	69	0.00
81	Benzo(a)anthracene	40.000	39.188	2.0	69	0.00
82	3,3'-Dichlorobenzidine	40.000	40.607	-1.5	70	0.00
83	Chrysene	40.000	38.987	2.5	68	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.968	0.1	68	0.00
85 c	Di-n-octyl phthalate	40.000	41.030	-2.6	70	0.00
86 I	Perylene-d12	20.000	20.000	0.0	70	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.153	-0.4	72	0.00
88	Benzo(b)fluoranthene	40.000	39.021	2.4	69	0.00
89	Benzo(k)fluoranthene	40.000	39.012	2.5	70	0.00
90 C	Benzo(a)pyrene	40.000	39.678	0.8	71	0.00
91	Dibenzo(a,h)anthracene	40.000	39.741	0.6	71	0.00
92	Benzo(g,h,i)perylene	40.000	39.660	0.9	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100121\
Data File : BP007285.D
Acq On : 01 Oct 2021 11:36
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Oct 01 12:52:08 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 12:49:05 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0