

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100122\
 Data File : BP011865.D
 Acq On : 01 Oct 2022 17:01
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020761

Quant Time: Oct 03 01:01:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	343700	20.000	ng/u1	0.00
20) Naphthalene-d8	10.704	136	1561041	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.539	164	1007076	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	2117694	20.000	ng/u1	0.00
79) Chrysene-d12	21.386	240	2064100	20.000	ng/u1	0.00
88) Perylene-d12	23.821	264	1133588	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	66768	8.568	ng/uL	0.00
4) Pyridine-d5	3.728	84	487518	21.357	ng/u1	0.00
7) Phenol-d5	7.057	99	640705	21.593	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	355242	21.599	ng/u1	0.00
11) 2-Chlorophenol-d4	7.422	132	522972	22.327	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	530352	21.622	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	272286	22.900	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	301292	23.684	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	548137	23.295	ng/u1	0.00
31) 4-Chloroaniline-d4	10.845	131	794829	22.157	ng/u1	0.00
46) Dimethylphthalate-d6	13.951	166	1573960	21.856	ng/u1	0.00
49) Acenaphthylene-d8	14.233	160	1868414	22.758	ng/u1	0.00
54) 4-Nitrophenol-d4	14.745	143	322678	21.637	ng/u1	0.00
60) Fluorene-d10	15.533	176	1376087	22.003	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.657	200	275646	21.505	ng/u1	0.00
73) Anthracene-d10	17.398	188	2151799	22.306	ng/u1	0.00
81) Pyrene-d10	19.633	212	2438303	23.325	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.662	264	1265823	22.307	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	70384	8.362	ng/uL	93
5) Pyridine	3.752	79	474273	21.407	ng/u1	96
6) Benzaldehyde	7.034	77	323872	24.918	ng/u1	99
8) Phenol	7.087	94	659442	21.428	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.316	93	530133	21.649	ng/u1	97
12) 2-Chlorophenol	7.457	128	557086	22.247	ng/u1	99
13) 2-Methylphenol	8.340	108	526337	21.708	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.428	45	514923	21.330	ng/u1	100
16) Acetophenone	8.722	105	815059	21.795	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.710	70	384743	21.859	ng/u1	99
18) 4-Methylphenol	8.675	108	584150	21.866	ng/u1	100
19) Hexachloroethane	8.975	117	208655	22.341	ng/u1	99
22) Nitrobenzene	9.104	77	568597	22.235	ng/u1	99
23) Isophorone	9.634	82	1143111	22.036	ng/u1	100
25) 2-Nitrophenol	9.816	139	333631	22.963	ng/u1	98
26) 2,4-Dimethylphenol	9.881	107	610078	22.352	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.116	93	731252	21.849	ng/u1	99
29) 2,4-Dichlorophenol	10.351	162	560932	23.007	ng/u1	99
30) Naphthalene	10.757	128	1851164	22.078	ng/u1	100
32) 4-Chloroaniline	10.869	127	811096	22.101	ng/u1	99
33) Hexachlorobutadiene	11.039	225	323451	22.413	ng/u1	97
34) Caprolactam	11.663	113	181123	20.991	ng/u1	96
35) 4-Chloro-3-methylphenol	11.992	107	583201	22.560	ng/u1	99
36) 2-Methylnaphthalene	12.363	142	1283921	21.927	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100122\
 Data File : BP011865.D
 Acq On : 01 Oct 2022 17:01
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020761

Quant Time: Oct 03 01:01:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	1287834	21.919	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.728	216	672332	23.021	ng/ul	100
40) Hexachlorocyclopentadiene	12.710	237	339858	20.370	ng/ul	95
41) 2,4,6-Trichlorophenol	12.975	196	458953	23.476	ng/ul	98
42) 2,4,5-Trichlorophenol	13.045	196	504779	23.703	ng/ul	99
43) 1,1'-Biphenyl	13.375	154	1666208	22.369	ng/ul	99
44) 2-Chloronaphthalene	13.416	162	1333715	22.647	ng/ul	100
45) 2-Nitroaniline	13.622	65	336013	23.374	ng/ul	99
47) Dimethylphthalate	13.998	163	1637392	21.695	ng/ul	99
48) 2,6-Dinitrotoluene	14.122	165	345511	23.490	ng/ul	99
50) Acenaphthylene	14.263	152	2054706	22.534	ng/ul	100
51) 3-Nitroaniline	14.451	138	387201	22.712	ng/ul	94
52) Acenaphthene	14.604	153	1430289	22.233	ng/ul	99
53) 2,4-Dinitrophenol	14.657	184	203025	20.359	ng/ul	96
55) 4-Nitrophenol	14.763	109	218134	21.668	ng/ul	98
56) Dibenzofuran	14.939	168	1944071	22.064	ng/ul	100
57) 2,4-Dinitrotoluene	14.904	165	499164	23.217	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.169	232	426505	23.054	ng/ul	97
59) Diethylphthalate	15.363	149	1645116	21.757	ng/ul	100
61) Fluorene	15.592	166	1598381	22.148	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.586	204	778173	21.782	ng/ul	100
63) 4-Nitroaniline	15.616	138	389959	23.970	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.669	198	289591	21.326	ng/ul	98
67) N-Nitrosodiphenylamine	15.804	169	1386253	22.191	ng/ul	99
68) 4-Bromophenyl-phenylether	16.486	248	484640	22.499	ng/ul	98
69) Hexachlorobenzene	16.598	284	556308	22.695	ng/ul	98
70) Atrazine	16.763	200	500522	22.205	ng/ul	99
71) Pentachlorophenol	16.945	266	355873	22.074	ng/ul	99
72) Phenanthrene	17.339	178	2580411	22.172	ng/ul	100
74) Anthracene	17.433	178	2614109	22.324	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.339	216	673244	23.033	ng/ul	99
76) Pentachlorobenzene	14.857	250	709964	22.731	ng/ul	99
77) Carbazole	17.704	167	2433538	23.002	ng/ul	99
78) Di-n-butylphthalate	18.251	149	2913719	22.489	ng/ul	100
80) Fluoranthene	19.304	202	3008037	23.513	ng/ul	99
82) Pyrene	19.657	202	3128538	23.488	ng/ul	100
83) Butylbenzylphthalate	20.527	149	1403226	24.375	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.304	252	1020936	21.485	ng/ul	100
85) Benzo(a)anthracene	21.368	228	3036283	22.494	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.286	149	2071130	23.884	ng/ul	98
87) Chrysene	21.421	228	2794500	22.060	ng/ul	99
89) Di-n-octyl phthalate	22.221	149	1743592	24.108	ng/ul	100
90) Benzo(b)fluoranthene	23.074	252	1634981	22.597	ng/ul#	96
91) Benzo(k)fluoranthene	23.121	252	1598624	22.944	ng/ul#	96
93) Benzo(a)pyrene	23.709	252	1408973	21.821	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.356	276	1605691	19.537	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.380	278	1451578	19.882	ng/ul#	92
96) Benzo(g,h,i)perylene	27.139	276	1412655	19.388	ng/ul#	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100122\
 Data File : BP011865.D
 Acq On : 01 Oct 2022 17:01
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020761

Quant Time: Oct 03 01:01:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.362 ng/uL
RT: 3.340 min Scan# 60
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion: 88 Resp: 70384
Ion Ratio Lower Upper
88 100
43 19.4 15.6 23.4
58 67.7 48.6 72.8



#5
Pyridine
Concen: 21.407 ng/uL
RT: 3.752 min Scan# 130
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion: 79 Resp: 474273
Ion Ratio Lower Upper
79 100
52 58.5 49.7 74.5
51 26.1 21.6 32.4









#6
Benzaldehyde
Concen: 24.918 ng/ul
RT: 7.034 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion: 77 Resp: 323872
Ion Ratio Lower Upper
77 100
105 108.3 87.0 130.6
106 103.0 83.8 125.6



#8
Phenol
Concen: 21.428 ng/ul
RT: 7.087 min Scan# 697
Delta R.T. 0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion: 94 Resp: 659442
Ion Ratio Lower Upper
94 100
65 24.0 19.5 29.3
66 30.9 25.0 37.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.649 ng/ul
RT: 7.316 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion: 93 Resp: 530133
Ion Ratio Lower Upper
93 100
63 67.1 56.2 84.4
95 32.9 25.9 38.9



#12
2-Chlorophenol
Concen: 22.247 ng/ul
RT: 7.457 min Scan# 760
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:128 Resp: 557086
Ion Ratio Lower Upper
128 100
64 38.6 31.1 46.7
130 33.2 25.5 38.3









#13
2-Methylphenol
Concen: 21.708 ng/ul
RT: 8.340 min Scan# 910
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:108 Resp: 526337
Ion Ratio Lower Upper
108 100
107 87.7 70.2 105.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.330 ng/ul
RT: 8.428 min Scan# 925
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion: 45 Resp: 514923
Ion Ratio Lower Upper
45 100
77 20.1 16.0 24.0
79 15.4 12.5 18.7









#16
Acetophenone
Concen: 21.795 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:105 Resp: 815059
Ion Ratio Lower Upper
105 100
77 70.1 55.4 83.0
51 19.1 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 21.859 ng/ul
RT: 8.710 min Scan# 973
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion: 70 Resp: 384743
Ion Ratio Lower Upper
70 100
42 41.1 32.7 49.1
101 10.3 8.4 12.6
130 25.4 19.4 29.2









#18
4-Methylphenol
Concen: 21.866 ng/ul
RT: 8.675 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:108 Resp: 584150
Ion Ratio Lower Upper
108 100
107 110.1 88.0 132.0



#19
Hexachloroethane
Concen: 22.341 ng/ul
RT: 8.975 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:117 Resp: 208655
Ion Ratio Lower Upper
117 100
201 102.1 82.4 123.6
199 63.8 51.6 77.4









#22
Nitrobenzene
Concen: 22.235 ng/ul
RT: 9.104 min Scan# 1040
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion: 77 Resp: 568597
Ion Ratio Lower Upper
77 100
123 53.1 41.6 62.4
65 12.3 9.7 14.5



#23
Isophorone
Concen: 22.036 ng/ul
RT: 9.634 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion: 82 Resp: 1143111
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 19.7 15.8 23.8









#25
2-Nitrophenol
Concen: 22.963 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:139 Resp: 333631
Ion Ratio Lower Upper
139 100
65 41.0 32.2 48.4
109 26.1 19.6 29.4



#26
2,4-Dimethylphenol
Concen: 22.352 ng/ul
RT: 9.881 min Scan# 1172
Delta R.T. 0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:107 Resp: 610078
Ion Ratio Lower Upper
107 100
121 53.4 43.0 64.6
122 92.9 74.9 112.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.849 ng/ul
 RT: 10.116 min Scan# 1212
 Delta R.T. -0.000 min
 Lab File: BP011865.D
 Acq: 01 Oct 2022 17:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020761


Tgt Ion: 93 Resp: 731252
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.3 39.5
 123 12.2 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 23.007 ng/ul
 RT: 10.351 min Scan# 1252
 Delta R.T. -0.000 min
 Lab File: BP011865.D
 Acq: 01 Oct 2022 17:01

Tgt Ion:162 Resp: 560932
 Ion Ratio Lower Upper
 162 100
 164 64.9 52.2 78.4
 98 32.6 27.0 40.4









#30
Naphthalene
Concen: 22.078 ng/ul
RT: 10.757 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:128 Resp: 1851164
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 22.101 ng/ul
RT: 10.869 min Scan# 1340
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:127 Resp: 811096
Ion Ratio Lower Upper
127 100
129 32.7 26.6 39.8









#33
Hexachlorobutadiene
Concen: 22.413 ng/ul
RT: 11.039 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:225 Resp: 323451
Ion Ratio Lower Upper
225 100
223 62.2 52.1 78.1
227 63.4 52.3 78.5



#34
Caprolactam
Concen: 20.991 ng/ul
RT: 11.663 min Scan# 1475
Delta R.T. 0.011 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:113 Resp: 181123
Ion Ratio Lower Upper
113 100
55 136.3 106.0 159.0
56 102.3 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 22.560 ng/ul
RT: 11.992 min Scan# 1531
Delta R.T. 0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:107 Resp: 583201
Ion Ratio Lower Upper
107 100
144 28.7 22.5 33.7
142 88.0 70.0 105.0



#36
2-Methylnaphthalene
Concen: 21.927 ng/ul
RT: 12.363 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:142 Resp: 1283921
Ion Ratio Lower Upper
142 100
141 89.0 70.6 106.0









#37
1-Methylnaphthalene
Concen: 21.919 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:142 Resp: 1287834
Ion Ratio Lower Upper
142 100
141 92.4 74.8 112.2
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 23.021 ng/ul
RT: 12.728 min Scan# 1656
Delta R.T. -0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:216 Resp: 672332
Ion Ratio Lower Upper
216 100
214 78.4 62.6 94.0
179 20.3 16.6 24.8
108 16.0 12.6 19.0









#40
Hexachlorocyclopentadiene
Concen: 20.370 ng/ul
RT: 12.710 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:237 Resp: 339858
Ion Ratio Lower Upper
237 100
235 59.5 51.0 76.6
272 13.3 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 23.476 ng/ul
RT: 12.975 min Scan# 1698
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:196 Resp: 458953
Ion Ratio Lower Upper
196 100
198 96.2 78.7 118.1
200 30.9 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 23.703 ng/ul
RT: 13.045 min Scan# 1709
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:196 Resp: 504779
Ion Ratio Lower Upper
196 100
198 95.6 76.6 114.8
200 30.8 25.6 38.4



#43
1,1'-Biphenyl
Concen: 22.369 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:154 Resp: 1666208
Ion Ratio Lower Upper
154 100
153 40.1 31.8 47.6
76 11.3 9.5 14.3









#44
2-Chloronaphthalene
Concen: 22.647 ng/ul
RT: 13.416 min Scan# 1773
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:162 Resp: 1333715
Ion Ratio Lower Upper
162 100
164 33.0 26.5 39.7
127 37.1 29.7 44.5



#45
2-Nitroaniline
Concen: 23.374 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion: 65 Resp: 336013
Ion Ratio Lower Upper
65 100
92 76.3 61.2 91.8
138 127.8 103.0 154.4









#47
Dimethylphthalate
Concen: 21.695 ng/ul
RT: 13.998 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:163 Resp: 1637392
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 23.490 ng/ul
RT: 14.122 min Scan# 1893
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:165 Resp: 345511
Ion Ratio Lower Upper
165 100
89 48.2 39.1 58.7
121 20.4 16.7 25.1









#50
Acenaphthylene
Concen: 22.534 ng/ul
RT: 14.263 min Scan# 1917
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:152 Resp: 2054706
Ion Ratio Lower Upper
152 100
151 20.2 16.0 24.0
153 13.3 10.8 16.2



#51
3-Nitroaniline
Concen: 22.712 ng/ul
RT: 14.451 min Scan# 1949
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:138 Resp: 387201
Ion Ratio Lower Upper
138 100
108 9.4 8.2 12.4
92 99.9 85.6 128.4









#52
Acenaphthene
Concen: 22.233 ng/ul
RT: 14.604 min Scan# 1975
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:153 Resp: 1430289
Ion Ratio Lower Upper
153 100
152 47.3 38.4 57.6
154 88.9 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 20.359 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. 0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:184 Resp: 203025
Ion Ratio Lower Upper
184 100
63 50.7 44.2 66.4
154 56.5 45.8 68.6









#55
4-Nitrophenol
Concen: 21.668 ng/ul
RT: 14.763 min Scan# 2000
Delta R.T. 0.011 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:109 Resp: 218134
Ion Ratio Lower Upper
109 100
139 159.6 125.0 187.6
65 125.3 101.9 152.9



#56
Dibenzofuran
Concen: 22.064 ng/ul
RT: 14.939 min Scan# 2032
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:168 Resp: 1944071
Ion Ratio Lower Upper
168 100
139 36.5 29.4 44.2





#57

2,4-Dinitrotoluene

Concen: 23.217 ng/ul

RT: 14.904 min Scan# 2026

Delta R.T. -0.000 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020761





Tgt Ion:165 Resp: 499164

Ion Ratio Lower Upper

165 100

63 35.2 28.2 42.2

182 2.7 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.054 ng/ul

RT: 15.169 min Scan# 2071

Delta R.T. 0.006 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01

Tgt Ion:232 Resp: 426505




Ion Ratio Lower Upper


232 100

131 42.8 36.2 54.2

130 2.1 1.7 2.5

166 29.8 25.0 37.4









#59
Diethylphthalate
Concen: 21.757 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:149 Resp: 1645116
Ion Ratio Lower Upper
149 100
177 22.3 17.8 26.8
150 12.3 9.9 14.9



#61
Fluorene
Concen: 22.148 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:166 Resp: 1598381
Ion Ratio Lower Upper
166 100
165 97.5 77.2 115.8
167 13.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 21.782 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:204 Resp: 778173
Ion Ratio Lower Upper
204 100
206 33.5 26.8 40.2
141 56.2 44.5 66.7



#63
4-Nitroaniline
Concen: 23.970 ng/ul
RT: 15.616 min Scan# 2147
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:138 Resp: 389959
Ion Ratio Lower Upper
138 100
92 45.9 35.9 53.9
108 65.4 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 21.326 ng/ul
RT: 15.669 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:198 Resp: 289591
Ion Ratio Lower Upper
198 100
182 4.3 3.7 5.5
77 21.7 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 22.191 ng/ul
RT: 15.804 min Scan# 2179
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:169 Resp: 1386253
Ion Ratio Lower Upper
169 100
168 66.0 53.7 80.5
167 36.6 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 22.499 ng/ul
RT: 16.486 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:248 Resp: 484640
Ion Ratio Lower Upper
248 100
250 97.3 78.6 118.0
141 66.3 56.0 84.0



#69
Hexachlorobenzene
Concen: 22.695 ng/ul
RT: 16.598 min Scan# 2314
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:284 Resp: 556308
Ion Ratio Lower Upper
284 100
142 33.6 27.3 40.9
249 30.2 25.2 37.8









#70
Atrazine
Concen: 22.205 ng/ul
RT: 16.763 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:200 Resp: 500522
Ion Ratio Lower Upper
200 100
173 26.9 21.6 32.4
215 49.2 38.9 58.3



#71
Pentachlorophenol
Concen: 22.074 ng/ul
RT: 16.945 min Scan# 2373
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:266 Resp: 355873
Ion Ratio Lower Upper
266 100
264 63.6 49.8 74.8
268 64.1 51.9 77.9









#72
Phenanthrene
Concen: 22.172 ng/ul
RT: 17.339 min Scan# 2440
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:178 Resp: 2580411
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 19.5 15.4 23.2



#74
Anthracene
Concen: 22.324 ng/ul
RT: 17.433 min Scan# 2456
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:178 Resp: 2614109
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.8 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.033 ng/uL
RT: 13.339 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:216 Resp: 673244
Ion Ratio Lower Upper
216 100
214 77.6 62.6 93.8
218 48.4 37.4 56.0



#76
Pentachlorobenzene
Concen: 22.731 ng/uL
RT: 14.857 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:250 Resp: 709964
Ion Ratio Lower Upper
250 100
248 63.6 51.4 77.0
252 62.7 51.5 77.3









#77
 Carbazole
 Concen: 23.002 ng/ul
 RT: 17.704 min Scan# 2502
 Delta R.T. -0.000 min
 Lab File: BP011865.D
 Acq: 01 Oct 2022 17:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020761


Tgt Ion:167 Resp: 2433538
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.1 25.7
 139 12.6 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 22.489 ng/ul
 RT: 18.251 min Scan# 2595
 Delta R.T. -0.000 min
 Lab File: BP011865.D
 Acq: 01 Oct 2022 17:01

Tgt Ion:149 Resp: 2913719
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 4.4 3.6 5.4









#80
Fluoranthene
Concen: 23.513 ng/ul
RT: 19.304 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:202 Resp: 3008037
Ion Ratio Lower Upper
202 100
101 11.9 9.8 14.6
100 8.9 7.4 11.0



#82
Pyrene
Concen: 23.488 ng/ul
RT: 19.657 min Scan# 2834
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:202 Resp: 3128538
Ion Ratio Lower Upper
202 100
101 14.4 11.6 17.4
100 11.6 9.4 14.0





#83

Butylbenzylphthalate

Concen: 24.375 ng/ul

RT: 20.527 min Scan# 2983

Delta R.T. -0.006 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020761





Tgt Ion:149 Resp: 1403226

Ion Ratio Lower Upper

149 100

91 64.8 50.2 75.4

206 20.6 16.9 25.3



#84

3,3'-Dichlorobenzidine

Concen: 21.485 ng/ul

RT: 21.304 min Scan# 3114

Delta R.T. -0.000 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01




Tgt Ion:252 Resp: 1020936


Ion Ratio Lower Upper

252 100

254 64.3 51.4 77.0

126 11.1 8.9 13.3





#85

Benzo(a)anthracene

Concen: 22.494 ng/ul

RT: 21.368 min Scan# 3111

Delta R.T. -0.000 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020761





Tgt Ion:228 Resp: 3036283

Ion Ratio Lower Upper

228 100

229 19.7 15.8 23.6

226 28.0 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.884 ng/ul

RT: 21.286 min Scan# 3111

Delta R.T. -0.000 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01




Tgt Ion:149 Resp: 2071130


Ion Ratio Lower Upper

149 100

167 27.5 21.2 31.8

279 4.2 2.9 4.3









#87
Chrysene
Concen: 22.060 ng/ul
RT: 21.421 min Scan# 3134
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:228 Resp: 2794500
Ion Ratio Lower Upper
228 100
226 31.2 25.3 37.9
229 20.3 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 24.108 ng/ul
RT: 22.221 min Scan# 3270
Delta R.T. -0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:149 Resp: 1743592









#90
Benzo(b)fluoranthene
Concen: 22.597 ng/ul
RT: 23.074 min Scan# 3415
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01




Instrument :
BNA_P
ClientSampleId :
SSTD020761


Tgt Ion:252 Resp: 1634981
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 6.8 9.2 13.8#



#91
Benzo(k)fluoranthene
Concen: 22.944 ng/ul
RT: 23.121 min Scan# 3423
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:252 Resp: 1598624
Ion Ratio Lower Upper
252 100
253 22.0 18.0 27.0
125 7.2 9.2 13.8#





#93

Benzo(a)pyrene

Concen: 21.821 ng/ul

RT: 23.709 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01

Instrument :

BNA_P

ClientSampleId :

SSTD020761





Tgt Ion:252 Resp: 1408973

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 7.9 9.8 14.8#



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.537 ng/ul

RT: 26.356 min Scan# 3973

Delta R.T. -0.006 min

Lab File: BP011865.D

Acq: 01 Oct 2022 17:01




Tgt Ion:276 Resp: 1605691


Ion Ratio Lower Upper

276 100

138 17.1 21.4 32.0#

277 25.3 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 19.882 ng/ul
RT: 26.380 min Scan# 3977
Delta R.T. -0.000 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Instrument :
BNA_P
ClientSampleId :
SSTD020761

Tgt Ion:278 Resp: 1451578
Ion Ratio Lower Upper
278 100
139 11.3 15.0 22.6#
279 24.2 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 19.388 ng/ul
RT: 27.139 min Scan# 4106
Delta R.T. -0.006 min
Lab File: BP011865.D
Acq: 01 Oct 2022 17:01

Tgt Ion:276 Resp: 1412655
Ion Ratio Lower Upper
276 100
138 15.5 19.8 29.6#
277 23.4 18.6 27.8

