

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100223\
 Data File : BP017480.D
 Acq On : 02 Oct 2023 16:50
 Operator : MA/JU
 Sample : 04562-09 50X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCJL9

Quant Time: Oct 02 23:14:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 30 04:02:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	100871	20.000	ng/ul	0.00
20) Naphthalene-d8	10.787	136	399567	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.640	164	237359	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.428	188	520986	20.000	ng/ul	0.00
79) Chrysene-d12	21.557	240	465700	20.000	ng/ul	0.00
88) Perylene-d12	24.127	264	513130	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	524	0.213	ng/uL	0.00
4) Pyridine-d5	3.722	84	316	0.046	ng/ul	0.00
7) Phenol-d5	7.111	99	1177	0.142	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	2822	0.564	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	3245	0.494	ng/ul	0.00
15) 4-Methylphenol-d8	8.687	113	1611	0.250	ng/ul	0.02
21) Nitrobenzene-d5	9.163	128	1762	0.610	ng/ul	0.00
24) 2-Nitrophenol-d4	9.881	143	1485	0.473	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.422	165	2973	0.502	ng/ul	0.02
31) 4-Chloroaniline-d4	10.975	131	1312	0.153	ng/ul	0.02
46) Dimethylphthalate-d6	14.057	166	14252	0.838	ng/ul	0.00
49) Acenaphthylene-d8	14.340	160	13492	0.690	ng/ul	0.00
54) 4-Nitrophenol-d4	14.857	143	11	0.004	ng/ul	-0.02
60) Fluorene-d10	15.645	176	11681	0.798	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.787	200	79	0.027	ng/ul	0.00
73) Anthracene-d10	17.528	188	18945	0.820	ng/ul	0.00
81) Pyrene-d10	19.780	212	23075	0.916	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.127	264	382621	15.487	ng/ul	0.17

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100223\
Data File : BP017480.D
Acq On : 02 Oct 2023 16:50
Operator : MA/JU
Sample : 04562-09 50X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCJL9

Quant Time: Oct 02 23:14:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 30 04:02:20 2023
Response via : Initial Calibration

