

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100223\
 Data File : BP017486.D
 Acq On : 02 Oct 2023 20:15
 Operator : MA/JU
 Sample : 04571-16
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BBJ31

Quant Time: Oct 02 23:16:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 30 04:02:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.945	152	148384	20.000	ng/ul	0.00
20) Naphthalene-d8	10.786	136	620675	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	347215	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.427	188	662893	20.000	ng/ul	0.00
79) Chrysene-d12	21.556	240	631553	20.000	ng/ul	0.00
88) Perylene-d12	24.127	264	706949	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	20018	5.541	ng/uL	0.00
4) Pyridine-d5	3.734	84	121587	12.039	ng/ul	0.00
7) Phenol-d5	7.104	99	105955	8.694	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	316240	42.999	ng/ul	0.00
11) 2-Chlorophenol-d4	7.463	132	321077	33.218	ng/ul	0.00
15) 4-Methylphenol-d8	8.669	113	202521	21.375	ng/ul	0.00
21) Nitrobenzene-d5	9.157	128	212151	47.280	ng/ul	0.00
24) 2-Nitrophenol-d4	9.869	143	221606	45.484	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	381132	41.405	ng/ul	0.00
31) 4-Chloroaniline-d4	10.951	131	383638	28.771	ng/ul	0.00
46) Dimethylphthalate-d6	14.057	166	1322263	53.160	ng/ul	0.00
49) Acenaphthylene-d8	14.339	160	1392264	48.668	ng/ul	0.00
54) 4-Nitrophenol-d4	14.880	143	19054	4.397	ng/ul	0.00
60) Fluorene-d10	15.645	176	1040821	48.605	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.780	200	150351	40.242	ng/ul	0.00
73) Anthracene-d10	17.527	188	1571368	53.448	ng/ul	0.00
81) Pyrene-d10	19.780	212	1877732	54.980	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.956	264	1875068	55.088	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100223\
Data File : BP017486.D
Acq On : 02 Oct 2023 20:15
Operator : MA/JU
Sample : 04571-16
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BBJ31

Quant Time: Oct 02 23:16:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 30 04:02:20 2023
Response via : Initial Calibration

