

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011867.D
 Acq On : 03 Oct 2022 12:16
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762

Quant Time: Oct 03 23:29:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	255673	20.000	ng/u1	0.00
20) Naphthalene-d8	10.705	136	1017865	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.540	164	571661	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1116527	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	1195328	20.000	ng/u1	0.00
88) Perylene-d12	23.816	264	1296231	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	49565	8.550	ng/uL	0.00
4) Pyridine-d5	3.728	84	321251	18.919	ng/u1	0.00
7) Phenol-d5	7.058	99	390242	17.680	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	222914	18.220	ng/u1	0.00
11) 2-Chlorophenol-d4	7.422	132	328889	18.876	ng/u1	0.00
15) 4-Methylphenol-d8	8.605	113	300594	16.474	ng/u1	0.00
21) Nitrobenzene-d5	9.058	128	154522	19.931	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	158556	19.115	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	291960	19.029	ng/u1	0.00
31) 4-Chloroaniline-d4	10.840	131	418590	17.896	ng/u1	0.00
46) Dimethylphthalate-d6	13.951	166	760990	18.615	ng/u1	0.00
49) Acenaphthylene-d8	14.234	160	947358	20.328	ng/u1	0.00
54) 4-Nitrophenol-d4	14.740	143	141507	16.716	ng/u1	0.00
60) Fluorene-d10	15.534	176	679431	19.138	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	107990	15.980	ng/u1	0.00
73) Anthracene-d10	17.392	188	1035831	20.366	ng/u1	0.00
81) Pyrene-d10	19.628	212	1201130	19.841	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.657	264	1302720	20.076	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	51551	8.233	ng/uL	97
5) Pyridine	3.746	79	323775	19.645	ng/u1	99
6) Benzaldehyde	7.034	77	204585	21.159	ng/u1	99
8) Phenol	7.081	94	411150	17.959	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.316	93	335209	18.402	ng/u1	98
12) 2-Chlorophenol	7.452	128	354510	19.031	ng/u1	98
13) 2-Methylphenol	8.340	108	308419	17.100	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.422	45	316739	17.638	ng/u1	100
16) Acetophenone	8.722	105	481252	17.300	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.705	70	216866	16.563	ng/u1	99
18) 4-Methylphenol	8.669	108	333012	16.757	ng/u1	98
19) Hexachloroethane	8.975	117	139758	20.116	ng/u1	99
22) Nitrobenzene	9.099	77	337152	20.220	ng/u1	98
23) Isophorone	9.634	82	610175	18.039	ng/u1	100
25) 2-Nitrophenol	9.816	139	182636	19.278	ng/u1	97
26) 2,4-Dimethylphenol	9.875	107	344970	19.384	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.116	93	417786	19.144	ng/u1	99
29) 2,4-Dichlorophenol	10.352	162	307227	19.326	ng/u1	98
30) Naphthalene	10.752	128	1093553	20.003	ng/u1	100
32) 4-Chloroaniline	10.863	127	433808	18.129	ng/u1	99
33) Hexachlorobutadiene	11.034	225	195981	20.827	ng/u1	97
34) Caprolactam	11.657	113	79460	14.123	ng/u1	99
35) 4-Chloro-3-methylphenol	11.993	107	289724	17.188	ng/u1	99
36) 2-Methylnaphthalene	12.363	142	708898	18.567	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011867.D
 Acq On : 03 Oct 2022 12:16
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762

Quant Time: Oct 03 23:29:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	704566	18.391	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.728	216	360595	21.751	ng/ul	99
40) Hexachlorocyclopentadiene	12.704	237	186725	19.716	ng/ul	99
41) 2,4,6-Trichlorophenol	12.969	196	217985	19.643	ng/ul	95
42) 2,4,5-Trichlorophenol	13.040	196	238952	19.767	ng/ul	98
43) 1,1'-Biphenyl	13.375	154	884284	20.913	ng/ul	100
44) 2-Chloronaphthalene	13.410	162	710861	21.264	ng/ul	99
45) 2-Nitroaniline	13.622	65	156361	19.161	ng/ul	99
47) Dimethylphthalate	13.998	163	803692	18.759	ng/ul	99
48) 2,6-Dinitrotoluene	14.116	165	153726	18.412	ng/ul	98
50) Acenaphthylene	14.263	152	1057686	20.435	ng/ul	99
51) 3-Nitroaniline	14.445	138	172692	17.845	ng/ul	99
52) Acenaphthene	14.604	153	740564	20.280	ng/ul	99
53) 2,4-Dinitrophenol	14.651	184	72401	12.790	ng/ul	98
55) 4-Nitrophenol	14.757	109	100079	17.513	ng/ul	95
56) Dibenzofuran	14.934	168	996025	19.914	ng/ul	98
57) 2,4-Dinitrotoluene	14.904	165	219361	17.974	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.163	232	189395	18.035	ng/ul	99
59) Diethylphthalate	15.363	149	778490	18.138	ng/ul	99
61) Fluorene	15.587	166	802291	19.585	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.581	204	389681	19.215	ng/ul	97
63) 4-Nitroaniline	15.610	138	173056	18.739	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.663	198	121473	16.967	ng/ul	94
67) N-Nitrosodiphenylamine	15.798	169	678528	20.601	ng/ul	99
68) 4-Bromophenyl-phenylether	16.481	248	222956	19.632	ng/ul	98
69) Hexachlorobenzene	16.592	284	259713	20.096	ng/ul	98
70) Atrazine	16.757	200	222496	18.722	ng/ul	98
71) Pentachlorophenol	16.945	266	142549	16.770	ng/ul	99
72) Phenanthrene	17.339	178	1263480	20.591	ng/ul	100
74) Anthracene	17.428	178	1275678	20.663	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.334	216	357609	23.205	ng/uL	99
76) Pentachlorobenzene	14.857	250	354106	21.503	ng/uL	100
77) Carbazole	17.704	167	1172890	21.027	ng/ul	100
78) Di-n-butylphthalate	18.251	149	1299475	19.023	ng/ul	99
80) Fluoranthene	19.304	202	1467704	19.811	ng/ul	98
82) Pyrene	19.657	202	1564758	20.286	ng/ul	97
83) Butylbenzylphthalate	20.528	149	620866	18.624	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.298	252	524309	19.053	ng/ul	99
85) Benzo(a)anthracene	21.363	228	1583429	20.257	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.286	149	927224	18.464	ng/ul	97
87) Chrysene	21.416	228	1476278	20.124	ng/ul	98
89) Di-n-octyl phthalate	22.222	149	1612787	19.502	ng/ul	100
90) Benzo(b)fluoranthene	23.069	252	1651782	19.965	ng/ul	99
91) Benzo(k)fluoranthene	23.116	252	1647531	20.679	ng/ul	99
93) Benzo(a)pyrene	23.704	252	1496096	20.263	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.351	276	1874597	19.947	ng/ul	99
95) Dibenzo(a,h)anthracene	26.374	278	1688429	20.225	ng/ul	99
96) Benzo(g,h,i)perylene	27.133	276	1652888	19.839	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011867.D
 Acq On : 03 Oct 2022 12:16
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762

Quant Time: Oct 03 23:29:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.233 ng/uL
 RT: 3.340 min Scan# 60
 Delta R.T. 0.000 min
 Lab File: BP011867.D
 Acq: 03 Oct 2022 12:16




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762


Tgt Ion: 88 Resp: 51551
 Ion Ratio Lower Upper
 88 100
 43 18.3 15.6 23.4
 58 62.9 48.6 72.8



#5
 Pyridine
 Concen: 19.645 ng/uL
 RT: 3.746 min Scan# 129
 Delta R.T. -0.006 min
 Lab File: BP011867.D
 Acq: 03 Oct 2022 12:16

Tgt Ion: 79 Resp: 323775
 Ion Ratio Lower Upper
 79 100
 52 61.3 49.7 74.5
 51 26.9 21.6 32.4









#6
Benzaldehyde
Concen: 21.159 ng/ul
RT: 7.034 min Scan# 688
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion: 77 Resp: 204585
Ion Ratio Lower Upper
77 100
105 108.1 87.0 130.6
106 103.0 83.8 125.6



#8
Phenol
Concen: 17.959 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion: 94 Resp: 411150
Ion Ratio Lower Upper
94 100
65 24.1 19.5 29.3
66 31.5 25.0 37.4





#10

Bis(2-Chloroethyl)ether

Concen: 18.402 ng/ul

RT: 7.316 min Scan# 736

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion: 93 Resp: 335209

Ion Ratio Lower Upper

93 100

63 68.3 56.2 84.4

95 33.2 25.9 38.9



#12

2-Chlorophenol

Concen: 19.031 ng/ul

RT: 7.452 min Scan# 759

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:128 Resp: 354510


Ion Ratio Lower Upper

128 100

64 40.2 31.1 46.7

130 33.0 25.5 38.3





#13

2-Methylphenol

Concen: 17.100 ng/ul

RT: 8.340 min Scan# 910

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :





SSTD020762

Tgt Ion:108 Resp: 308419

Ion Ratio Lower Upper

108 100

107 88.4 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 17.638 ng/ul

RT: 8.422 min Scan# 924

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion: 45 Resp: 316739


Ion Ratio Lower Upper

45 100

77 19.8 16.0 24.0

79 15.6 12.5 18.7









#16
Acetophenone
Concen: 17.300 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:105 Resp: 481252
Ion Ratio Lower Upper
105 100
77 71.7 55.4 83.0
51 18.7 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 16.563 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion: 70 Resp: 216866
Ion Ratio Lower Upper
70 100
42 40.6 32.7 49.1
101 10.2 8.4 12.6
130 23.7 19.4 29.2









#18
4-Methylphenol
Concen: 16.757 ng/ul
RT: 8.669 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:108 Resp: 333012
Ion Ratio Lower Upper
108 100
107 107.6 88.0 132.0



#19
Hexachloroethane
Concen: 20.116 ng/ul
RT: 8.975 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:117 Resp: 139758
Ion Ratio Lower Upper
117 100
201 103.7 82.4 123.6
199 64.9 51.6 77.4









#22
Nitrobenzene
Concen: 20.220 ng/ul
RT: 9.099 min Scan# 1039
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion: 77 Resp: 337152
Ion Ratio Lower Upper
77 100
123 50.7 41.6 62.4
65 12.5 9.7 14.5



#23
Isophorone
Concen: 18.039 ng/ul
RT: 9.634 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion: 82 Resp: 610175
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.2
138 20.0 15.8 23.8









#25
2-Nitrophenol
Concen: 19.278 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:139 Resp: 182636
Ion Ratio Lower Upper
139 100
65 41.7 32.2 48.4
109 26.2 19.6 29.4



#26
2,4-Dimethylphenol
Concen: 19.384 ng/ul
RT: 9.875 min Scan# 1171
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:107 Resp: 344970
Ion Ratio Lower Upper
107 100
121 54.3 43.0 64.6
122 91.2 74.9 112.3





#27

Bis(2-Chloroethoxy)methane

Concen: 19.144 ng/ul

RT: 10.116 min Scan# 1212

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion: 93 Resp: 417786

Ion Ratio Lower Upper

93 100

95 32.0 26.3 39.5

123 11.8 9.5 14.3



#29

2,4-Dichlorophenol

Concen: 19.326 ng/ul

RT: 10.352 min Scan# 1252

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion: 162 Resp: 307227


Ion Ratio Lower Upper

162 100

164 67.3 52.2 78.4

98 33.4 27.0 40.4









#30
Naphthalene
Concen: 20.003 ng/ul
RT: 10.752 min Scan# 1320
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:128 Resp: 1093553
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 18.129 ng/ul
RT: 10.863 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:127 Resp: 433808
Ion Ratio Lower Upper
127 100
129 32.6 26.6 39.8









#33
Hexachlorobutadiene
Concen: 20.827 ng/ul
RT: 11.034 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:225 Resp: 195981
Ion Ratio Lower Upper
225 100
223 62.7 52.1 78.1
227 63.0 52.3 78.5



#34
Caprolactam
Concen: 14.123 ng/ul
RT: 11.657 min Scan# 1474
Delta R.T. 0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:113 Resp: 79460
Ion Ratio Lower Upper
113 100
55 134.5 106.0 159.0
56 98.7 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 17.188 ng/ul
RT: 11.993 min Scan# 1531
Delta R.T. 0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:107 Resp: 289724
Ion Ratio Lower Upper
107 100
144 27.5 22.5 33.7
142 87.1 70.0 105.0



#36
2-Methylnaphthalene
Concen: 18.567 ng/ul
RT: 12.363 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:142 Resp: 708898
Ion Ratio Lower Upper
142 100
141 89.0 70.6 106.0









#37
1-Methylnaphthalene
Concen: 18.391 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:142 Resp: 704566
Ion Ratio Lower Upper
142 100
141 91.3 74.8 112.2
116 3.6 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.751 ng/ul
RT: 12.728 min Scan# 1656
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:216 Resp: 360595
Ion Ratio Lower Upper
216 100
214 79.7 62.6 94.0
179 21.2 16.6 24.8
108 16.3 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 19.716 ng/ul

RT: 12.704 min Scan# 1652

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:237 Resp: 186725

Ion Ratio Lower Upper

237 100

235 62.9 51.0 76.6

272 13.0 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 19.643 ng/ul

RT: 12.969 min Scan# 1697

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:196 Resp: 217985


Ion Ratio Lower Upper

196 100

198 91.6 78.7 118.1

200 30.3 24.5 36.7





#42

2,4,5-Trichlorophenol

Concen: 19.767 ng/ul

RT: 13.040 min Scan# 1709

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:196 Resp: 238952

Ion Ratio Lower Upper

196 100

198 97.8 76.6 114.8

200 31.3 25.6 38.4



#43

1,1'-Biphenyl

Concen: 20.913 ng/ul

RT: 13.375 min Scan# 1766

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:154 Resp: 884284


Ion Ratio Lower Upper

154 100

153 39.6 31.8 47.6

76 11.9 9.5 14.3









#44
2-Chloronaphthalene
Concen: 21.264 ng/ul
RT: 13.410 min Scan# 1772
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:162 Resp: 710861
Ion Ratio Lower Upper
162 100
164 32.7 26.5 39.7
127 37.8 29.7 44.5



#45
2-Nitroaniline
Concen: 19.161 ng/ul
RT: 13.622 min Scan# 1808
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion: 65 Resp: 156361
Ion Ratio Lower Upper
65 100
92 76.4 61.2 91.8
138 129.6 103.0 154.4





#47

Dimethylphthalate

Concen: 18.759 ng/ul

RT: 13.998 min Scan# 1872

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:163 Resp: 803692

Ion Ratio Lower Upper

163 100

194 4.0 3.3 4.9

164 10.0 8.5 12.7



#48

2,6-Dinitrotoluene

Concen: 18.412 ng/ul

RT: 14.116 min Scan# 1892

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:165 Resp: 153726


Ion Ratio Lower Upper

165 100

89 50.7 39.1 58.7

121 20.8 16.7 25.1









#50
Acenaphthylene
Concen: 20.435 ng/ul
RT: 14.263 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:152 Resp: 1057686
Ion Ratio Lower Upper
152 100
151 20.1 16.0 24.0
153 13.1 10.8 16.2



#51
3-Nitroaniline
Concen: 17.845 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:138 Resp: 172692
Ion Ratio Lower Upper
138 100
108 10.8 8.2 12.4
92 107.8 85.6 128.4









#52
Acenaphthene
Concen: 20.280 ng/ul
RT: 14.604 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:153 Resp: 740564
Ion Ratio Lower Upper
153 100
152 47.4 38.4 57.6
154 87.8 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 12.790 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:184 Resp: 72401
Ion Ratio Lower Upper
184 100
63 54.6 44.2 66.4
154 55.5 45.8 68.6









#55
4-Nitrophenol
Concen: 17.513 ng/ul
RT: 14.757 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:109 Resp: 100079
Ion Ratio Lower Upper
109 100
139 163.5 125.0 187.6
65 122.9 101.9 152.9



#56
Dibenzofuran
Concen: 19.914 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:168 Resp: 996025
Ion Ratio Lower Upper
168 100
139 35.7 29.4 44.2









#57
2,4-Dinitrotoluene
Concen: 17.974 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:165 Resp: 219361
Ion Ratio Lower Upper
165 100
63 34.9 28.2 42.2
182 2.6 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.035 ng/ul
RT: 15.163 min Scan# 2070
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:232 Resp: 189395
Ion Ratio Lower Upper
232 100
131 45.7 36.2 54.2
130 2.4 1.7 2.5
166 31.3 25.0 37.4









#59
Diethylphthalate
Concen: 18.138 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:149 Resp: 778490
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 12.2 9.9 14.9



#61
Fluorene
Concen: 19.585 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:166 Resp: 802291
Ion Ratio Lower Upper
166 100
165 96.9 77.2 115.8
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 19.215 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:204 Resp: 389681
Ion Ratio Lower Upper
204 100
206 32.8 26.8 40.2
141 58.9 44.5 66.7



#63
4-Nitroaniline
Concen: 18.739 ng/ul
RT: 15.610 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:138 Resp: 173056
Ion Ratio Lower Upper
138 100
92 49.0 35.9 53.9
108 65.9 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 16.967 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:198 Resp: 121473
Ion Ratio Lower Upper
198 100
182 5.0 3.7 5.5
77 24.3 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 20.601 ng/ul
RT: 15.798 min Scan# 2178
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:169 Resp: 678528
Ion Ratio Lower Upper
169 100
168 66.3 53.7 80.5
167 37.2 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 19.632 ng/ul
RT: 16.481 min Scan# 2294
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:248 Resp: 222956
Ion Ratio Lower Upper
248 100
250 97.7 78.6 118.0
141 73.3 56.0 84.0



#69
Hexachlorobenzene
Concen: 20.096 ng/ul
RT: 16.592 min Scan# 2313
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:284 Resp: 259713
Ion Ratio Lower Upper
284 100
142 36.6 27.3 40.9
249 31.5 25.2 37.8





#70

Atrazine

Concen: 18.722 ng/ul

RT: 16.757 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:200 Resp: 222496

Ion Ratio Lower Upper

200 100

173 26.4 21.6 32.4

215 49.9 38.9 58.3



#71

Pentachlorophenol

Concen: 16.770 ng/ul

RT: 16.945 min Scan# 2373

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:266 Resp: 142549


Ion Ratio Lower Upper

266 100

264 62.6 49.8 74.8

268 62.9 51.9 77.9









#72
Phenanthrene
Concen: 20.591 ng/ul
RT: 17.339 min Scan# 2440
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:178 Resp: 1263480
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 19.4 15.4 23.2



#74
Anthracene
Concen: 20.663 ng/ul
RT: 17.428 min Scan# 2455
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:178 Resp: 1275678
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 18.4 15.4 23.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.205 ng/uL

RT: 13.334 min Scan# 1759

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:216 Resp: 357609

Ion Ratio Lower Upper

216 100

214 79.3 62.6 93.8

218 47.0 37.4 56.0



#76

Pentachlorobenzene

Concen: 21.503 ng/uL

RT: 14.857 min Scan# 2018

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:250 Resp: 354106


Ion Ratio Lower Upper

250 100

248 63.8 51.4 77.0

252 64.5 51.5 77.3





#77

Carbazole

Concen: 21.027 ng/ul

RT: 17.704 min Scan# 2502

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:167 Resp: 1172890

Ion Ratio Lower Upper

167 100

166 21.5 17.1 25.7

139 13.0 10.1 15.1



#78

Di-n-butylphthalate

Concen: 19.023 ng/ul

RT: 18.251 min Scan# 2595

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:149 Resp: 1299475


Ion Ratio Lower Upper

149 100

150 9.0 7.5 11.3

104 4.1 3.6 5.4









#80
Fluoranthene
Concen: 19.811 ng/ul
RT: 19.304 min Scan# 2774
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:202 Resp: 1467704
Ion Ratio Lower Upper
202 100
101 11.7 9.8 14.6
100 8.4 7.4 11.0



#82
Pyrene
Concen: 20.286 ng/ul
RT: 19.657 min Scan# 2834
Delta R.T. 0.000 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:202 Resp: 1564758
Ion Ratio Lower Upper
202 100
101 13.1 11.6 17.4
100 10.9 9.4 14.0









#83
Butylbenzylphthalate
Concen: 18.624 ng/ul
RT: 20.528 min Scan# 2982
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:149 Resp: 620866
Ion Ratio Lower Upper
149 100
91 64.9 50.2 75.4
206 20.4 16.9 25.3



#84
3,3'-Dichlorobenzidine
Concen: 19.053 ng/ul
RT: 21.298 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:252 Resp: 524309
Ion Ratio Lower Upper
252 100
254 63.6 51.4 77.0
126 11.4 8.9 13.3





#85

Benzo(a)anthracene

Concen: 20.257 ng/ul

RT: 21.363 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:228 Resp: 1583429

Ion Ratio Lower Upper

228 100

229 19.4 15.8 23.6

226 27.4 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.464 ng/ul

RT: 21.286 min Scan# 3111

Delta R.T. 0.000 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:149 Resp: 927224


Ion Ratio Lower Upper

149 100

167 27.9 21.2 31.8

279 4.2 2.9 4.3









#87
Chrysene
Concen: 20.124 ng/ul
RT: 21.416 min Scan# 3133
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:228 Resp: 1476278
Ion Ratio Lower Upper
228 100
226 30.2 25.3 37.9
229 20.0 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 19.502 ng/ul
RT: 22.222 min Scan# 3270
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:149 Resp: 1612787









#90
Benzo(b)fluoranthene
Concen: 19.965 ng/ul
RT: 23.069 min Scan# 3414
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:252 Resp: 1651782
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 11.2 9.2 13.8



#91
Benzo(k)fluoranthene
Concen: 20.679 ng/ul
RT: 23.116 min Scan# 3422
Delta R.T. -0.006 min
Lab File: BP011867.D
Acq: 03 Oct 2022 12:16

Tgt Ion:252 Resp: 1647531
Ion Ratio Lower Upper
252 100
253 21.7 18.0 27.0
125 11.5 9.2 13.8





#93

Benzo(a)pyrene

Concen: 20.263 ng/ul

RT: 23.704 min Scan# 3522

Delta R.T. -0.012 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:252 Resp: 1496096

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 12.6 9.8 14.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.947 ng/ul

RT: 26.351 min Scan# 3972

Delta R.T. -0.012 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16




Tgt Ion:276 Resp: 1874597


Ion Ratio Lower Upper

276 100

138 26.1 21.4 32.0

277 24.7 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 20.225 ng/ul

RT: 26.374 min Scan# 3976

Delta R.T. -0.006 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:278 Resp: 1688429

Ion Ratio Lower Upper

278 100

139 17.8 15.0 22.6

279 23.8 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 19.839 ng/ul

RT: 27.133 min Scan# 4105

Delta R.T. -0.012 min

Lab File: BP011867.D

Acq: 03 Oct 2022 12:16

Tgt Ion:276 Resp: 1652888

Ion Ratio Lower Upper

276 100

138 23.9 19.8 29.6

277 23.7 18.6 27.8

