

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011884.D
 Acq On : 03 Oct 2022 22:30
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Oct 04 00:40:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	258711	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	1062862	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.533	164	642448	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1330073	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	1377140	20.000	ng/u1	-0.01
88) Perylene-d12	23.809	264	1419839	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	48340	8.241	ng/uL	0.00
4) Pyridine-d5	3.728	84	329297	19.165	ng/u1	0.00
7) Phenol-d5	7.051	99	403288	18.057	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	226465	18.293	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	340203	19.296	ng/u1	0.00
15) 4-Methylphenol-d8	8.598	113	316736	17.155	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	160196	19.788	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	169012	19.513	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	325464	20.315	ng/u1	0.00
31) 4-Chloroaniline-d4	10.839	131	448418	18.360	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	875394	19.054	ng/u1	0.00
49) Acenaphthylene-d8	14.227	160	1061425	20.266	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	176169	18.518	ng/u1	0.00
60) Fluorene-d10	15.527	176	785943	19.699	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	138293	17.178	ng/u1	0.00
73) Anthracene-d10	17.392	188	1210755	19.984	ng/u1	0.00
81) Pyrene-d10	19.621	212	1408941	20.201	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.651	264	1432467	20.154	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	50212	7.925	ng/uL	97
5) Pyridine	3.746	79	320944	19.245	ng/u1	98
6) Benzaldehyde	7.028	77	210524	21.518	ng/u1	96
8) Phenol	7.081	94	422943	18.258	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.310	93	337773	18.325	ng/u1	97
12) 2-Chlorophenol	7.451	128	364805	19.354	ng/u1	98
13) 2-Methylphenol	8.334	108	321934	17.639	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.422	45	320785	17.653	ng/u1	100
16) Acetophenone	8.716	105	498023	17.692	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.704	70	224959	16.980	ng/u1	99
18) 4-Methylphenol	8.663	108	350412	17.426	ng/u1	98
19) Hexachloroethane	8.969	117	138072	19.640	ng/u1	97
22) Nitrobenzene	9.098	77	349416	20.068	ng/u1	99
23) Isophorone	9.628	82	633193	17.927	ng/u1	99
25) 2-Nitrophenol	9.810	139	197126	19.927	ng/u1	99
26) 2,4-Dimethylphenol	9.869	107	365043	19.643	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.110	93	433138	19.008	ng/u1	99
29) 2,4-Dichlorophenol	10.345	162	336902	20.295	ng/u1	99
30) Naphthalene	10.751	128	1145136	20.059	ng/u1	99
32) 4-Chloroaniline	10.863	127	468039	18.731	ng/u1	98
33) Hexachlorobutadiene	11.034	225	204440	20.807	ng/u1	97
34) Caprolactam	11.651	113	109035	18.559	ng/u1	98
35) 4-Chloro-3-methylphenol	11.986	107	325393	18.487	ng/u1	100
36) 2-Methylnaphthalene	12.357	142	764936	19.187	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011884.D
 Acq On : 03 Oct 2022 22:30
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Oct 04 00:40:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	763762	19.092	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.728	216	394634	21.181	ng/u1	99
40) Hexachlorocyclopentadiene	12.704	237	201408	18.923	ng/u1	98
41) 2,4,6-Trichlorophenol	12.969	196	254216	20.384	ng/u1	99
42) 2,4,5-Trichlorophenol	13.039	196	279616	20.582	ng/u1	99
43) 1,1'-Biphenyl	13.369	154	974936	20.517	ng/u1	99
44) 2-Chloronaphthalene	13.410	162	785956	20.920	ng/u1	99
45) 2-Nitroaniline	13.616	65	177098	19.311	ng/u1	99
47) Dimethylphthalate	13.992	163	918208	19.071	ng/u1	99
48) 2,6-Dinitrotoluene	14.110	165	175845	18.740	ng/u1	96
50) Acenaphthylene	14.257	152	1183906	20.353	ng/u1	100
51) 3-Nitroaniline	14.445	138	200577	18.442	ng/u1	98
52) Acenaphthene	14.598	153	834159	20.326	ng/u1	99
53) 2,4-Dinitrophenol	14.645	184	118493	18.626	ng/u1	98
55) 4-Nitrophenol	14.751	109	120353	18.740	ng/u1	96
56) Dibenzofuran	14.933	168	1142232	20.321	ng/u1	98
57) 2,4-Dinitrotoluene	14.898	165	262876	19.166	ng/u1	100
58) 2,3,4,6-Tetrachlorophenol	15.163	232	231451	19.611	ng/u1	97
59) Diethylphthalate	15.357	149	902596	18.712	ng/u1	100
61) Fluorene	15.586	166	923316	20.056	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.580	204	448681	19.687	ng/u1	98
63) 4-Nitroaniline	15.610	138	204418	19.697	ng/u1	98
66) 4,6-Dinitro-2-methylph...	15.663	198	151125	17.720	ng/u1	97
67) N-Nitrosodiphenylamine	15.792	169	786326	20.041	ng/u1	99
68) 4-Bromophenyl-phenylether	16.474	248	267662	19.784	ng/u1	98
69) Hexachlorobenzene	16.592	284	305403	19.837	ng/u1	98
70) Atrazine	16.751	200	261822	18.494	ng/u1	100
71) Pentachlorophenol	16.939	266	187680	18.535	ng/u1	99
72) Phenanthrene	17.333	178	1480678	20.257	ng/u1	99
74) Anthracene	17.421	178	1493978	20.314	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.333	216	391732	21.338	ng/uL	99
76) Pentachlorobenzene	14.851	250	407413	20.768	ng/uL	97
77) Carbazole	17.698	167	1385178	20.846	ng/u1	100
78) Di-n-butylphthalate	18.245	149	1514877	18.616	ng/u1	100
80) Fluoranthene	19.298	202	1718774	20.137	ng/u1	100
82) Pyrene	19.651	202	1820039	20.480	ng/u1	99
83) Butylbenzylphthalate	20.521	149	706397	18.392	ng/u1	95
84) 3,3'-Dichlorobenzidine	21.298	252	600604	18.944	ng/u1	99
85) Benzo(a)anthracene	21.362	228	1813787	20.140	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.280	149	1052673	18.195	ng/u1	99
87) Chrysene	21.415	228	1689569	19.991	ng/u1	99
89) Di-n-octyl phthalate	22.215	149	1791028	19.771	ng/u1	100
90) Benzo(b)fluoranthene	23.062	252	1842555	20.332	ng/u1	100
91) Benzo(k)fluoranthene	23.115	252	1854148	21.246	ng/u1	99
93) Benzo(a)pyrene	23.703	252	1663512	20.569	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.344	276	2021434	19.637	ng/u1	99
95) Dibenzo(a,h)anthracene	26.362	278	1812046	19.816	ng/u1	99
96) Benzo(g,h,i)perylene	27.121	276	1773436	19.433	ng/u1	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011884.D
 Acq On : 03 Oct 2022 22:30
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Oct 04 00:40:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.925 ng/uL
RT: 3.340 min Scan# 60
Delta R.T. -0.000 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 88 Resp: 50212
Ion Ratio Lower Upper
88 100
43 19.2 15.6 23.4
58 64.1 48.6 72.8



#5
Pyridine
Concen: 19.245 ng/uL
RT: 3.746 min Scan# 129
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion: 79 Resp: 320944
Ion Ratio Lower Upper
79 100
52 60.0 49.7 74.5
51 26.5 21.6 32.4









#6
Benzaldehyde
Concen: 21.518 ng/ul
RT: 7.028 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:	77	Resp:	210524
Ion	Ratio	Lower	Upper
77	100		
105	103.9	87.0	130.6
106	101.9	83.8	125.6



#8
Phenol
Concen: 18.258 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. -0.000 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:	94	Resp:	422943
Ion	Ratio	Lower	Upper
94	100		
65	24.1	19.5	29.3
66	30.3	25.0	37.4









#10
Bis(2-Chloroethyl)ether
Concen: 18.325 ng/ul
RT: 7.310 min Scan# 736
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 93 Resp: 337773
Ion Ratio Lower Upper
93 100
63 67.6 56.2 84.4
95 33.0 25.9 38.9



#12
2-Chlorophenol
Concen: 19.354 ng/ul
RT: 7.451 min Scan# 759
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:128 Resp: 364805
Ion Ratio Lower Upper
128 100
64 37.7 31.1 46.7
130 32.8 25.5 38.3









#13
2-Methylphenol
Concen: 17.639 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 321934
Ion Ratio Lower Upper
108 100
107 89.1 70.2 105.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.653 ng/ul
RT: 8.422 min Scan# 924
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion: 45 Resp: 320785
Ion Ratio Lower Upper
45 100
77 20.1 16.0 24.0
79 15.6 12.5 18.7









#16
Acetophenone
Concen: 17.692 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:105 Resp: 498023
Ion Ratio Lower Upper
105 100
77 70.3 55.4 83.0
51 19.5 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 16.980 ng/ul
RT: 8.704 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion: 70 Resp: 224959
Ion Ratio Lower Upper
70 100
42 39.6 32.7 49.1
101 10.2 8.4 12.6
130 24.6 19.4 29.2









#18
4-Methylphenol
Concen: 17.426 ng/ul
RT: 8.663 min Scan# 966
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 350412
Ion Ratio Lower Upper
108 100
107 111.7 88.0 132.0



#19
Hexachloroethane
Concen: 19.640 ng/ul
RT: 8.969 min Scan# 1017
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:117 Resp: 138072
Ion Ratio Lower Upper
117 100
201 100.1 82.4 123.6
199 62.4 51.6 77.4









#22
Nitrobenzene
Concen: 20.068 ng/ul
RT: 9.098 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 349416
Ion Ratio Lower Upper
77 100
123 52.4 41.6 62.4
65 12.4 9.7 14.5



#23
Isophorone
Concen: 17.927 ng/ul
RT: 9.628 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion: 82 Resp: 633193
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 20.2 15.8 23.8









#25
2-Nitrophenol
Concen: 19.927 ng/ul
RT: 9.810 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:139	Resp: 197126
Ion Ratio	Lower Upper
139 100	
65 39.7	32.2 48.4
109 24.9	19.6 29.4



#26
2,4-Dimethylphenol
Concen: 19.643 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:107	Resp: 365043
Ion Ratio	Lower Upper
107 100	
121 53.0	43.0 64.6
122 89.4	74.9 112.3









#27
Bis(2-Chloroethoxy)methane
Concen: 19.008 ng/ul
RT: 10.110 min Scan# 1211
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 93 Resp: 433138
Ion Ratio Lower Upper
93 100
95 32.4 26.3 39.5
123 11.9 9.5 14.3



#29
2,4-Dichlorophenol
Concen: 20.295 ng/ul
RT: 10.345 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:162 Resp: 336902
Ion Ratio Lower Upper
162 100
164 65.6 52.2 78.4
98 31.7 27.0 40.4









#30
Naphthalene
Concen: 20.059 ng/ul
RT: 10.751 min Scan# 1320
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:128 Resp: 1145136
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 12.8 10.5 15.7



#32
4-Chloroaniline
Concen: 18.731 ng/ul
RT: 10.863 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:127 Resp: 468039
Ion Ratio Lower Upper
127 100
129 32.1 26.6 39.8









#33
Hexachlorobutadiene
Concen: 20.807 ng/ul
RT: 11.034 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:225 Resp: 204440
Ion Ratio Lower Upper
225 100
223 61.9 52.1 78.1
227 64.2 52.3 78.5



#34
Caprolactam
Concen: 18.559 ng/ul
RT: 11.651 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:113 Resp: 109035
Ion Ratio Lower Upper
113 100
55 135.6 106.0 159.0
56 99.2 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 18.487 ng/ul
RT: 11.986 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:107 Resp: 325393
Ion Ratio Lower Upper
107 100
144 28.3 22.5 33.7
142 87.9 70.0 105.0



#36
2-Methylnaphthalene
Concen: 19.187 ng/ul
RT: 12.357 min Scan# 1593
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:142 Resp: 764936
Ion Ratio Lower Upper
142 100
141 89.5 70.6 106.0









#37
1-Methylnaphthalene
Concen: 19.092 ng/ul
RT: 12.575 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:142 Resp: 763762
Ion Ratio Lower Upper
142 100
141 92.9 74.8 112.2
116 3.6 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.181 ng/ul
RT: 12.728 min Scan# 1656
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:216 Resp: 394634
Ion Ratio Lower Upper
216 100
214 78.8 62.6 94.0
179 19.8 16.6 24.8
108 15.2 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 18.923 ng/ul

RT: 12.704 min Scan# 1652

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:237 Resp: 201408

Ion Ratio Lower Upper

237 100

235 61.9 51.0 76.6

272 13.1 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 20.384 ng/ul

RT: 12.969 min Scan# 1697

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion:196 Resp: 254216


Ion Ratio Lower Upper

196 100

198 97.3 78.7 118.1

200 31.2 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 20.582 ng/ul
RT: 13.039 min Scan# 1709
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:	196	Resp:	279616
Ion Ratio	Lower	Upper	
196	100		
198	96.2	76.6	114.8
200	30.3	25.6	38.4



#43
1,1'-Biphenyl
Concen: 20.517 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:	154	Resp:	974936
Ion Ratio	Lower	Upper	
154	100		
153	40.3	31.8	47.6
76	11.9	9.5	14.3





#44

2-Chloronaphthalene

Concen: 20.920 ng/ul

RT: 13.410 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:162 Resp: 785956

Ion Ratio Lower Upper

162 100

164 32.2 26.5 39.7

127 37.1 29.7 44.5



#45

2-Nitroaniline

Concen: 19.311 ng/ul

RT: 13.616 min Scan# 1807

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion: 65 Resp: 177098


Ion Ratio Lower Upper

65 100

92 77.0 61.2 91.8

138 126.4 103.0 154.4









#47
Dimethylphthalate
Concen: 19.071 ng/ul
RT: 13.992 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:163 Resp: 918208
Ion Ratio Lower Upper
163 100
194 3.9 3.3 4.9
164 9.9 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 18.740 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:165 Resp: 175845
Ion Ratio Lower Upper
165 100
89 52.0 39.1 58.7
121 21.7 16.7 25.1









#50
Acenaphthylene
Concen: 20.353 ng/ul
RT: 14.257 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:152 Resp: 1183906
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 13.3 10.8 16.2



#51
3-Nitroaniline
Concen: 18.442 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:138 Resp: 200577
Ion Ratio Lower Upper
138 100
108 10.4 8.2 12.4
92 105.2 85.6 128.4









#52
Acenaphthene
Concen: 20.326 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:153 Resp: 834159
Ion Ratio Lower Upper
153 100
152 47.3 38.4 57.6
154 87.4 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 18.626 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:184 Resp: 118493
Ion Ratio Lower Upper
184 100
63 53.4 44.2 66.4
154 57.8 45.8 68.6









#55
4-Nitrophenol
Concen: 18.740 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:109 Resp: 120353
Ion Ratio Lower Upper
109 100
139 163.5 125.0 187.6
65 129.5 101.9 152.9



#56
Dibenzofuran
Concen: 20.321 ng/ul
RT: 14.933 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:168 Resp: 1142232
Ion Ratio Lower Upper
168 100
139 35.4 29.4 44.2









#57
2,4-Dinitrotoluene
Concen: 19.166 ng/ul
RT: 14.898 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:165 Resp: 262876
Ion Ratio Lower Upper
165 100
63 35.2 28.2 42.2
182 2.6 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.611 ng/ul
RT: 15.163 min Scan# 2070
Delta R.T. -0.000 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:232 Resp: 231451
Ion Ratio Lower Upper
232 100
131 42.0 36.2 54.2
130 2.1 1.7 2.5
166 30.5 25.0 37.4









#59
Diethylphthalate
Concen: 18.712 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:149 Resp: 902596
Ion Ratio Lower Upper
149 100
177 22.0 17.8 26.8
150 12.4 9.9 14.9



#61
Fluorene
Concen: 20.056 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:166 Resp: 923316
Ion Ratio Lower Upper
166 100
165 99.1 77.2 115.8
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 19.687 ng/ul
RT: 15.580 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:204 Resp: 448681
Ion Ratio Lower Upper
204 100
206 31.4 26.8 40.2
141 55.9 44.5 66.7



#63
4-Nitroaniline
Concen: 19.697 ng/ul
RT: 15.610 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:138 Resp: 204418
Ion Ratio Lower Upper
138 100
92 47.3 35.9 53.9
108 65.0 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 17.720 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:198 Resp: 151125
Ion Ratio Lower Upper
198 100
182 4.7 3.7 5.5
77 22.3 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 20.041 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:169 Resp: 786326
Ion Ratio Lower Upper
169 100
168 67.8 53.7 80.5
167 36.9 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 19.784 ng/ul
RT: 16.474 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:248 Resp: 267662
Ion Ratio Lower Upper
248 100
250 97.2 78.6 118.0
141 72.8 56.0 84.0



#69
Hexachlorobenzene
Concen: 19.837 ng/ul
RT: 16.592 min Scan# 2313
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:284 Resp: 305403
Ion Ratio Lower Upper
284 100
142 34.7 27.3 40.9
249 30.4 25.2 37.8





#70

Atrazine

Concen: 18.494 ng/ul

RT: 16.751 min Scan# 2341

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:200 Resp: 261822

Ion Ratio Lower Upper

200 100

173 26.5 21.6 32.4

215 48.6 38.9 58.3



#71

Pentachlorophenol

Concen: 18.535 ng/ul

RT: 16.939 min Scan# 2372

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion:266 Resp: 187680


Ion Ratio Lower Upper

266 100

264 62.9 49.8 74.8

268 64.7 51.9 77.9





#72

Phenanthrene

Concen: 20.257 ng/ul

RT: 17.333 min Scan# 2440

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:178 Resp: 1480678

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 19.6 15.4 23.2



#74

Anthracene

Concen: 20.314 ng/ul

RT: 17.421 min Scan# 2454

Delta R.T. -0.012 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion:178 Resp: 1493978


Ion Ratio Lower Upper

178 100

179 15.7 12.3 18.5

176 18.8 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.338 ng/uL
RT: 13.333 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:216 Resp: 391732
Ion Ratio Lower Upper
216 100
214 77.7 62.6 93.8
218 48.3 37.4 56.0



#76
Pentachlorobenzene
Concen: 20.768 ng/uL
RT: 14.851 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:250 Resp: 407413
Ion Ratio Lower Upper
250 100
248 60.8 51.4 77.0
252 62.4 51.5 77.3









#77
Carbazole
Concen: 20.846 ng/ul
RT: 17.698 min Scan# 2501
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:167 Resp: 1385178
Ion Ratio Lower Upper
167 100
166 21.3 17.1 25.7
139 12.5 10.1 15.1



#78
Di-n-butylphthalate
Concen: 18.616 ng/ul
RT: 18.245 min Scan# 2594
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:149 Resp: 1514877
Ion Ratio Lower Upper
149 100
150 9.3 7.5 11.3
104 4.6 3.6 5.4









#80
Fluoranthene
Concen: 20.137 ng/ul
RT: 19.298 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:202 Resp: 1718774
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.6
100 9.1 7.4 11.0



#82
Pyrene
Concen: 20.480 ng/ul
RT: 19.651 min Scan# 2833
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:202 Resp: 1820039
Ion Ratio Lower Upper
202 100
101 13.9 11.6 17.4
100 11.6 9.4 14.0





#83

Butylbenzylphthalate

Concen: 18.392 ng/ul

RT: 20.521 min Scan# 2982

Delta R.T. -0.012 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:149 Resp: 706397

Ion Ratio Lower Upper

149 100

91 67.0 50.2 75.4

206 20.0 16.9 25.3



#84

3,3'-Dichlorobenzidine

Concen: 18.944 ng/ul

RT: 21.298 min Scan# 3113

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion:252 Resp: 600604


Ion Ratio Lower Upper

252 100

254 63.7 51.4 77.0

126 10.7 8.9 13.3





#85

Benzo(a)anthracene

Concen: 20.140 ng/ul

RT: 21.362 min Scan# 3110

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 1813787

Ion Ratio Lower Upper

228 100

229 19.5 15.8 23.6

226 27.3 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.195 ng/ul

RT: 21.280 min Scan# 3110

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion:149 Resp: 1052673


Ion Ratio Lower Upper

149 100

167 27.2 21.2 31.8

279 3.9 2.9 4.3





#87

Chrysene

Concen: 19.991 ng/ul

RT: 21.415 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 1689569

Ion Ratio Lower Upper

228 100

226 30.7 25.3 37.9

229 19.9 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 19.771 ng/ul




RT: 22.215 min Scan# 3269


Delta R.T. -0.012 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Tgt Ion:149 Resp: 1791028









#90
Benzo(b)fluoranthene
Concen: 20.332 ng/ul
RT: 23.062 min Scan# 3414
Delta R.T. -0.012 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:252 Resp: 1842555
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.5 9.2 13.8



#91
Benzo(k)fluoranthene
Concen: 21.246 ng/ul
RT: 23.115 min Scan# 3422
Delta R.T. -0.006 min
Lab File: BP011884.D
Acq: 03 Oct 2022 22:30

Tgt Ion:252 Resp: 1854148
Ion Ratio Lower Upper
252 100
253 22.2 18.0 27.0
125 10.9 9.2 13.8





#93

Benzo(a)pyrene

Concen: 20.569 ng/ul

RT: 23.703 min Scan# 3522

Delta R.T. -0.012 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:252 Resp: 1663512

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 11.7 9.8 14.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.637 ng/ul

RT: 26.344 min Scan# 3971

Delta R.T. -0.018 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30




Tgt Ion:276 Resp: 2021434


Ion Ratio Lower Upper

276 100

138 27.2 21.4 32.0

277 24.8 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 19.816 ng/ul

RT: 26.362 min Scan# 3974

Delta R.T. -0.018 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:278 Resp: 1812046

Ion Ratio Lower Upper

278 100

139 18.3 15.0 22.6

279 24.1 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 19.433 ng/ul

RT: 27.121 min Scan# 4103

Delta R.T. -0.024 min

Lab File: BP011884.D

Acq: 03 Oct 2022 22:30

Tgt Ion:276 Resp: 1773436

Ion Ratio Lower Upper

276 100

138 24.6 19.8 29.6

277 23.9 18.6 27.8

