

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011889.D
 Acq On : 04 Oct 2022 01:20
 Operator : CG/JU
 Sample : PB147997BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS997

Quant Time: Oct 04 03:56:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	247156	20.000	ng/u1	0.00
20) Naphthalene-d8	10.699	136	1017452	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	606098	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1262136	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	1337321	20.000	ng/u1	-0.01
88) Perylene-d12	23.810	264	1352169	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	37951	6.772	ng/uL	0.00
4) Pyridine-d5	3.722	84	482804	29.412	ng/u1	-0.01
7) Phenol-d5	7.052	99	682907	32.006	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	378254	31.982	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.416	132	570878	33.893	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	527825	29.925	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	268230	34.611	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	288203	34.759	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	535035	34.887	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	723057	30.926	ng/u1	-0.01
46) Dimethylphthalate-d6	13.945	166	1443748	33.310	ng/u1	0.00
49) Acenaphthylene-d8	14.228	160	1728375	34.980	ng/u1	0.00
54) 4-Nitrophenol-d4	14.734	143	309943	34.533	ng/u1	0.00
60) Fluorene-d10	15.528	176	1289925	34.270	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	249371	32.643	ng/u1	0.00
73) Anthracene-d10	17.392	188	1989800	34.610	ng/u1	0.00
81) Pyrene-d10	19.622	212	2407975	35.554	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.651	264	2529492	37.370	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	77197	12.754	ng/uL	97
5) Pyridine	3.740	79	489372	30.716	ng/u1	99
6) Benzaldehyde	7.028	77	369485m	39.531	ng/u1	
8) Phenol	7.081	94	656237	29.653	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.311	93	519485	29.500	ng/u1	98
12) 2-Chlorophenol	7.452	128	562754	31.252	ng/u1	100
13) 2-Methylphenol	8.334	108	493663	28.313	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.422	45	495525	28.544	ng/u1	100
16) Acetophenone	8.716	105	750010	27.890	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.705	70	335598	26.515	ng/u1	99
18) 4-Methylphenol	8.663	108	533330	27.762	ng/u1	100
19) Hexachloroethane	8.969	117	210299	31.313	ng/u1	100
22) Nitrobenzene	9.099	77	537099	32.225	ng/u1	99
23) Isophorone	9.628	82	978764	28.948	ng/u1	99
25) 2-Nitrophenol	9.810	139	302481	31.942	ng/u1	99
26) 2,4-Dimethylphenol	9.869	107	474964	26.699	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.110	93	656626	30.101	ng/u1	99
29) 2,4-Dichlorophenol	10.346	162	512891	32.276	ng/u1	99
30) Naphthalene	10.746	128	1725270	31.570	ng/u1	99
32) 4-Chloroaniline	10.863	127	669030	27.970	ng/u1	98
33) Hexachlorobutadiene	11.034	225	303694	32.288	ng/u1	99
34) Caprolactam	11.651	113	148212	26.353	ng/u1	97
35) 4-Chloro-3-methylphenol	11.987	107	505180	29.983	ng/u1	100
36) 2-Methylnaphthalene	12.357	142	1150497	30.146	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011889.D
 Acq On : 04 Oct 2022 01:20
 Operator : CG/JU
 Sample : PB147997BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS997

Quant Time: Oct 04 03:56:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	1161010	30.318	ng/u1	98
39) 1,2,4,5-Tetrachloroben...	12.722	216	590850	33.615	ng/u1	100
40) Hexachlorocyclopentadiene	12.704	237	247915	24.689	ng/u1	97
41) 2,4,6-Trichlorophenol	12.963	196	387422	32.927	ng/u1	98
42) 2,4,5-Trichlorophenol	13.040	196	428002	33.394	ng/u1	99
43) 1,1'-Biphenyl	13.369	154	1449703	32.337	ng/u1	100
44) 2-Chloronaphthalene	13.410	162	1173456	33.108	ng/u1	99
45) 2-Nitroaniline	13.616	65	282729	32.679	ng/u1	98
47) Dimethylphthalate	13.993	163	1400585	30.834	ng/u1	99
48) 2,6-Dinitrotoluene	14.116	165	281913	31.846	ng/u1	98
50) Acenaphthylene	14.257	152	1778498	32.409	ng/u1	99
51) 3-Nitroaniline	14.445	138	319163	31.106	ng/u1	96
52) Acenaphthene	14.598	153	1248495	32.247	ng/u1	99
53) 2,4-Dinitrophenol	14.645	184	161397	26.892	ng/u1	97
55) 4-Nitrophenol	14.751	109	194346	32.076	ng/u1	96
56) Dibenzofuran	14.934	168	1700045	32.059	ng/u1	98
57) 2,4-Dinitrotoluene	14.898	165	415437	32.106	ng/u1	98
58) 2,3,4,6-Tetrachlorophenol	15.157	232	355657	31.943	ng/u1	99
59) Diethylphthalate	15.357	149	1373226	30.177	ng/u1	100
61) Fluorene	15.587	166	1389353	31.988	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.581	204	676970	31.485	ng/u1	98
63) 4-Nitroaniline	15.610	138	340217	34.747	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.663	198	245134	30.289	ng/u1	100
67) N-Nitrosodiphenylamine	15.792	169	1180167	31.698	ng/u1	99
68) 4-Bromophenyl-phenylether	16.475	248	404174	31.482	ng/u1	98
69) Hexachlorobenzene	16.586	284	468153	32.045	ng/u1	98
70) Atrazine	16.751	200	391756	29.161	ng/u1	100
71) Pentachlorophenol	16.939	266	296958	30.905	ng/u1	99
72) Phenanthrene	17.334	178	2269886	32.725	ng/u1	100
74) Anthracene	17.422	178	2239422	32.089	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.334	216	618060	35.478	ng/uL	98
76) Pentachlorobenzene	14.851	250	547463	29.410	ng/uL	98
77) Carbazole	17.698	167	2147585	34.059	ng/u1	100
78) Di-n-butylphthalate	18.245	149	2384550	30.881	ng/u1	99
80) Fluoranthene	19.298	202	2681896	32.357	ng/u1	99
82) Pyrene	19.651	202	2855125	33.084	ng/u1	98
83) Butylbenzylphthalate	20.522	149	1139089	30.540	ng/u1	98
84) 3,3'-Dichlorobenzidine	21.292	252	984927	31.992	ng/u1	99
85) Benzo(a)anthracene	21.363	228	2853978	32.634	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.280	149	1663327	29.606	ng/u1	98
87) Chrysene	21.416	228	2668154	32.510	ng/u1	98
89) Di-n-octyl phthalate	22.216	149	2900895	33.626	ng/u1	100
90) Benzo(b)fluoranthene	23.063	252	2915836	33.785	ng/u1	100
91) Benzo(k)fluoranthene	23.116	252	2948527	35.478	ng/u1	99
93) Benzo(a)pyrene	23.704	252	2644753	34.339	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.345	276	3201338	32.656	ng/u1	99
95) Dibenzo(a,h)anthracene	26.362	278	2900428	33.305	ng/u1	99
96) Benzo(g,h,i)perylene	27.127	276	2826287	32.519	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011889.D
 Acq On : 04 Oct 2022 01:20
 Operator : CG/JU
 Sample : PB147997BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS997

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022


Quant Time: Oct 04 03:56:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 12.754 ng/uL
RT: 3.334 min Scan# 59
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion: 88 Resp: 77197
Ion Ratio Lower Upper
88 100
43 19.5 15.6 23.4
58 63.9 48.6 72.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#5
Pyridine
Concen: 30.716 ng/uL
RT: 3.740 min Scan# 128
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion: 79 Resp: 489372
Ion Ratio Lower Upper
79 100
52 61.0 49.7 74.5
51 26.6 21.6 32.4





#6
Benzaldehyde
Concen: 39.531 ng/ul m
RT: 7.028 min Scan# 687
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument : BNA_P
Client SampleId : SLCS997



Tgt Ion: 77 Resp: 369485
Ion Ratio Lower Upper
77 100
105 107.1 87.0 130.6
106 103.8 83.8 125.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#8
Phenol
Concen: 29.653 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. 0.000 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion: 94 Resp: 656237
Ion Ratio Lower Upper
94 100
65 24.4 19.5 29.3
66 31.4 25.0 37.4





#10
Bis(2-Chloroethyl)ether
Concen: 29.500 ng/ul
RT: 7.311 min Scan# 735
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion: 93 Resp: 519485
Ion Ratio Lower Upper
93 100
63 67.8 56.2 84.4
95 32.5 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#12
2-Chlorophenol
Concen: 31.252 ng/ul
RT: 7.452 min Scan# 759
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion:128 Resp: 562754
Ion Ratio Lower Upper
128 100
64 38.9 31.1 46.7
130 32.4 25.5 38.3





#13
2-Methylphenol
Concen: 28.313 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion: 108 Resp: 493663
Ion Ratio Lower Upper
108 100
107 87.8 70.2 105.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.544 ng/ul
RT: 8.422 min Scan# 924
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion: 45 Resp: 495525
Ion Ratio Lower Upper
45 100
77 20.2 16.0 24.0
79 15.5 12.5 18.7





#16
Acetophenone
Concen: 27.890 ng/ul
RT: 8.716 min Scan# 974
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion: 105 Resp: 750010
Ion Ratio Lower Upper
105 100
77 71.6 55.4 83.0
51 19.3 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#17
N-Nitroso-di-n-propylamine
Concen: 26.515 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion: 70 Resp: 335598
Ion Ratio Lower Upper
70 100
42 41.1 32.7 49.1
101 11.0 8.4 12.6
130 24.6 19.4 29.2





#18
4-Methylphenol
Concen: 27.762 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:108 Resp: 533330
Ion Ratio Lower Upper
108 100
107 109.8 88.0 132.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#19
Hexachloroethane
Concen: 31.313 ng/ul
RT: 8.969 min Scan# 1017
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion:117 Resp: 210299
Ion Ratio Lower Upper
117 100
201 102.8 82.4 123.6
199 64.0 51.6 77.4





#22
Nitrobenzene
Concen: 32.225 ng/ul
RT: 9.099 min Scan# 1039
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion: 77 Resp: 537099
Ion Ratio Lower Upper
77 100
123 52.6 41.6 62.4
65 12.8 9.7 14.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#23
Isophorone
Concen: 28.948 ng/ul
RT: 9.628 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20



Tgt Ion: 82 Resp: 978764
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.2
138 19.6 15.8 23.8









#25
2-Nitrophenol
Concen: 31.942 ng/ul
RT: 9.810 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Instrument :
BNA_P
ClientSampleId :
SLCS997

Tgt Ion:139 Resp: 302481
Ion Ratio Lower Upper
139 100
65 40.6 32.2 48.4
109 25.2 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#26
2,4-Dimethylphenol
Concen: 26.699 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:107 Resp: 474964
Ion Ratio Lower Upper
107 100
121 53.3 43.0 64.6
122 92.5 74.9 112.3









#27
Bis(2-Chloroethoxy)methane
Concen: 30.101 ng/ul
RT: 10.110 min Scan# 1211
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Instrument :
BNA_P
ClientSampleId :
SLCS997

Tgt Ion: 93 Resp: 656626
Ion Ratio Lower Upper
93 100
95 32.5 26.3 39.5
123 11.8 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#29
2,4-Dichlorophenol
Concen: 32.276 ng/ul
RT: 10.346 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:162 Resp: 512891
Ion Ratio Lower Upper
162 100
164 66.1 52.2 78.4
98 33.3 27.0 40.4





#30
Naphthalene
Concen: 31.570 ng/ul
RT: 10.746 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:128 Resp: 1725270
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 12.7 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#32
4-Chloroaniline
Concen: 27.970 ng/ul
RT: 10.863 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:127 Resp: 669030
Ion Ratio Lower Upper
127 100
129 32.3 26.6 39.8





#33
Hexachlorobutadiene
Concen: 32.288 ng/ul
RT: 11.034 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:225 Resp: 303694
Ion Ratio Lower Upper
225 100
223 63.1 52.1 78.1
227 65.1 52.3 78.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#34
Caprolactam
Concen: 26.353 ng/ul
RT: 11.651 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:113 Resp: 148212
Ion Ratio Lower Upper
113 100
55 135.0 106.0 159.0
56 103.0 78.7 118.1





#35
4-Chloro-3-methylphenol
Concen: 29.983 ng/ul
RT: 11.987 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:107 Resp: 505180
Ion Ratio Lower Upper
107 100
144 27.6 22.5 33.7
142 87.2 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#36
2-Methylnaphthalene
Concen: 30.146 ng/ul
RT: 12.357 min Scan# 1593
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:142 Resp: 1150497
Ion Ratio Lower Upper
142 100
141 88.2 70.6 106.0





#37
1-Methylnaphthalene
Concen: 30.318 ng/ul
RT: 12.575 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:142 Resp: 1161010
Ion Ratio Lower Upper
142 100
141 91.6 74.8 112.2
116 3.6 2.9 4.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.615 ng/ul
RT: 12.722 min Scan# 1655
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion:216 Resp: 590850
Ion Ratio Lower Upper
216 100
214 78.3 62.6 94.0
179 20.8 16.6 24.8
108 16.3 12.6 19.0





#40
Hexachlorocyclopentadiene
Concen: 24.689 ng/ul
RT: 12.704 min Scan# 1652
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:237 Resp: 247915
Ion Ratio Lower Upper
237 100
235 61.4 51.0 76.6
272 12.6 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#41
2,4,6-Trichlorophenol
Concen: 32.927 ng/ul
RT: 12.963 min Scan# 1696
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20



Tgt Ion:196 Resp: 387422
Ion Ratio Lower Upper
196 100
198 96.0 78.7 118.1
200 31.0 24.5 36.7





#42

2,4,5-Trichlorophenol

Concen: 33.394 ng/ul

RT: 13.040 min Scan# 1709

Delta R.T. -0.006 min

Lab File: BP011889.D

Acq: 04 Oct 2022 01:20

Instrument :

BNA_P

ClientSampleId :

SLCS997

Tgt Ion:196 Resp: 428002

Ion Ratio Lower Upper

196 100





198 96.4 76.6 114.8

200 30.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#43

1,1'-Biphenyl

Concen: 32.337 ng/ul

RT: 13.369 min Scan# 1765

Delta R.T. -0.006 min

Lab File: BP011889.D

Acq: 04 Oct 2022 01:20




Tgt Ion:154 Resp: 1449703


Ion Ratio Lower Upper

154 100

153 39.6 31.8 47.6


76 11.8 9.5 14.3





#44
2-Chloronaphthalene
Concen: 33.108 ng/ul
RT: 13.410 min Scan# 1772
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:162 Resp: 1173456
Ion Ratio Lower Upper
162 100
164 33.6 26.5 39.7
127 36.7 29.7 44.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#45
2-Nitroaniline
Concen: 32.679 ng/ul
RT: 13.616 min Scan# 1807
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion: 65 Resp: 282729
Ion Ratio Lower Upper
65 100
92 75.8 61.2 91.8
138 124.7 103.0 154.4





#47
Dimethylphthalate
Concen: 30.834 ng/ul
RT: 13.993 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:163 Resp: 1400585
Ion Ratio Lower Upper
163 100
194 3.9 3.3 4.9
164 10.0 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#48
2,6-Dinitrotoluene
Concen: 31.846 ng/ul
RT: 14.116 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:165 Resp: 281913
Ion Ratio Lower Upper
165 100
89 47.2 39.1 58.7
121 20.2 16.7 25.1





#50
Acenaphthylene
Concen: 32.409 ng/ul
RT: 14.257 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:152 Resp: 1778498
Ion Ratio Lower Upper
152 100
151 19.6 16.0 24.0
153 13.2 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#51
3-Nitroaniline
Concen: 31.106 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:138 Resp: 319163
Ion Ratio Lower Upper
138 100
108 9.2 8.2 12.4
92 102.5 85.6 128.4





#52
Acenaphthene
Concen: 32.247 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:153 Resp: 1248495
Ion Ratio Lower Upper
153 100
152 46.7 38.4 57.6
154 88.2 70.6 105.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#53
2,4-Dinitrophenol
Concen: 26.892 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:184 Resp: 161397
Ion Ratio Lower Upper
184 100
63 57.1 44.2 66.4
154 59.2 45.8 68.6





#55
4-Nitrophenol
Concen: 32.076 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:109 Resp: 194346
Ion Ratio Lower Upper
109 100
139 164.1 125.0 187.6
65 128.2 101.9 152.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#56
Dibenzofuran
Concen: 32.059 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:168 Resp: 1700045
Ion Ratio Lower Upper
168 100
139 35.4 29.4 44.2





#57
2,4-Dinitrotoluene
Concen: 32.106 ng/ul
RT: 14.898 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:165 Resp: 415437
Ion Ratio Lower Upper
165 100
63 34.3 28.2 42.2
182 2.6 2.2 3.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 31.943 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion:232 Resp: 355657
Ion Ratio Lower Upper
232 100
131 45.8 36.2 54.2
130 2.2 1.7 2.5
166 32.3 25.0 37.4





#59
Diethylphthalate
Concen: 30.177 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:149 Resp: 1373226
Ion Ratio Lower Upper
149 100
177 22.4 17.8 26.8
150 12.5 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#61
Fluorene
Concen: 31.988 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:166 Resp: 1389353
Ion Ratio Lower Upper
166 100
165 99.2 77.2 115.8
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 31.485 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Instrument :
BNA_P
ClientSampleId :
SLCS997

Tgt Ion:204 Resp: 676970
Ion Ratio Lower Upper
204 100
206 32.7 26.8 40.2
141 54.2 44.5 66.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#63
4-Nitroaniline
Concen: 34.747 ng/ul
RT: 15.610 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:138 Resp: 340217
Ion Ratio Lower Upper
138 100
92 46.5 35.9 53.9
108 64.9 51.6 77.4





#66
4,6-Dinitro-2-methylphenol
Concen: 30.289 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:198 Resp: 245134
Ion Ratio Lower Upper
198 100
182 3.9 3.7 5.5
77 20.7 16.6 24.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#67
N-Nitrosodiphenylamine
Concen: 31.698 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:169 Resp: 1180167
Ion Ratio Lower Upper
169 100
168 67.5 53.7 80.5
167 36.9 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 31.482 ng/ul
RT: 16.475 min Scan# 2293
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Instrument :
BNA_P
ClientSampleId :
SLCS997

Tgt Ion:248 Resp: 404174
Ion Ratio Lower Upper
248 100
250 97.3 78.6 118.0
141 72.7 56.0 84.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#69
Hexachlorobenzene
Concen: 32.045 ng/ul
RT: 16.586 min Scan# 2312
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:284 Resp: 468153
Ion Ratio Lower Upper
284 100
142 36.2 27.3 40.9
249 31.2 25.2 37.8





#70

Atrazine

Concen: 29.161 ng/ul

RT: 16.751 min Scan# 2340

Delta R.T. -0.006 min

Lab File: BP011889.D

Acq: 04 Oct 2022 01:20

Instrument :

BNA_P

ClientSampleId :





SLCS997

Tgt Ion:200	Resp: 391756
Ion Ratio	Lower Upper
200 100	
173 26.8	21.6 32.4
215 48.3	38.9 58.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#71

Pentachlorophenol

Concen: 30.905 ng/ul




RT: 16.939 min Scan# 2372


Delta R.T. -0.006 min

Lab File: BP011889.D

Acq: 04 Oct 2022 01:20


Tgt Ion:266	Resp: 296958
Ion Ratio	Lower Upper
266 100	
264 63.2	49.8 74.8
268 64.3	51.9 77.9





#72
Phenanthrene
Concen: 32.725 ng/ul
RT: 17.334 min Scan# 2439
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument : BNA_P
ClientSampleId : SLCS997



Tgt Ion:178 Resp: 2269886
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 19.2 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#74
Anthracene
Concen: 32.089 ng/ul
RT: 17.422 min Scan# 2454
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:178 Resp: 2239422
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 18.9 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.478 ng/uL
RT: 13.334 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Instrument :
BNA_P
ClientSampleId :
SLCS997

Tgt Ion:216 Resp: 618060
Ion Ratio Lower Upper
216 100
214 78.4 62.6 93.8
218 49.2 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#76
Pentachlorobenzene
Concen: 29.410 ng/uL
RT: 14.851 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:250 Resp: 547463
Ion Ratio Lower Upper
250 100
248 62.7 51.4 77.0
252 63.3 51.5 77.3





#77
 Carbazole
 Concen: 34.059 ng/ul
 RT: 17.698 min Scan# 2501
 Delta R.T. -0.006 min
 Lab File: BP011889.D
 Acq: 04 Oct 2022 01:20




Instrument :
 BNA_P
 ClientSampleId :
 SLCS997




Tgt Ion:167 Resp: 2147585
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 12.6 10.1 15.1

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022




#78
 Di-n-butylphthalate
 Concen: 30.881 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. -0.006 min
 Lab File: BP011889.D
 Acq: 04 Oct 2022 01:20




Tgt Ion:149 Resp: 2384550
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 4.5 3.6 5.4





#80
Fluoranthene
Concen: 32.357 ng/u1
RT: 19.298 min Scan# 2773
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:202 Resp: 2681896
Ion Ratio Lower Upper
202 100
101 11.7 9.8 14.6
100 8.6 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#82
Pyrene
Concen: 33.084 ng/u1
RT: 19.651 min Scan# 2833
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:202 Resp: 2855125
Ion Ratio Lower Upper
202 100
101 13.8 11.6 17.4
100 11.0 9.4 14.0





#83
Butylbenzylphthalate
Concen: 30.540 ng/ul
RT: 20.522 min Scan# 2981
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:149 Resp: 1139089
Ion Ratio Lower Upper
149 100
91 63.9 50.2 75.4
206 20.2 16.9 25.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#84
3,3'-Dichlorobenzidine
Concen: 31.992 ng/ul
RT: 21.292 min Scan# 3112
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion:252 Resp: 984927
Ion Ratio Lower Upper
252 100
254 63.9 51.4 77.0
126 11.8 8.9 13.3





#85
Benzo(a)anthracene
Concen: 32.634 ng/ul
RT: 21.363 min Scan# 3124
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997




Tgt Ion:228 Resp: 2853978
Ion Ratio Lower Upper
228 100
229 20.0 15.8 23.6
226 27.9 22.0 33.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 29.606 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Tgt Ion:149 Resp: 1663327
Ion Ratio Lower Upper
149 100
167 27.4 21.2 31.8
279 4.1 2.9 4.3





#87
Chrysene
Concen: 32.510 ng/ul
RT: 21.416 min Scan# 3133
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:228 Resp: 2668154
Ion Ratio Lower Upper
228 100
226 29.9 25.3 37.9
229 20.0 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#89
Di-n-octyl phthalate
Concen: 33.626 ng/ul
RT: 22.216 min Scan# 3269
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20


Tgt Ion:149 Resp: 2900895





#90
Benzo(b)fluoranthene
Concen: 33.785 ng/ul
RT: 23.063 min Scan# 3413
Delta R.T. -0.012 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:252 Resp: 2915836
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.8
125 11.4 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#91
Benzo(k)fluoranthene
Concen: 35.478 ng/ul
RT: 23.116 min Scan# 3422
Delta R.T. -0.006 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:252 Resp: 2948527
Ion Ratio Lower Upper
252 100
253 22.1 18.0 27.0
125 10.8 9.2 13.8





#93

Benzo(a)pyrene

Concen: 34.339 ng/ul

RT: 23.704 min Scan# 3522

Delta R.T. -0.012 min

Lab File: BP011889.D

Acq: 04 Oct 2022 01:20

Instrument :

BNA_P

ClientSampleId :

SLCS997

Tgt Ion:252 Resp: 264475

Ion Ratio Lower Upper

252 100





253 22.2 17.7 26.5

125 11.8 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 32.656 ng/ul

RT: 26.345 min Scan# 3971

Delta R.T. -0.018 min

Lab File: BP011889.D

Acq: 04 Oct 2022 01:20




Tgt Ion:276 Resp: 3201338


Ion Ratio Lower Upper

276 100

138 27.2 21.4 32.0


277 24.9 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 33.305 ng/ul
RT: 26.362 min Scan# 3974
Delta R.T. -0.018 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20




Instrument :
BNA_P
ClientSampleId :
SLCS997



Tgt Ion:278 Resp: 2900428
Ion Ratio Lower Upper
278 100
139 17.8 15.0 22.6
279 23.9 19.0 28.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#96
Benzo(g,h,i)perylene
Concen: 32.519 ng/ul
RT: 27.127 min Scan# 4104
Delta R.T. -0.018 min
Lab File: BP011889.D
Acq: 04 Oct 2022 01:20

Tgt Ion:276 Resp: 2826287
Ion Ratio Lower Upper
276 100
138 24.1 19.8 29.6
277 23.7 18.6 27.8

