

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011891.D
 Acq On : 04 Oct 2022 02:28
 Operator : CG/JU
 Sample : PB147998BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS998

Quant Time: Oct 04 04:22:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	263249	20.000	ng/u1	0.00
20) Naphthalene-d8	10.699	136	1036950	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	586342	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1187808	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	1350925	20.000	ng/u1	-0.01
88) Perylene-d12	23.810	264	1383721	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	37944	6.357	ng/uL	0.00
4) Pyridine-d5	3.723	84	468536	26.798	ng/u1	-0.01
7) Phenol-d5	7.052	99	644767	28.371	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	364917	28.968	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.416	132	540829	30.146	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	488514	26.003	ng/u1	0.00
21) Nitrobenzene-d5	9.052	128	252498	31.969	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.775	143	264476	31.298	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.316	165	487630	31.198	ng/u1	0.00
31) 4-Chloroaniline-d4	10.840	131	651975	27.361	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	1246213	29.722	ng/u1	0.00
49) Acenaphthylene-d8	14.228	160	1525707	31.919	ng/u1	0.00
54) 4-Nitrophenol-d4	14.734	143	265993	30.635	ng/u1	0.00
60) Fluorene-d10	15.528	176	1126761	30.943	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	206404	28.709	ng/u1	0.00
73) Anthracene-d10	17.387	188	1698612	31.394	ng/u1	-0.01
81) Pyrene-d10	19.622	212	2137598	31.244	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.651	264	2342434	33.817	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	75796	11.757	ng/uL	95
5) Pyridine	3.740	79	473149	27.883	ng/u1	97
6) Benzaldehyde	7.028	77	347103m	34.866	ng/u1	
8) Phenol	7.075	94	617868	26.212	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.311	93	497024	26.499	ng/u1	98
12) 2-Chlorophenol	7.452	128	532394	27.758	ng/u1	99
13) 2-Methylphenol	8.334	108	459649	24.751	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.422	45	472975	25.580	ng/u1	99
16) Acetophenone	8.716	105	707291	24.693	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.705	70	313391	23.246	ng/u1	99
18) 4-Methylphenol	8.663	108	491240	24.008	ng/u1	100
19) Hexachloroethane	8.969	117	203363	28.429	ng/u1	99
22) Nitrobenzene	9.093	77	500926	29.489	ng/u1	98
23) Isophorone	9.628	82	891037	25.858	ng/u1	99
25) 2-Nitrophenol	9.810	139	280565	29.070	ng/u1	99
26) 2,4-Dimethylphenol	9.869	107	427717	23.591	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.110	93	608134	27.354	ng/u1	99
29) 2,4-Dichlorophenol	10.346	162	467465	28.864	ng/u1	99
30) Naphthalene	10.746	128	1609068	28.890	ng/u1	100
32) 4-Chloroaniline	10.863	127	606097	24.863	ng/u1	98
33) Hexachlorobutadiene	11.028	225	288982	30.146	ng/u1	97
34) Caprolactam	11.652	113	123638	21.570	ng/u1	95
35) 4-Chloro-3-methylphenol	11.987	107	442040	25.742	ng/u1	99
36) 2-Methylnaphthalene	12.357	142	1055007	27.124	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011891.D
 Acq On : 04 Oct 2022 02:28
 Operator : CG/JU
 Sample : PB147998BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS998

Quant Time: Oct 04 04:22:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	1058122	27.111	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.722	216	537785	31.627	ng/u1	99
40) Hexachlorocyclopentadiene	12.698	237	232101	23.893	ng/u1	99
41) 2,4,6-Trichlorophenol	12.963	196	339428	29.820	ng/u1	97
42) 2,4,5-Trichlorophenol	13.034	196	373478	30.122	ng/u1	99
43) 1,1'-Biphenyl	13.369	154	1309383	30.192	ng/u1	100
44) 2-Chloronaphthalene	13.410	162	1048675	30.584	ng/u1	100
45) 2-Nitroaniline	13.616	65	239223	28.582	ng/u1	99
47) Dimethylphthalate	13.993	163	1210873	27.556	ng/u1	98
48) 2,6-Dinitrotoluene	14.110	165	238608	27.862	ng/u1	97
50) Acenaphthylene	14.257	152	1565160	29.483	ng/u1	99
51) 3-Nitroaniline	14.445	138	269342	27.135	ng/u1	96
52) Acenaphthene	14.598	153	1097663	29.306	ng/u1	99
53) 2,4-Dinitrophenol	14.645	184	136729	23.549	ng/u1	98
55) 4-Nitrophenol	14.751	109	167164	28.520	ng/u1	97
56) Dibenzofuran	14.934	168	1495611	29.154	ng/u1	98
57) 2,4-Dinitrotoluene	14.898	165	345992	27.640	ng/u1	98
58) 2,3,4,6-Tetrachlorophenol	15.157	232	304230	28.245	ng/u1	99
59) Diethylphthalate	15.357	149	1178567	26.772	ng/u1	100
61) Fluorene	15.587	166	1208855	28.770	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.581	204	586413	28.192	ng/u1	99
63) 4-Nitroaniline	15.610	138	283878	29.970	ng/u1	98
66) 4,6-Dinitro-2-methylph...	15.663	198	207508	27.245	ng/u1	99
67) N-Nitrosodiphenylamine	15.792	169	1024958	29.252	ng/u1	99
68) 4-Bromophenyl-phenylether	16.475	248	345883	28.628	ng/u1	98
69) Hexachlorobenzene	16.587	284	400908	29.159	ng/u1	98
70) Atrazine	16.751	200	331573	26.226	ng/u1	99
71) Pentachlorophenol	16.939	266	250693	27.723	ng/u1	97
72) Phenanthrene	17.334	178	1959120	30.012	ng/u1	100
74) Anthracene	17.422	178	1942125	29.570	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.328	216	554598	33.827	ng/uL	99
76) Pentachlorobenzene	14.851	250	480506	27.428	ng/uL	98
77) Carbazole	17.698	167	1865679	31.439	ng/u1	100
78) Di-n-butylphthalate	18.245	149	2010354	27.664	ng/u1	99
80) Fluoranthene	19.298	202	2378372	28.406	ng/u1	99
82) Pyrene	19.651	202	2538873	29.123	ng/u1	99
83) Butylbenzylphthalate	20.522	149	1012416	26.871	ng/u1	98
84) 3,3'-Dichlorobenzidine	21.292	252	895469	28.793	ng/u1	98
85) Benzo(a)anthracene	21.357	228	2641395	29.899	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.280	149	1484333	26.154	ng/u1	98
87) Chrysene	21.410	228	2498017	30.130	ng/u1	99
89) Di-n-octyl phthalate	22.216	149	2605275	29.511	ng/u1	100
90) Benzo(b)fluoranthene	23.063	252	2721923	30.819	ng/u1	99
91) Benzo(k)fluoranthene	23.116	252	2734377	32.151	ng/u1	99
93) Benzo(a)pyrene	23.698	252	2455456	31.154	ng/u1	100
94) Indeno(1,2,3-cd)pyrene	26.345	276	3007117	29.975	ng/u1	99
95) Dibenzo(a,h)anthracene	26.363	278	2722628	30.551	ng/u1	98
96) Benzo(g,h,i)perylene	27.127	276	2639749	29.680	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011891.D
 Acq On : 04 Oct 2022 02:28
 Operator : CG/JU
 Sample : PB147998BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :

BNA_P

ClientSampleId :


SLCS998


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022


Quant Time: Oct 04 04:22:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 11.757 ng/uL
RT: 3.334 min Scan# 59
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion: 88 Resp: 75796
Ion Ratio Lower Upper
88 100
43 21.4 15.6 23.4
58 65.0 48.6 72.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#5
Pyridine
Concen: 27.883 ng/uL
RT: 3.740 min Scan# 128
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion: 79 Resp: 473149
Ion Ratio Lower Upper
79 100
52 59.6 49.7 74.5
51 26.3 21.6 32.4





#6
Benzaldehyde
Concen: 34.866 ng/ul m
RT: 7.028 min Scan# 687
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument : BNA_P
ClientSampleId : SLCS998



Tgt Ion: 77 Resp: 347103
Ion Ratio Lower Upper
77 100
105 107.6 87.0 130.6
106 106.5 83.8 125.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#8
Phenol
Concen: 26.212 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion: 94 Resp: 617868
Ion Ratio Lower Upper
94 100
65 24.3 19.5 29.3
66 31.9 25.0 37.4





#10
Bis(2-Chloroethyl)ether
Concen: 26.499 ng/ul
RT: 7.311 min Scan# 735
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion: 93 Resp: 497024
Ion Ratio Lower Upper
93 100
63 68.3 56.2 84.4
95 31.9 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#12
2-Chlorophenol
Concen: 27.758 ng/ul
RT: 7.452 min Scan# 759
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Tgt Ion:128 Resp: 532394
Ion Ratio Lower Upper
128 100
64 38.3 31.1 46.7
130 32.6 25.5 38.3





#13
2-Methylphenol
Concen: 24.751 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion:108 Resp: 459649
Ion Ratio Lower Upper
108 100
107 87.4 70.2 105.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.580 ng/ul
RT: 8.422 min Scan# 924
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Tgt Ion: 45 Resp: 472975
Ion Ratio Lower Upper
45 100
77 19.1 16.0 24.0
79 15.3 12.5 18.7





#16
Acetophenone
Concen: 24.693 ng/ul
RT: 8.716 min Scan# 911
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion: 105 Resp: 707291
Ion Ratio Lower Upper
105 100
77 69.3 55.4 83.0
51 19.3 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#17
N-Nitroso-di-n-propylamine
Concen: 23.246 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28



Tgt Ion: 70 Resp: 313391
Ion Ratio Lower Upper
70 100
42 40.4 32.7 49.1
101 10.7 8.4 12.6
130 25.3 19.4 29.2









#18
4-Methylphenol
Concen: 24.008 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:108 Resp: 491240
Ion Ratio Lower Upper
108 100
107 110.4 88.0 132.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#19
Hexachloroethane
Concen: 28.429 ng/ul
RT: 8.969 min Scan# 1017
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:117 Resp: 203363
Ion Ratio Lower Upper
117 100
201 104.5 82.4 123.6
199 65.3 51.6 77.4





#22
Nitrobenzene
Concen: 29.489 ng/ul
RT: 9.093 min Scan# 1038
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion: 77 Resp: 500926
Ion Ratio Lower Upper
77 100
123 50.5 41.6 62.4
65 11.8 9.7 14.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#23
Isophorone
Concen: 25.858 ng/ul
RT: 9.628 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28



Tgt Ion: 82 Resp: 891037
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.2
138 20.1 15.8 23.8









#25
2-Nitrophenol
Concen: 29.070 ng/ul
RT: 9.810 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:139 Resp: 28056
Ion Ratio Lower Upper
139 100
65 40.2 32.2 48.4
109 25.4 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#26
2,4-Dimethylphenol
Concen: 23.591 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:107 Resp: 427717
Ion Ratio Lower Upper
107 100
121 54.2 43.0 64.6
122 93.6 74.9 112.3









#27
Bis(2-Chloroethoxy)methane
Concen: 27.354 ng/ul
RT: 10.110 min Scan# 1211
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion: 93 Resp: 608134
Ion Ratio Lower Upper
93 100
95 32.2 26.3 39.5
123 12.2 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#29
2,4-Dichlorophenol
Concen: 28.864 ng/ul
RT: 10.346 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:162 Resp: 467465
Ion Ratio Lower Upper
162 100
164 65.0 52.2 78.4
98 32.7 27.0 40.4





#30
Naphthalene
Concen: 28.890 ng/ul
RT: 10.746 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:128 Resp: 1609068
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.0 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#32
4-Chloroaniline
Concen: 24.863 ng/ul
RT: 10.863 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:127 Resp: 606097
Ion Ratio Lower Upper
127 100
129 31.9 26.6 39.8





#33
Hexachlorobutadiene
Concen: 30.146 ng/ul
RT: 11.028 min Scan# 1368
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:225 Resp: 288982
Ion Ratio Lower Upper
225 100
223 61.9 52.1 78.1
227 63.7 52.3 78.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#34
Caprolactam
Concen: 21.570 ng/ul
RT: 11.652 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:113 Resp: 123638
Ion Ratio Lower Upper
113 100
55 138.5 106.0 159.0
56 102.7 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 25.742 ng/ul
RT: 11.987 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:107 Resp: 442040
Ion Ratio Lower Upper
107 100
144 28.3 22.5 33.7
142 88.0 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#36
2-Methylnaphthalene
Concen: 27.124 ng/ul
RT: 12.357 min Scan# 1593
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:142 Resp: 1055007
Ion Ratio Lower Upper
142 100
141 88.5 70.6 106.0





#37
1-Methylnaphthalene
Concen: 27.111 ng/ul
RT: 12.575 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion:142 Resp: 105812
Ion Ratio Lower Upper
142 100
141 92.2 74.8 112.2
116 3.5 2.9 4.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.627 ng/ul
RT: 12.722 min Scan# 1655
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Tgt Ion:216 Resp: 537785
Ion Ratio Lower Upper
216 100
214 77.4 62.6 94.0
179 21.0 16.6 24.8
108 16.5 12.6 19.0





#40
Hexachlorocyclopentadiene
Concen: 23.893 ng/ul
RT: 12.698 min Scan# 1652
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion:237 Resp: 232101
Ion Ratio Lower Upper
237 100
235 63.1 51.0 76.6
272 13.0 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#41
2,4,6-Trichlorophenol
Concen: 29.820 ng/ul
RT: 12.963 min Scan# 1696
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Tgt Ion:196 Resp: 339428
Ion Ratio Lower Upper
196 100
198 95.3 78.7 118.1
200 30.8 24.5 36.7





#42
2,4,5-Trichlorophenol
Concen: 30.122 ng/ul
RT: 13.034 min Scan# 1708
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion:196 Resp: 373478
Ion Ratio Lower Upper
196 100
198 95.6 76.6 114.8
200 30.8 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#43
1,1'-Biphenyl
Concen: 30.192 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Tgt Ion:154 Resp: 1309383
Ion Ratio Lower Upper
154 100
153 39.6 31.8 47.6
76 11.7 9.5 14.3





#44
2-Chloronaphthalene
Concen: 30.584 ng/ul
RT: 13.410 min Scan# 1772
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument : BNA_P
ClientSampleId : SLCS998



Tgt Ion:162 Resp: 1048675
Ion Ratio Lower Upper
162 100
164 33.1 26.5 39.7
127 36.9 29.7 44.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#45
2-Nitroaniline
Concen: 28.582 ng/ul
RT: 13.616 min Scan# 1807
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion: 65 Resp: 239223
Ion Ratio Lower Upper
65 100
92 76.3 61.2 91.8
138 130.6 103.0 154.4





#47
Dimethylphthalate
Concen: 27.556 ng/ul
RT: 13.993 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:163 Resp: 1210873
Ion Ratio Lower Upper
163 100
194 3.8 3.3 4.9
164 9.9 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#48
2,6-Dinitrotoluene
Concen: 27.862 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:165 Resp: 238608
Ion Ratio Lower Upper
165 100
89 52.2 39.1 58.7
121 20.8 16.7 25.1





#50
Acenaphthylene
Concen: 29.483 ng/ul
RT: 14.257 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion:152 Resp: 1565160
Ion Ratio Lower Upper
152 100
151 19.9 16.0 24.0
153 13.1 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#51
3-Nitroaniline
Concen: 27.135 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Tgt Ion:138 Resp: 269342
Ion Ratio Lower Upper
138 100
108 9.9 8.2 12.4
92 102.1 85.6 128.4





#52
Acenaphthene
Concen: 29.306 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:153 Resp: 1097663
Ion Ratio Lower Upper
153 100
152 47.4 38.4 57.6
154 88.5 70.6 105.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#53
2,4-Dinitrophenol
Concen: 23.549 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:184 Resp: 136729
Ion Ratio Lower Upper
184 100
63 53.7 44.2 66.4
154 58.0 45.8 68.6





#55
4-Nitrophenol
Concen: 28.520 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:109 Resp: 167164
Ion Ratio Lower Upper
109 100
139 162.6 125.0 187.6
65 127.1 101.9 152.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#56
Dibenzofuran
Concen: 29.154 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:168 Resp: 1495611
Ion Ratio Lower Upper
168 100
139 35.6 29.4 44.2





#57
2,4-Dinitrotoluene
Concen: 27.640 ng/ul
RT: 14.898 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:165 Resp: 345992
Ion Ratio Lower Upper
165 100
63 36.3 28.2 42.2
182 2.9 2.2 3.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 28.245 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:232 Resp: 304230
Ion Ratio Lower Upper
232 100
131 45.5 36.2 54.2
130 2.3 1.7 2.5
166 30.8 25.0 37.4









#59
Diethylphthalate
Concen: 26.772 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:149 Resp: 117856
Ion Ratio Lower Upper
149 100
177 22.2 17.8 26.8
150 12.3 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#61
Fluorene
Concen: 28.770 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:166 Resp: 1208855
Ion Ratio Lower Upper
166 100
165 97.2 77.2 115.8
167 12.8 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 28.192 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion: 204 Resp: 586413
Ion Ratio Lower Upper
204 100
206 32.3 26.8 40.2
141 56.1 44.5 66.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#63
4-Nitroaniline
Concen: 29.970 ng/ul
RT: 15.610 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion: 138 Resp: 283878
Ion Ratio Lower Upper
138 100
92 47.1 35.9 53.9
108 64.7 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 27.245 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:198 Resp: 207508
Ion Ratio Lower Upper
198 100
182 4.4 3.7 5.5
77 20.0 16.6 24.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#67
N-Nitrosodiphenylamine
Concen: 29.252 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:169 Resp: 1024958
Ion Ratio Lower Upper
169 100
168 67.6 53.7 80.5
167 36.6 29.9 44.9





#68
4-Bromophenyl-phenylether
Concen: 28.628 ng/ul
RT: 16.475 min Scan# 2293
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:248 Resp: 345883
Ion Ratio Lower Upper
248 100
250 99.9 78.6 118.0
141 72.9 56.0 84.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#69
Hexachlorobenzene
Concen: 29.159 ng/ul
RT: 16.587 min Scan# 2312
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:284 Resp: 400908
Ion Ratio Lower Upper
284 100
142 36.3 27.3 40.9
249 31.4 25.2 37.8





#70

Atrazine

Concen: 26.226 ng/ul

RT: 16.751 min Scan# 2340

Delta R.T. -0.006 min

Lab File: BP011891.D

Acq: 04 Oct 2022 02:28

Instrument :

BNA_P

ClientSampleId :





SLCS998

Tgt Ion:200	Resp:	331573
Ion Ratio	Lower	Upper
200	100	
173	26.7	21.6 32.4
215	49.9	38.9 58.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#71

Pentachlorophenol

Concen: 27.723 ng/ul




RT: 16.939 min Scan# 2372


Delta R.T. -0.006 min

Lab File: BP011891.D

Acq: 04 Oct 2022 02:28


Tgt Ion:266	Resp:	250693
Ion Ratio	Lower	Upper
266	100	
264	65.1	49.8 74.8
268	63.3	51.9 77.9





#72
Phenanthrene
Concen: 30.012 ng/ul
RT: 17.334 min Scan# 2439
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument : BNA_P
ClientSampleId : SLCS998



Tgt Ion:178 Resp: 1959120
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 19.2 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#74
Anthracene
Concen: 29.570 ng/ul
RT: 17.422 min Scan# 2454
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:178 Resp: 1942125
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.6 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 33.827 ng/uL
RT: 13.328 min Scan# 1758
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:216 Resp: 554598
Ion Ratio Lower Upper
216 100
214 78.3 62.6 93.8
218 47.4 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#76
Pentachlorobenzene
Concen: 27.428 ng/uL
RT: 14.851 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:250 Resp: 480506
Ion Ratio Lower Upper
250 100
248 61.8 51.4 77.0
252 63.3 51.5 77.3





#77
 Carbazole
 Concen: 31.439 ng/ul
 RT: 17.698 min Scan# 2501
 Delta R.T. -0.006 min
 Lab File: BP011891.D
 Acq: 04 Oct 2022 02:28




Instrument :
 BNA_P
 ClientSampleId :
 SLCS998




Tgt Ion:167 Resp: 1865679
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 12.4 10.1 15.1

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
 Supervised By :mohammad ahmed 10/06/2022




#78
 Di-n-butylphthalate
 Concen: 27.664 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. -0.006 min
 Lab File: BP011891.D
 Acq: 04 Oct 2022 02:28




Tgt Ion:149 Resp: 2010354
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.5 11.3
 104 4.4 3.6 5.4





#80
Fluoranthene
Concen: 28.406 ng/u1
RT: 19.298 min Scan# 2773
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:202 Resp: 2378372
Ion Ratio Lower Upper
202 100
101 11.8 9.8 14.6
100 8.8 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#82
Pyrene
Concen: 29.123 ng/u1
RT: 19.651 min Scan# 2833
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28


Tgt Ion:202 Resp: 2538873
Ion Ratio Lower Upper
202 100
101 14.0 11.6 17.4
100 11.4 9.4 14.0





#83
Butylbenzylphthalate
Concen: 26.871 ng/ul
RT: 20.522 min Scan# 2981
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998




Tgt Ion:149 Resp: 1012416
Ion Ratio Lower Upper
149 100
91 64.0 50.2 75.4
206 19.6 16.9 25.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022




#84
3,3'-Dichlorobenzidine
Concen: 28.793 ng/ul
RT: 21.292 min Scan# 3112
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28



Tgt Ion:252 Resp: 895469
Ion Ratio Lower Upper
252 100
254 65.6 51.4 77.0
126 12.0 8.9 13.3









#85
Benzo(a)anthracene
Concen: 29.899 ng/ul
RT: 21.357 min Scan# 3112
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :
BNA_P
ClientSampleId :
SLCS998

Tgt Ion:228 Resp: 2641395
Ion Ratio Lower Upper
228 100
229 20.3 15.8 23.6
226 28.0 22.0 33.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 26.154 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:149 Resp: 1484333
Ion Ratio Lower Upper
149 100
167 27.5 21.2 31.8
279 4.1 2.9 4.3





#87

Chrysene

Concen: 30.130 ng/ul

RT: 21.410 min Scan# 3133

Delta R.T. -0.012 min

Lab File: BP011891.D

Acq: 04 Oct 2022 02:28

Instrument :

BNA_P

ClientSampleId :

SLCS998

Tgt Ion:228 Resp: 249801

Ion Ratio Lower Upper

228 100





226 30.5 25.3 37.9

229 19.9 16.2 24.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#89

Di-n-octyl phthalate

Concen: 29.511 ng/ul




RT: 22.216 min Scan# 3269


Delta R.T. -0.012 min

Lab File: BP011891.D

Acq: 04 Oct 2022 02:28


Tgt Ion:149 Resp: 2605275





#90
Benzo(b)fluoranthene
Concen: 30.819 ng/ul
RT: 23.063 min Scan# 3413
Delta R.T. -0.012 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28




Instrument :
BNA_P
ClientSampleId :
SLCS998



Tgt Ion:252 Resp: 2721923
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 11.3 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#91
Benzo(k)fluoranthene
Concen: 32.151 ng/ul
RT: 23.116 min Scan# 3422
Delta R.T. -0.006 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Tgt Ion:252 Resp: 2734377
Ion Ratio Lower Upper
252 100
253 22.0 18.0 27.0
125 11.0 9.2 13.8





#93

Benzo(a)pyrene

Concen: 31.154 ng/ul

RT: 23.698 min Scan# 3522

Delta R.T. -0.018 min

Lab File: BP011891.D

Acq: 04 Oct 2022 02:28

Instrument :

BNA_P

ClientSampleId :

SLCS998

Tgt Ion:252 Resp: 2455456

Ion Ratio Lower Upper

252 100





253 22.0 17.7 26.5

125 12.5 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.975 ng/ul

RT: 26.345 min Scan# 3971

Delta R.T. -0.018 min

Lab File: BP011891.D

Acq: 04 Oct 2022 02:28




Tgt Ion:276 Resp: 3007117


Ion Ratio Lower Upper

276 100

138 27.4 21.4 32.0

277 25.0 19.9 29.9






#95
Dibenzo(a,h)anthracene
Concen: 30.551 ng/ul
RT: 26.363 min Scan# 3974
Delta R.T. -0.018 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28

Instrument :

BNA_P

ClientSampleId :

SLCS998






Tgt Ion:278 Resp: 2722628
Ion Ratio Lower Upper
278 100
139 17.8 15.0 22.6
279 24.3 19.0 28.6


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/04/2022
Supervised By :mohammad ahmed 10/06/2022



#96
Benzo(g,h,i)perylene
Concen: 29.680 ng/ul
RT: 27.127 min Scan# 4104
Delta R.T. -0.018 min
Lab File: BP011891.D
Acq: 04 Oct 2022 02:28



Tgt Ion:276 Resp: 2639749
Ion Ratio Lower Upper
276 100
138 24.1 19.8 29.6
277 24.0 18.6 27.8

