

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011902.D
 Acq On : 04 Oct 2022 08:43
 Operator : CG/JU
 Sample : N4873-08MSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EXZ54MSD

Quant Time: Oct 05 04:24:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	253074	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	994079	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.534	164	558828	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1130379	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	1298770	20.000	ng/u1	0.00
88) Perylene-d12	23.815	264	1339760	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	25839	4.503	ng/uL	0.00
4) Pyridine-d5	3.728	84	137577	8.185	ng/u1	0.00
7) Phenol-d5	7.052	99	127046	5.815	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	360001	29.727	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.416	132	402119	23.316	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	228798	12.668	ng/u1	0.00
21) Nitrobenzene-d5	9.052	128	257644	34.027	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.775	143	256732	31.692	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.316	165	435493	29.064	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	570632	24.980	ng/u1	-0.01
46) Dimethylphthalate-d6	13.945	166	1467996	36.735	ng/u1	0.00
49) Acenaphthylene-d8	14.228	160	1593369	34.975	ng/u1	0.00
54) 4-Nitrophenol-d4	14.734	143	62785	7.587	ng/u1	0.00
60) Fluorene-d10	15.528	176	1264720	36.442	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	258070	37.719	ng/u1	0.00
73) Anthracene-d10	17.392	188	2085977	40.511	ng/u1	0.00
81) Pyrene-d10	19.627	212	2725156	41.431	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.657	264	2953554	44.039	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	32336	5.217	ng/uL	95
5) Pyridine	3.746	79	154227	9.454	ng/u1	99
6) Benzaldehyde	7.028	77	394099	41.178	ng/u1	99
8) Phenol	7.075	94	141112	6.227	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.310	93	511250	28.354	ng/u1	97
12) 2-Chlorophenol	7.446	128	418441	22.694	ng/u1	99
13) 2-Methylphenol	8.334	108	275304	15.420	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.416	45	477426	26.858	ng/u1	99
16) Acetophenone	8.716	105	742475	26.964	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.699	70	333076	25.700	ng/u1	98
18) 4-Methylphenol	8.663	108	251619	12.791	ng/u1	96
19) Hexachloroethane	8.969	117	183975	26.752	ng/u1	99
22) Nitrobenzene	9.093	77	527934	32.420	ng/u1	99
23) Isophorone	9.622	82	941536	28.501	ng/u1	100
25) 2-Nitrophenol	9.804	139	278804	30.134	ng/u1	98
26) 2,4-Dimethylphenol	9.869	107	305451	17.574	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.104	93	639145	29.989	ng/u1	100
29) 2,4-Dichlorophenol	10.340	162	435070	28.022	ng/u1	99
30) Naphthalene	10.746	128	1629517	30.519	ng/u1	100
32) 4-Chloroaniline	10.857	127	563247	24.101	ng/u1	98
33) Hexachlorobutadiene	11.028	225	274420	29.861	ng/u1	98
34) Caprolactam	11.657	113	16382	2.981	ng/u1	94
35) 4-Chloro-3-methylphenol	11.981	107	386000	23.448	ng/u1	99
36) 2-Methylnaphthalene	12.357	142	1081745	29.011	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011902.D
 Acq On : 04 Oct 2022 08:43
 Operator : CG/JU
 Sample : N4873-08MSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EXZ54MSD

Quant Time: Oct 05 04:24:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	1101435	29.438	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.722	216	553459	34.151	ng/ul	99
40) Hexachlorocyclopentadiene	12.704	237	197168	21.297	ng/ul	98
41) 2,4,6-Trichlorophenol	12.963	196	358393	33.037	ng/ul	97
42) 2,4,5-Trichlorophenol	13.034	196	397248	33.617	ng/ul	99
43) 1,1'-Biphenyl	13.369	154	1379203	33.367	ng/ul	100
44) 2-Chloronaphthalene	13.410	162	1106183	33.850	ng/ul	99
45) 2-Nitroaniline	13.616	65	292918	36.720	ng/ul	99
47) Dimethylphthalate	13.992	163	1489999	35.577	ng/ul	99
48) 2,6-Dinitrotoluene	14.110	165	304653	37.326	ng/ul	96
50) Acenaphthylene	14.257	152	1705184	33.702	ng/ul	100
51) 3-Nitroaniline	14.445	138	322079	34.045	ng/ul	94
52) Acenaphthene	14.598	153	1215258	34.044	ng/ul	99
53) 2,4-Dinitrophenol	14.645	184	174542	31.542	ng/ul	99
55) 4-Nitrophenol	14.751	109	41993	7.517	ng/ul	98
56) Dibenzofuran	14.934	168	1692158	34.609	ng/ul	97
57) 2,4-Dinitrotoluene	14.898	165	456585	38.271	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.157	232	366306	35.682	ng/ul	99
59) Diethylphthalate	15.357	149	1495537	35.644	ng/ul	100
61) Fluorene	15.586	166	1429875	35.706	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.581	204	681908	34.397	ng/ul	99
63) 4-Nitroaniline	15.610	138	363256	40.239	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.663	198	266450	36.761	ng/ul	99
67) N-Nitrosodiphenylamine	15.792	169	1260579	37.804	ng/ul	99
68) 4-Bromophenyl-phenylether	16.475	248	430005	37.399	ng/ul	98
69) Hexachlorobenzene	16.586	284	509963	38.976	ng/ul	97
70) Atrazine	16.751	200	437280	36.344	ng/ul	98
71) Pentachlorophenol	16.939	266	315096	36.615	ng/ul	100
72) Phenanthrene	17.333	178	2507481	40.364	ng/ul	100
74) Anthracene	17.428	178	2461132	39.376	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.328	216	579348	37.132	ng/ul	99
76) Pentachlorobenzene	14.851	250	544160	32.639	ng/ul	99
77) Carbazole	17.698	167	2454323	43.460	ng/ul	99
78) Di-n-butylphthalate	18.245	149	2707000	39.142	ng/ul	100
80) Fluoranthene	19.298	202	3135517	38.952	ng/ul	100
82) Pyrene	19.657	202	3340758	39.861	ng/ul	97
83) Butylbenzylphthalate	20.521	149	1386092	38.266	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.298	252	1145660	38.317	ng/ul	98
85) Benzo(a)anthracene	21.363	228	3460338	40.742	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.280	149	2033360	37.267	ng/ul	98
87) Chrysene	21.416	228	3228963	40.511	ng/ul	99
89) Di-n-octyl phthalate	22.215	149	3616114	42.305	ng/ul	100
90) Benzo(b)fluoranthene	23.068	252	3674811	42.974	ng/ul	100
91) Benzo(k)fluoranthene	23.115	252	3598442	43.699	ng/ul	100
93) Benzo(a)pyrene	23.704	252	3226288	42.277	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.351	276	3975274	40.926	ng/ul	100
95) Dibenzo(a,h)anthracene	26.374	278	3593034	41.640	ng/ul	99
96) Benzo(g,h,i)perylene	27.133	276	3525072	40.935	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 5.217 ng/uL
RT: 3.334 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion: 88 Resp: 32336
Ion Ratio Lower Upper
88 100
43 20.2 15.6 23.4
58 65.2 48.6 72.8



#5
Pyridine
Concen: 9.454 ng/uL
RT: 3.746 min Scan# 129
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion: 79 Resp: 154227
Ion Ratio Lower Upper
79 100
52 60.9 49.7 74.5
51 26.9 21.6 32.4









#6
Benzaldehyde
Concen: 41.178 ng/ul
RT: 7.028 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion: 77 Resp: 394099
Ion Ratio Lower Upper
77 100
105 108.7 87.0 130.6
106 103.5 83.8 125.6



#8
Phenol
Concen: 6.227 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion: 94 Resp: 141112
Ion Ratio Lower Upper
94 100
65 25.0 19.5 29.3
66 31.5 25.0 37.4









#10
Bis(2-Chloroethyl)ether
Concen: 28.354 ng/ul
RT: 7.310 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion: 93 Resp: 511250
Ion Ratio Lower Upper
93 100
63 66.7 56.2 84.4
95 32.5 25.9 38.9



#12
2-Chlorophenol
Concen: 22.694 ng/ul
RT: 7.446 min Scan# 758
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:128 Resp: 418441
Ion Ratio Lower Upper
128 100
64 39.7 31.1 46.7
130 32.7 25.5 38.3





#13

2-Methylphenol

Concen: 15.420 ng/ul

RT: 8.334 min Scan# 909

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :





EXZ54MSD

Tgt Ion:108 Resp: 275304

Ion Ratio Lower Upper

108 100

107 88.2 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 26.858 ng/ul

RT: 8.416 min Scan# 923

Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion: 45 Resp: 477426


Ion Ratio Lower Upper

45 100

77 20.6 16.0 24.0

79 16.0 12.5 18.7









#16
Acetophenone
Concen: 26.964 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:105 Resp: 742475
Ion Ratio Lower Upper
105 100
77 69.7 55.4 83.0
51 19.1 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 25.700 ng/ul
RT: 8.699 min Scan# 971
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion: 70 Resp: 333076
Ion Ratio Lower Upper
70 100
42 39.5 32.7 49.1
101 9.9 8.4 12.6
130 25.1 19.4 29.2









#18
4-Methylphenol
Concen: 12.791 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:108 Resp: 251619
Ion Ratio Lower Upper
108 100
107 113.7 88.0 132.0



#19
Hexachloroethane
Concen: 26.752 ng/ul
RT: 8.969 min Scan# 1017
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:117 Resp: 183975
Ion Ratio Lower Upper
117 100
201 103.9 82.4 123.6
199 66.4 51.6 77.4









#22
Nitrobenzene
Concen: 32.420 ng/ul
RT: 9.093 min Scan# 1039
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion: 77 Resp: 527934
Ion Ratio Lower Upper
77 100
123 52.5 41.6 62.4
65 12.1 9.7 14.5



#23
Isophorone
Concen: 28.501 ng/ul
RT: 9.622 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion: 82 Resp: 941536
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 19.6 15.8 23.8









#25
2-Nitrophenol
Concen: 30.134 ng/ul
RT: 9.804 min Scan# 1159
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:139 Resp: 278804
Ion Ratio Lower Upper
139 100
65 41.5 32.2 48.4
109 25.8 19.6 29.4



#26
2,4-Dimethylphenol
Concen: 17.574 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:107 Resp: 305451
Ion Ratio Lower Upper
107 100
121 52.7 43.0 64.6
122 90.2 74.9 112.3





#27

Bis(2-Chloroethoxy)methane

Concen: 29.989 ng/ul

RT: 10.104 min Scan# 11

Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion: 93 Resp: 639145

Ion Ratio Lower Upper

93 100

95 32.7 26.3 39.5

123 12.0 9.5 14.3



#29

2,4-Dichlorophenol

Concen: 28.022 ng/ul

RT: 10.340 min Scan# 1250

Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion:162 Resp: 435070


Ion Ratio Lower Upper

162 100

164 66.3 52.2 78.4

98 33.0 27.0 40.4









#30
Naphthalene
Concen: 30.519 ng/ul
RT: 10.746 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:128 Resp: 1629517
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 24.101 ng/ul
RT: 10.857 min Scan# 1338
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:127 Resp: 563247
Ion Ratio Lower Upper
127 100
129 32.3 26.6 39.8









#33
Hexachlorobutadiene
Concen: 29.861 ng/ul
RT: 11.028 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:225 Resp: 274420
Ion Ratio Lower Upper
225 100
223 64.9 52.1 78.1
227 63.1 52.3 78.5



#34
Caprolactam
Concen: 2.981 ng/ul
RT: 11.657 min Scan# 1474
Delta R.T. 0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:113 Resp: 16382
Ion Ratio Lower Upper
113 100
55 136.7 106.0 159.0
56 107.1 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 23.448 ng/ul
RT: 11.981 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:107 Resp: 386000
Ion Ratio Lower Upper
107 100
144 28.8 22.5 33.7
142 87.8 70.0 105.0



#36
2-Methylnaphthalene
Concen: 29.011 ng/ul
RT: 12.357 min Scan# 1593
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:142 Resp: 1081745
Ion Ratio Lower Upper
142 100
141 89.4 70.6 106.0









#37
1-Methylnaphthalene
Concen: 29.438 ng/ul
RT: 12.575 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:142 Resp: 1101435
Ion Ratio Lower Upper
142 100
141 93.3 74.8 112.2
116 3.4 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.151 ng/ul
RT: 12.722 min Scan# 1655
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:216 Resp: 553459
Ion Ratio Lower Upper
216 100
214 79.0 62.6 94.0
179 20.6 16.6 24.8
108 16.2 12.6 19.0









#40
Hexachlorocyclopentadiene
Concen: 21.297 ng/ul
RT: 12.704 min Scan# 1652
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:237 Resp: 197168
Ion Ratio Lower Upper
237 100
235 61.5 51.0 76.6
272 12.9 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 33.037 ng/ul
RT: 12.963 min Scan# 1696
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:196 Resp: 358393
Ion Ratio Lower Upper
196 100
198 94.7 78.7 118.1
200 30.2 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 33.617 ng/ul
RT: 13.034 min Scan# 1708
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:196 Resp: 397248
Ion Ratio Lower Upper
196 100
198 95.2 76.6 114.8
200 30.9 25.6 38.4



#43
1,1'-Biphenyl
Concen: 33.367 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:154 Resp: 1379203
Ion Ratio Lower Upper
154 100
153 39.9 31.8 47.6
76 11.7 9.5 14.3









#44
2-Chloronaphthalene
Concen: 33.850 ng/ul
RT: 13.410 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:162 Resp: 1106183
Ion Ratio Lower Upper
162 100
164 32.8 26.5 39.7
127 36.6 29.7 44.5



#45
2-Nitroaniline
Concen: 36.720 ng/ul
RT: 13.616 min Scan# 1807
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion: 65 Resp: 292918
Ion Ratio Lower Upper
65 100
92 77.4 61.2 91.8
138 126.9 103.0 154.4









#47
Dimethylphthalate
Concen: 35.577 ng/ul
RT: 13.992 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:163 Resp: 1489999
Ion Ratio Lower Upper
163 100
194 3.9 3.3 4.9
164 10.1 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 37.326 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:165 Resp: 304653
Ion Ratio Lower Upper
165 100
89 52.5 39.1 58.7
121 20.7 16.7 25.1









#50
Acenaphthylene
Concen: 33.702 ng/ul
RT: 14.257 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:152 Resp: 1705184
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 13.2 10.8 16.2



#51
3-Nitroaniline
Concen: 34.045 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:138 Resp: 322079
Ion Ratio Lower Upper
138 100
108 9.5 8.2 12.4
92 100.7 85.6 128.4









#52
Acenaphthene
Concen: 34.044 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:153 Resp: 1215258
Ion Ratio Lower Upper
153 100
152 47.8 38.4 57.6
154 89.6 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 31.542 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:184 Resp: 174542
Ion Ratio Lower Upper
184 100
63 53.9 44.2 66.4
154 56.7 45.8 68.6









#55
4-Nitrophenol
Concen: 7.517 ng/ul
RT: 14.751 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:109 Resp: 41993
Ion Ratio Lower Upper
109 100
139 158.1 125.0 187.6
65 130.4 101.9 152.9



#56
Dibenzofuran
Concen: 34.609 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:168 Resp: 1692158
Ion Ratio Lower Upper
168 100
139 35.2 29.4 44.2









#57
2,4-Dinitrotoluene
Concen: 38.271 ng/ul
RT: 14.898 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:165 Resp: 456585
Ion Ratio Lower Upper
165 100
63 35.2 28.2 42.2
182 2.6 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.682 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:232 Resp: 366306
Ion Ratio Lower Upper
232 100
131 44.7 36.2 54.2
130 2.4 1.7 2.5
166 31.4 25.0 37.4









#59
Diethylphthalate
Concen: 35.644 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:149 Resp: 1495537
Ion Ratio Lower Upper
149 100
177 22.3 17.8 26.8
150 12.4 9.9 14.9



#61
Fluorene
Concen: 35.706 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:166 Resp: 1429875
Ion Ratio Lower Upper
166 100
165 96.6 77.2 115.8
167 13.1 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 34.397 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:204 Resp: 681908
Ion Ratio Lower Upper
204 100
206 33.2 26.8 40.2
141 54.7 44.5 66.7



#63
4-Nitroaniline
Concen: 40.239 ng/ul
RT: 15.610 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:138 Resp: 363256
Ion Ratio Lower Upper
138 100
92 46.0 35.9 53.9
108 64.7 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 36.761 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:198 Resp: 266450
Ion Ratio Lower Upper
198 100
182 4.2 3.7 5.5
77 21.3 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 37.804 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:169 Resp: 1260579
Ion Ratio Lower Upper
169 100
168 67.8 53.7 80.5
167 37.4 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 37.399 ng/ul
RT: 16.475 min Scan# 211
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:248 Resp: 430005
Ion Ratio Lower Upper
248 100
250 97.1 78.6 118.0
141 72.4 56.0 84.0



#69
Hexachlorobenzene
Concen: 38.976 ng/ul
RT: 16.586 min Scan# 2312
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:284 Resp: 509963
Ion Ratio Lower Upper
284 100
142 36.2 27.3 40.9
249 30.2 25.2 37.8









#70
Atrazine
Concen: 36.344 ng/ul
RT: 16.751 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:200 Resp: 437280
Ion Ratio Lower Upper
200 100
173 27.1 21.6 32.4
215 46.5 38.9 58.3



#71
Pentachlorophenol
Concen: 36.615 ng/ul
RT: 16.939 min Scan# 2372
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:266 Resp: 315096
Ion Ratio Lower Upper
266 100
264 62.6 49.8 74.8
268 64.8 51.9 77.9









#72
Phenanthrene
Concen: 40.364 ng/ul
RT: 17.333 min Scan# 2439
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:178 Resp: 2507481
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 19.5 15.4 23.2



#74
Anthracene
Concen: 39.376 ng/ul
RT: 17.428 min Scan# 2455
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:178 Resp: 2461132
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.0 15.4 23.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 37.132 ng/uL

RT: 13.328 min Scan# 1759

Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion:216 Resp: 579348

Ion Ratio Lower Upper

216 100

214 78.2 62.6 93.8

218 47.8 37.4 56.0



#76

Pentachlorobenzene

Concen: 32.639 ng/uL

RT: 14.851 min Scan# 2017

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion:250 Resp: 544160


Ion Ratio Lower Upper

250 100

248 63.3 51.4 77.0

252 63.9 51.5 77.3





#77

Carbazole

Concen: 43.460 ng/ul

RT: 17.698 min Scan# 2501

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion:167 Resp: 2454323

Ion Ratio Lower Upper

167 100

166 21.9 17.1 25.7

139 12.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 39.142 ng/ul

RT: 18.245 min Scan# 2594

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion:149 Resp: 2707000


Ion Ratio Lower Upper

149 100

150 9.3 7.5 11.3

104 4.4 3.6 5.4









#80
Fluoranthene
Concen: 38.952 ng/ul
RT: 19.298 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:202 Resp: 3135517
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.6
100 9.3 7.4 11.0



#82
Pyrene
Concen: 39.861 ng/ul
RT: 19.657 min Scan# 2834
Delta R.T. -0.000 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:202 Resp: 3340758
Ion Ratio Lower Upper
202 100
101 13.2 11.6 17.4
100 10.5 9.4 14.0









#83
Butylbenzylphthalate
Concen: 38.266 ng/ul
RT: 20.521 min Scan# 2981
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43




Instrument :
BNA_P
ClientSampleId :
EXZ54MSD


Tgt Ion:149 Resp: 1386092
Ion Ratio Lower Upper
149 100
91 67.3 50.2 75.4
206 20.4 16.9 25.3



#84
3,3'-Dichlorobenzidine
Concen: 38.317 ng/ul
RT: 21.298 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:252 Resp: 1145660
Ion Ratio Lower Upper
252 100
254 65.8 51.4 77.0
126 11.4 8.9 13.3





#85

Benzo(a)anthracene

Concen: 40.742 ng/ul

RT: 21.363 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion:228 Resp: 3460338

Ion Ratio Lower Upper

228 100

229 20.0 15.8 23.6

226 28.1 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 37.267 ng/ul

RT: 21.280 min Scan# 3110

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion:149 Resp: 2033360


Ion Ratio Lower Upper

149 100

167 27.4 21.2 31.8

279 4.0 2.9 4.3





#87

Chrysene

Concen: 40.511 ng/ul

RT: 21.416 min Scan# 3133

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion:228 Resp: 3228963

Ion Ratio Lower Upper

228 100

226 30.9 25.3 37.9

229 20.1 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 42.305 ng/ul




RT: 22.215 min Scan# 3269


Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Tgt Ion:149 Resp: 3616114





#90

Benzo(b)fluoranthene

Concen: 42.974 ng/ul

RT: 23.068 min Scan# 3413

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion:252 Resp: 3674811

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.8

125 11.3 9.2 13.8



#91

Benzo(k)fluoranthene

Concen: 43.699 ng/ul

RT: 23.115 min Scan# 3422

Delta R.T. -0.006 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion:252 Resp: 3598442


Ion Ratio Lower Upper

252 100

253 22.5 18.0 27.0

125 11.3 9.2 13.8





#93

Benzo(a)pyrene

Concen: 42.277 ng/ul

RT: 23.704 min Scan# 3

Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43

Instrument :

BNA_P

ClientSampleId :

EXZ54MSD





Tgt Ion:252 Resp: 3226288

Ion Ratio Lower Upper

252 100

253 22.5 17.7 26.5

125 12.6 9.8 14.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 40.926 ng/ul

RT: 26.351 min Scan# 3972

Delta R.T. -0.012 min

Lab File: BP011902.D

Acq: 04 Oct 2022 08:43




Tgt Ion:276 Resp: 3975274


Ion Ratio Lower Upper

276 100

138 26.9 21.4 32.0

277 25.0 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 41.640 ng/ul
RT: 26.374 min Scan# 3974
Delta R.T. -0.006 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Instrument :
BNA_P
ClientSampleId :
EXZ54MSD

Tgt Ion:278 Resp: 3593034
Ion Ratio Lower Upper
278 100
139 17.6 15.0 22.6
279 24.1 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 40.935 ng/ul
RT: 27.133 min Scan# 4105
Delta R.T. -0.012 min
Lab File: BP011902.D
Acq: 04 Oct 2022 08:43

Tgt Ion:276 Resp: 3525072
Ion Ratio Lower Upper
276 100
138 24.3 19.8 29.6
277 24.5 18.6 27.8

