

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
 Data File : BP011916.D
 Acq On : 04 Oct 2022 17:17
 Operator : CG/JU
 Sample : N4873-18
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EXZ63

Quant Time: Oct 05 05:55:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 01 00:53:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	238890	20.000	ng/ul	0.00
20) Naphthalene-d8	10.693	136	957917	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.528	164	554252	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.286	188	1137879	20.000	ng/ul	-0.01
79) Chrysene-d12	21.374	240	1247144	20.000	ng/ul	-0.01
88) Perylene-d12	23.804	264	1314711	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	26832	4.954	ng/uL	0.00
4) Pyridine-d5	3.752	84	111856	7.050	ng/ul	0.02
7) Phenol-d5	7.052	99	123573	5.992	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	336975	29.478	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.416	132	395050	24.266	ng/ul	0.00
15) 4-Methylphenol-d8	8.593	113	231686	13.590	ng/ul	-0.01
21) Nitrobenzene-d5	9.052	128	234795	32.180	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.775	143	238224	30.517	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.310	165	407072	28.193	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.834	131	456926	20.758	ng/ul	-0.01
46) Dimethylphthalate-d6	13.940	166	1276283	32.201	ng/ul	-0.01
49) Acenaphthylene-d8	14.222	160	1494301	33.072	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.739	143	36606	4.460	ng/ul	0.00
60) Fluorene-d10	15.528	176	1167092	33.907	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	179344	26.040	ng/ul	0.00
73) Anthracene-d10	17.386	188	1979107	38.183	ng/ul	-0.01
81) Pyrene-d10	19.622	212	2510662	39.750	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.651	264	2713177	41.226	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100322\
Data File : BP011916.D
Acq On : 04 Oct 2022 17:17
Operator : CG/JU
Sample : N4873-18
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EXZ63

Quant Time: Oct 05 05:55:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 01 00:53:59 2022
Response via : Initial Calibration

