

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100423\
 Data File : BP017536.D
 Acq On : 04 Oct 2023 17:06
 Operator : MA/JU
 Sample : PB156005BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK005

Quant Time: Oct 05 00:38:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 04 16:33:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	72186	20.000	ng/ul	0.00
20) Naphthalene-d8	10.781	136	287280	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	170613	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.422	188	383298	20.000	ng/ul	0.00
79) Chrysene-d12	21.563	240	375278	20.000	ng/ul	0.00
88) Perylene-d12	24.145	264	418383	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	15888	8.799	ng/uL	0.00
4) Pyridine-d5	3.728	84	225430	47.699	ng/ul	0.00
7) Phenol-d5	7.104	99	277174	45.637	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	172752	46.614	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	225551	47.164	ng/ul	0.00
15) 4-Methylphenol-d8	8.669	113	215270	44.275	ng/ul	0.00
21) Nitrobenzene-d5	9.151	128	104948	47.485	ng/ul	0.00
24) 2-Nitrophenol-d4	9.869	143	116177	46.606	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	204408	44.232	ng/ul	0.00
31) 4-Chloroaniline-d4	10.951	131	282581	42.925	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	682041	51.249	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	714229	48.670	ng/ul	0.00
54) 4-Nitrophenol-d4	14.881	143	96096	41.233	ng/ul	0.00
60) Fluorene-d10	15.639	176	555814	49.239	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.780	200	98400	41.179	ng/ul	0.00
73) Anthracene-d10	17.527	188	905847	51.055	ng/ul	0.00
81) Pyrene-d10	19.780	212	1116125	53.523	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.974	264	1132942	53.658	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100423\
Data File : BP017536.D
Acq On : 04 Oct 2023 17:06
Operator : MA/JU
Sample : PB156005BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK005

Quant Time: Oct 05 00:38:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 04 16:33:47 2023
Response via : Initial Calibration

