

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100423\
 Data File : BP017540.D
 Acq On : 04 Oct 2023 19:24
 Operator : MA/JU
 Sample : PB156008BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK008

Quant Time: Oct 05 00:51:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 04 16:33:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	76375	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	305425	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	188317	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.422	188	424868	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	420440	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	484976	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	15158	7.935	ng/uL	0.00
4) Pyridine-d5	3.728	84	212292	42.455	ng/u1	0.00
7) Phenol-d5	7.099	99	256421	39.904	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	167403	42.693	ng/u1	0.00
11) 2-Chlorophenol-d4	7.463	132	208563	41.220	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	204212	39.697	ng/u1	0.00
21) Nitrobenzene-d5	9.152	128	99832	42.487	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	109663	41.379	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.393	165	193721	39.429	ng/u1	0.00
31) 4-Chloroaniline-d4	10.952	131	281028	40.153	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	670755	45.663	ng/u1	0.00
49) Acenaphthylene-d8	14.334	160	693831	42.835	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	99178	38.555	ng/u1	0.00
60) Fluorene-d10	15.634	176	550030	44.145	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.781	200	93594	35.335	ng/u1	0.00
73) Anthracene-d10	17.522	188	903279	45.929	ng/u1	0.00
81) Pyrene-d10	19.775	212	1122260	48.037	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.963	264	1163390	47.534	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100423\
 Data File : BP017540.D
 Acq On : 04 Oct 2023 19:24
 Operator : MA/JU
 Sample : PB156008BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK008

Quant Time: Oct 05 00:51:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 04 16:33:47 2023
 Response via : Initial Calibration

