

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
 Data File : BPCCC940.D
 Acq On : 05 Oct 2022 21:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020767

Quant Time: Oct 05 23:26:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 05 23:18:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	247566	20.000	ng/u1	0.00
20) Naphthalene-d8	10.687	136	980072	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	569862	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.287	188	1167261	20.000	ng/u1	0.00
79) Chrysene-d12	21.369	240	1199458	20.000	ng/u1	0.00
88) Perylene-d12	23.798	264	1205648	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	49973	8.903	ng/uL	0.00
4) Pyridine-d5	3.723	84	324671	19.746	ng/u1	0.00
7) Phenol-d5	7.046	99	395527	18.506	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.211	67	222971	18.822	ng/u1	0.00
11) 2-Chlorophenol-d4	7.411	132	335226	19.869	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	303769	17.194	ng/u1	0.00
21) Nitrobenzene-d5	9.046	128	151977	20.358	ng/u1	0.00
24) 2-Nitrophenol-d4	9.769	143	157211	19.684	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	304384	20.604	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	429664	19.078	ng/u1	0.00
46) Dimethylphthalate-d6	13.940	166	797512	19.570	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	974170	20.970	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	157232	18.632	ng/u1	0.00
60) Fluorene-d10	15.522	176	713862	20.171	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.640	200	122293	17.310	ng/u1	0.00
73) Anthracene-d10	17.381	188	1103746	20.758	ng/u1	0.00
81) Pyrene-d10	19.616	212	1274930	20.988	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.639	264	1256564	20.820	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	54029	8.912	ng/uL	98
5) Pyridine	3.740	79	324222	20.317	ng/u1	98
6) Benzaldehyde	7.022	77	202676	21.648	ng/u1	99
8) Phenol	7.069	94	413701	18.663	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.305	93	331333	18.784	ng/u1	98
12) 2-Chlorophenol	7.440	128	358941	19.900	ng/u1	99
13) 2-Methylphenol	8.328	108	308587	17.669	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.411	45	321114	18.467	ng/u1	99
16) Acetophenone	8.711	105	477358	17.722	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.693	70	208080	16.413	ng/u1	99
18) 4-Methylphenol	8.658	108	336852	17.505	ng/u1	97
19) Hexachloroethane	8.963	117	140059	20.820	ng/u1	98
22) Nitrobenzene	9.093	77	341333	21.260	ng/u1	98
23) Isophorone	9.622	82	586493	18.008	ng/u1	99
25) 2-Nitrophenol	9.805	139	183042	20.066	ng/u1	97
26) 2,4-Dimethylphenol	9.863	107	343286	20.033	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.105	93	408363	19.434	ng/u1	100
29) 2,4-Dichlorophenol	10.334	162	316668	20.688	ng/u1	100
30) Naphthalene	10.740	128	1105728	21.005	ng/u1	99
32) 4-Chloroaniline	10.857	127	444465	19.290	ng/u1	99
33) Hexachlorobutadiene	11.022	225	195817	21.612	ng/u1	97
34) Caprolactam	11.640	113	107372	19.820	ng/u1	94
35) 4-Chloro-3-methylphenol	11.981	107	301560	18.580	ng/u1	99
36) 2-Methylnaphthalene	12.352	142	717739	19.524	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
 Data File : BPCCC940.D
 Acq On : 05 Oct 2022 21:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020767

Quant Time: Oct 05 23:26:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 05 23:18:55 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.569	142	716916	19.435	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.716	216	372666	22.550	ng/ul	99
40) Hexachlorocyclopentadiene	12.693	237	202715	21.472	ng/ul	98
41) 2,4,6-Trichlorophenol	12.957	196	228315	20.639	ng/ul	99
42) 2,4,5-Trichlorophenol	13.028	196	252852	20.983	ng/ul	99
43) 1,1'-Biphenyl	13.363	154	908067	21.544	ng/ul	100
44) 2-Chloronaphthalene	13.404	162	734775	22.049	ng/ul	100
45) 2-Nitroaniline	13.610	65	162464	19.972	ng/ul	97
47) Dimethylphthalate	13.987	163	839549	19.658	ng/ul	99
48) 2,6-Dinitrotoluene	14.104	165	157726	18.950	ng/ul	94
50) Acenaphthylene	14.251	152	1099533	21.311	ng/ul	99
51) 3-Nitroaniline	14.440	138	176060	18.250	ng/ul	98
52) Acenaphthene	14.593	153	772801	21.230	ng/ul	99
53) 2,4-Dinitrophenol	14.640	184	108241	19.182	ng/ul	99
55) 4-Nitrophenol	14.746	109	108229	18.999	ng/ul	93
56) Dibenzofuran	14.928	168	1051531	21.090	ng/ul	98
57) 2,4-Dinitrotoluene	14.893	165	235116	19.326	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.151	232	209802	20.041	ng/ul#	98
59) Diethylphthalate	15.351	149	813444	19.012	ng/ul	100
61) Fluorene	15.575	166	860013	21.060	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.575	204	415573	20.557	ng/ul	99
63) 4-Nitroaniline	15.604	138	177001	19.227	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.657	198	133546	17.843	ng/ul	97
67) N-Nitrosodiphenylamine	15.787	169	718491	20.866	ng/ul	99
68) 4-Bromophenyl-phenylether	16.469	248	242278	20.406	ng/ul	97
69) Hexachlorobenzene	16.581	284	278308	20.598	ng/ul	97
70) Atrazine	16.745	200	233820	18.820	ng/ul	100
71) Pentachlorophenol	16.934	266	164665	18.530	ng/ul	99
72) Phenanthrene	17.328	178	1359790	21.198	ng/ul	100
74) Anthracene	17.416	178	1373585	21.282	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.322	216	368136	22.849	ng/uL	98
76) Pentachlorobenzene	14.845	250	369435	21.459	ng/uL	99
77) Carbazole	17.692	167	1262727	21.653	ng/ul	100
78) Di-n-butylphthalate	18.239	149	1348878	18.888	ng/ul	99
80) Fluoranthene	19.292	202	1564057	21.039	ng/ul	99
82) Pyrene	19.645	202	1660163	21.449	ng/ul	99
83) Butylbenzylphthalate	20.516	149	618941	18.502	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.292	252	525919	19.046	ng/ul	100
85) Benzo(a)anthracene	21.357	228	1649191	21.025	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.275	149	904002	17.940	ng/ul	100
87) Chrysene	21.410	228	1558795	21.176	ng/ul	98
89) Di-n-octyl phthalate	22.210	149	1518381	19.739	ng/ul	100
90) Benzo(b)fluoranthene	23.057	252	1691069	21.975	ng/ul	99
91) Benzo(k)fluoranthene	23.104	252	1619014	21.848	ng/ul	100
93) Benzo(a)pyrene	23.692	252	1469513	21.398	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.333	276	1786217	20.435	ng/ul	99
95) Dibenzo(a,h)anthracene	26.345	278	1611514	20.754	ng/ul	99
96) Benzo(g,h,i)perylene	27.104	276	1562322	20.161	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
Data File : BPCCC940.D
Acq On : 05 Oct 2022 21:23
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020767

Quant Time: Oct 05 23:26:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 05 23:18:55 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.912 ng/uL
 RT: 3.334 min Scan# 59
 Delta R.T. 0.000 min
 Lab File: BPCCC940.D
 Acq: 05 Oct 2022 21:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020767


Tgt Ion: 88 Resp: 54029
 Ion Ratio Lower Upper
 88 100
 43 18.6 15.6 23.4
 58 62.3 48.6 72.8



#5
 Pyridine
 Concen: 20.317 ng/uL
 RT: 3.740 min Scan# 128
 Delta R.T. 0.000 min
 Lab File: BPCCC940.D
 Acq: 05 Oct 2022 21:23

Tgt Ion: 79 Resp: 324222
 Ion Ratio Lower Upper
 79 100
 52 59.6 49.7 74.5
 51 26.8 21.6 32.4









#6
Benzaldehyde
Concen: 21.648 ng/ul
RT: 7.022 min Scan# 61
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion: 77 Resp: 202676
Ion Ratio Lower Upper
77 100
105 108.5 87.0 130.6
106 103.1 83.8 125.6



#8
Phenol
Concen: 18.663 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion: 94 Resp: 413701
Ion Ratio Lower Upper
94 100
65 24.0 19.5 29.3
66 31.4 25.0 37.4









#10
Bis(2-Chloroethyl)ether
Concen: 18.784 ng/ul
RT: 7.305 min Scan# 71
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion: 93 Resp: 331333
Ion Ratio Lower Upper
93 100
63 68.7 56.2 84.4
95 32.0 25.9 38.9



#12
2-Chlorophenol
Concen: 19.900 ng/ul
RT: 7.440 min Scan# 757
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:128 Resp: 358941
Ion Ratio Lower Upper
128 100
64 39.3 31.1 46.7
130 31.6 25.5 38.3





#13

2-Methylphenol

Concen: 17.669 ng/ul

RT: 8.328 min Scan# 908

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :





SSTD020767

Tgt Ion:108 Resp: 308587

Ion Ratio Lower Upper

108 100

107 87.5 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.467 ng/ul

RT: 8.411 min Scan# 922

Delta R.T. -0.006 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23




Tgt Ion: 45 Resp: 321114


Ion Ratio Lower Upper

45 100

77 19.3 16.0 24.0

79 15.3 12.5 18.7









#16
Acetophenone
Concen: 17.722 ng/ul
RT: 8.711 min Scan# 91
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:105 Resp: 477358
Ion Ratio Lower Upper
105 100
77 71.3 55.4 83.0
51 19.5 15.2 22.8



#17
N-Nitroso-di-n-propylamine
Concen: 16.413 ng/ul
RT: 8.693 min Scan# 970
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion: 70 Resp: 208080
Ion Ratio Lower Upper
70 100
42 41.2 32.7 49.1
101 10.3 8.4 12.6
130 25.3 19.4 29.2









#18
4-Methylphenol
Concen: 17.505 ng/ul
RT: 8.658 min Scan# 90
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:108 Resp: 336852
Ion Ratio Lower Upper
108 100
107 113.1 88.0 132.0



#19
Hexachloroethane
Concen: 20.820 ng/ul
RT: 8.963 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:117 Resp: 140059
Ion Ratio Lower Upper
117 100
201 100.0 82.4 123.6
199 63.3 51.6 77.4









#22
Nitrobenzene
Concen: 21.260 ng/ul
RT: 9.093 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion: 77 Resp: 341333
Ion Ratio Lower Upper
77 100
123 51.0 41.6 62.4
65 12.7 9.7 14.5



#23
Isophorone
Concen: 18.008 ng/ul
RT: 9.622 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion: 82 Resp: 586493
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 20.3 15.8 23.8









#25
2-Nitrophenol
Concen: 20.066 ng/ul
RT: 9.805 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:139 Resp: 183042
Ion Ratio Lower Upper
139 100
65 41.9 32.2 48.4
109 26.1 19.6 29.4



#26
2,4-Dimethylphenol
Concen: 20.033 ng/ul
RT: 9.863 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:107 Resp: 343286
Ion Ratio Lower Upper
107 100
121 54.7 43.0 64.6
122 93.2 74.9 112.3





#27

Bis(2-Chloroethoxy)methane

Concen: 19.434 ng/ul

RT: 10.105 min Scan# 1210

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion: 93 Resp: 408363

Ion Ratio Lower Upper

93 100

95 33.1 26.3 39.5

123 12.1 9.5 14.3



#29

2,4-Dichlorophenol

Concen: 20.688 ng/ul

RT: 10.334 min Scan# 1249

Delta R.T. -0.006 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23




Tgt Ion:162 Resp: 316668


Ion Ratio Lower Upper

162 100

164 65.4 52.2 78.4

98 33.8 27.0 40.4









#30
Naphthalene
Concen: 21.005 ng/ul
RT: 10.740 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:128 Resp: 1105728
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 19.290 ng/ul
RT: 10.857 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:127 Resp: 444465
Ion Ratio Lower Upper
127 100
129 33.8 26.6 39.8









#33
Hexachlorobutadiene
Concen: 21.612 ng/ul
RT: 11.022 min Scan# 1367
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:225 Resp: 195817
Ion Ratio Lower Upper
225 100
223 62.3 52.1 78.1
227 63.4 52.3 78.5



#34
Caprolactam
Concen: 19.820 ng/ul
RT: 11.640 min Scan# 1471
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:113 Resp: 107372
Ion Ratio Lower Upper
113 100
55 138.3 106.0 159.0
56 105.7 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 18.580 ng/ul
RT: 11.981 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:107 Resp: 301560
Ion Ratio Lower Upper
107 100
144 28.4 22.5 33.7
142 86.9 70.0 105.0



#36
2-Methylnaphthalene
Concen: 19.524 ng/ul
RT: 12.352 min Scan# 1592
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:142 Resp: 717739
Ion Ratio Lower Upper
142 100
141 87.7 70.6 106.0









#37
1-Methylnaphthalene
Concen: 19.435 ng/ul
RT: 12.569 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:142 Resp: 716916
Ion Ratio Lower Upper
142 100
141 91.9 74.8 112.2
116 3.4 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.550 ng/ul
RT: 12.716 min Scan# 1654
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:216 Resp: 372666
Ion Ratio Lower Upper
216 100
214 78.7 62.6 94.0
179 20.4 16.6 24.8
108 16.3 12.6 19.0









#40
Hexachlorocyclopentadiene
Concen: 21.472 ng/ul
RT: 12.693 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:237 Resp: 202715
Ion Ratio Lower Upper
237 100
235 61.9 51.0 76.6
272 12.4 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 20.639 ng/ul
RT: 12.957 min Scan# 1695
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:196 Resp: 228315
Ion Ratio Lower Upper
196 100
198 98.0 78.7 118.1
200 32.2 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 20.983 ng/ul
RT: 13.028 min Scan# 1707
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:196 Resp: 252852
Ion Ratio Lower Upper
196 100
198 97.1 76.6 114.8
200 31.6 25.6 38.4



#43
1,1'-Biphenyl
Concen: 21.544 ng/ul
RT: 13.363 min Scan# 1764
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:154 Resp: 908067
Ion Ratio Lower Upper
154 100
153 39.7 31.8 47.6
76 12.0 9.5 14.3









#44
2-Chloronaphthalene
Concen: 22.049 ng/ul
RT: 13.404 min Scan# 1771
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:162 Resp: 734775
Ion Ratio Lower Upper
162 100
164 32.9 26.5 39.7
127 37.4 29.7 44.5



#45
2-Nitroaniline
Concen: 19.972 ng/ul
RT: 13.610 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion: 65 Resp: 162464
Ion Ratio Lower Upper
65 100
92 75.9 61.2 91.8
138 122.8 103.0 154.4









#47
Dimethylphthalate
Concen: 19.658 ng/ul
RT: 13.987 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:163 Resp: 839549
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.5 12.7



#48
2,6-Dinitrotoluene
Concen: 18.950 ng/ul
RT: 14.104 min Scan# 1890
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:165 Resp: 157726
Ion Ratio Lower Upper
165 100
89 54.3 39.1 58.7
121 22.0 16.7 25.1









#50
Acenaphthylene
Concen: 21.311 ng/ul
RT: 14.251 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:152 Resp: 1099533
Ion Ratio Lower Upper
152 100
151 19.4 16.0 24.0
153 13.1 10.8 16.2



#51
3-Nitroaniline
Concen: 18.250 ng/ul
RT: 14.440 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:138 Resp: 176060
Ion Ratio Lower Upper
138 100
108 10.1 8.2 12.4
92 104.3 85.6 128.4









#52
Acenaphthene
Concen: 21.230 ng/ul
RT: 14.593 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:153 Resp: 772801
Ion Ratio Lower Upper
153 100
152 47.5 38.4 57.6
154 89.4 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 19.182 ng/ul
RT: 14.640 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:184 Resp: 108241
Ion Ratio Lower Upper
184 100
63 56.3 44.2 66.4
154 57.6 45.8 68.6









#55
4-Nitrophenol
Concen: 18.999 ng/ul
RT: 14.746 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:109 Resp: 108229
Ion Ratio Lower Upper
109 100
139 166.1 125.0 187.6
65 133.8 101.9 152.9



#56
Dibenzofuran
Concen: 21.090 ng/ul
RT: 14.928 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:168 Resp: 1051531
Ion Ratio Lower Upper
168 100
139 35.9 29.4 44.2





#57

2,4-Dinitrotoluene

Concen: 19.326 ng/ul

RT: 14.893 min Scan# 2024

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion:165 Resp: 235116

Ion Ratio Lower Upper

165 100

63 37.2 28.2 42.2

182 2.5 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.041 ng/ul

RT: 15.151 min Scan# 2068

Delta R.T. -0.006 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Tgt Ion:232 Resp: 209802




Ion Ratio Lower Upper


232 100

131 46.8 36.2 54.2

130 2.6 1.7 2.5#

166 32.1 25.0 37.4









#59
Diethylphthalate
Concen: 19.012 ng/ul
RT: 15.351 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:149 Resp: 813444
Ion Ratio Lower Upper
149 100
177 22.4 17.8 26.8
150 12.3 9.9 14.9



#61
Fluorene
Concen: 21.060 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:166 Resp: 860013
Ion Ratio Lower Upper
166 100
165 97.1 77.2 115.8
167 13.9 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 20.557 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:204 Resp: 415573
Ion Ratio Lower Upper
204 100
206 33.0 26.8 40.2
141 54.3 44.5 66.7



#63
4-Nitroaniline
Concen: 19.227 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:138 Resp: 177001
Ion Ratio Lower Upper
138 100
92 47.5 35.9 53.9
108 65.8 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 17.843 ng/ul
RT: 15.657 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:198 Resp: 133546
Ion Ratio Lower Upper
198 100
182 4.9 3.7 5.5
77 22.2 16.6 24.8



#67
N-Nitrosodiphenylamine
Concen: 20.866 ng/ul
RT: 15.787 min Scan# 2176
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:169 Resp: 718491
Ion Ratio Lower Upper
169 100
168 65.6 53.7 80.5
167 36.9 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 20.406 ng/ul
RT: 16.469 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:248 Resp: 242278
Ion Ratio Lower Upper
248 100
250 96.3 78.6 118.0
141 74.3 56.0 84.0



#69
Hexachlorobenzene
Concen: 20.598 ng/ul
RT: 16.581 min Scan# 2311
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:284 Resp: 278308
Ion Ratio Lower Upper
284 100
142 36.8 27.3 40.9
249 31.2 25.2 37.8





#70

Atrazine

Concen: 18.820 ng/ul

RT: 16.745 min Scan# 21

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion:200 Resp: 233820

Ion Ratio Lower Upper

200 100

173 26.6 21.6 32.4

215 48.5 38.9 58.3



#71

Pentachlorophenol

Concen: 18.530 ng/ul

RT: 16.934 min Scan# 2371

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23




Tgt Ion:266 Resp: 164665


Ion Ratio Lower Upper

266 100

264 62.9 49.8 74.8

268 64.5 51.9 77.9









#72
Phenanthrene
Concen: 21.198 ng/ul
RT: 17.328 min Scan# 2438
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:178 Resp: 1359790
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 19.1 15.4 23.2



#74
Anthracene
Concen: 21.282 ng/ul
RT: 17.416 min Scan# 2453
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:178 Resp: 1373585
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.9 15.4 23.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.849 ng/uL

RT: 13.322 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion:216 Resp: 368136

Ion Ratio Lower Upper

216 100

214 79.4 62.6 93.8

218 48.0 37.4 56.0



#76

Pentachlorobenzene

Concen: 21.459 ng/uL

RT: 14.845 min Scan# 2016

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23




Tgt Ion:250 Resp: 369435


Ion Ratio Lower Upper

250 100

248 62.6 51.4 77.0

252 64.1 51.5 77.3









#77
 Carbazole
 Concen: 21.653 ng/ul
 RT: 17.692 min Scan# 2500
 Delta R.T. 0.000 min
 Lab File: BPCCC940.D
 Acq: 05 Oct 2022 21:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020767


Tgt Ion:167 Resp: 1262727
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.1 25.7
 139 12.7 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 18.888 ng/ul
 RT: 18.239 min Scan# 2593
 Delta R.T. 0.000 min
 Lab File: BPCCC940.D
 Acq: 05 Oct 2022 21:23

Tgt Ion:149 Resp: 1348878
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.5 11.3
 104 4.5 3.6 5.4





#80

Fluoranthene

Concen: 21.039 ng/ul

RT: 19.292 min Scan# 2772

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion:202 Resp: 1564057

Ion Ratio Lower Upper

202 100

101 11.9 9.8 14.6

100 9.1 7.4 11.0



#82

Pyrene

Concen: 21.449 ng/ul

RT: 19.645 min Scan# 2832

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23




Tgt Ion:202 Resp: 1660163


Ion Ratio Lower Upper

202 100

101 14.0 11.6 17.4

100 11.4 9.4 14.0









#83
Butylbenzylphthalate
Concen: 18.502 ng/ul
RT: 20.516 min Scan# 2981
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:149 Resp: 618941
Ion Ratio Lower Upper
149 100
91 66.8 50.2 75.4
206 19.4 16.9 25.3



#84
3,3'-Dichlorobenzidine
Concen: 19.046 ng/ul
RT: 21.292 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:252 Resp: 525919
Ion Ratio Lower Upper
252 100
254 64.1 51.4 77.0
126 11.1 8.9 13.3









#85
Benzo(a)anthracene
Concen: 21.025 ng/ul
RT: 21.357 min Scan# 3123
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:228 Resp: 1649191
Ion Ratio Lower Upper
228 100
229 19.5 15.8 23.6
226 26.9 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 17.940 ng/ul
RT: 21.275 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:149 Resp: 904002
Ion Ratio Lower Upper
149 100
167 26.2 21.2 31.8
279 3.7 2.9 4.3









#87
Chrysene
Concen: 21.176 ng/ul
RT: 21.410 min Scan# 3132
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:228 Resp: 1558795
Ion Ratio Lower Upper
228 100
226 30.1 25.3 37.9
229 19.5 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 19.739 ng/ul
RT: 22.210 min Scan# 3268
Delta R.T. 0.000 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:149 Resp: 1518381





#90

Benzo(b)fluoranthene

Concen: 21.975 ng/ul

RT: 23.057 min Scan# 3412

Delta R.T. 0.000 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion:252 Resp: 1691069

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 10.8 9.2 13.8



#91

Benzo(k)fluoranthene

Concen: 21.848 ng/ul

RT: 23.104 min Scan# 3420

Delta R.T. -0.006 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23




Tgt Ion:252 Resp: 1619014


Ion Ratio Lower Upper

252 100

253 22.5 18.0 27.0

125 11.5 9.2 13.8









#93
Benzo(a)pyrene
Concen: 21.398 ng/ul
RT: 23.692 min Scan# 31
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23




Instrument :
BNA_P
ClientSampleId :
SSTD020767


Tgt Ion:252 Resp: 1469513
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 12.3 9.8 14.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.435 ng/ul
RT: 26.333 min Scan# 3969
Delta R.T. -0.006 min
Lab File: BPCCC940.D
Acq: 05 Oct 2022 21:23

Tgt Ion:276 Resp: 1786217
Ion Ratio Lower Upper
276 100
138 27.4 21.4 32.0
277 25.0 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 20.754 ng/ul

RT: 26.345 min Scan# 3973

Delta R.T. -0.012 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Instrument :

BNA_P

ClientSampleId :

SSTD020767





Tgt Ion:278 Resp: 1611514

Ion Ratio Lower Upper

278 100

139 18.0 15.0 22.6

279 23.2 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 20.161 ng/ul

RT: 27.104 min Scan# 4100

Delta R.T. -0.012 min

Lab File: BPCCC940.D

Acq: 05 Oct 2022 21:23

Tgt Ion:276 Resp: 1562322

Ion Ratio Lower Upper

276 100

138 23.8 19.8 29.6

277 24.0 18.6 27.8

