

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP100522\  
 Data File : BP011972.D  
 Acq On : 06 Oct 2022 16:15  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 53 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTD020769

Quant Time: Oct 06 23:07:03 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP093022.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed Oct 05 23:18:55 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| -----                         |        |      |          |        |        |          |
| Internal Standards            |        |      |          |        |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.881  | 152  | 241403   | 20.000 | ng/u1  | 0.00     |
| 20) Naphthalene-d8            | 10.687 | 136  | 967585   | 20.000 | ng/u1  | 0.00     |
| 38) Acenaphthene-d10          | 14.528 | 164  | 565620   | 20.000 | ng/u1  | 0.00     |
| 64) Phenanthrene-d10          | 17.281 | 188  | 1152913  | 20.000 | ng/u1  | 0.00     |
| 79) Chrysene-d12              | 21.369 | 240  | 1177779  | 20.000 | ng/u1  | 0.00     |
| 88) Perylene-d12              | 23.798 | 264  | 1207455  | 20.000 | ng/u1  | 0.00     |
| System Monitoring Compounds   |        |      |          |        |        |          |
| 3) 1,4-Dioxane-d8             | 3.299  | 96   | 49941    | 9.124  | ng/uL  | 0.00     |
| 4) Pyridine-d5                | 3.722  | 84   | 332204   | 20.720 | ng/u1  | 0.00     |
| 7) Phenol-d5                  | 7.046  | 99   | 396807   | 19.040 | ng/u1  | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.211  | 67   | 223086   | 19.312 | ng/u1  | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.411  | 132  | 328440   | 19.964 | ng/u1  | 0.00     |
| 15) 4-Methylphenol-d8         | 8.593  | 113  | 307256   | 17.835 | ng/u1  | 0.00     |
| 21) Nitrobenzene-d5           | 9.046  | 128  | 154807   | 21.005 | ng/u1  | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.769  | 143  | 161940   | 20.538 | ng/u1  | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.310 | 165  | 304135   | 20.853 | ng/u1  | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.828 | 131  | 423588   | 19.051 | ng/u1  | 0.00     |
| 46) Dimethylphthalate-d6      | 13.940 | 166  | 800844   | 19.799 | ng/u1  | 0.00     |
| 49) Acenaphthylene-d8         | 14.222 | 160  | 974612   | 21.136 | ng/u1  | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.728 | 143  | 158387   | 18.910 | ng/u1  | 0.00     |
| 60) Fluorene-d10              | 15.522 | 176  | 716307   | 20.392 | ng/u1  | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.640 | 200  | 124785   | 17.882 | ng/u1  | 0.00     |
| 73) Anthracene-d10            | 17.381 | 188  | 1103594  | 21.014 | ng/u1  | 0.00     |
| 81) Pyrene-d10                | 19.616 | 212  | 1275213  | 21.379 | ng/u1  | 0.00     |
| 92) Benzo(a)pyrene-d12        | 23.639 | 264  | 1279341  | 21.166 | ng/u1  | 0.00     |
| Target Compounds              |        |      |          |        |        |          |
|                               |        |      |          |        | Qvalue |          |
| 2) 1,4-Dioxane                | 3.334  | 88   | 52401    | 8.864  | ng/uL  | 99       |
| 5) Pyridine                   | 3.740  | 79   | 324141   | 20.830 | ng/u1  | 100      |
| 6) Benzaldehyde               | 7.022  | 77   | 197302   | 21.612 | ng/u1  | 99       |
| 8) Phenol                     | 7.069  | 94   | 412599   | 19.088 | ng/u1  | 98       |
| 10) Bis(2-Chloroethyl)ether   | 7.305  | 93   | 327932   | 19.066 | ng/u1  | 100      |
| 12) 2-Chlorophenol            | 7.440  | 128  | 355230   | 20.197 | ng/u1  | 99       |
| 13) 2-Methylphenol            | 8.328  | 108  | 312981   | 18.378 | ng/u1  | 98       |
| 14) 2,2'-oxybis(1-Chloropr... | 8.405  | 45   | 326221   | 19.239 | ng/u1  | 96       |
| 16) Acetophenone              | 8.710  | 105  | 482700   | 18.377 | ng/u1  | 100      |
| 17) N-Nitroso-di-n-propyla... | 8.693  | 70   | 217158   | 17.566 | ng/u1  | 99       |
| 18) 4-Methylphenol            | 8.658  | 108  | 336429   | 17.930 | ng/u1  | 100      |
| 19) Hexachloroethane          | 8.963  | 117  | 138932   | 21.179 | ng/u1  | 98       |
| 22) Nitrobenzene              | 9.087  | 77   | 346419   | 21.856 | ng/u1  | 99       |
| 23) Isophorone                | 9.616  | 82   | 611636   | 19.022 | ng/u1  | 100      |
| 25) 2-Nitrophenol             | 9.799  | 139  | 187324   | 20.801 | ng/u1  | 95       |
| 26) 2,4-Dimethylphenol        | 9.863  | 107  | 346858   | 20.503 | ng/u1  | 99       |
| 27) Bis(2-Chloroethoxy)met... | 10.099 | 93   | 409791   | 19.754 | ng/u1  | 99       |
| 29) 2,4-Dichlorophenol        | 10.334 | 162  | 318852   | 21.099 | ng/u1  | 98       |
| 30) Naphthalene               | 10.740 | 128  | 1104844  | 21.259 | ng/u1  | 99       |
| 32) 4-Chloroaniline           | 10.852 | 127  | 439899   | 19.339 | ng/u1  | 100      |
| 33) Hexachlorobutadiene       | 11.022 | 225  | 193036   | 21.581 | ng/u1  | 98       |
| 34) Caprolactam               | 11.640 | 113  | 107153   | 20.035 | ng/u1  | 90       |
| 35) 4-Chloro-3-methylphenol   | 11.981 | 107  | 309063   | 19.288 | ng/u1  | 99       |
| 36) 2-Methylnaphthalene       | 12.351 | 142  | 721151   | 19.870 | ng/u1  | 99       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP100522\  
 Data File : BP011972.D  
 Acq On : 06 Oct 2022 16:15  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 53 Sample Multiplier: 1


Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTD020769


Quant Time: Oct 06 23:07:03 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP093022.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed Oct 05 23:18:55 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 37) 1-Methylnaphthalene       | 12.569 | 142  | 718672   | 19.734 | ng/ul  | 99       |
| 39) 1,2,4,5-Tetrachloroben... | 12.716 | 216  | 368723   | 22.479 | ng/ul  | 99       |
| 40) Hexachlorocyclopentadiene | 12.693 | 237  | 196208   | 20.938 | ng/ul  | 97       |
| 41) 2,4,6-Trichlorophenol     | 12.957 | 196  | 235338   | 21.433 | ng/ul  | 98       |
| 42) 2,4,5-Trichlorophenol     | 13.028 | 196  | 255893   | 21.395 | ng/ul  | 99       |
| 43) 1,1'-Biphenyl             | 13.357 | 154  | 908230   | 21.709 | ng/ul  | 99       |
| 44) 2-Chloronaphthalene       | 13.398 | 162  | 727880   | 22.006 | ng/ul  | 98       |
| 45) 2-Nitroaniline            | 13.610 | 65   | 172457   | 21.360 | ng/ul  | 98       |
| 47) Dimethylphthalate         | 13.987 | 163  | 840968   | 19.839 | ng/ul  | 99       |
| 48) 2,6-Dinitrotoluene        | 14.104 | 165  | 161105   | 19.501 | ng/ul  | 93       |
| 50) Acenaphthylene            | 14.245 | 152  | 1096284  | 21.407 | ng/ul  | 100      |
| 51) 3-Nitroaniline            | 14.434 | 138  | 177972   | 18.587 | ng/ul  | 99       |
| 52) Acenaphthene              | 14.592 | 153  | 769720   | 21.304 | ng/ul  | 99       |
| 53) 2,4-Dinitrophenol         | 14.640 | 184  | 116092   | 20.727 | ng/ul  | 98       |
| 55) 4-Nitrophenol             | 14.745 | 109  | 110063   | 19.466 | ng/ul  | 96       |
| 56) Dibenzofuran              | 14.928 | 168  | 1041617  | 21.048 | ng/ul  | 99       |
| 57) 2,4-Dinitrotoluene        | 14.892 | 165  | 238758   | 19.772 | ng/ul  | 98       |
| 58) 2,3,4,6-Tetrachlorophenol | 15.151 | 232  | 212774   | 20.478 | ng/ul  | 98       |
| 59) Diethylphthalate          | 15.351 | 149  | 812134   | 19.124 | ng/ul  | 100      |
| 61) Fluorene                  | 15.575 | 166  | 847172   | 20.901 | ng/ul  | 99       |
| 62) 4-Chlorophenyl-phenyle... | 15.569 | 204  | 406625   | 20.265 | ng/ul  | 98       |
| 63) 4-Nitroaniline            | 15.604 | 138  | 177500   | 19.426 | ng/ul  | 98       |
| 66) 4,6-Dinitro-2-methylph... | 15.651 | 198  | 136155   | 18.417 | ng/ul# | 93       |
| 67) N-Nitrosodiphenylamine    | 15.787 | 169  | 715340   | 21.033 | ng/ul  | 99       |
| 68) 4-Bromophenyl-phenylether | 16.469 | 248  | 240477   | 20.506 | ng/ul  | 98       |
| 69) Hexachlorobenzene         | 16.581 | 284  | 280425   | 21.013 | ng/ul  | 97       |
| 70) Atrazine                  | 16.745 | 200  | 238290   | 19.418 | ng/ul  | 99       |
| 71) Pentachlorophenol         | 16.928 | 266  | 169905   | 19.358 | ng/ul  | 99       |
| 72) Phenanthrene              | 17.328 | 178  | 1348731  | 21.287 | ng/ul  | 99       |
| 74) Anthracene                | 17.416 | 178  | 1368004  | 21.459 | ng/ul  | 99       |
| 75) 1,2,3,4-Tetrachloroben... | 13.322 | 216  | 363837   | 22.864 | ng/uL  | 99       |
| 76) Pentachlorobenzene        | 14.845 | 250  | 369771   | 21.746 | ng/uL  | 99       |
| 77) Carbazole                 | 17.692 | 167  | 1249692  | 21.696 | ng/ul  | 100      |
| 78) Di-n-butylphthalate       | 18.239 | 149  | 1375860  | 19.506 | ng/ul  | 99       |
| 80) Fluoranthene              | 19.292 | 202  | 1559423  | 21.363 | ng/ul  | 100      |
| 82) Pyrene                    | 19.645 | 202  | 1654244  | 21.766 | ng/ul  | 99       |
| 83) Butylbenzylphthalate      | 20.516 | 149  | 642769   | 19.568 | ng/ul  | 96       |
| 84) 3,3'-Dichlorobenzidine    | 21.286 | 252  | 527870   | 19.469 | ng/ul  | 99       |
| 85) Benzo(a)anthracene        | 21.351 | 228  | 1638866  | 21.278 | ng/ul  | 100      |
| 86) Bis(2-ethylhexyl)phtha... | 21.275 | 149  | 920274   | 18.599 | ng/ul  | 99       |
| 87) Chrysene                  | 21.410 | 228  | 1539410  | 21.298 | ng/ul  | 98       |
| 89) Di-n-octyl phthalate      | 22.210 | 149  | 1568466  | 20.360 | ng/ul  | 100      |
| 90) Benzo(b)fluoranthene      | 23.057 | 252  | 1659109  | 21.528 | ng/ul  | 100      |
| 91) Benzo(k)fluoranthene      | 23.104 | 252  | 1621531  | 21.849 | ng/ul  | 99       |
| 93) Benzo(a)pyrene            | 23.692 | 252  | 1479220  | 21.508 | ng/ul  | 100      |
| 94) Indeno(1,2,3-cd)pyrene    | 26.327 | 276  | 1796623  | 20.523 | ng/ul  | 100      |
| 95) Dibenzo(a,h)anthracene    | 26.351 | 278  | 1623751  | 20.880 | ng/ul  | 99       |
| 96) Benzo(g,h,i)perylene      | 27.104 | 276  | 1594046  | 20.539 | ng/ul  | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 06 23:07:03 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP093022.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Wed Oct 05 23:18:55 2022  
Response via : Initial Calibration









#2  
 1,4-Dioxane  
 Concen: 8.864 ng/uL  
 RT: 3.334 min Scan# 59  
 Delta R.T. 0.000 min  
 Lab File: BP011972.D  
 Acq: 06 Oct 2022 16:15




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTD020769


Tgt Ion: 88 Resp: 52401  
 Ion Ratio Lower Upper  
 88 100  
 43 18.3 15.6 23.4  
 58 61.0 48.6 72.8



#5  
 Pyridine  
 Concen: 20.830 ng/uL  
 RT: 3.740 min Scan# 128  
 Delta R.T. 0.000 min  
 Lab File: BP011972.D  
 Acq: 06 Oct 2022 16:15

Tgt Ion: 79 Resp: 324141  
 Ion Ratio Lower Upper  
 79 100  
 52 61.8 49.7 74.5  
 51 27.1 21.6 32.4









#6  
Benzaldehyde  
Concen: 21.612 ng/ul  
RT: 7.022 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion: 77 Resp: 197302  
Ion Ratio Lower Upper  
77 100  
105 108.2 87.0 130.6  
106 103.9 83.8 125.6



#8  
Phenol  
Concen: 19.088 ng/ul  
RT: 7.069 min Scan# 694  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion: 94 Resp: 412599  
Ion Ratio Lower Upper  
94 100  
65 25.1 19.5 29.3  
66 32.4 25.0 37.4









#10  
Bis(2-Chloroethyl)ether  
Concen: 19.066 ng/ul  
RT: 7.305 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion: 93 Resp: 327932  
Ion Ratio Lower Upper  
93 100  
63 70.2 56.2 84.4  
95 32.2 25.9 38.9



#12  
2-Chlorophenol  
Concen: 20.197 ng/ul  
RT: 7.440 min Scan# 757  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion: 128 Resp: 355230  
Ion Ratio Lower Upper  
128 100  
64 39.5 31.1 46.7  
130 32.8 25.5 38.3





#13

2-Methylphenol

Concen: 18.378 ng/ul

RT: 8.328 min Scan# 908

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :





SSTD020769

Tgt Ion:108 Resp: 312981

Ion Ratio Lower Upper

108 100

107 89.1 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.239 ng/ul

RT: 8.405 min Scan# 921

Delta R.T. -0.012 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion: 45 Resp: 326221


Ion Ratio Lower Upper

45 100

77 18.4 16.0 24.0

79 13.9 12.5 18.7









#16  
Acetophenone  
Concen: 18.377 ng/ul  
RT: 8.710 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:105 Resp: 482700  
Ion Ratio Lower Upper  
105 100  
77 69.6 55.4 83.0  
51 19.1 15.2 22.8



#17  
N-Nitroso-di-n-propylamine  
Concen: 17.566 ng/ul  
RT: 8.693 min Scan# 970  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion: 70 Resp: 217158  
Ion Ratio Lower Upper  
70 100  
42 41.3 32.7 49.1  
101 10.6 8.4 12.6  
130 24.1 19.4 29.2









#18  
4-Methylphenol  
Concen: 17.930 ng/ul  
RT: 8.658 min Scan# 90  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:108 Resp: 336429  
Ion Ratio Lower Upper  
108 100  
107 109.9 88.0 132.0



#19  
Hexachloroethane  
Concen: 21.179 ng/ul  
RT: 8.963 min Scan# 1016  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:117 Resp: 138932  
Ion Ratio Lower Upper  
117 100  
201 100.1 82.4 123.6  
199 63.5 51.6 77.4









#22  
Nitrobenzene  
Concen: 21.856 ng/ul  
RT: 9.087 min Scan# 1038  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion: 77 Resp: 346419  
Ion Ratio Lower Upper  
77 100  
123 50.8 41.6 62.4  
65 12.0 9.7 14.5



#23  
Isophorone  
Concen: 19.022 ng/ul  
RT: 9.616 min Scan# 1127  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion: 82 Resp: 611636  
Ion Ratio Lower Upper  
82 100  
95 6.6 5.4 8.2  
138 19.6 15.8 23.8









#25  
2-Nitrophenol  
Concen: 20.801 ng/ul  
RT: 9.799 min Scan# 1158  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:139 Resp: 187324  
Ion Ratio Lower Upper  
139 100  
65 44.3 32.2 48.4  
109 25.7 19.6 29.4



#26  
2,4-Dimethylphenol  
Concen: 20.503 ng/ul  
RT: 9.863 min Scan# 1169  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:107 Resp: 346858  
Ion Ratio Lower Upper  
107 100  
121 54.1 43.0 64.6  
122 92.4 74.9 112.3









#27  
 Bis(2-Chloroethoxy)methane  
 Concen: 19.754 ng/ul  
 RT: 10.099 min Scan# 1209  
 Delta R.T. -0.006 min  
 Lab File: BP011972.D  
 Acq: 06 Oct 2022 16:15




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTD020769


Tgt Ion: 93 Resp: 409791  
 Ion Ratio Lower Upper  
 93 100  
 95 32.1 26.3 39.5  
 123 11.5 9.5 14.3



#29  
 2,4-Dichlorophenol  
 Concen: 21.099 ng/ul  
 RT: 10.334 min Scan# 1249  
 Delta R.T. -0.006 min  
 Lab File: BP011972.D  
 Acq: 06 Oct 2022 16:15

Tgt Ion: 162 Resp: 318852  
 Ion Ratio Lower Upper  
 162 100  
 164 64.2 52.2 78.4  
 98 32.6 27.0 40.4









#30  
Naphthalene  
Concen: 21.259 ng/ul  
RT: 10.740 min Scan# 1319  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:128 Resp: 1104844  
Ion Ratio Lower Upper  
128 100  
129 11.0 8.9 13.3  
127 12.5 10.5 15.7



#32  
4-Chloroaniline  
Concen: 19.339 ng/ul  
RT: 10.852 min Scan# 1337  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:127 Resp: 439899  
Ion Ratio Lower Upper  
127 100  
129 33.1 26.6 39.8





#33

Hexachlorobutadiene

Concen: 21.581 ng/ul

RT: 11.022 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:225 Resp: 193036

Ion Ratio Lower Upper

225 100

223 64.4 52.1 78.1

227 63.4 52.3 78.5



#34

Caprolactam

Concen: 20.035 ng/ul

RT: 11.640 min Scan# 1471

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:113 Resp: 107153


Ion Ratio Lower Upper

113 100

55 143.7 106.0 159.0

56 109.4 78.7 118.1









#35  
4-Chloro-3-methylphenol  
Concen: 19.288 ng/ul  
RT: 11.981 min Scan# 1529  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:107 Resp: 309063  
Ion Ratio Lower Upper  
107 100  
144 28.4 22.5 33.7  
142 86.7 70.0 105.0



#36  
2-Methylnaphthalene  
Concen: 19.870 ng/ul  
RT: 12.351 min Scan# 1592  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:142 Resp: 721151  
Ion Ratio Lower Upper  
142 100  
141 87.0 70.6 106.0









#37  
1-Methylnaphthalene  
Concen: 19.734 ng/ul  
RT: 12.569 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:142 Resp: 718672  
Ion Ratio Lower Upper  
142 100  
141 92.1 74.8 112.2  
116 3.6 2.9 4.3



#39  
1,2,4,5-Tetrachlorobenzene  
Concen: 22.479 ng/ul  
RT: 12.716 min Scan# 1654  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:216 Resp: 368723  
Ion Ratio Lower Upper  
216 100  
214 77.0 62.6 94.0  
179 20.1 16.6 24.8  
108 16.2 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 20.938 ng/ul

RT: 12.693 min Scan# 1650

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:237 Resp: 196208

Ion Ratio Lower Upper

237 100

235 61.5 51.0 76.6

272 12.6 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 21.433 ng/ul

RT: 12.957 min Scan# 1695

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:196 Resp: 235338


Ion Ratio Lower Upper

196 100

198 96.9 78.7 118.1

200 31.9 24.5 36.7









#42  
2,4,5-Trichlorophenol  
Concen: 21.395 ng/ul  
RT: 13.028 min Scan# 1707  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:196 Resp: 255893  
Ion Ratio Lower Upper  
196 100  
198 95.8 76.6 114.8  
200 30.4 25.6 38.4



#43  
1,1'-Biphenyl  
Concen: 21.709 ng/ul  
RT: 13.357 min Scan# 1763  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:154 Resp: 908230  
Ion Ratio Lower Upper  
154 100  
153 40.0 31.8 47.6  
76 12.4 9.5 14.3





#44

2-Chloronaphthalene

Concen: 22.006 ng/ul

RT: 13.398 min Scan# 1771

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:162 Resp: 727880

Ion Ratio Lower Upper

162 100

164 32.6 26.5 39.7

127 38.4 29.7 44.5



#45

2-Nitroaniline

Concen: 21.360 ng/ul

RT: 13.610 min Scan# 1806

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion: 65 Resp: 172457


Ion Ratio Lower Upper

65 100

92 75.9 61.2 91.8

138 125.0 103.0 154.4









#47  
Dimethylphthalate  
Concen: 19.839 ng/ul  
RT: 13.987 min Scan# 1870  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:163 Resp: 840968  
Ion Ratio Lower Upper  
163 100  
194 4.1 3.3 4.9  
164 10.0 8.5 12.7



#48  
2,6-Dinitrotoluene  
Concen: 19.501 ng/ul  
RT: 14.104 min Scan# 1890  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:165 Resp: 161105  
Ion Ratio Lower Upper  
165 100  
89 54.8 39.1 58.7  
121 23.1 16.7 25.1









#50  
Acenaphthylene  
Concen: 21.407 ng/ul  
RT: 14.245 min Scan# 1915  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:152 Resp: 1096284  
Ion Ratio Lower Upper  
152 100  
151 19.8 16.0 24.0  
153 13.3 10.8 16.2



#51  
3-Nitroaniline  
Concen: 18.587 ng/ul  
RT: 14.434 min Scan# 1946  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:138 Resp: 177972  
Ion Ratio Lower Upper  
138 100  
108 9.6 8.2 12.4  
92 108.5 85.6 128.4









#52  
Acenaphthene  
Concen: 21.304 ng/ul  
RT: 14.592 min Scan# 1773  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:153 Resp: 769720  
Ion Ratio Lower Upper  
153 100  
152 47.3 38.4 57.6  
154 87.6 70.6 105.8



#53  
2,4-Dinitrophenol  
Concen: 20.727 ng/ul  
RT: 14.640 min Scan# 1981  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:184 Resp: 116092  
Ion Ratio Lower Upper  
184 100  
63 57.0 44.2 66.4  
154 56.7 45.8 68.6









#55  
4-Nitrophenol  
Concen: 19.466 ng/ul  
RT: 14.745 min Scan# 1999  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:109 Resp: 110063  
Ion Ratio Lower Upper  
109 100  
139 159.7 125.0 187.6  
65 132.9 101.9 152.9



#56  
Dibenzofuran  
Concen: 21.048 ng/ul  
RT: 14.928 min Scan# 2030  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:168 Resp: 1041617  
Ion Ratio Lower Upper  
168 100  
139 36.2 29.4 44.2









#57  
2,4-Dinitrotoluene  
Concen: 19.772 ng/ul  
RT: 14.892 min Scan# 2024  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:165 Resp: 238758  
Ion Ratio Lower Upper  
165 100  
63 36.5 28.2 42.2  
182 2.5 2.2 3.4



#58  
2,3,4,6-Tetrachlorophenol  
Concen: 20.478 ng/ul  
RT: 15.151 min Scan# 2068  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:232 Resp: 212774  
Ion Ratio Lower Upper  
232 100  
131 44.1 36.2 54.2  
130 2.5 1.7 2.5  
166 30.2 25.0 37.4









#59  
Diethylphthalate  
Concen: 19.124 ng/ul  
RT: 15.351 min Scan# 2102  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:149 Resp: 812134  
Ion Ratio Lower Upper  
149 100  
177 22.4 17.8 26.8  
150 12.4 9.9 14.9



#61  
Fluorene  
Concen: 20.901 ng/ul  
RT: 15.575 min Scan# 2140  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:166 Resp: 847172  
Ion Ratio Lower Upper  
166 100  
165 98.0 77.2 115.8  
167 13.6 10.7 16.1









#62  
4-Chlorophenyl-phenylether  
Concen: 20.265 ng/ul  
RT: 15.569 min Scan# 2139  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:204 Resp: 406625  
Ion Ratio Lower Upper  
204 100  
206 32.8 26.8 40.2  
141 57.2 44.5 66.7



#63  
4-Nitroaniline  
Concen: 19.426 ng/ul  
RT: 15.604 min Scan# 2145  
Delta R.T. 0.000 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:138 Resp: 177500  
Ion Ratio Lower Upper  
138 100  
92 46.5 35.9 53.9  
108 66.2 51.6 77.4









#66  
4,6-Dinitro-2-methylphenol  
Concen: 18.417 ng/ul  
RT: 15.651 min Scan# 2154  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:198 Resp: 136155  
Ion Ratio Lower Upper  
198 100  
182 5.8 3.7 5.5#  
77 24.5 16.6 24.8



#67  
N-Nitrosodiphenylamine  
Concen: 21.033 ng/ul  
RT: 15.787 min Scan# 2176  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:169 Resp: 715340  
Ion Ratio Lower Upper  
169 100  
168 66.6 53.7 80.5  
167 36.1 29.9 44.9









#68  
4-Bromophenyl-phenylether  
Concen: 20.506 ng/ul  
RT: 16.469 min Scan# 2111  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:248 Resp: 240477  
Ion Ratio Lower Upper  
248 100  
250 99.1 78.6 118.0  
141 74.0 56.0 84.0



#69  
Hexachlorobenzene  
Concen: 21.013 ng/ul  
RT: 16.581 min Scan# 2311  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:284 Resp: 280425  
Ion Ratio Lower Upper  
284 100  
142 36.3 27.3 40.9  
249 30.7 25.2 37.8





#70

Atrazine

Concen: 19.418 ng/ul

RT: 16.745 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:200 Resp: 238290

Ion Ratio Lower Upper

200 100

173 26.5 21.6 32.4

215 48.3 38.9 58.3



#71

Pentachlorophenol

Concen: 19.358 ng/ul

RT: 16.928 min Scan# 2370

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:266 Resp: 169905


Ion Ratio Lower Upper

266 100

264 64.0 49.8 74.8

268 64.9 51.9 77.9





#72

Phenanthrene

Concen: 21.287 ng/ul

RT: 17.328 min Scan# 2438

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:178 Resp: 1348731

Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 19.5 15.4 23.2



#74

Anthracene

Concen: 21.459 ng/ul

RT: 17.416 min Scan# 2453

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:178 Resp: 1368004


Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 18.7 15.4 23.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.864 ng/uL

RT: 13.322 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:216 Resp: 363837

Ion Ratio Lower Upper

216 100

214 77.9 62.6 93.8

218 47.8 37.4 56.0



#76

Pentachlorobenzene

Concen: 21.746 ng/uL

RT: 14.845 min Scan# 2016

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:250 Resp: 369771


Ion Ratio Lower Upper

250 100

248 62.9 51.4 77.0

252 63.4 51.5 77.3





#77

Carbazole

Concen: 21.696 ng/ul

RT: 17.692 min Scan# 2500

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:167 Resp: 1249692

Ion Ratio Lower Upper

167 100

166 21.4 17.1 25.7

139 12.3 10.1 15.1



#78

Di-n-butylphthalate

Concen: 19.506 ng/ul

RT: 18.239 min Scan# 2593

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:149 Resp: 1375860


Ion Ratio Lower Upper

149 100

150 9.1 7.5 11.3

104 4.6 3.6 5.4





#80

Fluoranthene

Concen: 21.363 ng/ul

RT: 19.292 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:202 Resp: 1559423

Ion Ratio Lower Upper

202 100

101 12.0 9.8 14.6

100 9.0 7.4 11.0



#82

Pyrene

Concen: 21.766 ng/ul

RT: 19.645 min Scan# 2832

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:202 Resp: 1654244


Ion Ratio Lower Upper

202 100

101 14.1 11.6 17.4

100 11.5 9.4 14.0





#83

Butylbenzylphthalate

Concen: 19.568 ng/ul

RT: 20.516 min Scan# 2981

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:149 Resp: 642769

Ion Ratio Lower Upper

149 100

91 66.4 50.2 75.4

206 19.6 16.9 25.3



#84

3,3'-Dichlorobenzidine

Concen: 19.469 ng/ul

RT: 21.286 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:252 Resp: 527870


Ion Ratio Lower Upper

252 100

254 64.1 51.4 77.0

126 12.5 8.9 13.3





#85

Benzo(a)anthracene

Concen: 21.278 ng/ul

RT: 21.351 min Scan# 3123

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:228 Resp: 1638866

Ion Ratio Lower Upper

228 100

229 19.9 15.8 23.6

226 27.4 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.599 ng/ul

RT: 21.275 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:149 Resp: 920274


Ion Ratio Lower Upper

149 100

167 27.0 21.2 31.8

279 3.9 2.9 4.3





#87

Chrysene

Concen: 21.298 ng/ul

RT: 21.410 min Scan# 3132

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:228 Resp: 1539410

Ion Ratio Lower Upper

228 100

226 30.4 25.3 37.9

229 19.8 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 20.360 ng/ul




RT: 22.210 min Scan# 3268


Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Tgt Ion:149 Resp: 1568466





#90

Benzo(b)fluoranthene

Concen: 21.528 ng/ul

RT: 23.057 min Scan# 3412

Delta R.T. 0.000 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15

Instrument :

BNA\_P

ClientSampleId :

SSTD020769





Tgt Ion:252 Resp: 1659109

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 11.2 9.2 13.8



#91

Benzo(k)fluoranthene

Concen: 21.849 ng/ul

RT: 23.104 min Scan# 3420

Delta R.T. -0.006 min

Lab File: BP011972.D

Acq: 06 Oct 2022 16:15




Tgt Ion:252 Resp: 1621531


Ion Ratio Lower Upper

252 100

253 21.8 18.0 27.0

125 11.2 9.2 13.8









#93  
Benzo(a)pyrene  
Concen: 21.508 ng/ul  
RT: 23.692 min Scan# 31  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15




Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769


Tgt Ion:252 Resp: 1479220  
Ion Ratio Lower Upper  
252 100  
253 22.0 17.7 26.5  
125 12.6 9.8 14.8



#94  
Indeno(1,2,3-cd)pyrene  
Concen: 20.523 ng/ul  
RT: 26.327 min Scan# 3968  
Delta R.T. -0.012 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:276 Resp: 1796623  
Ion Ratio Lower Upper  
276 100  
138 26.2 21.4 32.0  
277 24.9 19.9 29.9









#95  
Dibenzo(a,h)anthracene  
Concen: 20.880 ng/ul  
RT: 26.351 min Scan# 3973  
Delta R.T. -0.006 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Instrument :  
BNA\_P  
ClientSampleId :  
SSTD020769

Tgt Ion:278 Resp: 1623751  
Ion Ratio Lower Upper  
278 100  
139 17.8 15.0 22.6  
279 24.0 19.0 28.6



#96  
Benzo(g,h,i)perylene  
Concen: 20.539 ng/ul  
RT: 27.104 min Scan# 4100  
Delta R.T. -0.012 min  
Lab File: BP011972.D  
Acq: 06 Oct 2022 16:15

Tgt Ion:276 Resp: 1594046  
Ion Ratio Lower Upper  
276 100  
138 24.2 19.8 29.6  
277 23.7 18.6 27.8

