

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
 Data File : BP011978.D
 Acq On : 06 Oct 2022 19:44
 Operator : CG/JU
 Sample : N4923-12
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E10041

Quant Time: Oct 06 23:08:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 05 23:18:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	250377	20.000	ng/ul	0.00
20) Naphthalene-d8	10.687	136	1030597	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.522	164	614777	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.281	188	1286200	20.000	ng/ul	0.00
79) Chrysene-d12	21.369	240	1280389	20.000	ng/ul	0.00
88) Perylene-d12	23.798	264	1289723	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	24727	4.356	ng/uL	0.00
4) Pyridine-d5	3.734	84	72787	4.377	ng/ul	0.01
7) Phenol-d5	7.046	99	107412	4.969	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	314388	26.240	ng/ul	0.00
11) 2-Chlorophenol-d4	7.410	132	340796	19.973	ng/ul	0.00
15) 4-Methylphenol-d8	8.593	113	184660	10.335	ng/ul	0.00
21) Nitrobenzene-d5	9.046	128	220231	28.055	ng/ul	0.00
24) 2-Nitrophenol-d4	9.769	143	217703	25.921	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.304	165	357808	23.033	ng/ul	0.00
31) 4-Chloroaniline-d4	10.828	131	387034	16.343	ng/ul	0.00
46) Dimethylphthalate-d6	13.934	166	1279559	29.105	ng/ul	0.00
49) Acenaphthylene-d8	14.216	160	1441974	28.772	ng/ul	0.00
54) 4-Nitrophenol-d4	14.745	143	28487	3.129	ng/ul	0.02
60) Fluorene-d10	15.516	176	1200661	31.448	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	174917	22.469	ng/ul	0.00
73) Anthracene-d10	17.381	188	1936581	33.054	ng/ul	0.00
81) Pyrene-d10	19.616	212	2308729	35.604	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.639	264	2338365	36.219	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
Data File : BP011978.D
Acq On : 06 Oct 2022 19:44
Operator : CG/JU
Sample : N4923-12
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E10041

Quant Time: Oct 06 23:08:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 05 23:18:55 2022
Response via : Initial Calibration

