

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
 Data File : BP011928.D
 Acq On : 05 Oct 2022 13:56
 Operator : CG/JU
 Sample : N4890-03MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EY0E1MS

Quant Time: Oct 05 23:21:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 05 23:18:55 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/06/2022
 Supervised By :mohammad ahmed 10/10/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	247519	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	995173	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	577878	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	1201641	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	1285110	20.000	ng/u1	0.00
88) Perylene-d12	23.804	264	1302407	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	24479	4.362	ng/uL	0.00
4) Pyridine-d5	3.722	84	166688	10.140	ng/u1	0.00
7) Phenol-d5	7.046	99	145232	6.797	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	373346	31.521	ng/u1	0.00
11) 2-Chlorophenol-d4	7.410	132	415820	24.651	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	254074	14.384	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	264698	34.920	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	263024	32.433	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	470253	31.349	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	616542	26.960	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	1621912	39.248	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1830317	38.852	ng/u1	0.00
54) 4-Nitrophenol-d4	14.734	143	71277	8.329	ng/u1	0.00
60) Fluorene-d10	15.522	176	1439144	40.101	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	267391	36.764	ng/u1	0.00
73) Anthracene-d10	17.386	188	2281070	41.673	ng/u1	0.00
81) Pyrene-d10	19.622	212	2774343	42.627	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.645	264	2861042	43.883	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	31074	5.126	ng/uL	92
5) Pyridine	3.746	79	166626	10.443	ng/u1	97
6) Benzaldehyde	7.022	77	370623	39.594	ng/u1	99
8) Phenol	7.075	94	149469	6.744	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.305	93	489643	27.765	ng/u1	99
12) 2-Chlorophenol	7.446	128	408708	22.664	ng/u1	99
13) 2-Methylphenol	8.328	108	277570	15.896	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.416	45	475066	27.325	ng/u1	99
16) Acetophenone	8.710	105	717034	26.624	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.693	70	323669	25.535	ng/u1	99
18) 4-Methylphenol	8.657	108	262391	13.638	ng/u1	99
19) Hexachloroethane	8.963	117	197815	29.411	ng/u1	98
22) Nitrobenzene	9.093	77	511457	31.373	ng/u1	99
23) Isophorone	9.622	82	953143	28.821	ng/u1	99
25) 2-Nitrophenol	9.804	139	265309	28.643	ng/u1	98
26) 2,4-Dimethylphenol	9.863	107	342399	19.678	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.104	93	627236	29.398	ng/u1	99
29) 2,4-Dichlorophenol	10.340	162	434571	27.959	ng/u1	99
30) Naphthalene	10.740	128	1650981	30.887	ng/u1	99
32) 4-Chloroaniline	10.857	127	561045	23.981	ng/u1	99
33) Hexachlorobutadiene	11.022	225	288003	31.305	ng/u1	97
34) Caprolactam	11.651	113	18065m	3.284	ng/u1	
35) 4-Chloro-3-methylphenol	11.981	107	413227	25.074	ng/u1	100
36) 2-Methylnaphthalene	12.351	142	1109078	29.711	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
 Data File : BP011928.D
 Acq On : 05 Oct 2022 13:56
 Operator : CG/JU
 Sample : N4890-03MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EY0E1MS


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/06/2022
 Supervised By :mohammad ahmed 10/10/2022

Quant Time: Oct 05 23:21:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 05 23:18:55 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.569	142	1132302	30.230	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.716	216	578558	34.523	ng/ul	99
40) Hexachlorocyclopentadiene	12.698	237	227882	23.803	ng/ul	98
41) 2,4,6-Trichlorophenol	12.957	196	378159	33.710	ng/ul	98
42) 2,4,5-Trichlorophenol	13.028	196	412633	33.767	ng/ul	97
43) 1,1'-Biphenyl	13.363	154	1460480	34.169	ng/ul	100
44) 2-Chloronaphthalene	13.404	162	1162287	34.394	ng/ul	100
45) 2-Nitroaniline	13.616	65	293846	35.622	ng/ul	95
47) Dimethylphthalate	13.987	163	1507048	34.798	ng/ul	99
48) 2,6-Dinitrotoluene	14.110	165	308512	36.553	ng/ul	99
50) Acenaphthylene	14.251	152	1815532	34.700	ng/ul	99
51) 3-Nitroaniline	14.439	138	314900	32.189	ng/ul	97
52) Acenaphthene	14.592	153	1297074	35.138	ng/ul	99
53) 2,4-Dinitrophenol	14.645	184	162270	28.358	ng/ul	97
55) 4-Nitrophenol	14.745	109	44677	7.734	ng/ul	98
56) Dibenzofuran	14.928	168	1796089	35.524	ng/ul	97
57) 2,4-Dinitrotoluene	14.898	165	446847	36.220	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.157	232	360027	33.915	ng/ul	98
59) Diethylphthalate	15.351	149	1500019	34.573	ng/ul	99
61) Fluorene	15.581	166	1502879	36.292	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.575	204	721591	35.199	ng/ul	99
63) 4-Nitroaniline	15.604	138	352595	37.770	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.657	198	252942	32.828	ng/ul	98
67) N-Nitrosodiphenylamine	15.792	169	1146924	32.356	ng/ul	100
68) 4-Bromophenyl-phenylether	16.475	248	446673	36.544	ng/ul	98
69) Hexachlorobenzene	16.586	284	517688	37.220	ng/ul	99
70) Atrazine	16.745	200	428645	33.514	ng/ul	99
71) Pentachlorophenol	16.933	266	290916	31.801	ng/ul	98
72) Phenanthrene	17.328	178	2501079	37.874	ng/ul	100
74) Anthracene	17.422	178	2496243	37.569	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.328	216	618917	37.316	ng/ul	99
76) Pentachlorobenzene	14.845	250	580113	32.732	ng/ul	99
77) Carbazole	17.692	167	2354105	39.213	ng/ul	99
78) Di-n-butylphthalate	18.239	149	2665917	36.262	ng/ul	100
80) Fluoranthene	19.292	202	2989603	37.534	ng/ul	99
82) Pyrene	19.651	202	3163115	38.142	ng/ul	97
83) Butylbenzylphthalate	20.521	149	1261273	35.190	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.292	252	788928	26.667	ng/ul	100
85) Benzo(a)anthracene	21.357	228	3214189	38.246	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.280	149	1866245	34.567	ng/ul#	97
87) Chrysene	21.410	228	3009390	38.157	ng/ul	99
89) Di-n-octyl phthalate	22.210	149	3168056	38.126	ng/ul	100
90) Benzo(b)fluoranthene	23.063	252	3372635	40.571	ng/ul	99
91) Benzo(k)fluoranthene	23.110	252	3194262	39.903	ng/ul	99
93) Benzo(a)pyrene	23.698	252	2900621	39.100	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.345	276	3586919	37.987	ng/ul	99
95) Dibenzo(a,h)anthracene	26.356	278	3247101	38.710	ng/ul	98
96) Benzo(g,h,i)perylene	27.121	276	3172114	37.893	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 5.126 ng/uL
RT: 3.334 min Scan# 59
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion: 88 Resp: 31074
Ion Ratio Lower Upper
88 100
43 21.2 15.6 23.4
58 67.7 48.6 72.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#5
Pyridine
Concen: 10.443 ng/uL
RT: 3.746 min Scan# 129
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion: 79 Resp: 166626
Ion Ratio Lower Upper
79 100
52 59.0 49.7 74.5
51 26.4 21.6 32.4





#6
Benzaldehyde
Concen: 39.594 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion: 77 Resp: 37062
Ion Ratio Lower Upper
77 100
105 108.0 87.0 130.6
106 104.0 83.8 125.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#8
Phenol
Concen: 6.744 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion: 94 Resp: 149469
Ion Ratio Lower Upper
94 100
65 24.9 19.5 29.3
66 31.2 25.0 37.4





#10
Bis(2-Chloroethyl)ether
Concen: 27.765 ng/ul
RT: 7.305 min Scan# 734
Delta R.T. -0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion: 93 Resp: 48964
Ion Ratio Lower Upper
93 100
63 69.9 56.2 84.4
95 32.8 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#12
2-Chlorophenol
Concen: 22.664 ng/ul
RT: 7.446 min Scan# 758
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Tgt Ion:128 Resp: 408708
Ion Ratio Lower Upper
128 100
64 39.3 31.1 46.7
130 32.3 25.5 38.3









#13
2-Methylphenol
Concen: 15.896 ng/ul
RT: 8.328 min Scan# 908
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion:108 Resp: 277570
Ion Ratio Lower Upper
108 100
107 88.1 70.2 105.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.325 ng/ul
RT: 8.416 min Scan# 923
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Tgt Ion: 45 Resp: 475066
Ion Ratio Lower Upper
45 100
77 19.3 16.0 24.0
79 15.1 12.5 18.7









#16
Acetophenone
Concen: 26.624 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion: 105 Resp: 717034
Ion Ratio Lower Upper
105 100
77 70.2 55.4 83.0
51 18.9 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#17
N-Nitroso-di-n-propylamine
Concen: 25.535 ng/ul
RT: 8.693 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion: 70 Resp: 323669
Ion Ratio Lower Upper
70 100
42 41.3 32.7 49.1
101 10.0 8.4 12.6
130 24.3 19.4 29.2





#18
4-Methylphenol
Concen: 13.638 ng/ul
RT: 8.657 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:108 Resp: 26239
Ion Ratio Lower Upper
108 100
107 109.2 88.0 132.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#19
Hexachloroethane
Concen: 29.411 ng/ul
RT: 8.963 min Scan# 1016
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:117 Resp: 197815
Ion Ratio Lower Upper
117 100
201 100.3 82.4 123.6
199 64.1 51.6 77.4





#22
Nitrobenzene
Concen: 31.373 ng/ul
RT: 9.093 min Scan# 1038
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion: 77 Resp: 51145
Ion Ratio Lower Upper
77 100
123 52.5 41.6 62.4
65 12.1 9.7 14.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#23
Isophorone
Concen: 28.821 ng/ul
RT: 9.622 min Scan# 1128
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Tgt Ion: 82 Resp: 953143
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.2
138 20.1 15.8 23.8









#25
2-Nitrophenol
Concen: 28.643 ng/ul
RT: 9.804 min Scan# 1159
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion:139 Resp: 265309
Ion Ratio Lower Upper
139 100
65 41.8 32.2 48.4
109 25.1 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#26
2,4-Dimethylphenol
Concen: 19.678 ng/ul
RT: 9.863 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:107 Resp: 342399
Ion Ratio Lower Upper
107 100
121 53.9 43.0 64.6
122 92.2 74.9 112.3





#27
Bis(2-Chloroethoxy)methane
Concen: 29.398 ng/ul
RT: 10.104 min Scan# 1210
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion: 93 Resp: 627230
Ion Ratio Lower Upper
93 100
95 32.4 26.3 39.5
123 12.2 9.5 14.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#29
2,4-Dichlorophenol
Concen: 27.959 ng/ul
RT: 10.340 min Scan# 1250
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:162 Resp: 434571
Ion Ratio Lower Upper
162 100
164 65.7 52.2 78.4
98 32.6 27.0 40.4





#30
Naphthalene
Concen: 30.887 ng/ul
RT: 10.740 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:128 Resp: 165098
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.7 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#32
4-Chloroaniline
Concen: 23.981 ng/ul
RT: 10.857 min Scan# 1338
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:127 Resp: 561045
Ion Ratio Lower Upper
127 100
129 32.4 26.6 39.8





#33
Hexachlorobutadiene
Concen: 31.305 ng/ul
RT: 11.022 min Scan# 1367
Delta R.T. -0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:225 Resp: 288000
Ion Ratio Lower Upper
225 100
223 62.4 52.1 78.1
227 63.7 52.3 78.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#34
Caprolactam
Concen: 3.284 ng/ul m
RT: 11.651 min Scan# 1473
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Tgt Ion:113 Resp: 18065
Ion Ratio Lower Upper
113 100
55 136.4 106.0 159.0
56 98.6 78.7 118.1









#35
4-Chloro-3-methylphenol
Concen: 25.074 ng/ul
RT: 11.981 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion:107 Resp: 41322
Ion Ratio Lower Upper
107 100
144 28.1 22.5 33.7
142 87.2 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#36
2-Methylnaphthalene
Concen: 29.711 ng/ul
RT: 12.351 min Scan# 1592
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:142 Resp: 1109078
Ion Ratio Lower Upper
142 100
141 88.4 70.6 106.0





#37
1-Methylnaphthalene
Concen: 30.230 ng/ul
RT: 12.569 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:142 Resp: 113230
Ion Ratio Lower Upper
142 100
141 91.3 74.8 112.2
116 3.5 2.9 4.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.523 ng/ul
RT: 12.716 min Scan# 1654
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:216 Resp: 578558
Ion Ratio Lower Upper
216 100
214 79.2 62.6 94.0
179 20.6 16.6 24.8
108 16.4 12.6 19.0





#40
Hexachlorocyclopentadiene
Concen: 23.803 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:237 Resp: 227882
Ion Ratio Lower Upper
237 100
235 61.9 51.0 76.6
272 13.5 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#41
2,4,6-Trichlorophenol
Concen: 33.710 ng/ul
RT: 12.957 min Scan# 1695
Delta R.T. -0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:196 Resp: 378159
Ion Ratio Lower Upper
196 100
198 95.9 78.7 118.1
200 30.5 24.5 36.7





#42
2,4,5-Trichlorophenol
Concen: 33.767 ng/ul
RT: 13.028 min Scan# 1707
Delta R.T. -0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:196 Resp: 41263
Ion Ratio Lower Upper
196 100
198 93.4 76.6 114.8
200 30.2 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#43
1,1'-Biphenyl
Concen: 34.169 ng/ul
RT: 13.363 min Scan# 1764
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:154 Resp: 1460480
Ion Ratio Lower Upper
154 100
153 39.9 31.8 47.6
76 11.8 9.5 14.3





#44
2-Chloronaphthalene
Concen: 34.394 ng/ul
RT: 13.404 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:162 Resp: 116228
Ion Ratio Lower Upper
162 100
164 32.7 26.5 39.7
127 37.2 29.7 44.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#45
2-Nitroaniline
Concen: 35.622 ng/ul
RT: 13.616 min Scan# 1807
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion: 65 Resp: 293846
Ion Ratio Lower Upper
65 100
92 78.5 61.2 91.8
138 135.5 103.0 154.4





#47
Dimethylphthalate
Concen: 34.798 ng/ul
RT: 13.987 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:163 Resp: 1507048
Ion Ratio Lower Upper
163 100
194 3.9 3.3 4.9
164 10.2 8.5 12.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#48
2,6-Dinitrotoluene
Concen: 36.553 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:165 Resp: 308512
Ion Ratio Lower Upper
165 100
89 48.3 39.1 58.7
121 20.4 16.7 25.1





#50
Acenaphthylene
Concen: 34.700 ng/ul
RT: 14.251 min Scan# 1915
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:152 Resp: 181553
Ion Ratio Lower Upper
152 100
151 19.7 16.0 24.0
153 13.2 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#51
3-Nitroaniline
Concen: 32.189 ng/ul
RT: 14.439 min Scan# 1947
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:138 Resp: 314900
Ion Ratio Lower Upper
138 100
108 9.6 8.2 12.4
92 104.2 85.6 128.4





#52
Acenaphthene
Concen: 35.138 ng/ul
RT: 14.592 min Scan# 1973
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:153 Resp: 1297074
Ion Ratio Lower Upper
153 100
152 47.2 38.4 57.6
154 87.7 70.6 105.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#53
2,4-Dinitrophenol
Concen: 28.358 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:184 Resp: 162270
Ion Ratio Lower Upper
184 100
63 50.7 44.2 66.4
154 57.0 45.8 68.6





#55
4-Nitrophenol
Concen: 7.734 ng/ul
RT: 14.745 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:109 Resp: 4467
Ion Ratio Lower Upper
109 100
139 154.1 125.0 187.6
65 125.2 101.9 152.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#56
Dibenzofuran
Concen: 35.524 ng/ul
RT: 14.928 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Tgt Ion:168 Resp: 1796089
Ion Ratio Lower Upper
168 100
139 35.1 29.4 44.2









#57
2,4-Dinitrotoluene
Concen: 36.220 ng/ul
RT: 14.898 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion:165 Resp: 44684
Ion Ratio Lower Upper
165 100
63 33.3 28.2 42.2
182 2.9 2.2 3.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 33.915 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:232 Resp: 360027
Ion Ratio Lower Upper
232 100
131 44.5 36.2 54.2
130 2.0 1.7 2.5
166 29.9 25.0 37.4





#59
Diethylphthalate
Concen: 34.573 ng/ul
RT: 15.351 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:149 Resp: 1500019
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 12.6 9.9 14.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#61
Fluorene
Concen: 36.292 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:166 Resp: 1502879
Ion Ratio Lower Upper
166 100
165 96.9 77.2 115.8
167 13.5 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 35.199 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion: 204 Resp: 72159
Ion Ratio Lower Upper
204 100
206 32.8 26.8 40.2
141 55.4 44.5 66.7

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 10/06/2022
Supervised By : mohammad ahmed 10/10/2022




#63
4-Nitroaniline
Concen: 37.770 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56



Tgt Ion: 138 Resp: 352595
Ion Ratio Lower Upper
138 100
92 45.1 35.9 53.9
108 64.4 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 32.828 ng/ul
RT: 15.657 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion:198 Resp: 25294
Ion Ratio Lower Upper
198 100
182 4.4 3.7 5.5
77 21.6 16.6 24.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#67
N-Nitrosodiphenylamine
Concen: 32.356 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:169 Resp: 1146924
Ion Ratio Lower Upper
169 100
168 67.3 53.7 80.5
167 36.8 29.9 44.9





#68
4-Bromophenyl-phenylether
Concen: 36.544 ng/ul
RT: 16.475 min Scan# 2293
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:248 Resp: 446673
Ion Ratio Lower Upper
248 100
250 97.2 78.6 118.0
141 67.0 56.0 84.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#69
Hexachlorobenzene
Concen: 37.220 ng/ul
RT: 16.586 min Scan# 2312
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:284 Resp: 517688
Ion Ratio Lower Upper
284 100
142 34.0 27.3 40.9
249 31.0 25.2 37.8





#70
Atrazine
Concen: 33.514 ng/ul
RT: 16.745 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:200 Resp: 42864
Ion Ratio Lower Upper
200 100
173 26.8 21.6 32.4
215 48.0 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#71
Pentachlorophenol
Concen: 31.801 ng/ul
RT: 16.933 min Scan# 2371
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:266 Resp: 290916
Ion Ratio Lower Upper
266 100
264 64.5 49.8 74.8
268 64.1 51.9 77.9





#72
Phenanthrene
Concen: 37.874 ng/ul
RT: 17.328 min Scan# 2438
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:178 Resp: 2501079
Ion Ratio Lower Upper
178 100
179 15.6 12.5 18.7
176 19.6 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#74
Anthracene
Concen: 37.569 ng/ul
RT: 17.422 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Tgt Ion:178 Resp: 2496243
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 18.7 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.316 ng/uL
RT: 13.328 min Scan# 1758
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56

Instrument :
BNA_P
ClientSampleId :
EY0E1MS

Tgt Ion:216 Resp: 61891
Ion Ratio Lower Upper
216 100
214 78.3 62.6 93.8
218 48.2 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#76
Pentachlorobenzene
Concen: 32.732 ng/uL
RT: 14.845 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:250 Resp: 580113
Ion Ratio Lower Upper
250 100
248 62.6 51.4 77.0
252 64.1 51.5 77.3





#77
 Carbazole
 Concen: 39.213 ng/ul
 RT: 17.692 min Scan# 2500
 Delta R.T. -0.000 min
 Lab File: BP011928.D
 Acq: 05 Oct 2022 13:56




Instrument :
 BNA_P
 ClientSampleId :
 EY0E1MS




Tgt Ion:167 Resp: 235410
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.1 25.7
 139 12.7 10.1 15.1

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
 Supervised By :mohammad ahmed 10/10/2022




#78
 Di-n-butylphthalate
 Concen: 36.262 ng/ul
 RT: 18.239 min Scan# 2593
 Delta R.T. -0.000 min
 Lab File: BP011928.D
 Acq: 05 Oct 2022 13:56




Tgt Ion:149 Resp: 2665917
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 4.3 3.6 5.4





#80
Fluoranthene
Concen: 37.534 ng/u1
RT: 19.292 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:202 Resp: 298960
Ion Ratio Lower Upper
202 100
101 12.0 9.8 14.6
100 8.9 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#82
Pyrene
Concen: 38.142 ng/u1
RT: 19.651 min Scan# 2833
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:202 Resp: 3163115
Ion Ratio Lower Upper
202 100
101 13.1 11.6 17.4
100 10.5 9.4 14.0





#83
Butylbenzylphthalate
Concen: 35.190 ng/ul
RT: 20.521 min Scan# 2981
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:149 Resp: 126127
Ion Ratio Lower Upper
149 100
91 63.4 50.2 75.4
206 21.1 16.9 25.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#84
3,3'-Dichlorobenzidine
Concen: 26.667 ng/ul
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:252 Resp: 788928
Ion Ratio Lower Upper
252 100
254 64.4 51.4 77.0
126 11.0 8.9 13.3





#85
Benzo(a)anthracene
Concen: 38.246 ng/ul
RT: 21.357 min Scan# 3123
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:228 Resp: 3214189
Ion Ratio Lower Upper
228 100
229 20.1 15.8 23.6
226 28.2 22.0 33.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 34.567 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:149 Resp: 1866245
Ion Ratio Lower Upper
149 100
167 27.9 21.2 31.8
279 4.4 2.9 4.3





#87
Chrysene
Concen: 38.157 ng/ul
RT: 21.410 min Scan# 3132
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:228 Resp: 3009390
Ion Ratio Lower Upper
228 100
226 30.9 25.3 37.9
229 20.1 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#89
Di-n-octyl phthalate
Concen: 38.126 ng/ul
RT: 22.210 min Scan# 3268
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:149 Resp: 3168056





#90
Benzo(b)fluoranthene
Concen: 40.571 ng/ul
RT: 23.063 min Scan# 3412
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS



Tgt Ion:252 Resp: 337263
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 10.6 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#91
Benzo(k)fluoranthene
Concen: 39.903 ng/ul
RT: 23.110 min Scan# 3421
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56


Tgt Ion:252 Resp: 3194262
Ion Ratio Lower Upper
252 100
253 22.6 18.0 27.0
125 10.6 9.2 13.8





#93
Benzo(a)pyrene
Concen: 39.100 ng/ul
RT: 23.698 min Scan# 3521
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:252 Resp: 290062
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 12.0 9.8 14.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 37.987 ng/ul
RT: 26.345 min Scan# 3971
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Tgt Ion:276 Resp: 3586919
Ion Ratio Lower Upper
276 100
138 27.6 21.4 32.0
277 25.5 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 38.710 ng/ul
RT: 26.356 min Scan# 3973
Delta R.T. -0.000 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56




Instrument :
BNA_P
ClientSampleId :
EY0E1MS




Tgt Ion:278 Resp: 324710
Ion Ratio Lower Upper
278 100
139 17.6 15.0 22.6
279 24.1 19.0 28.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/06/2022
Supervised By :mohammad ahmed 10/10/2022



#96
Benzo(g,h,i)perylene
Concen: 37.893 ng/ul
RT: 27.121 min Scan# 4103
Delta R.T. 0.006 min
Lab File: BP011928.D
Acq: 05 Oct 2022 13:56



Tgt Ion:276 Resp: 3172114
Ion Ratio Lower Upper
276 100
138 24.0 19.8 29.6
277 23.8 18.6 27.8

