

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
 Data File : BP011932.D
 Acq On : 05 Oct 2022 16:13
 Operator : CG/JU
 Sample : N4890-07
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EY0E0

Quant Time: Oct 05 23:23:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 05 23:18:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	222947	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	913152	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	547044	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	1166041	20.000	ng/u1	0.00
79) Chrysene-d12	21.368	240	1163391	20.000	ng/u1	0.00
88) Perylene-d12	23.804	264	1177169	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	25640	5.072	ng/uL	0.00
4) Pyridine-d5	3.734	84	41663	2.814	ng/u1	0.01
7) Phenol-d5	7.046	99	130864	6.799	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	341131	31.976	ng/u1	0.00
11) 2-Chlorophenol-d4	7.410	132	388487	25.569	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	236576	14.869	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	231507	33.285	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	234614	31.528	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	407848	29.631	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	245815	11.715	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	1379845	35.273	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1579284	35.413	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	41880	5.170	ng/u1	0.01
60) Fluorene-d10	15.522	176	1287382	37.894	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	197377	27.966	ng/u1	0.00
73) Anthracene-d10	17.386	188	2240034	42.173	ng/u1	0.00
81) Pyrene-d10	19.616	212	2609481	44.289	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.639	264	2573592	43.674	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100522\
Data File : BP011932.D
Acq On : 05 Oct 2022 16:13
Operator : CG/JU
Sample : N4890-07
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EY0E0

Quant Time: Oct 05 23:23:05 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP093022.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 05 23:18:55 2022
Response via : Initial Calibration

