

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP100719\
 Data File : BP000545.D
 Acq On : 07 Oct 2019 16:06
 Operator : JU
 Sample : K5231-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 20190766-SS-COMPOSITE-6

Quant Time: Oct 09 12:32:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP100119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 01 18:03:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	267498	20.00	ng	-0.01
21) Naphthalene-d8	8.05	136	959944	20.00	ng	-0.01
39) Acenaphthene-d10	9.80	164	531597	20.00	ng	-0.01
64) Phenanthrene-d10	11.30	188	1012482	20.00	ng	0.00
76) Chrysene-d12	13.97	240	908181	20.00	ng	-0.01
87) Perylene-d12	15.45	264	989538	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	1940472	116.61	ng	0.00
7) Phenol-d6	6.40	99	2482065	123.77	ng	0.00
23) Nitrobenzene-d5	7.32	82	1348693	85.81	ng	-0.01
42) 2,4,6-Tribromophenol	10.60	330	710125	125.36	ng	0.00
45) 2-Fluorobiphenyl	9.13	172	2788350	84.87	ng	0.00
79) Terphenyl-d14	12.91	244	3389823	75.57	ng	0.00
Target Compounds						
10) Phenol	6.41	94	45611	2.221	ng	86
50) Dimethylphthalate	9.52	163	373507	10.112	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP100719\
Data File : BP000545.D
Acq On : 07 Oct 2019 16:06
Operator : JU
Sample : K5231-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
20190766-SS-COMPOSITE-6

Quant Time: Oct 09 12:32:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP100119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 01 18:03:42 2019
Response via : Initial Calibration

