

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017622.D
 Acq On : 09 Oct 2023 17:56
 Operator : MA/JU
 Sample : SSTD01067
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Quant Time: Oct 10 04:01:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	150765	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	604203	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.622	164	397160	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.416	188	947649	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	981638	20.000	ng/u1	0.00
88) Perylene-d12	24.133	264	1074740	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	12829	3.403	ng/uL	0.00
4) Pyridine-d5	3.722	84	89107	9.024	ng/u1	0.00
7) Phenol-d5	7.081	99	103561	8.199	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	69723	9.023	ng/u1	0.00
11) 2-Chlorophenol-d4	7.446	132	86135	8.619	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	82317	8.159	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	39849	8.619	ng/u1	0.00
24) 2-Nitrophenol-d4	9.852	143	41173	7.969	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	86378	8.889	ng/u1	0.00
31) 4-Chloroaniline-d4	10.940	131	116944	8.457	ng/u1	0.00
46) Dimethylphthalate-d6	14.040	166	277912	8.961	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	296624	8.676	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	29615	5.743	ng/u1	0.00
60) Fluorene-d10	15.628	176	247216	9.334	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.769	200	37151	6.509	ng/u1	0.00
73) Anthracene-d10	17.516	188	394673	8.965	ng/u1	0.00
81) Pyrene-d10	19.769	212	486831	8.912	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	482932	8.876	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	14624	3.771	ng/uL	95
5) Pyridine	3.740	79	95012	9.183	ng/u1	96
6) Benzaldehyde	7.087	77	67146	10.557	ng/u1	96
8) Phenol	7.111	94	110340	8.712	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.363	93	91128	8.887	ng/u1	96
12) 2-Chlorophenol	7.481	128	90679	8.945	ng/u1	98
13) 2-Methylphenol	8.381	108	77996	8.249	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.440	45	68893	5.109	ng/u1	99
16) Acetophenone	8.787	105	137582	8.659	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.752	70	69647	8.529	ng/u1	99
18) 4-Methylphenol	8.716	108	87837	8.560	ng/u1	100
19) Hexachloroethane	8.993	117	40275	8.908	ng/u1	96
22) Nitrobenzene	9.181	77	107910	9.020	ng/u1	97
23) Isophorone	9.693	82	183385	8.450	ng/u1	99
25) 2-Nitrophenol	9.887	139	46050	8.311	ng/u1	96
26) 2,4-Dimethylphenol	9.928	107	98284	8.900	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.181	93	118388	8.888	ng/u1	98
29) 2,4-Dichlorophenol	10.410	162	84531	8.923	ng/u1	93
30) Naphthalene	10.816	128	304296	9.383	ng/u1	99
32) 4-Chloroaniline	10.963	127	115752	8.767	ng/u1	100
33) Hexachlorobutadiene	11.040	225	66304	9.479	ng/u1	95
34) Caprolactam	11.793	113	20643	7.200	ng/u1	92
35) 4-Chloro-3-methylphenol	12.075	107	85759	8.579	ng/u1	98
36) 2-Methylnaphthalene	12.428	142	205254	9.308	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017622.D
 Acq On : 09 Oct 2023 17:56
 Operator : MA/JU
 Sample : SSTD01067
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Quant Time: Oct 10 04:01:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	209327	9.206	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.781	216	122718	9.346	ng/ul	99
40) Hexachlorocyclopentadiene	12.722	237	43276	6.036	ng/ul	94
41) 2,4,6-Trichlorophenol	13.046	196	70833	8.526	ng/ul	98
42) 2,4,5-Trichlorophenol	13.122	196	71935	8.234	ng/ul	96
43) 1,1'-Biphenyl	13.445	154	279134	9.154	ng/ul	99
44) 2-Chloronaphthalene	13.498	162	223887	9.199	ng/ul	97
45) 2-Nitroaniline	13.745	65	53020	7.962	ng/ul	93
47) Dimethylphthalate	14.087	163	279645	8.984	ng/ul	100
48) 2,6-Dinitrotoluene	14.240	165	46113	8.026	ng/ul	94
50) Acenaphthylene	14.351	152	346499	8.776	ng/ul	100
51) 3-Nitroaniline	14.587	138	42353	7.781	ng/ul	94
52) Acenaphthene	14.687	153	237569	9.084	ng/ul	98
53) 2,4-Dinitrophenol	14.792	184	49133	15.236	ng/ul	90
55) 4-Nitrophenol	14.887	109	38875	8.359	ng/ul	97
56) Dibenzofuran	15.028	168	345988	9.365	ng/ul	97
57) 2,4-Dinitrotoluene	15.028	165	73847	8.552	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.251	232	67625	8.745	ng/ul	96
59) Diethylphthalate	15.445	149	279554	8.947	ng/ul	99
61) Fluorene	15.687	166	281923	9.265	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.675	204	152028	9.629	ng/ul	97
63) 4-Nitroaniline	15.757	138	42250	8.294	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.787	198	42138	6.878	ng/ul	95
67) N-Nitrosodiphenylamine	15.904	169	235278	8.887	ng/ul	99
68) 4-Bromophenyl-phenylether	16.586	248	94196	9.234	ng/ul	96
69) Hexachlorobenzene	16.669	284	113756	9.478	ng/ul	94
70) Atrazine	16.869	200	83244	8.127	ng/ul	98
71) Pentachlorophenol	17.039	266	53796	7.552	ng/ul	94
72) Phenanthrene	17.457	178	467523	9.111	ng/ul	99
74) Anthracene	17.551	178	474835	9.112	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.398	216	124018	8.798	ng/uL	100
76) Pentachlorobenzene	14.916	250	128412	9.109	ng/uL	97
77) Carbazole	17.845	167	400167	8.574	ng/ul	98
78) Di-n-butylphthalate	18.351	149	451702	8.133	ng/ul	99
80) Fluoranthene	19.439	202	567894	8.945	ng/ul	100
82) Pyrene	19.798	202	611197	9.131	ng/ul	98
83) Butylbenzylphthalate	20.669	149	191166	7.460	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.480	252	155045	7.188	ng/ul	98
85) Benzo(a)anthracene	21.533	228	625111	9.054	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.410	149	265276	7.224	ng/ul	99
87) Chrysene	21.592	228	591334	9.197	ng/ul	100
89) Di-n-octyl phthalate	22.398	149	451328	6.820	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	603819	8.933	ng/ul	98
91) Benzo(k)fluoranthene	23.380	252	623735	9.187	ng/ul	100
93) Benzo(a)pyrene	24.010	252	557748	8.811	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.862	276	670084	8.714	ng/ul	98
95) Dibenzo(a,h)anthracene	26.892	278	542797	8.439	ng/ul	100
96) Benzo(g,h,i)perylene	27.715	276	547914	8.866	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017622.D
 Acq On : 09 Oct 2023 17:56
 Operator : MA/JU
 Sample : SSTD01067
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Quant Time: Oct 10 04:01:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.771 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:	88	Resp:	14624
Ion	Ratio	Lower	Upper
88	100		
43	30.9	24.3	36.5
58	72.3	53.4	80.0



#5
Pyridine
Concen: 9.183 ng/uL
RT: 3.740 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:	79	Resp:	95012
Ion	Ratio	Lower	Upper
79	100		
52	72.4	55.4	83.2
51	38.0	28.2	42.2









#6
Benzaldehyde
Concen: 10.557 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 77 Resp: 67146
Ion Ratio Lower Upper
77 100
105 87.9 75.0 112.6
106 90.5 71.8 107.6



#8
Phenol
Concen: 8.712 ng/ul
RT: 7.111 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion: 94 Resp: 110340
Ion Ratio Lower Upper
94 100
65 28.0 24.2 36.4
66 42.4 33.0 49.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 8.887 ng/ul
 RT: 7.363 min Scan# 727
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Tgt Ion: 93	Resp:	91128
Ion	Ratio	Lower Upper
93	100	
63	72.5	59.9 89.9
95	30.9	27.2 40.8



#12
 2-Chlorophenol
 Concen: 8.945 ng/ul
 RT: 7.481 min Scan# 747
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56

Tgt Ion: 128	Resp:	90679
Ion	Ratio	Lower Upper
128	100	
64	48.1	37.0 55.4
130	33.0	25.8 38.6





#13

2-Methylphenol

Concen: 8.249 ng/ul

RT: 8.381 min Scan# 900

Delta R.T. 0.006 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :





SSTD010632

Tgt Ion:108 Resp: 77996

Ion Ratio Lower Upper

108 100

107 95.2 73.4 110.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 5.109 ng/ul

RT: 8.440 min Scan# 910

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion: 45 Resp: 68893


Ion Ratio Lower Upper

45 100

77 17.2 13.4 20.2

79 12.3 10.4 15.6









#16
Acetophenone
Concen: 8.659 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:105 Resp: 137582
Ion Ratio Lower Upper
105 100
77 79.6 66.0 99.0
51 27.2 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 8.529 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion: 70 Resp: 69647
Ion Ratio Lower Upper
70 100
42 57.7 46.8 70.2
101 9.6 8.3 12.5
130 21.1 16.5 24.7









#18
4-Methylphenol
Concen: 8.560 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:108 Resp: 87837
Ion Ratio Lower Upper
108 100
107 111.1 89.0 133.6



#19
Hexachloroethane
Concen: 8.908 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:117 Resp: 40275
Ion Ratio Lower Upper
117 100
201 105.1 80.9 121.3
199 61.0 52.0 78.0









#22
Nitrobenzene
Concen: 9.020 ng/ul
RT: 9.181 min Scan# 1036
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion: 77 Resp: 107910
Ion Ratio Lower Upper
77 100
123 39.1 32.5 48.7
65 11.6 10.2 15.4



#23
Isophorone
Concen: 8.450 ng/ul
RT: 9.693 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion: 82 Resp: 183385
Ion Ratio Lower Upper
82 100
95 7.7 5.7 8.5
138 17.3 14.1 21.1









#25
2-Nitrophenol
Concen: 8.311 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:139 Resp: 46050
Ion Ratio Lower Upper
139 100
65 59.3 46.1 69.1
109 33.8 29.6 44.4



#26
2,4-Dimethylphenol
Concen: 8.900 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:107 Resp: 98284
Ion Ratio Lower Upper
107 100
121 48.2 38.0 57.0
122 83.1 65.6 98.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 8.888 ng/ul
 RT: 10.181 min Scan# 1206
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632


Tgt Ion: 93 Resp: 118388
 Ion Ratio Lower Upper
 93 100
 95 32.1 27.0 40.4
 123 10.5 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 8.923 ng/ul
 RT: 10.410 min Scan# 1245
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56

Tgt Ion: 162 Resp: 84531
 Ion Ratio Lower Upper
 162 100
 164 71.5 51.9 77.9
 98 38.0 28.6 43.0









#30
Naphthalene
Concen: 9.383 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:128 Resp: 304296
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 13.0 10.6 16.0



#32
4-Chloroaniline
Concen: 8.767 ng/ul
RT: 10.963 min Scan# 1339
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:127 Resp: 115752
Ion Ratio Lower Upper
127 100
129 33.0 26.5 39.7









#33
Hexachlorobutadiene
Concen: 9.479 ng/ul
RT: 11.040 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:225 Resp: 66304
Ion Ratio Lower Upper
225 100
223 64.7 49.4 74.2
227 66.0 48.6 72.8



#34
Caprolactam
Concen: 7.200 ng/ul
RT: 11.793 min Scan# 1480
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:113 Resp: 20643
Ion Ratio Lower Upper
113 100
55 146.1 124.2 186.2
56 115.2 101.2 151.8









#35
4-Chloro-3-methylphenol
Concen: 8.579 ng/ul
RT: 12.075 min Scan# 1528
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:107 Resp: 85759
Ion Ratio Lower Upper
107 100
144 25.3 20.6 31.0
142 78.8 65.0 97.4



#36
2-Methylnaphthalene
Concen: 9.308 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:142 Resp: 205254
Ion Ratio Lower Upper
142 100
141 87.4 72.1 108.1





#37

1-Methylnaphthalene

Concen: 9.206 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:142 Resp: 209327

Ion Ratio Lower Upper

142 100

141 95.0 72.6 108.8

116 4.1 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.346 ng/ul

RT: 12.781 min Scan# 1648

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Tgt Ion:216 Resp: 122718




Ion Ratio Lower Upper


216 100

214 79.6 63.4 95.2

179 21.5 17.0 25.6

108 19.6 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 6.036 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:237	Resp:	43276
Ion Ratio	Lower	Upper
237	100	
235	67.3	50.2 75.2
272	11.8	10.7 16.1



#41
2,4,6-Trichlorophenol
Concen: 8.526 ng/ul
RT: 13.046 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:196	Resp:	70833
Ion Ratio	Lower	Upper
196	100	
198	97.4	80.2 120.2
200	31.5	24.6 37.0









#42
2,4,5-Trichlorophenol
Concen: 8.234 ng/ul
RT: 13.122 min Scan# 1706
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:196 Resp: 71935
Ion Ratio Lower Upper
196 100
198 97.8 74.6 112.0
200 33.1 25.8 38.8



#43
1,1'-Biphenyl
Concen: 9.154 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:154 Resp: 279134
Ion Ratio Lower Upper
154 100
153 40.0 32.2 48.4
76 13.3 10.0 15.0









#44
2-Chloronaphthalene
Concen: 9.199 ng/ul
RT: 13.498 min Scan# 1770
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:162 Resp: 223887
Ion Ratio Lower Upper
162 100
164 31.9 26.9 40.3
127 39.5 30.3 45.5



#45
2-Nitroaniline
Concen: 7.962 ng/ul
RT: 13.745 min Scan# 1812
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion: 65 Resp: 53020
Ion Ratio Lower Upper
65 100
92 72.7 53.0 79.6
138 95.7 81.9 122.9









#47
Dimethylphthalate
Concen: 8.984 ng/ul
RT: 14.087 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:163 Resp: 279645
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.2
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 8.026 ng/ul
RT: 14.240 min Scan# 1896
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:165 Resp: 46113
Ion Ratio Lower Upper
165 100
89 72.0 53.9 80.9
121 25.6 17.4 26.2





#50

Acenaphthylene

Concen: 8.776 ng/ul

RT: 14.351 min Scan# 1915

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:152 Resp: 346499

Ion Ratio Lower Upper

152 100

151 20.1 16.1 24.1

153 12.9 10.2 15.4



#51

3-Nitroaniline

Concen: 7.781 ng/ul

RT: 14.587 min Scan# 1955

Delta R.T. 0.006 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:138 Resp: 42353


Ion Ratio Lower Upper

138 100

108 11.3 10.0 15.0

92 118.2 100.2 150.4





#52

Acenaphthene

Concen: 9.084 ng/ul

RT: 14.687 min Scan# 1972

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:153 Resp: 237569

Ion Ratio Lower Upper

153 100

152 49.0 38.0 57.0

154 89.3 69.7 104.5



#53

2,4-Dinitrophenol

Concen: 15.236 ng/ul

RT: 14.792 min Scan# 1990

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:184 Resp: 49133


Ion Ratio Lower Upper

184 100

63 67.3 60.5 90.7

154 61.7 56.1 84.1









#55
4-Nitrophenol
Concen: 8.359 ng/ul
RT: 14.887 min Scan# 2005
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:109 Resp: 38875
Ion Ratio Lower Upper
109 100
139 125.1 94.6 142.0
65 127.2 101.8 152.6



#56
Dibenzofuran
Concen: 9.365 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:168 Resp: 345988
Ion Ratio Lower Upper
168 100
139 38.9 29.8 44.8









#57
2,4-Dinitrotoluene
Concen: 8.552 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:165 Resp: 73847
Ion Ratio Lower Upper
165 100
63 75.1 53.0 79.4
182 2.6 2.6 3.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.745 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:232 Resp: 67625
Ion Ratio Lower Upper
232 100
131 43.1 32.4 48.6
130 2.4 2.1 3.1
166 29.5 21.7 32.5





#59

Diethylphthalate

Concen: 8.947 ng/ul

RT: 15.445 min Scan# 2101

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:149 Resp: 279554

Ion Ratio Lower Upper

149 100

177 22.2 17.7 26.5

150 13.2 9.7 14.5



#61

Fluorene

Concen: 9.265 ng/ul

RT: 15.687 min Scan# 2142

Delta R.T. 0.006 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:166 Resp: 281923


Ion Ratio Lower Upper

166 100

165 94.7 78.8 118.2

167 12.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 9.629 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:204 Resp: 152028
Ion Ratio Lower Upper
204 100
206 31.6 27.2 40.8
141 56.8 47.0 70.4



#63
4-Nitroaniline
Concen: 8.294 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:138 Resp: 42250
Ion Ratio Lower Upper
138 100
92 57.9 41.2 61.8
108 101.2 81.1 121.7









#66
4,6-Dinitro-2-methylphenol
Concen: 6.878 ng/ul
RT: 15.787 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:198 Resp: 42138
Ion Ratio Lower Upper
198 100
182 6.4 4.6 6.8
77 26.5 23.8 35.8



#67
N-Nitrosodiphenylamine
Concen: 8.887 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:169 Resp: 235278
Ion Ratio Lower Upper
169 100
168 66.7 53.0 79.4
167 36.3 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 9.234 ng/ul
RT: 16.586 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:248 Resp: 94196
Ion Ratio Lower Upper
248 100
250 92.9 79.3 118.9
141 67.8 54.2 81.2



#69
Hexachlorobenzene
Concen: 9.478 ng/ul
RT: 16.669 min Scan# 2309
Delta R.T. 0.006 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:284 Resp: 113756
Ion Ratio Lower Upper
284 100
142 29.7 28.2 42.4
249 31.6 26.2 39.4





#70

Atrazine

Concen: 8.127 ng/ul

RT: 16.869 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:200 Resp: 83244

Ion Ratio Lower Upper

200 100

173 25.0 19.5 29.3

215 49.1 37.8 56.8



#71

Pentachlorophenol

Concen: 7.552 ng/ul

RT: 17.039 min Scan# 2372

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:266 Resp: 53796


Ion Ratio Lower Upper

266 100

264 61.2 52.5 78.7

268 62.3 54.6 81.8





#72

Phenanthrene

Concen: 9.111 ng/ul

RT: 17.457 min Scan# 2443

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:178 Resp: 467523

Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 19.1 15.8 23.6



#74

Anthracene

Concen: 9.112 ng/ul

RT: 17.551 min Scan# 2459

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:178 Resp: 474835


Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 18.9 15.0 22.6









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 8.798 ng/uL
 RT: 13.398 min Scan# 1753
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56




Instrument :
 BNA_P
 ClientSampleId :
 SST010632


Tgt Ion:	216	Resp:	124018
Ion Ratio	Lower	Upper	
216	100		
214	76.6	61.1	91.7
218	46.8	37.1	55.7



#76
 Pentachlorobenzene
 Concen: 9.109 ng/uL
 RT: 14.916 min Scan# 2011
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56

Tgt Ion:	250	Resp:	128412
Ion Ratio	Lower	Upper	
250	100		
248	59.9	49.6	74.4
252	65.9	50.9	76.3





#77

Carbazole

Concen: 8.574 ng/ul

RT: 17.845 min Scan# 2509

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:167 Resp: 400167

Ion Ratio Lower Upper

167 100

166 22.3 17.1 25.7

139 13.4 10.0 15.0



#78

Di-n-butylphthalate

Concen: 8.133 ng/ul

RT: 18.351 min Scan# 2595

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:149 Resp: 451702


Ion Ratio Lower Upper

149 100

150 9.4 7.2 10.8

104 6.1 4.6 6.8









#80
Fluoranthene
Concen: 8.945 ng/ul
RT: 19.439 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56




Instrument :
BNA_P
ClientSampleId :
SSTD010632


Tgt Ion:202 Resp: 567894
Ion Ratio Lower Upper
202 100
101 9.4 7.4 11.0
100 7.3 5.9 8.9



#82
Pyrene
Concen: 9.131 ng/ul
RT: 19.798 min Scan# 2841
Delta R.T. 0.000 min
Lab File: BP017622.D
Acq: 09 Oct 2023 17:56

Tgt Ion:202 Resp: 611197
Ion Ratio Lower Upper
202 100
101 11.0 9.4 14.0
100 10.0 7.4 11.0





#83

Butylbenzylphthalate

Concen: 7.460 ng/ul

RT: 20.669 min Scan# 2989

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:149 Resp: 191166

Ion Ratio Lower Upper

149 100

91 73.1 57.1 85.7

206 21.4 17.6 26.4



#84

3,3'-Dichlorobenzidine

Concen: 7.188 ng/ul

RT: 21.480 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:252 Resp: 155045


Ion Ratio Lower Upper

252 100

254 62.9 51.5 77.3

126 10.0 7.8 11.8





#85

Benzo(a)anthracene

Concen: 9.054 ng/ul

RT: 21.533 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:228 Resp: 625111

Ion Ratio Lower Upper

228 100

229 19.2 15.5 23.3

226 26.9 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 7.224 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:149 Resp: 265276


Ion Ratio Lower Upper

149 100

167 26.8 21.7 32.5

279 4.3 3.6 5.4





#87

Chrysene

Concen: 9.197 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:228 Resp: 591334

Ion Ratio Lower Upper

228 100

226 29.4 23.8 35.6

229 19.5 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 6.820 ng/ul




RT: 22.398 min Scan# 3283


Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Tgt Ion:149 Resp: 451328





#90

Benzo(b)fluoranthene

Concen: 8.933 ng/ul

RT: 23.327 min Scan# 3441

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:252 Resp: 603819

Ion Ratio Lower Upper

252 100

253 22.1 17.0 25.6

125 9.6 6.6 10.0



#91

Benzo(k)fluoranthene

Concen: 9.187 ng/ul

RT: 23.380 min Scan# 3450

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:252 Resp: 623735


Ion Ratio Lower Upper

252 100

253 21.5 17.1 25.7

125 8.4 6.6 10.0





#93

Benzo(a)pyrene

Concen: 8.811 ng/ul

RT: 24.010 min Scan# 3557

Delta R.T. 0.000 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56

Instrument :

BNA_P

ClientSampleId :

SSTD010632





Tgt Ion:252 Resp: 557748

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 9.7 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.714 ng/ul

RT: 26.862 min Scan# 4042

Delta R.T. 0.006 min

Lab File: BP017622.D

Acq: 09 Oct 2023 17:56




Tgt Ion:276 Resp: 670084


Ion Ratio Lower Upper

276 100

138 17.8 14.5 21.7

277 22.9 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 8.439 ng/ul
 RT: 26.892 min Scan# 4047
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010632

Tgt Ion:	278	Resp:	542797
Ion	Ratio	Lower	Upper
278	100		
139	13.9	11.0	16.6
279	24.8	19.9	29.9



#96
 Benzo(g,h,i)perylene
 Concen: 8.866 ng/ul
 RT: 27.715 min Scan# 4187
 Delta R.T. 0.000 min
 Lab File: BP017622.D
 Acq: 09 Oct 2023 17:56

Tgt Ion:	276	Resp:	547914
Ion	Ratio	Lower	Upper
276	100		
138	18.5	15.8	23.6
277	23.8	19.0	28.6

