

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017623.D
 Acq On : 09 Oct 2023 18:30
 Operator : MA/JU
 Sample : SST02068
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020633

Quant Time: Oct 10 04:02:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	111453	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	451021	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.622	164	295409	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.416	188	688372	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	721446	20.000	ng/u1	0.00
88) Perylene-d12	24.127	264	789927	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	25567	9.173	ng/uL	0.00
4) Pyridine-d5	3.716	84	171608	23.509	ng/u1	0.00
7) Phenol-d5	7.081	99	204038	21.853	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	130134	22.782	ng/u1	0.00
11) 2-Chlorophenol-d4	7.446	132	169559	22.951	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	165388	22.174	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	78626	22.782	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	84883	22.010	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	167376	23.074	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	233404	22.613	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	537509	23.300	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	588284	23.132	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	68072	17.748	ng/u1	0.00
60) Fluorene-d10	15.628	176	465626	23.636	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.769	200	79576	19.192	ng/u1	0.00
73) Anthracene-d10	17.516	188	753719	23.569	ng/u1	0.00
81) Pyrene-d10	19.769	212	932868	23.237	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	928570	23.220	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	26379	9.201	ng/uL	100
5) Pyridine	3.740	79	178500	23.338	ng/u1	100
6) Benzaldehyde	7.087	77	117868	25.069	ng/u1	100
8) Phenol	7.110	94	207893	22.204	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.363	93	171463	22.620	ng/u1	100
12) 2-Chlorophenol	7.481	128	171630	22.901	ng/u1	100
13) 2-Methylphenol	8.375	108	152584	21.829	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.440	45	127976	12.839	ng/u1	100
16) Acetophenone	8.787	105	261838	22.293	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.752	70	135924	22.516	ng/u1	100
18) 4-Methylphenol	8.716	108	167563	22.089	ng/u1	100
19) Hexachloroethane	8.993	117	77303	23.130	ng/u1	100
22) Nitrobenzene	9.181	77	210012	23.515	ng/u1	100
23) Isophorone	9.693	82	360471	22.250	ng/u1	100
25) 2-Nitrophenol	9.887	139	93253	22.545	ng/u1	100
26) 2,4-Dimethylphenol	9.928	107	190524	23.112	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.181	93	225297	22.658	ng/u1	100
29) 2,4-Dichlorophenol	10.410	162	167020	23.619	ng/u1	100
30) Naphthalene	10.816	128	569142	23.509	ng/u1	100
32) 4-Chloroaniline	10.957	127	227980	23.132	ng/u1	100
33) Hexachlorobutadiene	11.040	225	126405	24.208	ng/u1	100
34) Caprolactam	11.787	113	45048	21.050	ng/u1	100
35) 4-Chloro-3-methylphenol	12.069	107	174558	23.393	ng/u1	100
36) 2-Methylnaphthalene	12.428	142	381310	23.165	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017623.D
 Acq On : 09 Oct 2023 18:30
 Operator : MA/JU
 Sample : SSTD02068
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633

Quant Time: Oct 10 04:02:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	400382	23.589	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.781	216	228975	23.444	ng/ul	100
40) Hexachlorocyclopentadiene	12.722	237	97357	18.256	ng/ul	100
41) 2,4,6-Trichlorophenol	13.045	196	141644	22.921	ng/ul	100
42) 2,4,5-Trichlorophenol	13.116	196	149104	22.947	ng/ul	100
43) 1,1'-Biphenyl	13.445	154	524344	23.118	ng/ul	100
44) 2-Chloronaphthalene	13.498	162	422420	23.334	ng/ul	100
45) 2-Nitroaniline	13.745	65	112277	22.668	ng/ul	100
47) Dimethylphthalate	14.087	163	533320	23.035	ng/ul	100
48) 2,6-Dinitrotoluene	14.234	165	95369	22.315	ng/ul	100
50) Acenaphthylene	14.351	152	674186	22.957	ng/ul	100
51) 3-Nitroaniline	14.581	138	89177	22.025	ng/ul	100
52) Acenaphthene	14.687	153	449802	23.124	ng/ul	100
53) 2,4-Dinitrophenol	14.792	184	33921	14.142	ng/ul	100
55) 4-Nitrophenol	14.881	109	62188	17.978	ng/ul	100
56) Dibenzofuran	15.028	168	645973	23.506	ng/ul	100
57) 2,4-Dinitrotoluene	15.028	165	148861	23.177	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.251	232	135554	23.567	ng/ul	100
59) Diethylphthalate	15.445	149	540197	23.245	ng/ul	100
61) Fluorene	15.681	166	527637	23.312	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.675	204	283081	24.105	ng/ul	100
63) 4-Nitroaniline	15.757	138	85645	22.605	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.786	198	87851	19.742	ng/ul	100
67) N-Nitrosodiphenylamine	15.904	169	442645	23.017	ng/ul	100
68) 4-Bromophenyl-phenylether	16.586	248	177188	23.912	ng/ul	100
69) Hexachlorobenzene	16.663	284	209825	24.067	ng/ul	100
70) Atrazine	16.869	200	166657	22.399	ng/ul	100
71) Pentachlorophenol	17.039	266	111472	21.542	ng/ul	100
72) Phenanthrene	17.457	178	875869	23.499	ng/ul	100
74) Anthracene	17.551	178	888322	23.468	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.398	216	235532	23.002	ng/uL	100
76) Pentachlorobenzene	14.916	250	242307	23.663	ng/uL	100
77) Carbazole	17.845	167	780379	23.018	ng/ul	100
78) Di-n-butylphthalate	18.351	149	904531	22.422	ng/ul	100
80) Fluoranthene	19.439	202	1071079	22.956	ng/ul	100
82) Pyrene	19.798	202	1128807	22.947	ng/ul	100
83) Butylbenzylphthalate	20.669	149	398242	21.145	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.480	252	318194	20.071	ng/ul	100
85) Benzo(a)anthracene	21.533	228	1183649	23.327	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.410	149	555928	20.600	ng/ul	100
87) Chrysene	21.592	228	1095533	23.183	ng/ul	100
89) Di-n-octyl phthalate	22.398	149	959664	19.729	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	1151254	23.173	ng/ul	100
91) Benzo(k)fluoranthene	23.380	252	1171830	23.483	ng/ul	100
93) Benzo(a)pyrene	24.010	252	1086327	23.348	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.856	276	1282058	22.684	ng/ul	100
95) Dibenzo(a,h)anthracene	26.892	278	1045584	22.117	ng/ul	100
96) Benzo(g,h,i)perylene	27.715	276	1039357	22.883	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.201 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion: 88 Resp: 26379
Ion Ratio Lower Upper
88 100
43 30.4 24.3 36.5
58 66.7 53.4 80.0



#5
Pyridine
Concen: 23.338 ng/uL
RT: 3.740 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion: 79 Resp: 178500
Ion Ratio Lower Upper
79 100
52 69.3 55.4 83.2
51 35.2 28.2 42.2









#6
Benzaldehyde
Concen: 25.069 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion: 77 Resp: 117868
Ion Ratio Lower Upper
77 100
105 93.8 75.0 112.6
106 89.7 71.8 107.6



#8
Phenol
Concen: 22.204 ng/ul
RT: 7.110 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion: 94 Resp: 207893
Ion Ratio Lower Upper
94 100
65 30.3 24.2 36.4
66 41.2 33.0 49.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 22.620 ng/ul
 RT: 7.363 min Scan# 727
 Delta R.T. 0.000 min
 Lab File: BP017623.D
 Acq: 09 Oct 2023 18:30

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633

Tgt Ion: 93 Resp: 171463




Ion	Ratio	Lower	Upper
93	100		
63	74.9	59.9	89.9
95	34.0	27.2	40.8




#12
 2-Chlorophenol
 Concen: 22.901 ng/ul
 RT: 7.481 min Scan# 747
 Delta R.T. 0.000 min
 Lab File: BP017623.D
 Acq: 09 Oct 2023 18:30

Tgt Ion: 128 Resp: 171630

Ion	Ratio	Lower	Upper
128	100		
64	46.2	37.0	55.4
130	32.2	25.8	38.6





#13

2-Methylphenol

Concen: 21.829 ng/ul

RT: 8.375 min Scan# 899

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :





SSTD020633

Tgt Ion:108 Resp: 152584

Ion Ratio Lower Upper

108 100

107 91.8 73.4 110.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.839 ng/ul

RT: 8.440 min Scan# 910

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion: 45 Resp: 127976


Ion Ratio Lower Upper

45 100

77 16.8 13.4 20.2

79 13.0 10.4 15.6









#16
Acetophenone
Concen: 22.293 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:105 Resp: 261838
Ion Ratio Lower Upper
105 100
77 82.5 66.0 99.0
51 27.4 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 22.516 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion: 70 Resp: 135924
Ion Ratio Lower Upper
70 100
42 58.5 46.8 70.2
101 10.4 8.3 12.5
130 20.6 16.5 24.7









#18
4-Methylphenol
Concen: 22.089 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:108 Resp: 167563
Ion Ratio Lower Upper
108 100
107 111.3 89.0 133.6



#19
Hexachloroethane
Concen: 23.130 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:117 Resp: 77303
Ion Ratio Lower Upper
117 100
201 101.1 80.9 121.3
199 65.0 52.0 78.0









#22
Nitrobenzene
Concen: 23.515 ng/ul
RT: 9.181 min Scan# 1036
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion: 77 Resp: 210012
Ion Ratio Lower Upper
77 100
123 40.6 32.5 48.7
65 12.8 10.2 15.4



#23
Isophorone
Concen: 22.250 ng/ul
RT: 9.693 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion: 82 Resp: 360471
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 17.6 14.1 21.1









#25
2-Nitrophenol
Concen: 22.545 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:139 Resp: 93253
Ion Ratio Lower Upper
139 100
65 57.6 46.1 69.1
109 37.0 29.6 44.4



#26
2,4-Dimethylphenol
Concen: 23.112 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:107 Resp: 190524
Ion Ratio Lower Upper
107 100
121 47.5 38.0 57.0
122 82.0 65.6 98.4





#27

Bis(2-Chloroethoxy)methane

Concen: 22.658 ng/ul

RT: 10.181 min Scan# 1206

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion: 93 Resp: 225297

Ion Ratio Lower Upper

93 100

95 33.7 27.0 40.4

123 11.1 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 23.619 ng/ul

RT: 10.410 min Scan# 1245

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:162 Resp: 167020


Ion Ratio Lower Upper

162 100

164 64.9 51.9 77.9

98 35.8 28.6 43.0









#30
Naphthalene
Concen: 23.509 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:128 Resp: 569142
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.6
127 13.3 10.6 16.0



#32
4-Chloroaniline
Concen: 23.132 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:127 Resp: 227980
Ion Ratio Lower Upper
127 100
129 33.1 26.5 39.7





#33

Hexachlorobutadiene

Concen: 24.208 ng/ul

RT: 11.040 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:225 Resp: 126405

Ion Ratio Lower Upper

225 100

223 61.8 49.4 74.2

227 60.7 48.6 72.8



#34

Caprolactam

Concen: 21.050 ng/ul

RT: 11.787 min Scan# 1479

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:113 Resp: 45048


Ion Ratio Lower Upper

113 100

55 155.2 124.2 186.2

56 126.5 101.2 151.8









#35
4-Chloro-3-methylphenol
Concen: 23.393 ng/ul
RT: 12.069 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:107 Resp: 174558
Ion Ratio Lower Upper
107 100
144 25.8 20.6 31.0
142 81.2 65.0 97.4



#36
2-Methylnaphthalene
Concen: 23.165 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:142 Resp: 381310
Ion Ratio Lower Upper
142 100
141 90.1 72.1 108.1





#37

1-Methylnaphthalene

Concen: 23.589 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:142 Resp: 400382

Ion Ratio Lower Upper

142 100

141 90.7 72.6 108.8

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 23.444 ng/ul

RT: 12.781 min Scan# 1648

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Tgt Ion:216 Resp: 228975




Ion Ratio Lower Upper


216 100

214 79.3 63.4 95.2

179 21.3 17.0 25.6

108 17.1 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 18.256 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:237 Resp: 97357
Ion Ratio Lower Upper
237 100
235 62.7 50.2 75.2
272 13.4 10.7 16.1



#41
2,4,6-Trichlorophenol
Concen: 22.921 ng/ul
RT: 13.045 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:196 Resp: 141644
Ion Ratio Lower Upper
196 100
198 100.2 80.2 120.2
200 30.8 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 22.947 ng/ul

RT: 13.116 min Scan# 1705

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:196 Resp: 149104

Ion Ratio Lower Upper

196 100

198 93.3 74.6 112.0

200 32.3 25.8 38.8



#43

1,1'-Biphenyl

Concen: 23.118 ng/ul

RT: 13.445 min Scan# 1761

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:154 Resp: 524344


Ion Ratio Lower Upper

154 100

153 40.3 32.2 48.4

76 12.5 10.0 15.0





#44

2-Chloronaphthalene

Concen: 23.334 ng/ul

RT: 13.498 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:162 Resp: 422420

Ion Ratio Lower Upper

162 100

164 33.6 26.9 40.3

127 37.9 30.3 45.5



#45

2-Nitroaniline

Concen: 22.668 ng/ul

RT: 13.745 min Scan# 1812

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion: 65 Resp: 112277


Ion Ratio Lower Upper

65 100

92 66.3 53.0 79.6

138 102.4 81.9 122.9





#47

Dimethylphthalate

Concen: 23.035 ng/ul

RT: 14.087 min Scan# 1870

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:163 Resp: 533320

Ion Ratio Lower Upper

163 100

194 4.3 3.4 5.2

164 9.9 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 22.315 ng/ul

RT: 14.234 min Scan# 1895

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:165 Resp: 95369


Ion Ratio Lower Upper

165 100

89 67.4 53.9 80.9

121 21.8 17.4 26.2





#50

Acenaphthylene

Concen: 22.957 ng/ul

RT: 14.351 min Scan# 1915

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:152 Resp: 674186

Ion Ratio Lower Upper

152 100

151 20.1 16.1 24.1

153 12.8 10.2 15.4



#51

3-Nitroaniline

Concen: 22.025 ng/ul

RT: 14.581 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:138 Resp: 89177


Ion Ratio Lower Upper

138 100

108 12.5 10.0 15.0

92 125.3 100.2 150.4









#52
Acenaphthene
Concen: 23.124 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:153 Resp: 449802
Ion Ratio Lower Upper
153 100
152 47.5 38.0 57.0
154 87.1 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 14.142 ng/ul
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:184 Resp: 33921
Ion Ratio Lower Upper
184 100
63 75.6 60.5 90.7
154 70.1 56.1 84.1









#55
4-Nitrophenol
Concen: 17.978 ng/ul
RT: 14.881 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:109 Resp: 62188
Ion Ratio Lower Upper
109 100
139 118.3 94.6 142.0
65 127.2 101.8 152.6



#56
Dibenzofuran
Concen: 23.506 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:168 Resp: 645973
Ion Ratio Lower Upper
168 100
139 37.3 29.8 44.8









#57
2,4-Dinitrotoluene
Concen: 23.177 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:165 Resp: 148861
Ion Ratio Lower Upper
165 100
63 66.2 53.0 79.4
182 3.2 2.6 3.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.567 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:232 Resp: 135554
Ion Ratio Lower Upper
232 100
131 40.5 32.4 48.6
130 2.6 2.1 3.1
166 27.1 21.7 32.5









#59
Diethylphthalate
Concen: 23.245 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:149 Resp: 540197
Ion Ratio Lower Upper
149 100
177 22.1 17.7 26.5
150 12.1 9.7 14.5



#61
Fluorene
Concen: 23.312 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:166 Resp: 527637
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 24.105 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:204 Resp: 283081
Ion Ratio Lower Upper
204 100
206 34.0 27.2 40.8
141 58.7 47.0 70.4



#63
4-Nitroaniline
Concen: 22.605 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:138 Resp: 85645
Ion Ratio Lower Upper
138 100
92 51.5 41.2 61.8
108 101.4 81.1 121.7









#66
4,6-Dinitro-2-methylphenol
Concen: 19.742 ng/ul
RT: 15.786 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:198 Resp: 87851
Ion Ratio Lower Upper
198 100
182 5.7 4.6 6.8
77 29.8 23.8 35.8



#67
N-Nitrosodiphenylamine
Concen: 23.017 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:169 Resp: 442645
Ion Ratio Lower Upper
169 100
168 66.2 53.0 79.4
167 37.0 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 23.912 ng/ul
RT: 16.586 min Scan# 2295
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:	248	Resp:	177188
Ion Ratio	Lower	Upper	
248	100		
250	99.1	79.3	118.9
141	67.7	54.2	81.2



#69
Hexachlorobenzene
Concen: 24.067 ng/ul
RT: 16.663 min Scan# 2308
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:	284	Resp:	209825
Ion Ratio	Lower	Upper	
284	100		
142	35.3	28.2	42.4
249	32.8	26.2	39.4





#70

Atrazine

Concen: 22.399 ng/ul

RT: 16.869 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:200 Resp: 166657

Ion Ratio Lower Upper

200 100

173 24.4 19.5 29.3

215 47.3 37.8 56.8



#71

Pentachlorophenol

Concen: 21.542 ng/ul

RT: 17.039 min Scan# 2372

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:266 Resp: 111472


Ion Ratio Lower Upper

266 100

264 65.6 52.5 78.7

268 68.2 54.6 81.8









#72
 Phenanthrene
 Concen: 23.499 ng/ul
 RT: 17.457 min Scan# 2443
 Delta R.T. 0.000 min
 Lab File: BP017623.D
 Acq: 09 Oct 2023 18:30

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633

Tgt Ion:178 Resp: 875869




Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.4	18.6
176	19.7	15.8	23.6




#74
 Anthracene
 Concen: 23.468 ng/ul
 RT: 17.551 min Scan# 2459
 Delta R.T. 0.000 min
 Lab File: BP017623.D
 Acq: 09 Oct 2023 18:30

Tgt Ion:178 Resp: 888322

Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.3	18.5
176	18.8	15.0	22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.002 ng/uL
RT: 13.398 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:216 Resp: 235532
Ion Ratio Lower Upper
216 100
214 76.4 61.1 91.7
218 46.4 37.1 55.7



#76
Pentachlorobenzene
Concen: 23.663 ng/uL
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:250 Resp: 242307
Ion Ratio Lower Upper
250 100
248 62.0 49.6 74.4
252 63.6 50.9 76.3









#77
 Carbazole
 Concen: 23.018 ng/ul
 RT: 17.845 min Scan# 2509
 Delta R.T. 0.000 min
 Lab File: BP017623.D
 Acq: 09 Oct 2023 18:30




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020633


Tgt Ion:167 Resp: 780379
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.1 25.7
 139 12.5 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 22.422 ng/ul
 RT: 18.351 min Scan# 2595
 Delta R.T. 0.000 min
 Lab File: BP017623.D
 Acq: 09 Oct 2023 18:30

Tgt Ion:149 Resp: 904531
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 5.7 4.6 6.8









#80
Fluoranthene
Concen: 22.956 ng/ul
RT: 19.439 min Scan# 2780
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30




Instrument :
BNA_P
ClientSampleId :
SSTD020633


Tgt Ion:202 Resp: 1071079
Ion Ratio Lower Upper
202 100
101 9.2 7.4 11.0
100 7.4 5.9 8.9



#82
Pyrene
Concen: 22.947 ng/ul
RT: 19.798 min Scan# 2841
Delta R.T. 0.000 min
Lab File: BP017623.D
Acq: 09 Oct 2023 18:30

Tgt Ion:202 Resp: 1128807
Ion Ratio Lower Upper
202 100
101 11.7 9.4 14.0
100 9.2 7.4 11.0





#83

Butylbenzylphthalate

Concen: 21.145 ng/ul

RT: 20.669 min Scan# 2989

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:149 Resp: 398242

Ion Ratio Lower Upper

149 100

91 71.4 57.1 85.7

206 22.0 17.6 26.4



#84

3,3'-Dichlorobenzidine

Concen: 20.071 ng/ul

RT: 21.480 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:252 Resp: 318194


Ion Ratio Lower Upper

252 100

254 64.4 51.5 77.3

126 9.8 7.8 11.8





#85

Benzo(a)anthracene

Concen: 23.327 ng/ul

RT: 21.533 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:228 Resp: 1183649

Ion Ratio Lower Upper

228 100

229 19.4 15.5 23.3

226 27.2 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.600 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:149 Resp: 555928


Ion Ratio Lower Upper

149 100

167 27.1 21.7 32.5

279 4.5 3.6 5.4





#87

Chrysene

Concen: 23.183 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:228 Resp: 1095533

Ion Ratio Lower Upper

228 100

226 29.7 23.8 35.6

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 19.729 ng/ul




RT: 22.398 min Scan# 3283


Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Tgt Ion:149 Resp: 959664





#90

Benzo(b)fluoranthene

Concen: 23.173 ng/ul

RT: 23.327 min Scan# 3441

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:252 Resp: 1151254

Ion Ratio Lower Upper

252 100

253 21.3 17.0 25.6

125 8.3 6.6 10.0



#91

Benzo(k)fluoranthene

Concen: 23.483 ng/ul

RT: 23.380 min Scan# 3450

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:252 Resp: 1171830


Ion Ratio Lower Upper

252 100

253 21.4 17.1 25.7

125 8.3 6.6 10.0





#93

Benzo(a)pyrene

Concen: 23.348 ng/ul

RT: 24.010 min Scan# 3557

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:252 Resp: 1086327

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 10.1 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.684 ng/ul

RT: 26.856 min Scan# 4041

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30




Tgt Ion:276 Resp: 1282058


Ion Ratio Lower Upper

276 100

138 18.1 14.5 21.7

277 24.1 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 22.117 ng/ul

RT: 26.892 min Scan# 4047

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Instrument :

BNA_P

ClientSampleId :

SSTD020633





Tgt Ion:278 Resp: 1045584

Ion Ratio Lower Upper

278 100

139 13.8 11.0 16.6

279 24.9 19.9 29.9



#96

Benzo(g,h,i)perylene

Concen: 22.883 ng/ul

RT: 27.715 min Scan# 4187

Delta R.T. 0.000 min

Lab File: BP017623.D

Acq: 09 Oct 2023 18:30

Tgt Ion:276 Resp: 1039357

Ion Ratio Lower Upper

276 100

138 19.7 15.8 23.6

277 23.8 19.0 28.6

