

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017624.D
 Acq On : 09 Oct 2023 19:04
 Operator : MA/JU
 Sample : SST04069
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040634

Quant Time: Oct 10 04:03:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	135020	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	556759	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.622	164	365405	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.410	188	830431	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	844821	20.000	ng/u1	0.00
88) Perylene-d12	24.127	264	923249	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	61165	18.115	ng/uL	0.00
4) Pyridine-d5	3.716	84	421166	47.626	ng/u1	0.00
7) Phenol-d5	7.081	99	517344	45.737	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	321437	46.450	ng/u1	0.00
11) 2-Chlorophenol-d4	7.446	132	427107	47.722	ng/u1	0.00
15) 4-Methylphenol-d8	8.651	113	419985	46.481	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	203671	47.806	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	235241	49.413	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	443027	49.474	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	601420	47.201	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	1341827	47.024	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	1499850	47.679	ng/u1	0.00
54) 4-Nitrophenol-d4	14.863	143	213193	44.938	ng/u1	0.00
60) Fluorene-d10	15.628	176	1163348	47.741	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.769	200	235252	47.033	ng/u1	0.00
73) Anthracene-d10	17.516	188	1876385	48.637	ng/u1	0.00
81) Pyrene-d10	19.769	212	2268111	48.247	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	2265632	48.474	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.316	88	61618	17.740	ng/uL	98
5) Pyridine	3.734	79	434324	46.874	ng/u1	100
6) Benzaldehyde	7.087	77	228325	40.085	ng/u1	98
8) Phenol	7.110	94	521622	45.987	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.363	93	421173	45.865	ng/u1	99
12) 2-Chlorophenol	7.481	128	419573	46.213	ng/u1	99
13) 2-Methylphenol	8.375	108	389854	46.038	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.440	45	315178	26.100	ng/u1	97
16) Acetophenone	8.787	105	656492	46.137	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.757	70	343033	46.905	ng/u1	99
18) 4-Methylphenol	8.716	108	426646	46.425	ng/u1	98
19) Hexachloroethane	8.993	117	181164	44.744	ng/u1	90
22) Nitrobenzene	9.181	77	526861	47.790	ng/u1	96
23) Isophorone	9.693	82	948582	47.431	ng/u1	98
25) 2-Nitrophenol	9.887	139	249534	48.871	ng/u1	95
26) 2,4-Dimethylphenol	9.928	107	489032	48.056	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.181	93	567374	46.223	ng/u1	98
29) 2,4-Dichlorophenol	10.404	162	423298	48.491	ng/u1	99
30) Naphthalene	10.816	128	1395045	46.680	ng/u1	99
32) 4-Chloroaniline	10.957	127	587098	48.256	ng/u1	97
33) Hexachlorobutadiene	11.040	225	310983	48.246	ng/u1	98
34) Caprolactam	11.798	113	130132	49.259	ng/u1	91
35) 4-Chloro-3-methylphenol	12.069	107	454229	49.313	ng/u1	99
36) 2-Methylnaphthalene	12.428	142	974604	47.965	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017624.D
 Acq On : 09 Oct 2023 19:04
 Operator : MA/JU
 Sample : SST04069
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040634

Quant Time: Oct 10 04:03:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	992302	47.359	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.781	216	572450	47.383	ng/ul	99
40) Hexachlorocyclopentadiene	12.722	237	293497	44.493	ng/ul	99
41) 2,4,6-Trichlorophenol	13.045	196	367350	48.058	ng/ul	96
42) 2,4,5-Trichlorophenol	13.116	196	391862	48.755	ng/ul	97
43) 1,1'-Biphenyl	13.451	154	1298634	46.288	ng/ul	100
44) 2-Chloronaphthalene	13.498	162	1040822	46.480	ng/ul	98
45) 2-Nitroaniline	13.745	65	303125	49.476	ng/ul	95
47) Dimethylphthalate	14.086	163	1342105	46.863	ng/ul	100
48) 2,6-Dinitrotoluene	14.239	165	270817	51.229	ng/ul	94
50) Acenaphthylene	14.351	152	1696689	46.708	ng/ul	99
51) 3-Nitroaniline	14.581	138	223280	44.583	ng/ul	95
52) Acenaphthene	14.686	153	1118357	46.481	ng/ul	99
53) 2,4-Dinitrophenol	14.786	184	125092	42.162	ng/ul	96
55) 4-Nitrophenol	14.875	109	193013	45.110	ng/ul	100
56) Dibenzofuran	15.028	168	1592127	46.838	ng/ul	100
57) 2,4-Dinitrotoluene	15.028	165	398886	50.208	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.251	232	355441	49.958	ng/ul	97
59) Diethylphthalate	15.451	149	1348205	46.901	ng/ul	100
61) Fluorene	15.686	166	1312908	46.896	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.675	204	688785	47.416	ng/ul	98
63) 4-Nitroaniline	15.757	138	202798	43.273	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.786	198	246999	46.010	ng/ul#	93
67) N-Nitrosodiphenylamine	15.904	169	1093182	47.120	ng/ul	99
68) 4-Bromophenyl-phenylether	16.586	248	437483	48.940	ng/ul	97
69) Hexachlorobenzene	16.669	284	517225	49.178	ng/ul	94
70) Atrazine	16.869	200	423163	47.145	ng/ul	98
71) Pentachlorophenol	17.039	266	308171	49.367	ng/ul	94
72) Phenanthrene	17.457	178	2138052	47.549	ng/ul	99
74) Anthracene	17.551	178	2184631	47.842	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.398	216	589227	47.699	ng/ul	97
76) Pentachlorobenzene	14.916	250	594005	48.086	ng/ul	98
77) Carbazole	17.839	167	1915824	46.843	ng/ul	99
78) Di-n-butylphthalate	18.351	149	2331163	47.900	ng/ul	99
80) Fluoranthene	19.439	202	2608257	47.738	ng/ul	99
82) Pyrene	19.798	202	2725075	47.306	ng/ul	99
83) Butylbenzylphthalate	20.669	149	1063156	48.206	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.480	252	828522	44.628	ng/ul	99
85) Benzo(a)anthracene	21.539	228	2846022	47.897	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.410	149	1504437	47.606	ng/ul	99
87) Chrysene	21.592	228	2644924	47.797	ng/ul	99
89) Di-n-octyl phthalate	22.398	149	2576445	45.319	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	2764200	47.605	ng/ul	99
91) Benzo(k)fluoranthene	23.380	252	2811375	48.204	ng/ul	99
93) Benzo(a)pyrene	24.015	252	2618808	48.157	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.868	276	3126949	47.338	ng/ul	98
95) Dibenzo(a,h)anthracene	26.892	278	2586010	46.802	ng/ul	99
96) Benzo(g,h,i)perylene	27.721	276	2505799	47.201	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP100923\
 Data File : BP017624.D
 Acq On : 09 Oct 2023 19:04
 Operator : MA/JU
 Sample : SST04069
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040634

Quant Time: Oct 10 04:03:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP100923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 10 03:51:30 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 17.740 ng/uL
RT: 3.316 min Scan# 39
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion: 88 Resp: 61618
Ion Ratio Lower Upper
88 100
43 31.3 24.3 36.5
58 65.3 53.4 80.0



#5
Pyridine
Concen: 46.874 ng/uL
RT: 3.734 min Scan# 110
Delta R.T. -0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion: 79 Resp: 434324
Ion Ratio Lower Upper
79 100
52 68.9 55.4 83.2
51 35.4 28.2 42.2









#6
Benzaldehyde
Concen: 40.085 ng/ul
RT: 7.087 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion: 77 Resp: 228325
Ion Ratio Lower Upper
77 100
105 95.8 75.0 112.6
106 88.4 71.8 107.6



#8
Phenol
Concen: 45.987 ng/ul
RT: 7.110 min Scan# 684
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion: 94 Resp: 521622
Ion Ratio Lower Upper
94 100
65 30.8 24.2 36.4
66 41.3 33.0 49.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 45.865 ng/ul
 RT: 7.363 min Scan# 727
 Delta R.T. -0.000 min
 Lab File: BP017624.D
 Acq: 09 Oct 2023 19:04

Instrument :
 BNA_P
 ClientSampleId :
 SST040634

Tgt Ion: 93 Resp: 421173




Ion	Ratio	Lower	Upper
93	100		
63	74.2	59.9	89.9
95	32.6	27.2	40.8




#12
 2-Chlorophenol
 Concen: 46.213 ng/ul
 RT: 7.481 min Scan# 747
 Delta R.T. -0.000 min
 Lab File: BP017624.D
 Acq: 09 Oct 2023 19:04

Tgt Ion:128 Resp: 419573

Ion	Ratio	Lower	Upper
128	100		
64	45.3	37.0	55.4
130	31.9	25.8	38.6





#13

2-Methylphenol

Concen: 46.038 ng/ul

RT: 8.375 min Scan# 899

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :





SSTD040634

Tgt Ion:108 Resp: 389854

Ion Ratio Lower Upper

108 100

107 91.3 73.4 110.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 26.100 ng/ul

RT: 8.440 min Scan# 910

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion: 45 Resp: 315178


Ion Ratio Lower Upper

45 100

77 16.1 13.4 20.2

79 11.5 10.4 15.6









#16
Acetophenone
Concen: 46.137 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:105 Resp: 656492
Ion Ratio Lower Upper
105 100
77 80.8 66.0 99.0
51 26.4 21.9 32.9



#17
N-Nitroso-di-n-propylamine
Concen: 46.905 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion: 70 Resp: 343033
Ion Ratio Lower Upper
70 100
42 58.2 46.8 70.2
101 9.9 8.3 12.5
130 22.0 16.5 24.7









#18
4-Methylphenol
Concen: 46.425 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:108 Resp: 426646
Ion Ratio Lower Upper
108 100
107 109.2 89.0 133.6



#19
Hexachloroethane
Concen: 44.744 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:117 Resp: 181164
Ion Ratio Lower Upper
117 100
201 113.1 80.9 121.3
199 70.1 52.0 78.0









#22
Nitrobenzene
Concen: 47.790 ng/ul
RT: 9.181 min Scan# 1036
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion: 77 Resp: 526861
Ion Ratio Lower Upper
77 100
123 43.3 32.5 48.7
65 13.1 10.2 15.4



#23
Isophorone
Concen: 47.431 ng/ul
RT: 9.693 min Scan# 1123
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion: 82 Resp: 948582
Ion Ratio Lower Upper
82 100
95 7.4 5.7 8.5
138 16.8 14.1 21.1









#25
2-Nitrophenol
Concen: 48.871 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:139 Resp: 249534
Ion Ratio Lower Upper
139 100
65 54.7 46.1 69.1
109 33.0 29.6 44.4



#26
2,4-Dimethylphenol
Concen: 48.056 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:107 Resp: 489032
Ion Ratio Lower Upper
107 100
121 48.8 38.0 57.0
122 81.9 65.6 98.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 46.223 ng/ul
 RT: 10.181 min Scan# 1206
 Delta R.T. -0.000 min
 Lab File: BP017624.D
 Acq: 09 Oct 2023 19:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040634


Tgt Ion: 93 Resp: 567374
 Ion Ratio Lower Upper
 93 100
 95 32.4 27.0 40.4
 123 11.6 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 48.491 ng/ul
 RT: 10.404 min Scan# 1244
 Delta R.T. -0.006 min
 Lab File: BP017624.D
 Acq: 09 Oct 2023 19:04

Tgt Ion: 162 Resp: 423298
 Ion Ratio Lower Upper
 162 100
 164 65.4 51.9 77.9
 98 36.4 28.6 43.0









#30
Naphthalene
Concen: 46.680 ng/ul
RT: 10.816 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:128 Resp: 1395045
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 48.256 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:127 Resp: 587098
Ion Ratio Lower Upper
127 100
129 31.4 26.5 39.7





#33

Hexachlorobutadiene

Concen: 48.246 ng/ul

RT: 11.040 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:225 Resp: 310983

Ion Ratio Lower Upper

225 100

223 62.1 49.4 74.2

227 63.6 48.6 72.8



#34

Caprolactam

Concen: 49.259 ng/ul

RT: 11.798 min Scan# 1481

Delta R.T. 0.012 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion:113 Resp: 130132


Ion Ratio Lower Upper

113 100

55 147.3 124.2 186.2

56 112.9 101.2 151.8









#35
4-Chloro-3-methylphenol
Concen: 49.313 ng/ul
RT: 12.069 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:107 Resp: 454229
Ion Ratio Lower Upper
107 100
144 25.2 20.6 31.0
142 80.5 65.0 97.4



#36
2-Methylnaphthalene
Concen: 47.965 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:142 Resp: 974604
Ion Ratio Lower Upper
142 100
141 91.7 72.1 108.1









#37
1-Methylnaphthalene
Concen: 47.359 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:142 Resp: 992302
Ion Ratio Lower Upper
142 100
141 91.7 72.6 108.8
116 4.2 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 47.383 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:216 Resp: 572450
Ion Ratio Lower Upper
216 100
214 80.8 63.4 95.2
179 21.6 17.0 25.6
108 16.7 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 44.493 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:237 Resp: 293497
Ion Ratio Lower Upper
237 100
235 61.9 50.2 75.2
272 13.0 10.7 16.1



#41
2,4,6-Trichlorophenol
Concen: 48.058 ng/ul
RT: 13.045 min Scan# 1693
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:196 Resp: 367350
Ion Ratio Lower Upper
196 100
198 95.9 80.2 120.2
200 31.4 24.6 37.0









#42
2,4,5-Trichlorophenol
Concen: 48.755 ng/ul
RT: 13.116 min Scan# 1705
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:196 Resp: 391862
Ion Ratio Lower Upper
196 100
198 96.2 74.6 112.0
200 30.6 25.8 38.8



#43
1,1'-Biphenyl
Concen: 46.288 ng/ul
RT: 13.451 min Scan# 1762
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:154 Resp: 1298634
Ion Ratio Lower Upper
154 100
153 40.2 32.2 48.4
76 12.0 10.0 15.0









#44
2-Chloronaphthalene
Concen: 46.480 ng/ul
RT: 13.498 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:162 Resp: 1040822
Ion Ratio Lower Upper
162 100
164 33.7 26.9 40.3
127 39.8 30.3 45.5



#45
2-Nitroaniline
Concen: 49.476 ng/ul
RT: 13.745 min Scan# 1812
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion: 65 Resp: 303125
Ion Ratio Lower Upper
65 100
92 70.1 53.0 79.6
138 108.0 81.9 122.9









#47
Dimethylphthalate
Concen: 46.863 ng/ul
RT: 14.086 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:163 Resp: 1342105
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.2
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 51.229 ng/ul
RT: 14.239 min Scan# 1896
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:165 Resp: 270817
Ion Ratio Lower Upper
165 100
89 61.3 53.9 80.9
121 20.6 17.4 26.2









#50
Acenaphthylene
Concen: 46.708 ng/ul
RT: 14.351 min Scan# 1915
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:152 Resp: 1696689
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1
153 13.3 10.2 15.4



#51
3-Nitroaniline
Concen: 44.583 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:138 Resp: 223280
Ion Ratio Lower Upper
138 100
108 11.1 10.0 15.0
92 120.1 100.2 150.4









#52
Acenaphthene
Concen: 46.481 ng/ul
RT: 14.686 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:153 Resp: 1118357
Ion Ratio Lower Upper
153 100
152 48.1 38.0 57.0
154 88.1 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 42.162 ng/ul
RT: 14.786 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:184 Resp: 125092
Ion Ratio Lower Upper
184 100
63 72.9 60.5 90.7
154 66.4 56.1 84.1









#55
4-Nitrophenol
Concen: 45.110 ng/ul
RT: 14.875 min Scan# 2004
Delta R.T. -0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:109 Resp: 193013
Ion Ratio Lower Upper
109 100
139 119.0 94.6 142.0
65 127.1 101.8 152.6



#56
Dibenzofuran
Concen: 46.838 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:168 Resp: 1592127
Ion Ratio Lower Upper
168 100
139 37.6 29.8 44.8









#57
2,4-Dinitrotoluene
Concen: 50.208 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:165 Resp: 398886
Ion Ratio Lower Upper
165 100
63 65.7 53.0 79.4
182 2.8 2.6 3.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 49.958 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:232 Resp: 355441
Ion Ratio Lower Upper
232 100
131 42.7 32.4 48.6
130 2.5 2.1 3.1
166 28.1 21.7 32.5









#59
Diethylphthalate
Concen: 46.901 ng/ul
RT: 15.451 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:149 Resp: 1348205
Ion Ratio Lower Upper
149 100
177 22.3 17.7 26.5
150 12.2 9.7 14.5



#61
Fluorene
Concen: 46.896 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:166 Resp: 1312908
Ion Ratio Lower Upper
166 100
165 97.6 78.8 118.2
167 13.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 47.416 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:204 Resp: 688785
Ion Ratio Lower Upper
204 100
206 32.8 27.2 40.8
141 56.9 47.0 70.4



#63
4-Nitroaniline
Concen: 43.273 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:138 Resp: 202798
Ion Ratio Lower Upper
138 100
92 54.8 41.2 61.8
108 104.9 81.1 121.7









#66
4,6-Dinitro-2-methylphenol
Concen: 46.010 ng/ul
RT: 15.786 min Scan# 2159
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:198 Resp: 246999
Ion Ratio Lower Upper
198 100
182 4.1 4.6 6.8#
77 25.7 23.8 35.8



#67
N-Nitrosodiphenylamine
Concen: 47.120 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:169 Resp: 1093182
Ion Ratio Lower Upper
169 100
168 67.6 53.0 79.4
167 37.1 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 48.940 ng/ul
RT: 16.586 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:248 Resp: 437483
Ion Ratio Lower Upper
248 100
250 94.8 79.3 118.9
141 69.5 54.2 81.2



#69
Hexachlorobenzene
Concen: 49.178 ng/ul
RT: 16.669 min Scan# 2309
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:284 Resp: 517225
Ion Ratio Lower Upper
284 100
142 30.4 28.2 42.4
249 30.5 26.2 39.4









#70
Atrazine
Concen: 47.145 ng/ul
RT: 16.869 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:200 Resp: 423163
Ion Ratio Lower Upper
200 100
173 25.9 19.5 29.3
215 48.2 37.8 56.8



#71
Pentachlorophenol
Concen: 49.367 ng/ul
RT: 17.039 min Scan# 2372
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:266 Resp: 308171
Ion Ratio Lower Upper
266 100
264 62.2 52.5 78.7
268 62.3 54.6 81.8









#72
Phenanthrene
Concen: 47.549 ng/ul
RT: 17.457 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:178 Resp: 2138052
Ion Ratio Lower Upper
178 100
179 15.3 12.4 18.6
176 19.0 15.8 23.6



#74
Anthracene
Concen: 47.842 ng/ul
RT: 17.551 min Scan# 2459
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:178 Resp: 2184631
Ion Ratio Lower Upper
178 100
179 15.8 12.3 18.5
176 18.7 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 47.699 ng/uL
RT: 13.398 min Scan# 1753
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:216 Resp: 589227
Ion Ratio Lower Upper
216 100
214 78.5 61.1 91.7
218 48.2 37.1 55.7



#76
Pentachlorobenzene
Concen: 48.086 ng/uL
RT: 14.916 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:250 Resp: 594005
Ion Ratio Lower Upper
250 100
248 63.8 49.6 74.4
252 64.2 50.9 76.3





#77

Carbazole

Concen: 46.843 ng/ul

RT: 17.839 min Scan# 2508

Delta R.T. -0.006 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:167 Resp: 1915824

Ion Ratio Lower Upper

167 100

166 21.7 17.1 25.7

139 13.4 10.0 15.0



#78

Di-n-butylphthalate

Concen: 47.900 ng/ul

RT: 18.351 min Scan# 2595

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion:149 Resp: 2331163


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 5.6 4.6 6.8





#80

Fluoranthene

Concen: 47.738 ng/ul

RT: 19.439 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNL_P

ClientSampleId :

SSTD040634





Tgt Ion:202 Resp: 2608257

Ion Ratio Lower Upper

202 100

101 9.6 7.4 11.0

100 7.5 5.9 8.9



#82

Pyrene

Concen: 47.306 ng/ul

RT: 19.798 min Scan# 2841

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion:202 Resp: 2725075


Ion Ratio Lower Upper

202 100

101 11.7 9.4 14.0

100 9.7 7.4 11.0





#83

Butylbenzylphthalate

Concen: 48.206 ng/ul

RT: 20.669 min Scan# 2989

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:149 Resp: 1063156

Ion Ratio Lower Upper

149 100

91 72.4 57.1 85.7

206 21.8 17.6 26.4



#84

3,3'-Dichlorobenzidine

Concen: 44.628 ng/ul

RT: 21.480 min Scan# 3127

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion:252 Resp: 828522


Ion Ratio Lower Upper

252 100

254 64.9 51.5 77.3

126 9.4 7.8 11.8





#85

Benzo(a)anthracene

Concen: 47.897 ng/ul

RT: 21.539 min Scan# 3136

Delta R.T. 0.006 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:228 Resp: 2846022

Ion Ratio Lower Upper

228 100

229 19.1 15.5 23.3

226 27.5 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 47.606 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion:149 Resp: 1504437


Ion Ratio Lower Upper

149 100

167 26.8 21.7 32.5

279 4.3 3.6 5.4





#87

Chrysene

Concen: 47.797 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:228 Resp: 2644924

Ion Ratio Lower Upper

228 100

226 30.1 23.8 35.6

229 19.7 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 45.319 ng/ul




RT: 22.398 min Scan# 3283


Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Tgt Ion:149 Resp: 2576445





#90

Benzo(b)fluoranthene

Concen: 47.605 ng/ul

RT: 23.327 min Scan# 3441

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:252 Resp: 2764200

Ion Ratio Lower Upper

252 100

253 21.7 17.0 25.6

125 8.7 6.6 10.0



#91

Benzo(k)fluoranthene

Concen: 48.204 ng/ul

RT: 23.380 min Scan# 3450

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04




Tgt Ion:252 Resp: 2811375


Ion Ratio Lower Upper

252 100

253 21.8 17.1 25.7

125 8.6 6.6 10.0









#93
Benzo(a)pyrene
Concen: 48.157 ng/ul
RT: 24.015 min Scan# 3557
Delta R.T. 0.006 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04




Instrument :
BNA_P
ClientSampleId :
SSTD040634


Tgt Ion:252 Resp: 2618808
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 9.3 8.1 12.1



#94
Indeno(1,2,3-cd)pyrene
Concen: 47.338 ng/ul
RT: 26.868 min Scan# 4043
Delta R.T. 0.012 min
Lab File: BP017624.D
Acq: 09 Oct 2023 19:04

Tgt Ion:276 Resp: 3126949
Ion Ratio Lower Upper
276 100
138 19.8 14.5 21.7
277 24.7 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 46.802 ng/ul

RT: 26.892 min Scan# 4047

Delta R.T. -0.000 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Instrument :

BNA_P

ClientSampleId :

SSTD040634





Tgt Ion:278 Resp: 2586010

Ion Ratio Lower Upper

278 100

139 13.5 11.0 16.6

279 24.4 19.9 29.9



#96

Benzo(g,h,i)perylene

Concen: 47.201 ng/ul

RT: 27.721 min Scan# 4188

Delta R.T. 0.006 min

Lab File: BP017624.D

Acq: 09 Oct 2023 19:04

Tgt Ion:276 Resp: 2505799

Ion Ratio Lower Upper

276 100

138 18.5 15.8 23.6

277 23.4 19.0 28.6

