

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101022\
 Data File : BP012026.D
 Acq On : 10 Oct 2022 16:40
 Operator : CG/JU
 Sample : SST01063
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010625

Quant Time: Oct 11 00:44:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 00:41:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	161787	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	665220	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	376590	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	775376	20.000	ng/u1	0.00
79) Chrysene-d12	21.357	240	831641	20.000	ng/u1	0.00
88) Perylene-d12	23.780	264	888377	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	6839	1.864	ng/uL	0.00
4) Pyridine-d5	3.711	84	53258	4.956	ng/u1	0.00
7) Phenol-d5	7.028	99	126692	9.071	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	78580	10.150	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	100350	9.101	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	94364	8.173	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	43974	8.679	ng/u1	0.00
24) 2-Nitrophenol-d4	9.752	143	42771	7.890	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	84863	8.463	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	130664	8.548	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	244347	9.073	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	293349	9.555	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	40584	7.277	ng/u1	0.00
60) Fluorene-d10	15.504	176	221765	9.482	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	29544	6.295	ng/u1	0.00
73) Anthracene-d10	17.369	188	339513	9.612	ng/u1	0.00
81) Pyrene-d10	19.604	212	394485	9.366	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.621	264	423082	9.514	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	7420	1.873	ng/uL#	95
5) Pyridine	3.734	79	56029	5.372	ng/u1#	95
6) Benzaldehyde	7.005	77	66959	10.944	ng/u1	99
8) Phenol	7.052	94	134898	9.312	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.287	93	109277	9.480	ng/u1	99
12) 2-Chlorophenol	7.422	128	110650	9.387	ng/u1	99
13) 2-Methylphenol	8.310	108	96474	8.453	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.393	45	128372	11.297	ng/u1	98
16) Acetophenone	8.693	105	158835	9.023	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.675	70	68899	8.316	ng/u1	97
18) 4-Methylphenol	8.640	108	104321	8.296	ng/u1	97
19) Hexachloroethane	8.940	117	47081	10.709	ng/u1	99
22) Nitrobenzene	9.069	77	112384	10.313	ng/u1	100
23) Isophorone	9.599	82	201427	9.112	ng/u1	100
25) 2-Nitrophenol	9.787	139	51610	8.336	ng/u1	99
26) 2,4-Dimethylphenol	9.846	107	107978	9.284	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.081	93	131504	9.220	ng/u1	99
29) 2,4-Dichlorophenol	10.316	162	90696	8.729	ng/u1	98
30) Naphthalene	10.722	128	352886	9.877	ng/u1	99
32) 4-Chloroaniline	10.840	127	136810	8.748	ng/u1	99
33) Hexachlorobutadiene	10.999	225	57884	9.413	ng/u1	99
34) Caprolactam	11.634	113	25172	6.846	ng/u1	97
35) 4-Chloro-3-methylphenol	11.963	107	89610	8.134	ng/u1	98
36) 2-Methylnaphthalene	12.334	142	221064	8.860	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101022\
 Data File : BP012026.D
 Acq On : 10 Oct 2022 16:40
 Operator : CG/JU
 Sample : SSTD01063
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD010625


Quant Time: Oct 11 00:44:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 00:41:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	222358	8.881	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.698	216	106666	9.767	ng/ul	98
40) Hexachlorocyclopentadiene	12.675	237	37377	5.991	ng/ul	97
41) 2,4,6-Trichlorophenol	12.946	196	63505	8.687	ng/ul	100
42) 2,4,5-Trichlorophenol	13.016	196	66498	8.350	ng/ul	97
43) 1,1'-Biphenyl	13.346	154	283728	10.186	ng/ul	99
44) 2-Chloronaphthalene	13.387	162	225016	10.218	ng/ul	99
45) 2-Nitroaniline	13.598	65	42516	7.909	ng/ul	98
47) Dimethylphthalate	13.969	163	261836	9.277	ng/ul	99
48) 2,6-Dinitrotoluene	14.093	165	35309	6.419	ng/ul	90
50) Acenaphthylene	14.234	152	340331	9.981	ng/ul	99
51) 3-Nitroaniline	14.428	138	40857	6.409	ng/ul	93
52) Acenaphthene	14.575	153	245476	10.204	ng/ul	100
53) 2,4-Dinitrophenol	14.634	184	18735	5.024	ng/ul	95
55) 4-Nitrophenol	14.734	109	31805	8.448	ng/ul	96
56) Dibenzofuran	14.910	168	334133	10.141	ng/ul	99
57) 2,4-Dinitrotoluene	14.881	165	60754	7.557	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.140	232	56911	8.226	ng/ul	99
59) Diethylphthalate	15.334	149	250725	8.868	ng/ul	98
61) Fluorene	15.563	166	271306	10.053	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.557	204	123975	9.280	ng/ul	99
63) 4-Nitroaniline	15.592	138	39338	6.466	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.645	198	33774	6.793	ng/ul#	98
67) N-Nitrosodiphenylamine	15.775	169	221709	9.693	ng/ul	98
68) 4-Bromophenyl-phenylether	16.457	248	69645	8.830	ng/ul	98
69) Hexachlorobenzene	16.569	284	81807	9.115	ng/ul	98
70) Atrazine	16.728	200	63119	7.648	ng/ul	97
71) Pentachlorophenol	16.916	266	42836	7.257	ng/ul	98
72) Phenanthrene	17.310	178	436977	10.255	ng/ul	99
74) Anthracene	17.404	178	433290	10.106	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.304	216	105982	9.903	ng/uL	99
76) Pentachlorobenzene	14.828	250	108464	9.484	ng/uL	98
77) Carbazole	17.675	167	388508	10.029	ng/ul	99
78) Di-n-butylphthalate	18.228	149	403161	8.499	ng/ul	100
80) Fluoranthene	19.281	202	480920	9.330	ng/ul	99
82) Pyrene	19.633	202	530698	9.889	ng/ul	98
83) Butylbenzylphthalate	20.504	149	180544	7.784	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.280	252	158255	8.266	ng/ul	97
85) Benzo(a)anthracene	21.339	228	526832	9.687	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.263	149	264862	7.581	ng/ul	99
87) Chrysene	21.392	228	512441	10.040	ng/ul	99
89) Di-n-octyl phthalate	22.192	149	464291	8.192	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	552383	9.742	ng/ul	99
91) Benzo(k)fluoranthene	23.086	252	541112	9.910	ng/ul	98
93) Benzo(a)pyrene	23.669	252	495573	9.794	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.304	276	622346	9.663	ng/ul	99
95) Dibenzo(a,h)anthracene	26.321	278	560778	9.801	ng/ul	99
96) Benzo(g,h,i)perylene	27.074	276	556577	9.747	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 11 00:44:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101022.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 11 00:41:15 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.873 ng/uL
RT: 3.317 min Scan# 50
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 88 Resp: 7420
Ion Ratio Lower Upper
88 100
43 17.0 17.2 25.8#
58 45.1 37.4 56.0



#5
Pyridine
Concen: 5.372 ng/uL
RT: 3.734 min Scan# 127
Delta R.T. 0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion: 79 Resp: 56029
Ion Ratio Lower Upper
79 100
52 56.9 44.2 66.4
51 34.9 22.9 34.3#









#6
Benzaldehyde
Concen: 10.944 ng/ul
RT: 7.005 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 77 Resp: 66959
Ion Ratio Lower Upper
77 100
105 100.0 79.8 119.8
106 98.6 77.6 116.4



#8
Phenol
Concen: 9.312 ng/ul
RT: 7.052 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion: 94 Resp: 134898
Ion Ratio Lower Upper
94 100
65 26.8 20.8 31.2
66 33.1 27.4 41.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 9.480 ng/ul
 RT: 7.287 min Scan# 731
 Delta R.T. 0.000 min
 Lab File: BP012026.D
 Acq: 10 Oct 2022 16:40

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010625

Tgt Ion: 93 Resp: 109277




Ion	Ratio	Lower	Upper
93	100		
63	76.4	61.5	92.3
95	32.8	26.6	40.0




#12
 2-Chlorophenol
 Concen: 9.387 ng/ul
 RT: 7.422 min Scan# 754
 Delta R.T. 0.000 min
 Lab File: BP012026.D
 Acq: 10 Oct 2022 16:40

Tgt Ion: 128 Resp: 110650

Ion	Ratio	Lower	Upper
128	100		
64	43.9	34.9	52.3
130	32.0	26.1	39.1





#13

2-Methylphenol

Concen: 8.453 ng/ul

RT: 8.310 min Scan# 905

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :





SSTD010625

Tgt Ion:108 Resp: 96474

Ion Ratio Lower Upper

108 100

107 88.4 72.4 108.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.297 ng/ul

RT: 8.393 min Scan# 919

Delta R.T. -0.006 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion: 45 Resp: 128372


Ion Ratio Lower Upper

45 100

77 15.6 13.4 20.0

79 11.6 9.9 14.9









#16
Acetophenone
Concen: 9.023 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:105 Resp: 158835
Ion Ratio Lower Upper
105 100
77 76.1 59.2 88.8
51 26.2 19.8 29.6



#17
N-Nitroso-di-n-propylamine
Concen: 8.316 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion: 70 Resp: 68899
Ion Ratio Lower Upper
70 100
42 48.4 39.9 59.9
101 9.7 7.0 10.6
130 18.7 16.8 25.2









#18
4-Methylphenol
Concen: 8.296 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:108 Resp: 104321
Ion Ratio Lower Upper
108 100
107 109.6 90.1 135.1



#19
Hexachloroethane
Concen: 10.709 ng/ul
RT: 8.940 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:117 Resp: 47081
Ion Ratio Lower Upper
117 100
201 92.8 74.2 111.4
199 59.9 46.3 69.5









#22
Nitrobenzene
Concen: 10.313 ng/ul
RT: 9.069 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion: 77 Resp: 112384
Ion Ratio Lower Upper
77 100
123 45.7 36.6 54.8
65 12.1 9.6 14.4



#23
Isophorone
Concen: 9.112 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion: 82 Resp: 201427
Ion Ratio Lower Upper
82 100
95 6.5 5.2 7.8
138 17.8 14.4 21.6









#25
2-Nitrophenol
Concen: 8.336 ng/ul
RT: 9.787 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:139 Resp: 51610
Ion Ratio Lower Upper
139 100
65 47.5 37.7 56.5
109 26.7 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 9.284 ng/ul
RT: 9.846 min Scan# 1166
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:107 Resp: 107978
Ion Ratio Lower Upper
107 100
121 53.5 41.1 61.7
122 90.6 71.1 106.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.220 ng/ul
 RT: 10.081 min Scan# 1206
 Delta R.T. -0.006 min
 Lab File: BP012026.D
 Acq: 10 Oct 2022 16:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010625


Tgt Ion: 93 Resp: 131504
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.2 39.2
 123 10.9 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 8.729 ng/ul
 RT: 10.316 min Scan# 1246
 Delta R.T. 0.000 min
 Lab File: BP012026.D
 Acq: 10 Oct 2022 16:40

Tgt Ion: 162 Resp: 90696
 Ion Ratio Lower Upper
 162 100
 164 63.9 52.5 78.7
 98 35.4 27.9 41.9









#30
Naphthalene
Concen: 9.877 ng/ul
RT: 10.722 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:128 Resp: 352886
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 13.0 10.1 15.1



#32
4-Chloroaniline
Concen: 8.748 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:127 Resp: 136810
Ion Ratio Lower Upper
127 100
129 31.5 25.8 38.8





#33

Hexachlorobutadiene

Concen: 9.413 ng/ul

RT: 10.999 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:225 Resp: 57884

Ion Ratio Lower Upper

225 100

223 63.3 49.8 74.8

227 62.7 51.0 76.4



#34

Caprolactam

Concen: 6.846 ng/ul

RT: 11.634 min Scan# 1470

Delta R.T. -0.006 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:113 Resp: 25172


Ion Ratio Lower Upper

113 100

55 161.1 132.9 199.3

56 119.8 97.7 146.5









#35
4-Chloro-3-methylphenol
Concen: 8.134 ng/ul
RT: 11.963 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:107 Resp: 89610
Ion Ratio Lower Upper
107 100
144 26.5 21.8 32.6
142 86.4 67.1 100.7



#36
2-Methylnaphthalene
Concen: 8.860 ng/ul
RT: 12.334 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:142 Resp: 221064
Ion Ratio Lower Upper
142 100
141 87.5 71.4 107.0









#37
1-Methylnaphthalene
Concen: 8.881 ng/ul
RT: 12.551 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:142 Resp: 222358
Ion Ratio Lower Upper
142 100
141 93.7 74.3 111.5
116 3.8 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.767 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:216 Resp: 106666
Ion Ratio Lower Upper
216 100
214 77.4 63.8 95.6
179 20.8 17.3 25.9
108 18.6 15.2 22.8









#40
Hexachlorocyclopentadiene
Concen: 5.991 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:237 Resp: 37377
Ion Ratio Lower Upper
237 100
235 63.2 52.4 78.6
272 11.8 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 8.687 ng/ul
RT: 12.946 min Scan# 1693
Delta R.T. 0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:196 Resp: 63505
Ion Ratio Lower Upper
196 100
198 94.0 75.2 112.8
200 29.0 23.4 35.2





#42

2,4,5-Trichlorophenol

Concen: 8.350 ng/ul

RT: 13.016 min Scan# 1705

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:196 Resp: 66498

Ion Ratio Lower Upper

196 100

198 98.3 76.5 114.7

200 30.3 25.3 37.9



#43

1,1'-Biphenyl

Concen: 10.186 ng/ul

RT: 13.346 min Scan# 1761

Delta R.T. 0.006 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:154 Resp: 283728


Ion Ratio Lower Upper

154 100

153 40.4 31.8 47.8

76 13.6 10.9 16.3









#44
2-Chloronaphthalene
Concen: 10.218 ng/ul
RT: 13.387 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:162 Resp: 225016
Ion Ratio Lower Upper
162 100
164 32.9 26.1 39.1
127 38.1 30.2 45.2



#45
2-Nitroaniline
Concen: 7.909 ng/ul
RT: 13.598 min Scan# 1804
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion: 65 Resp: 42516
Ion Ratio Lower Upper
65 100
92 71.6 57.3 85.9
138 108.4 88.8 133.2









#47
Dimethylphthalate
Concen: 9.277 ng/ul
RT: 13.969 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:163 Resp: 261836
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 6.419 ng/ul
RT: 14.093 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:165 Resp: 35309
Ion Ratio Lower Upper
165 100
89 68.5 48.2 72.2
121 27.1 19.3 28.9









#50
Acenaphthylene
Concen: 9.981 ng/ul
RT: 14.234 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:152 Resp: 340331
Ion Ratio Lower Upper
152 100
151 20.3 15.8 23.6
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 6.409 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. 0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:138 Resp: 40857
Ion Ratio Lower Upper
138 100
108 13.5 10.2 15.4
92 112.3 96.6 144.8









#52
Acenaphthene
Concen: 10.204 ng/ul
RT: 14.575 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:153 Resp: 245476
Ion Ratio Lower Upper
153 100
152 49.1 39.2 58.8
154 87.8 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 5.024 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:184 Resp: 18735
Ion Ratio Lower Upper
184 100
63 63.4 52.0 78.0
154 64.2 46.6 70.0









#55
4-Nitrophenol
Concen: 8.448 ng/ul
RT: 14.734 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:109 Resp: 31805
Ion Ratio Lower Upper
109 100
139 154.2 124.5 186.7
65 142.5 107.8 161.6



#56
Dibenzofuran
Concen: 10.141 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:168 Resp: 334133
Ion Ratio Lower Upper
168 100
139 36.6 30.0 45.0





#57

2,4-Dinitrotoluene

Concen: 7.557 ng/ul

RT: 14.881 min Scan# 2022

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:165 Resp: 60754

Ion Ratio Lower Upper

165 100

63 43.8 32.9 49.3

182 3.2 2.1 3.1#



#58

2,3,4,6-Tetrachlorophenol

Concen: 8.226 ng/ul

RT: 15.140 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Tgt Ion:232 Resp: 56911




Ion Ratio Lower Upper


232 100

131 49.3 39.3 58.9

130 2.2 1.8 2.8

166 31.4 25.7 38.5









#59
Diethylphthalate
Concen: 8.868 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:149 Resp: 250725
Ion Ratio Lower Upper
149 100
177 22.2 16.8 25.2
150 12.5 10.0 15.0



#61
Fluorene
Concen: 10.053 ng/ul
RT: 15.563 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:166 Resp: 271306
Ion Ratio Lower Upper
166 100
165 97.4 79.3 118.9
167 13.2 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 9.280 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:204 Resp: 123975
Ion Ratio Lower Upper
204 100
206 33.3 25.9 38.9
141 59.9 47.9 71.9



#63
4-Nitroaniline
Concen: 6.466 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:138 Resp: 39338
Ion Ratio Lower Upper
138 100
92 55.7 40.8 61.2
108 80.1 59.5 89.3









#66
4,6-Dinitro-2-methylphenol
Concen: 6.793 ng/ul
RT: 15.645 min Scan# 2152
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:198 Resp: 33774
Ion Ratio Lower Upper
198 100
182 9.3 5.6 8.4#
77 24.7 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 9.693 ng/ul
RT: 15.775 min Scan# 2174
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:169 Resp: 221709
Ion Ratio Lower Upper
169 100
168 67.4 52.2 78.2
167 36.9 30.1 45.1









#68
4-Bromophenyl-phenylether
Concen: 8.830 ng/ul
RT: 16.457 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:248 Resp: 69645
Ion Ratio Lower Upper
248 100
250 95.9 75.1 112.7
141 74.7 58.8 88.2



#69
Hexachlorobenzene
Concen: 9.115 ng/ul
RT: 16.569 min Scan# 2309
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:284 Resp: 81807
Ion Ratio Lower Upper
284 100
142 38.5 32.1 48.1
249 33.5 26.0 39.0





#70

Atrazine

Concen: 7.648 ng/ul

RT: 16.728 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:200 Resp: 63119

Ion Ratio Lower Upper

200 100

173 27.8 20.9 31.3

215 48.2 37.0 55.4



#71

Pentachlorophenol

Concen: 7.257 ng/ul

RT: 16.916 min Scan# 2368

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:266 Resp: 42836


Ion Ratio Lower Upper

266 100

264 63.3 49.7 74.5

268 60.5 49.7 74.5





#72

Phenanthrene

Concen: 10.255 ng/ul

RT: 17.310 min Scan# 2435

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:178 Resp: 436977

Ion Ratio Lower Upper

178 100

179 15.1 12.5 18.7

176 18.9 15.5 23.3



#74

Anthracene

Concen: 10.106 ng/ul

RT: 17.404 min Scan# 2451

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:178 Resp: 433290


Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 18.0 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.903 ng/uL

RT: 13.304 min Scan# 1754

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:216 Resp: 105982

Ion Ratio Lower Upper

216 100

214 80.9 64.0 96.0

218 48.2 38.2 57.2



#76

Pentachlorobenzene

Concen: 9.484 ng/uL

RT: 14.828 min Scan# 2013

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:250 Resp: 108464


Ion Ratio Lower Upper

250 100

248 61.8 50.9 76.3

252 63.7 51.5 77.3









#77
Carbazole
Concen: 10.029 ng/ul
RT: 17.675 min Scan# 2498
Delta R.T. -0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:167 Resp: 388508
Ion Ratio Lower Upper
167 100
166 21.9 17.2 25.8
139 13.2 10.7 16.1



#78
Di-n-butylphthalate
Concen: 8.499 ng/ul
RT: 18.228 min Scan# 2591
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:149 Resp: 403161
Ion Ratio Lower Upper
149 100
150 9.0 7.0 10.6
104 4.8 3.7 5.5





#80

Fluoranthene

Concen: 9.330 ng/ul

RT: 19.281 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:202 Resp: 480920

Ion Ratio Lower Upper

202 100

101 12.7 10.2 15.2

100 9.2 7.7 11.5



#82

Pyrene

Concen: 9.889 ng/ul

RT: 19.633 min Scan# 2830

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:202 Resp: 530698


Ion Ratio Lower Upper

202 100

101 14.8 11.6 17.4

100 11.2 9.8 14.8





#83

Butylbenzylphthalate

Concen: 7.784 ng/ul

RT: 20.504 min Scan# 2978

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:149 Resp: 180544

Ion Ratio Lower Upper

149 100

91 72.3 56.6 84.8

206 18.3 15.0 22.6



#84

3,3'-Dichlorobenzidine

Concen: 8.266 ng/ul

RT: 21.280 min Scan# 3110

Delta R.T. 0.006 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:252 Resp: 158255


Ion Ratio Lower Upper

252 100

254 65.5 50.5 75.7

126 12.2 11.0 16.4









#85
Benzo(a)anthracene
Concen: 9.687 ng/ul
RT: 21.339 min Scan# 3121
Delta R.T. -0.006 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40




Instrument :
BNA_P
ClientSampleId :
SSTD010625


Tgt Ion:228 Resp: 526832
Ion Ratio Lower Upper
228 100
229 19.8 15.5 23.3
226 27.4 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 7.581 ng/ul
RT: 21.263 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP012026.D
Acq: 10 Oct 2022 16:40

Tgt Ion:149 Resp: 264862
Ion Ratio Lower Upper
149 100
167 25.5 21.0 31.6
279 3.6 2.9 4.3





#87

Chrysene

Concen: 10.040 ng/ul

RT: 21.392 min Scan# 3129

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:228 Resp: 512441

Ion Ratio Lower Upper

228 100

226 29.8 24.1 36.1

229 19.3 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 8.192 ng/ul




RT: 22.192 min Scan# 3265


Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Tgt Ion:149 Resp: 464291





#90

Benzo(b)fluoranthene

Concen: 9.742 ng/ul

RT: 23.039 min Scan# 3409

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:252 Resp: 552383

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8

125 12.6 9.9 14.9



#91

Benzo(k)fluoranthene

Concen: 9.910 ng/ul

RT: 23.086 min Scan# 3417

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:252 Resp: 541112


Ion Ratio Lower Upper

252 100

253 22.6 17.4 26.0

125 12.3 9.4 14.2





#93

Benzo(a)pyrene

Concen: 9.794 ng/ul

RT: 23.669 min Scan# 3516

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:252 Resp: 495573

Ion Ratio Lower Upper

252 100

253 21.0 17.3 25.9

125 13.8 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.663 ng/ul

RT: 26.304 min Scan# 3964

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40




Tgt Ion:276 Resp: 622346


Ion Ratio Lower Upper

276 100

138 30.3 24.1 36.1

277 24.1 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 9.801 ng/ul

RT: 26.321 min Scan# 3966

Delta R.T. 0.006 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Instrument :

BNA_P

ClientSampleId :

SSTD010625





Tgt Ion:278 Resp: 560778

Ion Ratio Lower Upper

278 100

139 18.9 15.4 23.0

279 23.8 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 9.747 ng/ul

RT: 27.074 min Scan# 4095

Delta R.T. 0.000 min

Lab File: BP012026.D

Acq: 10 Oct 2022 16:40

Tgt Ion:276 Resp: 556577

Ion Ratio Lower Upper

276 100

138 26.8 20.9 31.3

277 23.9 19.0 28.4

