

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101022\
 Data File : BP012027.D
 Acq On : 10 Oct 2022 17:14
 Operator : CG/JU
 Sample : SST02064
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020626

Quant Time: Oct 11 00:45:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 00:41:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	156934	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	620385	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	346547	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	720318	20.000	ng/u1	0.00
79) Chrysene-d12	21.357	240	834377	20.000	ng/u1	0.00
88) Perylene-d12	23.780	264	886017	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	13713	3.854	ng/uL	0.00
4) Pyridine-d5	3.705	84	114164	10.953	ng/u1	0.00
7) Phenol-d5	7.028	99	268833	19.843	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	164655	21.926	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	215756	20.174	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	199840	17.844	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	97909	20.720	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	97076	19.201	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	186606	19.955	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	278071	19.505	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	491234	19.822	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	617160	21.845	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	102811	20.034	ng/u1	0.00
60) Fluorene-d10	15.504	176	449402	20.881	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	79807	18.305	ng/u1	0.00
73) Anthracene-d10	17.369	188	743060	22.646	ng/u1	0.00
81) Pyrene-d10	19.604	212	876126	20.734	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.621	264	949217	21.401	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	14269	3.713	ng/uL	100
5) Pyridine	3.728	79	118948	11.758	ng/u1	100
6) Benzaldehyde	7.005	77	156642	26.394	ng/u1	100
8) Phenol	7.057	94	289090	20.573	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.287	93	227421	20.339	ng/u1	100
12) 2-Chlorophenol	7.422	128	236384	20.674	ng/u1	100
13) 2-Methylphenol	8.310	108	206381	18.642	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.399	45	267156	24.236	ng/u1	100
16) Acetophenone	8.693	105	333757	19.546	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.675	70	145858	18.149	ng/u1	100
18) 4-Methylphenol	8.640	108	224437	18.399	ng/u1	100
19) Hexachloroethane	8.940	117	98452	23.087	ng/u1	100
22) Nitrobenzene	9.075	77	244509	24.059	ng/u1	100
23) Isophorone	9.599	82	421898	20.464	ng/u1	100
25) 2-Nitrophenol	9.787	139	114999	19.916	ng/u1	100
26) 2,4-Dimethylphenol	9.846	107	230222	21.224	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.087	93	273605	20.570	ng/u1	100
29) 2,4-Dichlorophenol	10.316	162	196630	20.293	ng/u1	100
30) Naphthalene	10.722	128	725043	21.759	ng/u1	100
32) 4-Chloroaniline	10.840	127	289523	19.851	ng/u1	100
33) Hexachlorobutadiene	10.998	225	117724	20.527	ng/u1	100
34) Caprolactam	11.640	113	57146	16.664	ng/u1	100
35) 4-Chloro-3-methylphenol	11.963	107	194996	18.980	ng/u1	100
36) 2-Methylnaphthalene	12.334	142	456023	19.597	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101022\
 Data File : BP012027.D
 Acq On : 10 Oct 2022 17:14
 Operator : CG/JU
 Sample : SST02064
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020626

Quant Time: Oct 11 00:45:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 00:41:15 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	459772	19.690	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.698	216	219905	21.881	ng/ul	100
40) Hexachlorocyclopentadiene	12.675	237	90322	15.732	ng/ul	100
41) 2,4,6-Trichlorophenol	12.940	196	138085	20.526	ng/ul	100
42) 2,4,5-Trichlorophenol	13.016	196	149282	20.371	ng/ul	100
43) 1,1'-Biphenyl	13.340	154	569493	22.218	ng/ul	100
44) 2-Chloronaphthalene	13.387	162	461985	22.797	ng/ul	100
45) 2-Nitroaniline	13.598	65	100370	20.290	ng/ul	100
47) Dimethylphthalate	13.969	163	526234	20.262	ng/ul	100
48) 2,6-Dinitrotoluene	14.092	165	86291	17.049	ng/ul	100
50) Acenaphthylene	14.234	152	705672	22.490	ng/ul	100
51) 3-Nitroaniline	14.422	138	104482	17.810	ng/ul	100
52) Acenaphthene	14.575	153	489336	22.105	ng/ul	100
53) 2,4-Dinitrophenol	14.634	184	50208	14.631	ng/ul	100
55) 4-Nitrophenol	14.734	109	80084	23.117	ng/ul	100
56) Dibenzofuran	14.910	168	659378	21.747	ng/ul	100
57) 2,4-Dinitrotoluene	14.881	165	147510	19.938	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.139	232	121714	19.119	ng/ul	100
59) Diethylphthalate	15.334	149	514374	19.769	ng/ul	100
61) Fluorene	15.563	166	540959	21.783	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.557	204	244306	19.872	ng/ul	100
63) 4-Nitroaniline	15.592	138	109556	19.570	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.645	198	88415	19.142	ng/ul	100
67) N-Nitrosodiphenylamine	15.775	169	444803	20.933	ng/ul	100
68) 4-Bromophenyl-phenylether	16.457	248	140074	19.118	ng/ul	100
69) Hexachlorobenzene	16.569	284	165565	19.857	ng/ul	100
70) Atrazine	16.733	200	147173	19.196	ng/ul	100
71) Pentachlorophenol	16.916	266	98383	17.941	ng/ul	100
72) Phenanthrene	17.310	178	874869	22.100	ng/ul	100
74) Anthracene	17.404	178	934678	23.467	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.304	216	217233	21.849	ng/ul	100
76) Pentachlorobenzene	14.828	250	215524	20.287	ng/ul	100
77) Carbazole	17.680	167	875374	24.325	ng/ul	100
78) Di-n-butylphthalate	18.227	149	913354	20.725	ng/ul	100
80) Fluoranthene	19.280	202	1077588	20.838	ng/ul	100
82) Pyrene	19.633	202	1157826	21.504	ng/ul	100
83) Butylbenzylphthalate	20.504	149	416379	17.893	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.274	252	363827	18.941	ng/ul	100
85) Benzo(a)anthracene	21.345	228	1175624	21.546	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.263	149	598885	17.085	ng/ul	100
87) Chrysene	21.392	228	1116986	21.813	ng/ul	100
89) Di-n-octyl phthalate	22.192	149	1049903	18.573	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	1227236	21.701	ng/ul	100
91) Benzo(k)fluoranthene	23.086	252	1236096	22.698	ng/ul	100
93) Benzo(a)pyrene	23.668	252	1112673	22.047	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.304	276	1412658	21.991	ng/ul	100
95) Dibenzo(a,h)anthracene	26.315	278	1264412	22.158	ng/ul	100
96) Benzo(g,h,i)perylene	27.074	276	1271672	22.330	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101022\
 Data File : BP012027.D
 Acq On : 10 Oct 2022 17:14
 Operator : CG/JU
 Sample : SSTD02064
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020626

Quant Time: Oct 11 00:45:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 11 00:41:15 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.713 ng/uL
RT: 3.317 min Scan# 50
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion: 88 Resp: 14269
Ion Ratio Lower Upper
88 100
43 21.5 17.2 25.8
58 46.7 37.4 56.0



#5
Pyridine
Concen: 11.758 ng/uL
RT: 3.728 min Scan# 126
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion: 79 Resp: 118948
Ion Ratio Lower Upper
79 100
52 55.3 44.2 66.4
51 28.6 22.9 34.3









#6
Benzaldehyde
Concen: 26.394 ng/ul
RT: 7.005 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion: 77 Resp: 156642
Ion Ratio Lower Upper
77 100
105 99.8 79.8 119.8
106 97.0 77.6 116.4



#8
Phenol
Concen: 20.573 ng/ul
RT: 7.057 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion: 94 Resp: 289090
Ion Ratio Lower Upper
94 100
65 26.0 20.8 31.2
66 34.2 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.339 ng/ul
RT: 7.287 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion: 93 Resp: 227421
Ion Ratio Lower Upper
93 100
63 76.9 61.5 92.3
95 33.3 26.6 40.0



#12
2-Chlorophenol
Concen: 20.674 ng/ul
RT: 7.422 min Scan# 754
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:128 Resp: 236384
Ion Ratio Lower Upper
128 100
64 43.6 34.9 52.3
130 32.6 26.1 39.1





#13

2-Methylphenol

Concen: 18.642 ng/ul

RT: 8.310 min Scan# 905

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :





SSTD020626

Tgt Ion:108 Resp: 206381

Ion Ratio Lower Upper

108 100

107 90.5 72.4 108.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.236 ng/ul

RT: 8.399 min Scan# 920

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion: 45 Resp: 267156


Ion Ratio Lower Upper

45 100

77 16.7 13.4 20.0

79 12.4 9.9 14.9









#16
Acetophenone
Concen: 19.546 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:105 Resp: 333757
Ion Ratio Lower Upper
105 100
77 74.0 59.2 88.8
51 24.7 19.8 29.6



#17
N-Nitroso-di-n-propylamine
Concen: 18.149 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion: 70 Resp: 145858
Ion Ratio Lower Upper
70 100
42 49.9 39.9 59.9
101 8.8 7.0 10.6
130 21.0 16.8 25.2









#18
4-Methylphenol
Concen: 18.399 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:108 Resp: 224437
Ion Ratio Lower Upper
108 100
107 112.6 90.1 135.1



#19
Hexachloroethane
Concen: 23.087 ng/ul
RT: 8.940 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:117 Resp: 98452
Ion Ratio Lower Upper
117 100
201 92.8 74.2 111.4
199 57.9 46.3 69.5









#22
Nitrobenzene
Concen: 24.059 ng/ul
RT: 9.075 min Scan# 1035
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion: 77 Resp: 244509
Ion Ratio Lower Upper
77 100
123 45.7 36.6 54.8
65 12.0 9.6 14.4



#23
Isophorone
Concen: 20.464 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion: 82 Resp: 421898
Ion Ratio Lower Upper
82 100
95 6.5 5.2 7.8
138 18.0 14.4 21.6









#25
2-Nitrophenol
Concen: 19.916 ng/ul
RT: 9.787 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:139 Resp: 114999
Ion Ratio Lower Upper
139 100
65 47.1 37.7 56.5
109 26.5 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 21.224 ng/ul
RT: 9.846 min Scan# 1166
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:107 Resp: 230222
Ion Ratio Lower Upper
107 100
121 51.4 41.1 61.7
122 88.9 71.1 106.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.570 ng/ul
 RT: 10.087 min Scan# 1207
 Delta R.T. 0.000 min
 Lab File: BP012027.D
 Acq: 10 Oct 2022 17:14




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020626


Tgt Ion: 93 Resp: 273605
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.2 39.2
 123 11.5 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 20.293 ng/ul
 RT: 10.316 min Scan# 1246
 Delta R.T. 0.000 min
 Lab File: BP012027.D
 Acq: 10 Oct 2022 17:14

Tgt Ion: 162 Resp: 196630
 Ion Ratio Lower Upper
 162 100
 164 65.6 52.5 78.7
 98 34.9 27.9 41.9









#30
Naphthalene
Concen: 21.759 ng/ul
RT: 10.722 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:128 Resp: 725043
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.6 10.1 15.1



#32
4-Chloroaniline
Concen: 19.851 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:127 Resp: 289523
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8





#33

Hexachlorobutadiene

Concen: 20.527 ng/ul

RT: 10.998 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:225 Resp: 117724

Ion Ratio Lower Upper

225 100

223 62.3 49.8 74.8

227 63.7 51.0 76.4



#34

Caprolactam

Concen: 16.664 ng/ul

RT: 11.640 min Scan# 1471

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:113 Resp: 57146


Ion Ratio Lower Upper

113 100

55 166.1 132.9 199.3

56 122.1 97.7 146.5









#35
4-Chloro-3-methylphenol
Concen: 18.980 ng/ul
RT: 11.963 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:107 Resp: 194996
Ion Ratio Lower Upper
107 100
144 27.2 21.8 32.6
142 83.9 67.1 100.7



#36
2-Methylnaphthalene
Concen: 19.597 ng/ul
RT: 12.334 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:142 Resp: 456023
Ion Ratio Lower Upper
142 100
141 89.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 19.690 ng/ul

RT: 12.551 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:142 Resp: 459772

Ion Ratio Lower Upper

142 100

141 92.9 74.3 111.5

116 3.5 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.881 ng/ul

RT: 12.698 min Scan# 1651

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Tgt Ion:216 Resp: 219905




Ion Ratio Lower Upper


216 100

214 79.7 63.8 95.6

179 21.6 17.3 25.9

108 19.0 15.2 22.8









#40
Hexachlorocyclopentadiene
Concen: 15.732 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:237 Resp: 90322
Ion Ratio Lower Upper
237 100
235 65.5 52.4 78.6
272 13.3 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 20.526 ng/ul
RT: 12.940 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:196 Resp: 138085
Ion Ratio Lower Upper
196 100
198 94.0 75.2 112.8
200 29.3 23.4 35.2









#42
2,4,5-Trichlorophenol
Concen: 20.371 ng/ul
RT: 13.016 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:196 Resp: 149282
Ion Ratio Lower Upper
196 100
198 95.6 76.5 114.7
200 31.6 25.3 37.9



#43
1,1'-Biphenyl
Concen: 22.218 ng/ul
RT: 13.340 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:154 Resp: 569493
Ion Ratio Lower Upper
154 100
153 39.8 31.8 47.8
76 13.6 10.9 16.3









#44
2-Chloronaphthalene
Concen: 22.797 ng/ul
RT: 13.387 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:162 Resp: 461985
Ion Ratio Lower Upper
162 100
164 32.6 26.1 39.1
127 37.7 30.2 45.2



#45
2-Nitroaniline
Concen: 20.290 ng/ul
RT: 13.598 min Scan# 1804
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion: 65 Resp: 100370
Ion Ratio Lower Upper
65 100
92 71.6 57.3 85.9
138 111.0 88.8 133.2









#47
Dimethylphthalate
Concen: 20.262 ng/ul
RT: 13.969 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:163 Resp: 526234
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 17.049 ng/ul
RT: 14.092 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:165 Resp: 86291
Ion Ratio Lower Upper
165 100
89 60.2 48.2 72.2
121 24.1 19.3 28.9









#50
Acenaphthylene
Concen: 22.490 ng/ul
RT: 14.234 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:152 Resp: 705672
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 17.810 ng/ul
RT: 14.422 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:138 Resp: 104482
Ion Ratio Lower Upper
138 100
108 12.8 10.2 15.4
92 120.7 96.6 144.8





#52

Acenaphthene

Concen: 22.105 ng/ul

RT: 14.575 min Scan# 1970

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:153 Resp: 489336

Ion Ratio Lower Upper

153 100

152 49.0 39.2 58.8

154 88.0 70.4 105.6



#53

2,4-Dinitrophenol

Concen: 14.631 ng/ul

RT: 14.634 min Scan# 1980

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:184 Resp: 50208


Ion Ratio Lower Upper

184 100

63 65.0 52.0 78.0

154 58.3 46.6 70.0









#55
4-Nitrophenol
Concen: 23.117 ng/ul
RT: 14.734 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:109 Resp: 80084
Ion Ratio Lower Upper
109 100
139 155.6 124.5 186.7
65 134.7 107.8 161.6



#56
Dibenzofuran
Concen: 21.747 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:168 Resp: 659378
Ion Ratio Lower Upper
168 100
139 37.5 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 19.938 ng/ul
RT: 14.881 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:165 Resp: 147510
Ion Ratio Lower Upper
165 100
63 41.1 32.9 49.3
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.119 ng/ul
RT: 15.139 min Scan# 2066
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:232 Resp: 121714
Ion Ratio Lower Upper
232 100
131 49.1 39.3 58.9
130 2.3 1.8 2.8
166 32.1 25.7 38.5









#59
Diethylphthalate
Concen: 19.769 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:149 Resp: 514374
Ion Ratio Lower Upper
149 100
177 21.0 16.8 25.2
150 12.5 10.0 15.0



#61
Fluorene
Concen: 21.783 ng/ul
RT: 15.563 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:166 Resp: 540959
Ion Ratio Lower Upper
166 100
165 99.1 79.3 118.9
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 19.872 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:204 Resp: 244306
Ion Ratio Lower Upper
204 100
206 32.4 25.9 38.9
141 59.9 47.9 71.9



#63
4-Nitroaniline
Concen: 19.570 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:138 Resp: 109556
Ion Ratio Lower Upper
138 100
92 51.0 40.8 61.2
108 74.4 59.5 89.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.142 ng/ul
RT: 15.645 min Scan# 2152
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:198 Resp: 88415
Ion Ratio Lower Upper
198 100
182 7.0 5.6 8.4
77 25.4 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 20.933 ng/ul
RT: 15.775 min Scan# 2174
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:169 Resp: 444803
Ion Ratio Lower Upper
169 100
168 65.2 52.2 78.2
167 37.6 30.1 45.1









#68
4-Bromophenyl-phenylether
Concen: 19.118 ng/ul
RT: 16.457 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:248 Resp: 140074
Ion Ratio Lower Upper
248 100
250 93.9 75.1 112.7
141 73.5 58.8 88.2



#69
Hexachlorobenzene
Concen: 19.857 ng/ul
RT: 16.569 min Scan# 2309
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:284 Resp: 165565
Ion Ratio Lower Upper
284 100
142 40.1 32.1 48.1
249 32.5 26.0 39.0





#70

Atrazine

Concen: 19.196 ng/ul

RT: 16.733 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:200 Resp: 147173

Ion Ratio Lower Upper

200 100

173 26.1 20.9 31.3

215 46.2 37.0 55.4



#71

Pentachlorophenol

Concen: 17.941 ng/ul

RT: 16.916 min Scan# 2368

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:266 Resp: 98383


Ion Ratio Lower Upper

266 100

264 62.1 49.7 74.5

268 62.1 49.7 74.5





#72

Phenanthrene

Concen: 22.100 ng/ul

RT: 17.310 min Scan# 2435

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:178 Resp: 874869

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.4 15.5 23.3



#74

Anthracene

Concen: 23.467 ng/ul

RT: 17.404 min Scan# 2451

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:178 Resp: 934678


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.0 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.849 ng/uL
RT: 13.304 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14




Instrument :
BNA_P
ClientSampleId :
SSTD020626


Tgt Ion:216 Resp: 217233
Ion Ratio Lower Upper
216 100
214 80.0 64.0 96.0
218 47.7 38.2 57.2



#76
Pentachlorobenzene
Concen: 20.287 ng/uL
RT: 14.828 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP012027.D
Acq: 10 Oct 2022 17:14

Tgt Ion:250 Resp: 215524
Ion Ratio Lower Upper
250 100
248 63.6 50.9 76.3
252 64.4 51.5 77.3





#77

Carbazole

Concen: 24.325 ng/ul

RT: 17.680 min Scan# 2498

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:167 Resp: 875374

Ion Ratio Lower Upper

167 100

166 21.5 17.2 25.8

139 13.4 10.7 16.1



#78

Di-n-butylphthalate

Concen: 20.725 ng/ul

RT: 18.227 min Scan# 2591

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:149 Resp: 913354


Ion Ratio Lower Upper

149 100

150 8.8 7.0 10.6

104 4.6 3.7 5.5





#80

Fluoranthene

Concen: 20.838 ng/ul

RT: 19.280 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:202 Resp: 1077588

Ion Ratio Lower Upper

202 100

101 12.7 10.2 15.2

100 9.6 7.7 11.5



#82

Pyrene

Concen: 21.504 ng/ul

RT: 19.633 min Scan# 2830

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:202 Resp: 1157826


Ion Ratio Lower Upper

202 100

101 14.5 11.6 17.4

100 12.3 9.8 14.8





#83

Butylbenzylphthalate

Concen: 17.893 ng/ul

RT: 20.504 min Scan# 2978

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:149 Resp: 416379

Ion Ratio Lower Upper

149 100

91 70.7 56.6 84.8

206 18.8 15.0 22.6



#84

3,3'-Dichlorobenzidine

Concen: 18.941 ng/ul

RT: 21.274 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:252 Resp: 363827


Ion Ratio Lower Upper

252 100

254 63.1 50.5 75.7

126 13.7 11.0 16.4





#85

Benzo(a)anthracene

Concen: 21.546 ng/ul

RT: 21.345 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:228 Resp: 1175624

Ion Ratio Lower Upper

228 100

229 19.4 15.5 23.3

226 27.2 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.085 ng/ul

RT: 21.263 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:149 Resp: 598885


Ion Ratio Lower Upper

149 100

167 26.3 21.0 31.6

279 3.6 2.9 4.3





#87

Chrysene

Concen: 21.813 ng/ul

RT: 21.392 min Scan# 3129

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:228 Resp: 1116986

Ion Ratio Lower Upper

228 100

226 30.1 24.1 36.1

229 19.7 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 18.573 ng/ul




RT: 22.192 min Scan# 3265


Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Tgt Ion:149 Resp: 1049903





#90

Benzo(b)fluoranthene

Concen: 21.701 ng/ul

RT: 23.039 min Scan# 3409

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:252 Resp: 1227236

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 12.4 9.9 14.9



#91

Benzo(k)fluoranthene

Concen: 22.698 ng/ul

RT: 23.086 min Scan# 3417

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:252 Resp: 1236096


Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 11.8 9.4 14.2





#93

Benzo(a)pyrene

Concen: 22.047 ng/ul

RT: 23.668 min Scan# 3516

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:252 Resp: 1112673

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 13.5 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.991 ng/ul

RT: 26.304 min Scan# 3964

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14




Tgt Ion:276 Resp: 1412658


Ion Ratio Lower Upper

276 100

138 30.1 24.1 36.1

277 25.0 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 22.158 ng/ul

RT: 26.315 min Scan# 3966

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Instrument :

BNA_P

ClientSampleId :

SSTD020626





Tgt Ion:278 Resp: 1264412

Ion Ratio Lower Upper

278 100

139 19.2 15.4 23.0

279 24.5 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 22.330 ng/ul

RT: 27.074 min Scan# 4095

Delta R.T. 0.000 min

Lab File: BP012027.D

Acq: 10 Oct 2022 17:14

Tgt Ion:276 Resp: 1271672

Ion Ratio Lower Upper

276 100

138 26.1 20.9 31.3

277 23.7 19.0 28.4

