

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101723\
 Data File : BP017700.D
 Acq On : 17 Oct 2023 16:21
 Operator : MA/JU
 Sample : 04889-10
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Oct 18 00:43:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 16:14:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	93880	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	357084	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.604	164	214846	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.398	188	524938	20.000	ng/ul	0.00
79) Chrysene-d12	21.539	240	575823	20.000	ng/ul	0.00
88) Perylene-d12	24.104	264	679043	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	284	0.112	ng/uL	0.00
4) Pyridine-d5	3.693	84	1567	0.223	ng/ul	0.00
7) Phenol-d5	7.058	99	11	0.001	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	78	0.015	ng/ul	-0.02
11) 2-Chlorophenol-d4	7.422	132	7	0.001	ng/ul	0.00
15) 4-Methylphenol-d8	8.628	113	15	0.002	ng/ul	0.00
21) Nitrobenzene-d5	9.140	128	95	0.032	ng/ul	0.02
24) 2-Nitrophenol-d4	9.822	143	8	0.002	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	47	0.007	ng/ul	0.00
31) 4-Chloroaniline-d4	10.904	131	30	0.003	ng/ul	0.00
46) Dimethylphthalate-d6	14.016	166	7	0.000	ng/ul	0.00
49) Acenaphthylene-d8	14.310	160	104	0.005	ng/ul	0.00
54) 4-Nitrophenol-d4	14.851	143	32	0.010	ng/ul	0.00
60) Fluorene-d10	15.610	176	22	0.001	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.892	200	7	0.002	ng/ul	0.14
73) Anthracene-d10	17.510	188	2457	0.091	ng/ul	0.01
81) Pyrene-d10	19.733	212	32	0.001	ng/ul	-0.02
92) Benzo(a)pyrene-d12	23.939	264	314	0.008	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101723\
Data File : BP017700.D
Acq On : 17 Oct 2023 16:21
Operator : MA/JU
Sample : 04889-10
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Oct 18 00:43:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 12 16:14:36 2023
Response via : Initial Calibration

