

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101723\
 Data File : BP017702.D
 Acq On : 17 Oct 2023 17:29
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020783

Quant Time: Oct 18 00:43:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 16:14:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	166538	20.000	ng/u1	0.00
20) Naphthalene-d8	10.734	136	687036	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.598	164	490623	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1183797	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	1531811	20.000	ng/u1	0.00
88) Perylene-d12	24.104	264	1870437	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	34168	7.564	ng/uL	0.00
4) Pyridine-d5	3.693	84	232749	18.640	ng/u1	0.00
7) Phenol-d5	7.057	99	288700	18.620	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	176710	18.794	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	239151	19.927	ng/u1	0.00
15) 4-Methylphenol-d8	8.622	113	238991	19.298	ng/u1	0.00
21) Nitrobenzene-d5	9.110	128	117733	20.402	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	133102	20.510	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.351	165	265447	21.763	ng/u1	0.00
31) 4-Chloroaniline-d4	10.910	131	360609	20.627	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	856973	20.168	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	950032	20.762	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	134718	18.957	ng/u1	0.00
60) Fluorene-d10	15.604	176	764668	21.156	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	160678	19.203	ng/u1	0.00
73) Anthracene-d10	17.492	188	1291751	21.137	ng/u1	0.00
81) Pyrene-d10	19.751	212	1733664	18.804	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.933	264	2162016	20.590	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	33671	7.029	ng/uL	98
5) Pyridine	3.716	79	242231	18.889	ng/u1	97
6) Benzaldehyde	7.063	77	167558	21.336	ng/u1	97
8) Phenol	7.081	94	291352	18.934	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.340	93	234551	18.996	ng/u1	98
12) 2-Chlorophenol	7.452	128	245416	20.467	ng/u1	96
13) 2-Methylphenol	8.351	108	218622	19.084	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.416	45	176140	11.056	ng/u1	98
16) Acetophenone	8.757	105	383704	19.912	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.728	70	193482	19.839	ng/u1	98
18) 4-Methylphenol	8.687	108	248053	19.921	ng/u1	98
19) Hexachloroethane	8.963	117	107483	20.065	ng/u1	86
22) Nitrobenzene	9.151	77	305155	20.587	ng/u1	97
23) Isophorone	9.663	82	521678	19.519	ng/u1	98
25) 2-Nitrophenol	9.857	139	145459	21.069	ng/u1	99
26) 2,4-Dimethylphenol	9.898	107	280924	20.478	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.157	93	326531	19.995	ng/u1	98
29) 2,4-Dichlorophenol	10.381	162	257362	21.868	ng/u1	96
30) Naphthalene	10.787	128	831162	20.810	ng/u1	99
32) 4-Chloroaniline	10.934	127	349776	20.924	ng/u1	98
33) Hexachlorobutadiene	11.010	225	203751	22.390	ng/u1	99
34) Caprolactam	11.763	113	71002	19.724	ng/u1	92
35) 4-Chloro-3-methylphenol	12.045	107	256530	20.660	ng/u1	97
36) 2-Methylnaphthalene	12.404	142	588731	21.523	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101723\
 Data File : BP017702.D
 Acq On : 17 Oct 2023 17:29
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020783


Quant Time: Oct 18 00:43:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 12 16:14:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	606421	21.842	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.757	216	379307	21.191	ng/ul	98
40) Hexachlorocyclopentadiene	12.698	237	170126	21.156	ng/ul	99
41) 2,4,6-Trichlorophenol	13.022	196	231741	21.025	ng/ul	98
42) 2,4,5-Trichlorophenol	13.092	196	246075	21.147	ng/ul	96
43) 1,1'-Biphenyl	13.428	154	812978	20.218	ng/ul	99
44) 2-Chloronaphthalene	13.475	162	665089	20.675	ng/ul	98
45) 2-Nitroaniline	13.722	65	172980	19.452	ng/ul	97
47) Dimethylphthalate	14.063	163	859450	20.482	ng/ul	100
48) 2,6-Dinitrotoluene	14.216	165	164095	21.129	ng/ul	99
50) Acenaphthylene	14.328	152	1061272	20.503	ng/ul	99
51) 3-Nitroaniline	14.563	138	152625	19.486	ng/ul	89
52) Acenaphthene	14.669	153	717423	20.837	ng/ul	99
53) 2,4-Dinitrophenol	14.775	184	79650	16.609	ng/ul	93
55) 4-Nitrophenol	14.857	109	128077	19.236	ng/ul	95
56) Dibenzofuran	15.004	168	1046805	21.194	ng/ul	100
57) 2,4-Dinitrotoluene	15.010	165	257756	21.501	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.228	232	242321	22.512	ng/ul	97
59) Diethylphthalate	15.428	149	866404	20.584	ng/ul	99
61) Fluorene	15.663	166	865685	21.258	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.657	204	470187	21.715	ng/ul	99
63) 4-Nitroaniline	15.739	138	159028	20.431	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.769	198	170938	19.506	ng/ul	91
67) N-Nitrosodiphenylamine	15.886	169	724703	20.488	ng/ul	100
68) 4-Bromophenyl-phenylether	16.563	248	315709	22.117	ng/ul	97
69) Hexachlorobenzene	16.645	284	340232	20.027	ng/ul	96
70) Atrazine	16.851	200	297792	21.754	ng/ul	98
71) Pentachlorophenol	17.022	266	221107	21.491	ng/ul	96
72) Phenanthrene	17.433	178	1456899	20.919	ng/ul	99
74) Anthracene	17.533	178	1495181	21.224	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.381	216	389648	20.861	ng/ul	99
76) Pentachlorobenzene	14.898	250	428401	22.349	ng/ul	97
77) Carbazole	17.822	167	1347562	21.123	ng/ul	99
78) Di-n-butylphthalate	18.333	149	1511985	19.245	ng/ul	100
80) Fluoranthene	19.421	202	1928497	18.304	ng/ul	99
82) Pyrene	19.780	202	2082427	18.919	ng/ul	98
83) Butylbenzylphthalate	20.651	149	733358	17.126	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.463	252	773476	19.713	ng/ul	98
85) Benzo(a)anthracene	21.521	228	2419482	20.599	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.398	149	1049019	16.201	ng/ul	99
87) Chrysene	21.574	228	2249979	20.758	ng/ul	99
89) Di-n-octyl phthalate	22.380	149	1892367	15.495	ng/ul	100
90) Benzo(b)fluoranthene	23.304	252	2531229	19.873	ng/ul	99
91) Benzo(k)fluoranthene	23.357	252	2628117	20.740	ng/ul	99
93) Benzo(a)pyrene	23.986	252	2451684	20.671	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.827	276	3150387	21.659	ng/ul	98
95) Dibenzo(a,h)anthracene	26.856	278	2630185	21.575	ng/ul	97
96) Benzo(g,h,i)perylene	27.680	276	2545401	22.007	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 18 00:43:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 12 16:14:36 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.029 ng/uL
RT: 3.299 min Scan# 30
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion: 88 Resp: 33671
Ion Ratio Lower Upper
88 100
43 31.1 27.3 40.9
58 67.3 53.3 79.9



#5
Pyridine
Concen: 18.889 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion: 79 Resp: 242231
Ion Ratio Lower Upper
79 100
52 65.7 55.8 83.6
51 36.2 28.7 43.1









#6
Benzaldehyde
Concen: 21.336 ng/ul
RT: 7.063 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion: 77 Resp: 167558
Ion Ratio Lower Upper
77 100
105 93.6 74.9 112.3
106 85.4 72.2 108.2



#8
Phenol
Concen: 18.934 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. -0.012 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion: 94 Resp: 291352
Ion Ratio Lower Upper
94 100
65 32.0 24.2 36.4
66 42.0 31.8 47.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.996 ng/ul
RT: 7.340 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion: 93 Resp: 234551
Ion Ratio Lower Upper
93 100
63 72.7 59.4 89.2
95 32.3 25.3 37.9



#12
2-Chlorophenol
Concen: 20.467 ng/ul
RT: 7.452 min Scan# 742
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:128 Resp: 245416
Ion Ratio Lower Upper
128 100
64 42.3 36.9 55.3
130 32.9 25.8 38.8





#13

2-Methylphenol

Concen: 19.084 ng/ul

RT: 8.351 min Scan# 897

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :





SSTD020783

Tgt Ion:108 Resp: 218622

Ion Ratio Lower Upper

108 100

107 90.7 70.2 105.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.056 ng/ul

RT: 8.416 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion: 45 Resp: 176140


Ion Ratio Lower Upper

45 100

77 17.5 13.0 19.6

79 12.2 9.5 14.3









#16
Acetophenone
Concen: 19.912 ng/ul
RT: 8.757 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:105 Resp: 383704
Ion Ratio Lower Upper
105 100
77 84.2 64.3 96.5
51 28.7 22.6 33.8



#17
N-Nitroso-di-n-propylamine
Concen: 19.839 ng/ul
RT: 8.728 min Scan# 959
Delta R.T. -0.000 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion: 70 Resp: 193482
Ion Ratio Lower Upper
70 100
42 58.2 46.6 70.0
101 10.7 7.8 11.6
130 23.0 16.4 24.6









#18
4-Methylphenol
Concen: 19.921 ng/ul
RT: 8.687 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:108 Resp: 248053
Ion Ratio Lower Upper
108 100
107 113.8 89.4 134.2



#19
Hexachloroethane
Concen: 20.065 ng/ul
RT: 8.963 min Scan# 999
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:117 Resp: 107483
Ion Ratio Lower Upper
117 100
201 125.8 85.4 128.2
199 73.0 53.7 80.5









#22
Nitrobenzene
Concen: 20.587 ng/ul
RT: 9.151 min Scan# 1101
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion: 77 Resp: 305155
Ion Ratio Lower Upper
77 100
123 43.9 33.8 50.6
65 14.1 10.6 16.0



#23
Isophorone
Concen: 19.519 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion: 82 Resp: 521678
Ion Ratio Lower Upper
82 100
95 8.0 6.0 9.0
138 18.3 13.9 20.9









#25
2-Nitrophenol
Concen: 21.069 ng/ul
RT: 9.857 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:139 Resp: 145459
Ion Ratio Lower Upper
139 100
65 53.3 43.4 65.2
109 35.5 28.1 42.1



#26
2,4-Dimethylphenol
Concen: 20.478 ng/ul
RT: 9.898 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:107 Resp: 280924
Ion Ratio Lower Upper
107 100
121 51.0 37.6 56.4
122 87.4 67.0 100.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.995 ng/ul
 RT: 10.157 min Scan# 1202
 Delta R.T. -0.006 min
 Lab File: BP017702.D
 Acq: 17 Oct 2023 17:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020783


Tgt Ion: 93 Resp: 326531
 Ion Ratio Lower Upper
 93 100
 95 33.4 25.9 38.9
 123 11.6 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 21.868 ng/ul
 RT: 10.381 min Scan# 1240
 Delta R.T. -0.006 min
 Lab File: BP017702.D
 Acq: 17 Oct 2023 17:29

Tgt Ion: 162 Resp: 257362
 Ion Ratio Lower Upper
 162 100
 164 66.9 50.5 75.7
 98 33.4 27.8 41.8









#30
Naphthalene
Concen: 20.810 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:128 Resp: 831162
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 20.924 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:127 Resp: 349776
Ion Ratio Lower Upper
127 100
129 32.0 26.6 40.0





#33

Hexachlorobutadiene

Concen: 22.390 ng/ul

RT: 11.010 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:225 Resp: 203751

Ion Ratio Lower Upper

225 100

223 62.3 50.2 75.2

227 63.2 51.0 76.4



#34

Caprolactam

Concen: 19.724 ng/ul

RT: 11.763 min Scan# 1475

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:113 Resp: 71002


Ion Ratio Lower Upper

113 100

55 147.8 126.9 190.3

56 114.8 97.9 146.9









#35
4-Chloro-3-methylphenol
Concen: 20.660 ng/ul
RT: 12.045 min Scan# 1525
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:107 Resp: 256530
Ion Ratio Lower Upper
107 100
144 27.1 20.1 30.1
142 81.8 63.3 94.9



#36
2-Methylnaphthalene
Concen: 21.523 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:142 Resp: 588731
Ion Ratio Lower Upper
142 100
141 88.1 72.8 109.2





#37

1-Methylnaphthalene

Concen: 21.842 ng/ul

RT: 12.628 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:142 Resp: 606421

Ion Ratio Lower Upper

142 100

141 91.9 74.3 111.5

116 4.4 3.3 4.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.191 ng/ul

RT: 12.757 min Scan# 1644

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Tgt Ion:216 Resp: 379307




Ion Ratio Lower Upper


216 100

214 76.3 59.8 89.8

179 19.9 16.6 25.0

108 15.5 12.8 19.2





#40

Hexachlorocyclopentadiene

Concen: 21.156 ng/ul

RT: 12.698 min Scan# 1634

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:237 Resp: 170126

Ion Ratio Lower Upper

237 100

235 63.0 50.4 75.6

272 12.5 11.3 16.9



#41

2,4,6-Trichlorophenol

Concen: 21.025 ng/ul

RT: 13.022 min Scan# 1689

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:196 Resp: 231741


Ion Ratio Lower Upper

196 100

198 95.1 77.3 115.9

200 29.6 25.4 38.2





#42

2,4,5-Trichlorophenol

Concen: 21.147 ng/ul

RT: 13.092 min Scan# 1701

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:196 Resp: 246075

Ion Ratio Lower Upper

196 100

198 95.1 73.0 109.4

200 31.2 23.9 35.9



#43

1,1'-Biphenyl

Concen: 20.218 ng/ul

RT: 13.428 min Scan# 1758

Delta R.T. -0.000 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:154 Resp: 812978


Ion Ratio Lower Upper

154 100

153 40.3 32.5 48.7

76 11.5 10.4 15.6





#44

2-Chloronaphthalene

Concen: 20.675 ng/ul

RT: 13.475 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:162 Resp: 665089

Ion Ratio Lower Upper

162 100

164 32.2 26.5 39.7

127 38.1 31.4 47.0



#45

2-Nitroaniline

Concen: 19.452 ng/ul

RT: 13.722 min Scan# 1808

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion: 65 Resp: 172980


Ion Ratio Lower Upper

65 100

92 66.9 55.3 82.9

138 105.9 82.8 124.2









#47
Dimethylphthalate
Concen: 20.482 ng/ul
RT: 14.063 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:163 Resp: 859450
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.129 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:165 Resp: 164095
Ion Ratio Lower Upper
165 100
89 62.4 49.0 73.4
121 20.7 16.5 24.7





#50

Acenaphthylene

Concen: 20.503 ng/ul

RT: 14.328 min Scan# 1911

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:152 Resp: 1061272

Ion Ratio Lower Upper

152 100

151 20.4 15.8 23.8

153 13.3 10.5 15.7



#51

3-Nitroaniline

Concen: 19.486 ng/ul

RT: 14.563 min Scan# 1951

Delta R.T. -0.000 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:138 Resp: 152625


Ion Ratio Lower Upper

138 100

108 10.0 8.2 12.2

92 110.0 99.0 148.6





#52

Acenaphthene

Concen: 20.837 ng/ul

RT: 14.669 min Scan# 1969

Delta R.T. -0.000 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:153 Resp: 717423

Ion Ratio Lower Upper

153 100

152 47.9 38.9 58.3

154 89.2 71.6 107.4



#53

2,4-Dinitrophenol

Concen: 16.609 ng/ul

RT: 14.775 min Scan# 1987

Delta R.T. -0.000 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:184 Resp: 79650


Ion Ratio Lower Upper

184 100

63 66.7 61.4 92.0

154 66.4 54.8 82.2









#55
4-Nitrophenol
Concen: 19.236 ng/ul
RT: 14.857 min Scan# 20
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:109 Resp: 128077
Ion Ratio Lower Upper
109 100
139 109.0 90.5 135.7
65 112.3 95.7 143.5



#56
Dibenzofuran
Concen: 21.194 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:168 Resp: 1046805
Ion Ratio Lower Upper
168 100
139 38.0 30.5 45.7





#57

2,4-Dinitrotoluene

Concen: 21.501 ng/ul

RT: 15.010 min Scan# 20

Delta R.T. -0.000 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:165 Resp: 257756

Ion Ratio Lower Upper

165 100

63 62.5 52.7 79.1

182 3.2 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.512 ng/ul

RT: 15.228 min Scan# 2064

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Tgt Ion:232 Resp: 242321




Ion Ratio Lower Upper


232 100

131 40.2 33.6 50.4

130 2.1 2.0 3.0

166 28.0 24.2 36.2









#59
Diethylphthalate
Concen: 20.584 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:149 Resp: 866404
Ion Ratio Lower Upper
149 100
177 22.0 17.9 26.9
150 12.4 10.1 15.1



#61
Fluorene
Concen: 21.258 ng/ul
RT: 15.663 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:166 Resp: 865685
Ion Ratio Lower Upper
166 100
165 98.2 79.6 119.4
167 13.6 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 21.715 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:204 Resp: 470187
Ion Ratio Lower Upper
204 100
206 33.4 26.6 40.0
141 53.9 43.9 65.9



#63
4-Nitroaniline
Concen: 20.431 ng/ul
RT: 15.739 min Scan# 2151
Delta R.T. -0.000 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:138 Resp: 159028
Ion Ratio Lower Upper
138 100
92 53.5 43.0 64.6
108 104.8 82.2 123.4









#66
4,6-Dinitro-2-methylphenol
Concen: 19.506 ng/ul
RT: 15.769 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:198 Resp: 170938
Ion Ratio Lower Upper
198 100
182 4.8 3.4 5.0
77 23.7 23.0 34.6



#67
N-Nitrosodiphenylamine
Concen: 20.488 ng/ul
RT: 15.886 min Scan# 2176
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:169 Resp: 724703
Ion Ratio Lower Upper
169 100
168 66.5 53.3 79.9
167 37.1 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 22.117 ng/ul
RT: 16.563 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:248 Resp: 315709
Ion Ratio Lower Upper
248 100
250 94.8 75.9 113.9
141 65.2 57.2 85.8



#69
Hexachlorobenzene
Concen: 20.027 ng/ul
RT: 16.645 min Scan# 2305
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:284 Resp: 340232
Ion Ratio Lower Upper
284 100
142 28.5 25.8 38.6
249 31.9 24.9 37.3





#70

Atrazine

Concen: 21.754 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:200 Resp: 297792

Ion Ratio Lower Upper

200 100

173 24.1 19.9 29.9

215 48.5 37.8 56.6



#71

Pentachlorophenol

Concen: 21.491 ng/ul

RT: 17.022 min Scan# 2369

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:266 Resp: 221107


Ion Ratio Lower Upper

266 100

264 66.2 49.2 73.8

268 66.1 51.8 77.6





#72

Phenanthrene

Concen: 20.919 ng/ul

RT: 17.433 min Scan# 2441

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:178 Resp: 1456899

Ion Ratio Lower Upper

178 100

179 15.6 12.1 18.1

176 19.0 15.3 22.9



#74

Anthracene

Concen: 21.224 ng/ul

RT: 17.533 min Scan# 2456

Delta R.T. -0.000 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:178 Resp: 1495181


Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 18.5 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.861 ng/uL
RT: 13.381 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:216 Resp: 389648
Ion Ratio Lower Upper
216 100
214 77.0 62.9 94.3
218 48.9 38.6 58.0



#76
Pentachlorobenzene
Concen: 22.349 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:250 Resp: 428401
Ion Ratio Lower Upper
250 100
248 60.5 50.8 76.2
252 63.5 49.9 74.9





#77

Carbazole

Concen: 21.123 ng/ul

RT: 17.822 min Scan# 2505

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:167 Resp: 1347562

Ion Ratio Lower Upper

167 100

166 21.7 17.0 25.6

139 12.8 10.2 15.2



#78

Di-n-butylphthalate

Concen: 19.245 ng/ul

RT: 18.333 min Scan# 2592

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:149 Resp: 1511985


Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.0

104 5.7 4.6 6.8









#80
Fluoranthene
Concen: 18.304 ng/ul
RT: 19.421 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:202 Resp: 1928497
Ion Ratio Lower Upper
202 100
101 8.4 7.0 10.6
100 6.6 5.8 8.6



#82
Pyrene
Concen: 18.919 ng/ul
RT: 19.780 min Scan# 2838
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:202 Resp: 2082427
Ion Ratio Lower Upper
202 100
101 10.1 8.8 13.2
100 8.6 7.4 11.0





#83

Butylbenzylphthalate

Concen: 17.126 ng/ul

RT: 20.651 min Scan# 2988

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:149 Resp: 733358

Ion Ratio Lower Upper

149 100

91 71.5 59.6 89.4

206 22.0 17.2 25.8



#84

3,3'-Dichlorobenzidine

Concen: 19.713 ng/ul

RT: 21.463 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:252 Resp: 773476


Ion Ratio Lower Upper

252 100

254 63.7 51.8 77.6

126 8.8 8.1 12.1









#85
Benzo(a)anthracene
Concen: 20.599 ng/ul
RT: 21.521 min Scan# 3134
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:228 Resp: 2419482
Ion Ratio Lower Upper
228 100
229 19.4 15.4 23.2
226 26.7 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 16.201 ng/ul
RT: 21.398 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:149 Resp: 1049019
Ion Ratio Lower Upper
149 100
167 27.2 21.7 32.5
279 5.5 3.7 5.5





#87

Chrysene

Concen: 20.758 ng/ul

RT: 21.574 min Scan# 3145

Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:228 Resp: 2249979

Ion Ratio Lower Upper

228 100

226 29.9 24.2 36.4

229 19.6 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 15.495 ng/ul




RT: 22.380 min Scan# 3280


Delta R.T. -0.006 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Tgt Ion:149 Resp: 1892367









#90
Benzo(b)fluoranthene
Concen: 19.873 ng/ul
RT: 23.304 min Scan# 3437
Delta R.T. -0.006 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29




Instrument :
BNA_P
ClientSampleId :
SSTD020783


Tgt Ion:252 Resp: 2531229
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 7.2 6.9 10.3



#91
Benzo(k)fluoranthene
Concen: 20.740 ng/ul
RT: 23.357 min Scan# 3446
Delta R.T. -0.012 min
Lab File: BP017702.D
Acq: 17 Oct 2023 17:29

Tgt Ion:252 Resp: 2628117
Ion Ratio Lower Upper
252 100
253 21.2 17.4 26.0
125 7.5 6.3 9.5





#93

Benzo(a)pyrene

Concen: 20.671 ng/ul

RT: 23.986 min Scan# 31

Delta R.T. -0.012 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:252 Resp: 2451684

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 8.6 7.3 10.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.659 ng/ul

RT: 26.827 min Scan# 4036

Delta R.T. -0.012 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29




Tgt Ion:276 Resp: 3150387


Ion Ratio Lower Upper

276 100

138 16.1 14.6 21.8

277 24.3 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 21.575 ng/ul

RT: 26.856 min Scan# 4045

Delta R.T. -0.012 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Instrument :

BNA_P

ClientSampleId :

SSTD020783





Tgt Ion:278 Resp: 2630185

Ion Ratio Lower Upper

278 100

139 11.6 10.8 16.2

279 23.5 19.8 29.8



#96

Benzo(g,h,i)perylene

Concen: 22.007 ng/ul

RT: 27.680 min Scan# 4181

Delta R.T. -0.018 min

Lab File: BP017702.D

Acq: 17 Oct 2023 17:29

Tgt Ion:276 Resp: 2545401

Ion Ratio Lower Upper

276 100

138 15.4 14.1 21.1

277 23.6 18.7 28.1

