

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101823\
 Data File : BP017716.D
 Acq On : 18 Oct 2023 18:58
 Operator : MA/JU
 Sample : SST01035
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010646

Quant Time: Oct 19 00:55:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 00:53:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	134462	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	507131	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	328713	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.369	188	731000	20.000	ng/u1	0.00
79) Chrysene-d12	21.510	240	704308	20.000	ng/u1	0.00
88) Perylene-d12	24.063	264	758404	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	14873	4.125	ng/uL	0.00
4) Pyridine-d5	3.699	84	86400	8.705	ng/u1	0.00
7) Phenol-d5	7.052	99	97813	7.987	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	64720	8.694	ng/u1	0.00
11) 2-Chlorophenol-d4	7.411	132	86764	9.079	ng/u1	0.00
15) 4-Methylphenol-d8	8.611	113	75166	7.733	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	37605	8.894	ng/u1	0.00
24) 2-Nitrophenol-d4	9.810	143	40566	8.557	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	83615	9.310	ng/u1	0.00
31) 4-Chloroaniline-d4	10.893	131	97890	7.762	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	270062	9.581	ng/u1	0.00
49) Acenaphthylene-d8	14.281	160	290183	9.488	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	21447	4.699	ng/u1	0.00
60) Fluorene-d10	15.587	176	233122	9.685	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.734	200	32692	6.481	ng/u1	0.00
73) Anthracene-d10	17.475	188	356791	9.490	ng/u1	0.00
81) Pyrene-d10	19.728	212	418401	9.788	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.892	264	397053	9.304	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	14695	3.790	ng/uL	97
5) Pyridine	3.717	79	91228	8.924	ng/u1	98
6) Benzaldehyde	7.052	77	64544	10.460	ng/u1	98
8) Phenol	7.075	94	98987	8.184	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.328	93	85113	8.681	ng/u1	99
12) 2-Chlorophenol	7.440	128	89525	9.398	ng/u1	97
13) 2-Methylphenol	8.340	108	71593	7.912	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.405	45	68921	5.494	ng/u1	97
16) Acetophenone	8.746	105	126548	8.368	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.711	70	62636	8.260	ng/u1	99
18) 4-Methylphenol	8.675	108	76398	7.800	ng/u1	96
19) Hexachloroethane	8.946	117	40259	9.417	ng/u1	91
22) Nitrobenzene	9.140	77	97115	9.004	ng/u1	100
23) Isophorone	9.652	82	170464	8.815	ng/u1	100
25) 2-Nitrophenol	9.846	139	44977	8.911	ng/u1	99
26) 2,4-Dimethylphenol	9.887	107	89624	8.990	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.140	93	108459	9.087	ng/u1	98
29) 2,4-Dichlorophenol	10.369	162	77512	8.990	ng/u1	97
30) Naphthalene	10.769	128	287585	9.787	ng/u1	100
32) 4-Chloroaniline	10.922	127	96320	7.980	ng/u1	99
33) Hexachlorobutadiene	10.993	225	67077	9.949	ng/u1	95
34) Caprolactam	11.757	113	16712	6.548	ng/u1	94
35) 4-Chloro-3-methylphenol	12.028	107	73694	8.273	ng/u1	97
36) 2-Methylnaphthalene	12.387	142	189964	9.468	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP101823\
 Data File : BP017716.D
 Acq On : 18 Oct 2023 18:58
 Operator : MA/JU
 Sample : SSTD01035
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010646


Quant Time: Oct 19 00:55:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 00:53:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.610	142	195746	9.573	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.740	216	115368	9.581	ng/ul	99
40) Hexachlorocyclopentadiene	12.681	237	48185	8.543	ng/ul	98
41) 2,4,6-Trichlorophenol	13.004	196	65805	8.954	ng/ul	99
42) 2,4,5-Trichlorophenol	13.081	196	65313	8.443	ng/ul	97
43) 1,1'-Biphenyl	13.410	154	261143	9.667	ng/ul	98
44) 2-Chloronaphthalene	13.457	162	210163	9.718	ng/ul	99
45) 2-Nitroaniline	13.704	65	43603	7.583	ng/ul	98
47) Dimethylphthalate	14.046	163	262648	9.428	ng/ul	100
48) 2,6-Dinitrotoluene	14.198	165	44726	8.757	ng/ul	95
50) Acenaphthylene	14.310	152	331943	9.590	ng/ul	99
51) 3-Nitroaniline	14.546	138	36882	7.293	ng/ul	98
52) Acenaphthene	14.646	153	222859	9.647	ng/ul	98
53) 2,4-Dinitrophenol	14.763	184	15006	4.868	ng/ul	96
55) 4-Nitrophenol	14.845	109	31657	7.366	ng/ul	94
56) Dibenzofuran	14.987	168	321610	9.730	ng/ul	98
57) 2,4-Dinitrotoluene	14.993	165	66826	8.481	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.210	232	61296	8.679	ng/ul#	93
59) Diethylphthalate	15.410	149	254043	9.169	ng/ul	100
61) Fluorene	15.645	166	262238	9.659	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.634	204	140000	9.707	ng/ul	100
63) 4-Nitroaniline	15.722	138	33682	6.800	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.745	198	37259	6.980	ng/ul#	92
67) N-Nitrosodiphenylamine	15.863	169	212165	9.712	ng/ul	100
68) 4-Bromophenyl-phenylether	16.545	248	81404	9.299	ng/ul	97
69) Hexachlorobenzene	16.622	284	94771	9.152	ng/ul	99
70) Atrazine	16.828	200	73001	8.758	ng/ul	96
71) Pentachlorophenol	16.998	266	46173	7.422	ng/ul	99
72) Phenanthrene	17.416	178	411601	9.577	ng/ul	99
74) Anthracene	17.510	178	417242	9.613	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.363	216	114936	9.834	ng/uL	98
76) Pentachlorobenzene	14.875	250	115383	9.742	ng/uL	98
77) Carbazole	17.804	167	328842	8.497	ng/ul	99
78) Di-n-butylphthalate	18.316	149	392991	8.307	ng/ul	100
80) Fluoranthene	19.398	202	484434	9.894	ng/ul	100
82) Pyrene	19.757	202	518545	10.113	ng/ul	99
83) Butylbenzylphthalate	20.633	149	161828	8.327	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.439	252	126335	7.248	ng/ul	98
85) Benzo(a)anthracene	21.492	228	508987	9.415	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.375	149	220272	7.566	ng/ul	99
87) Chrysene	21.551	228	497279	9.917	ng/ul	100
89) Di-n-octyl phthalate	22.351	149	371932	7.635	ng/ul	100
90) Benzo(b)fluoranthene	23.269	252	484422	9.328	ng/ul	100
91) Benzo(k)fluoranthene	23.322	252	506341	9.839	ng/ul	99
93) Benzo(a)pyrene	23.945	252	461868	9.588	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.762	276	541436	9.197	ng/ul	100
95) Dibenzo(a,h)anthracene	26.792	278	452649	9.201	ng/ul	100
96) Benzo(g,h,i)perylene	27.610	276	441097	9.420	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 19 00:55:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 19 00:53:23 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.790 ng/uL
RT: 3.299 min Scan# 30
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:	88	Resp:	14695
Ion	Ratio	Lower	Upper
88	100		
43	31.4	24.4	36.6
58	69.0	53.0	79.6



#5
Pyridine
Concen: 8.924 ng/uL
RT: 3.717 min Scan# 107
Delta R.T. 0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:	79	Resp:	91228
Ion	Ratio	Lower	Upper
79	100		
52	69.2	57.2	85.8
51	34.9	28.6	42.8









#6
Benzaldehyde
Concen: 10.460 ng/ul
RT: 7.052 min Scan# 674
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion: 77 Resp: 64544
Ion Ratio Lower Upper
77 100
105 92.0 75.4 113.2
106 91.3 72.3 108.5



#8
Phenol
Concen: 8.184 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion: 94 Resp: 98987
Ion Ratio Lower Upper
94 100
65 29.4 25.0 37.4
66 40.3 33.0 49.6









#10
Bis(2-Chloroethyl)ether
Concen: 8.681 ng/ul
RT: 7.328 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion: 93 Resp: 85113
Ion Ratio Lower Upper
93 100
63 76.1 60.7 91.1
95 33.2 27.8 41.6



#12
2-Chlorophenol
Concen: 9.398 ng/ul
RT: 7.440 min Scan# 740
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion: 128 Resp: 89525
Ion Ratio Lower Upper
128 100
64 43.9 36.7 55.1
130 31.0 26.2 39.2





#13

2-Methylphenol

Concen: 7.912 ng/ul

RT: 8.340 min Scan# 893

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :





SSTD010646

Tgt Ion:108 Resp: 71593

Ion Ratio Lower Upper

108 100

107 89.8 73.4 110.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 5.494 ng/ul

RT: 8.405 min Scan# 904

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion: 45 Resp: 68921


Ion Ratio Lower Upper

45 100

77 20.3 14.2 21.4

79 13.0 10.3 15.5









#16
Acetophenone
Concen: 8.368 ng/ul
RT: 8.746 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:105 Resp: 126548
Ion Ratio Lower Upper
105 100
77 81.4 64.4 96.6
51 27.4 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 8.260 ng/ul
RT: 8.711 min Scan# 956
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion: 70 Resp: 62636
Ion Ratio Lower Upper
70 100
42 55.6 44.6 66.8
101 9.4 7.5 11.3
130 20.3 17.3 25.9









#18
4-Methylphenol
Concen: 7.800 ng/ul
RT: 8.675 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:108 Resp: 76398
Ion Ratio Lower Upper
108 100
107 113.9 87.5 131.3



#19
Hexachloroethane
Concen: 9.417 ng/ul
RT: 8.946 min Scan# 996
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:117 Resp: 40259
Ion Ratio Lower Upper
117 100
201 108.1 96.3 144.5
199 67.0 57.7 86.5









#22
Nitrobenzene
Concen: 9.004 ng/ul
RT: 9.140 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion: 77 Resp: 97115
Ion Ratio Lower Upper
77 100
123 41.8 33.5 50.3
65 13.2 10.8 16.2



#23
Isophorone
Concen: 8.815 ng/ul
RT: 9.652 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion: 82 Resp: 170464
Ion Ratio Lower Upper
82 100
95 7.5 6.2 9.2
138 18.1 14.4 21.6









#25
2-Nitrophenol
Concen: 8.911 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:139 Resp: 44977
Ion Ratio Lower Upper
139 100
65 51.2 40.4 60.6
109 32.9 25.4 38.2



#26
2,4-Dimethylphenol
Concen: 8.990 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:107 Resp: 89624
Ion Ratio Lower Upper
107 100
121 48.5 40.6 61.0
122 85.6 66.9 100.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.087 ng/ul
 RT: 10.140 min Scan# 1199
 Delta R.T. -0.006 min
 Lab File: BP017716.D
 Acq: 18 Oct 2023 18:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010646


Tgt Ion: 93 Resp: 108459
 Ion Ratio Lower Upper
 93 100
 95 32.5 24.7 37.1
 123 11.5 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 8.990 ng/ul
 RT: 10.369 min Scan# 1238
 Delta R.T. 0.000 min
 Lab File: BP017716.D
 Acq: 18 Oct 2023 18:58

Tgt Ion:162 Resp: 77512
 Ion Ratio Lower Upper
 162 100
 164 64.2 53.8 80.8
 98 31.2 25.5 38.3









#30
Naphthalene
Concen: 9.787 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:128 Resp: 287585
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.4 10.5 15.7



#32
4-Chloroaniline
Concen: 7.980 ng/ul
RT: 10.922 min Scan# 1332
Delta R.T. 0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:127 Resp: 96320
Ion Ratio Lower Upper
127 100
129 31.8 25.1 37.7





#33

Hexachlorobutadiene

Concen: 9.949 ng/ul

RT: 10.993 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:225 Resp: 67077

Ion Ratio Lower Upper

225 100

223 65.1 50.6 75.8

227 68.3 50.1 75.1



#34

Caprolactam

Concen: 6.548 ng/ul

RT: 11.757 min Scan# 1474

Delta R.T. -0.012 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:113 Resp: 16712


Ion Ratio Lower Upper

113 100

55 138.7 112.1 168.1

56 96.4 87.0 130.6









#35
4-Chloro-3-methylphenol
Concen: 8.273 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:107 Resp: 73694
Ion Ratio Lower Upper
107 100
144 26.4 22.9 34.3
142 82.3 68.1 102.1



#36
2-Methylnaphthalene
Concen: 9.468 ng/ul
RT: 12.387 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:142 Resp: 189964
Ion Ratio Lower Upper
142 100
141 86.8 71.5 107.3





#37

1-Methylnaphthalene

Concen: 9.573 ng/ul

RT: 12.610 min Scan# 1619

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:142 Resp: 195746

Ion Ratio Lower Upper

142 100

141 91.2 72.5 108.7

116 4.0 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.581 ng/ul

RT: 12.740 min Scan# 1641

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Tgt Ion:216 Resp: 115368




Ion Ratio Lower Upper


216 100

214 78.3 62.2 93.2

179 20.4 15.8 23.8

108 16.0 12.6 18.8





#40

Hexachlorocyclopentadiene

Concen: 8.543 ng/ul

RT: 12.681 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:237 Resp: 48185

Ion Ratio Lower Upper

237 100

235 63.8 50.3 75.5

272 12.5 11.1 16.7



#41

2,4,6-Trichlorophenol

Concen: 8.954 ng/ul

RT: 13.004 min Scan# 1686

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:196 Resp: 65805


Ion Ratio Lower Upper

196 100

198 94.8 74.7 112.1

200 28.9 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 8.443 ng/ul
RT: 13.081 min Scan# 1699
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:196 Resp: 65313
Ion Ratio Lower Upper
196 100
198 98.7 76.0 114.0
200 31.6 24.3 36.5



#43
1,1'-Biphenyl
Concen: 9.667 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:154 Resp: 261143
Ion Ratio Lower Upper
154 100
153 39.6 32.6 49.0
76 12.0 9.8 14.8









#44
2-Chloronaphthalene
Concen: 9.718 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:162 Resp: 210163
Ion Ratio Lower Upper
162 100
164 32.4 25.9 38.9
127 38.2 29.9 44.9



#45
2-Nitroaniline
Concen: 7.583 ng/ul
RT: 13.704 min Scan# 1805
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion: 65 Resp: 43603
Ion Ratio Lower Upper
65 100
92 68.9 56.2 84.4
138 105.5 85.9 128.9









#47
Dimethylphthalate
Concen: 9.428 ng/ul
RT: 14.046 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:163 Resp: 262648
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.5 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 8.757 ng/ul
RT: 14.198 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:165 Resp: 44726
Ion Ratio Lower Upper
165 100
89 63.2 48.0 72.0
121 23.8 16.6 24.8





#50

Acenaphthylene

Concen: 9.590 ng/ul

RT: 14.310 min Scan# 1908

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:152 Resp: 331943

Ion Ratio Lower Upper

152 100

151 19.4 15.9 23.9

153 13.4 10.8 16.2



#51

3-Nitroaniline

Concen: 7.293 ng/ul

RT: 14.546 min Scan# 1948

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:138 Resp: 36882


Ion Ratio Lower Upper

138 100

108 9.8 9.3 13.9

92 116.9 95.2 142.8





#52

Acenaphthene

Concen: 9.647 ng/ul

RT: 14.646 min Scan# 1965

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:153 Resp: 222859

Ion Ratio Lower Upper

153 100

152 48.8 37.6 56.4

154 89.0 70.0 105.0



#53

2,4-Dinitrophenol

Concen: 4.868 ng/ul

RT: 14.763 min Scan# 1985

Delta R.T. 0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:184 Resp: 15006


Ion Ratio Lower Upper

184 100

63 59.1 50.6 76.0

154 61.4 47.9 71.9









#55
4-Nitrophenol
Concen: 7.366 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:109 Resp: 31657
Ion Ratio Lower Upper
109 100
139 106.3 92.9 139.3
65 113.9 94.5 141.7



#56
Dibenzofuran
Concen: 9.730 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:168 Resp: 321610
Ion Ratio Lower Upper
168 100
139 37.2 28.6 43.0









#57
2,4-Dinitrotoluene
Concen: 8.481 ng/ul
RT: 14.993 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:	165	Resp:	66826
Ion Ratio	Lower	Upper	
165	100		
63	61.8	51.1	76.7
182	2.7	2.7	4.1#



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.679 ng/ul
RT: 15.210 min Scan# 2061
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:	232	Resp:	61296
Ion Ratio	Lower	Upper	
232	100		
131	45.4	32.9	49.3
130	3.3	1.8	2.8#
166	33.7	23.3	34.9









#59
Diethylphthalate
Concen: 9.169 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:149 Resp: 254043
Ion Ratio Lower Upper
149 100
177 22.8 18.0 27.0
150 12.8 10.2 15.4



#61
Fluorene
Concen: 9.659 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:166 Resp: 262238
Ion Ratio Lower Upper
166 100
165 98.9 77.4 116.0
167 13.8 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 9.707 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument : BNA_P
ClientSampleId : SSTD010646


Tgt Ion:204 Resp: 140000
Ion Ratio Lower Upper
204 100
206 33.0 26.2 39.4
141 53.2 42.9 64.3



#63
4-Nitroaniline
Concen: 6.800 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:138 Resp: 33682
Ion Ratio Lower Upper
138 100
92 62.7 43.0 64.6
108 107.8 82.7 124.1









#66
4,6-Dinitro-2-methylphenol
Concen: 6.980 ng/ul
RT: 15.745 min Scan# 2153
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:198 Resp: 37259
Ion Ratio Lower Upper
198 100
182 6.3 3.9 5.9#
77 29.4 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 9.712 ng/ul
RT: 15.863 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:169 Resp: 212165
Ion Ratio Lower Upper
169 100
168 67.0 53.7 80.5
167 37.0 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 9.299 ng/ul
RT: 16.545 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:248 Resp: 81404
Ion Ratio Lower Upper
248 100
250 97.1 76.8 115.2
141 70.3 52.4 78.6



#69
Hexachlorobenzene
Concen: 9.152 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:284 Resp: 94771
Ion Ratio Lower Upper
284 100
142 34.9 27.6 41.4
249 31.3 24.5 36.7









#70
Atrazine
Concen: 8.758 ng/ul
RT: 16.828 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:200 Resp: 73001
Ion Ratio Lower Upper
200 100
173 23.0 19.8 29.6
215 45.3 38.5 57.7



#71
Pentachlorophenol
Concen: 7.422 ng/ul
RT: 16.998 min Scan# 2365
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:266 Resp: 46173
Ion Ratio Lower Upper
266 100
264 63.4 49.4 74.2
268 63.3 50.3 75.5





#72

Phenanthrene

Concen: 9.577 ng/ul

RT: 17.416 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:178 Resp: 411601

Ion Ratio Lower Upper

178 100

179 14.7 12.2 18.2

176 19.4 15.3 22.9



#74

Anthracene

Concen: 9.613 ng/ul

RT: 17.510 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:178 Resp: 417242


Ion Ratio Lower Upper

178 100

179 15.1 12.0 18.0

176 18.4 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.834 ng/uL

RT: 13.363 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:216 Resp: 114936

Ion Ratio Lower Upper

216 100

214 77.6 60.7 91.1

218 47.2 38.2 57.4



#76

Pentachlorobenzene

Concen: 9.742 ng/uL

RT: 14.875 min Scan# 2004

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:250 Resp: 115383


Ion Ratio Lower Upper

250 100

248 61.2 50.2 75.4

252 62.8 49.6 74.4









#77
Carbazole
Concen: 8.497 ng/ul
RT: 17.804 min Scan# 2503
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD010646


Tgt Ion:167 Resp: 328842
Ion Ratio Lower Upper
167 100
166 20.5 16.7 25.1
139 12.1 9.8 14.6



#78
Di-n-butylphthalate
Concen: 8.307 ng/ul
RT: 18.316 min Scan# 2589
Delta R.T. -0.006 min
Lab File: BP017716.D
Acq: 18 Oct 2023 18:58

Tgt Ion:149 Resp: 392991
Ion Ratio Lower Upper
149 100
150 9.1 7.2 10.8
104 5.5 4.3 6.5





#80

Fluoranthene

Concen: 9.894 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:202 Resp: 484434

Ion Ratio Lower Upper

202 100

101 8.8 7.1 10.7

100 7.1 5.7 8.5



#82

Pyrene

Concen: 10.113 ng/ul

RT: 19.757 min Scan# 2834

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:202 Resp: 518545


Ion Ratio Lower Upper

202 100

101 11.1 8.3 12.5

100 9.3 7.4 11.0





#83

Butylbenzylphthalate

Concen: 8.327 ng/ul

RT: 20.633 min Scan# 2983

Delta R.T. 0.000 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:149 Resp: 161828

Ion Ratio Lower Upper

149 100

91 68.3 58.8 88.2

206 22.5 17.5 26.3



#84

3,3'-Dichlorobenzidine

Concen: 7.248 ng/ul

RT: 21.439 min Scan# 3120

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:252 Resp: 126335


Ion Ratio Lower Upper

252 100

254 63.1 51.8 77.8

126 10.7 7.8 11.6





#85

Benzo(a)anthracene

Concen: 9.415 ng/ul

RT: 21.492 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:228 Resp: 508987

Ion Ratio Lower Upper

228 100

229 19.8 15.6 23.4

226 27.6 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 7.566 ng/ul

RT: 21.375 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:149 Resp: 220272


Ion Ratio Lower Upper

149 100

167 27.7 21.6 32.4

279 4.3 3.6 5.4





#87

Chrysene

Concen: 9.917 ng/ul

RT: 21.551 min Scan# 3140

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:228 Resp: 497279

Ion Ratio Lower Upper

228 100

226 30.0 24.0 36.0

229 19.0 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 7.635 ng/ul




RT: 22.351 min Scan# 3275


Delta R.T. -0.012 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Tgt Ion:149 Resp: 371932





#90

Benzo(b)fluoranthene

Concen: 9.328 ng/ul

RT: 23.269 min Scan# 3432

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:252 Resp: 484422

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 8.6 7.1 10.7



#91

Benzo(k)fluoranthene

Concen: 9.839 ng/ul

RT: 23.322 min Scan# 3440

Delta R.T. -0.006 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:252 Resp: 506341


Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 8.1 7.0 10.6





#93

Benzo(a)pyrene

Concen: 9.588 ng/ul

RT: 23.945 min Scan# 3

Delta R.T. -0.012 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD010646





Tgt Ion:252 Resp: 461868

Ion Ratio Lower Upper

252 100

253 21.2 17.4 26.2

125 9.5 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.197 ng/ul

RT: 26.762 min Scan# 4025

Delta R.T. -0.018 min

Lab File: BP017716.D

Acq: 18 Oct 2023 18:58




Tgt Ion:276 Resp: 541436


Ion Ratio Lower Upper

276 100

138 19.5 15.6 23.4

277 24.6 19.4 29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 9.201 ng/ul
 RT: 26.792 min Scan# 4033
 Delta R.T. -0.018 min
 Lab File: BP017716.D
 Acq: 18 Oct 2023 18:58

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010646

Tgt Ion:278 Resp: 452649
 Ion Ratio Lower Upper
 278 100
 139 14.7 11.4 17.2
 279 23.5 18.8 28.2



#96
 Benzo(g,h,i)perylene
 Concen: 9.420 ng/ul
 RT: 27.610 min Scan# 4169
 Delta R.T. -0.018 min
 Lab File: BP017716.D
 Acq: 18 Oct 2023 18:58

Tgt Ion:276 Resp: 441097
 Ion Ratio Lower Upper
 276 100
 138 19.0 15.0 22.6
 277 23.6 18.5 27.7

