

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102023\
 Data File : BP017739.D
 Acq On : 20 Oct 2023 12:11
 Operator : MA/JU
 Sample : PB156396BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK396

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/23/2023
 Supervised By :mohammad ahmed 10/23/2023

Quant Time: Oct 20 22:42:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	108461	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	419368	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.587	164	245473	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.375	188	517925	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.521	240	424478	20.000	ng/u1	0.00
88) Perylene-d12	24.080	264	451025	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	21414	7.345	ng/uL	0.00
4) Pyridine-d5	3.687	84	271734	35.598	ng/u1	0.00
7) Phenol-d5	7.046	99	300551	33.300	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	207982	37.313	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	260989	35.454	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	241071	34.012	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	116215	34.199	ng/u1	0.00
24) 2-Nitrophenol-d4	9.810	143	127050	33.838	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	220839	29.984	ng/u1	0.00
31) 4-Chloroaniline-d4	10.893	131	305420	31.687	ng/u1	0.00
46) Dimethylphthalate-d6	14.004	166	717592	34.823	ng/u1	0.00
49) Acenaphthylene-d8	14.281	160	793300	34.520	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	76677m	26.114	ng/u1	0.00
60) Fluorene-d10	15.592	176	611924	34.353	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	90409	26.501	ng/u1	0.00
73) Anthracene-d10	17.481	188	903625	34.141	ng/u1	0.00
81) Pyrene-d10	19.739	212	1056014	39.681	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.910	264	912447	36.322	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102023\
Data File : BP017739.D
Acq On : 20 Oct 2023 12:11
Operator : MA/JU
Sample : PB156396BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK396

Quant Time: Oct 20 22:42:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 19 01:32:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/23/2023
Supervised By :mohammad ahmed 10/23/2023

