

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102121\
 Data File : BP007545.D
 Acq On : 21 Oct 2021 09:21
 Operator : CG/JU
 Sample : PB140022BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK022

Quant Time: Oct 21 09:58:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 20 10:49:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	198950	20.000	ng/ul	0.00
20) Naphthalene-d8	10.758	136	810290	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.587	164	525864	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.340	188	1128850	20.000	ng/ul	0.00
79) Chrysene-d12	21.428	240	1164294	20.000	ng/ul	0.00
88) Perylene-d12	23.880	264	1184005	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.376	96	27553	5.470	ng/uL	0.00
4) Pyridine-d5	3.793	84	413018	29.953	ng/ul	0.00
7) Phenol-d5	7.111	99	474032	30.510	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	271512	29.299	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	377180	30.859	ng/ul	0.00
15) 4-Methylphenol-d8	8.658	113	377469	31.541	ng/ul	0.00
21) Nitrobenzene-d5	9.111	128	174763	34.081	ng/ul	0.00
24) 2-Nitrophenol-d4	9.834	143	183189	38.246	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	365135	31.254	ng/ul	0.00
31) 4-Chloroaniline-d4	10.887	131	519388	32.876	ng/ul	0.00
46) Dimethylphthalate-d6	13.999	166	1149487	31.595	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	1386954	30.335	ng/ul	0.00
54) 4-Nitrophenol-d4	14.775	143	183709	31.470	ng/ul	0.00
60) Fluorene-d10	15.581	176	988644	32.472	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.693	200	117442	25.964	ng/ul	0.00
73) Anthracene-d10	17.440	188	1624229	33.880	ng/ul	0.00
81) Pyrene-d10	19.675	212	1908187	32.784	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.716	264	1876026	31.844	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102121\
Data File : BP007545.D
Acq On : 21 Oct 2021 09:21
Operator : CG/JU
Sample : PB140022BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK022

Quant Time: Oct 21 09:58:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 20 10:49:33 2021
Response via : Initial Calibration

