

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102121\
 Data File : BP007576.D
 Acq On : 22 Oct 2021 06:31
 Operator : CG/JU
 Sample : PB140021BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK021

Quant Time: Oct 22 07:20:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 20 10:49:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	236776	20.000	ng/u1	0.00
20) Naphthalene-d8	10.751	136	952004	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	623600	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.333	188	1326946	20.000	ng/u1	0.00
79) Chrysene-d12	21.421	240	1350821	20.000	ng/u1	0.00
88) Perylene-d12	23.868	264	1385333	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	31257	5.214	ng/uL	0.00
4) Pyridine-d5	3.787	84	466312	28.415	ng/u1	0.00
7) Phenol-d5	7.104	99	566276	30.624	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	324043	29.382	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	457137	31.426	ng/u1	0.00
15) 4-Methylphenol-d8	8.651	113	450473	31.627	ng/u1	0.00
21) Nitrobenzene-d5	9.104	128	217665	36.129	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	234644	41.696	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	444772	32.404	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	602193	32.443	ng/u1	0.00
46) Dimethylphthalate-d6	13.992	166	1355401	31.416	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	1643698	30.316	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	219005	31.636	ng/u1	-0.01
60) Fluorene-d10	15.575	176	1159504	32.115	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	139068	26.155	ng/u1	0.00
73) Anthracene-d10	17.433	188	1852959	32.881	ng/u1	0.00
81) Pyrene-d10	19.669	212	2170709	32.145	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.709	264	2147954	31.161	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102121\
Data File : BP007576.D
Acq On : 22 Oct 2021 06:31
Operator : CG/JU
Sample : PB140021BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK021

Quant Time: Oct 22 07:20:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 20 10:49:33 2021
Response via : Initial Calibration

