

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102121\
 Data File : BP007593.D
 Acq On : 22 Oct 2021 16:57
 Operator : CG/JU
 Sample : M4281-03
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BFY59

Quant Time: Oct 22 18:14:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 20 10:49:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	104028	20.000	ng/ul	0.00
20) Naphthalene-d8	10.752	136	421408	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.581	164	275902	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.334	188	606854	20.000	ng/ul	0.00
79) Chrysene-d12	21.422	240	659008	20.000	ng/ul	0.00
88) Perylene-d12	23.869	264	670281	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.376	96	19873	7.545	ng/uL	0.00
4) Pyridine-d5	3.793	84	126138	17.495	ng/ul	0.00
7) Phenol-d5	7.111	99	109077	13.426	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	230763	47.624	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	266634	41.720	ng/ul	0.00
15) 4-Methylphenol-d8	8.652	113	187032	29.888	ng/ul	0.00
21) Nitrobenzene-d5	9.105	128	163506	61.310	ng/ul	0.00
24) 2-Nitrophenol-d4	9.828	143	161581	64.865	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	321288	52.880	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	376002	45.763	ng/ul	0.00
46) Dimethylphthalate-d6	13.993	166	1072089	56.166	ng/ul	0.00
49) Acenaphthylene-d8	14.275	160	1250109	52.113	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	43762	14.288	ng/ul	0.00
60) Fluorene-d10	15.575	176	918679	57.511	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.687	200	40191	16.528	ng/ul	0.00
73) Anthracene-d10	17.434	188	1558130	60.458	ng/ul	0.00
81) Pyrene-d10	19.669	212	1899130	57.646	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.710	264	1939658	58.157	ng/ul	0.00


Target Compounds					Qvalue
8) Phenol	7.134	94	9530	1.147 ng/ul#	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102121\
Data File : BP007593.D
Acq On : 22 Oct 2021 16:57
Operator : CG/JU
Sample : M4281-03
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BFY59

Quant Time: Oct 22 18:14:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP101421.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 20 10:49:33 2021
Response via : Initial Calibration





#8
Phenol
Concen: 1.147 ng/ul
RT: 7.134 min Scan# 654
Delta R.T. -0.006 min
Lab File: BP007593.D
Acq: 22 Oct 2021 16:57

Instrument :
BNA_P
ClientSampleId :
BFY59

Tgt Ion: 94 Resp: 9530

Ion	Ratio	Lower	Upper
94	100		
65	34.1	21.9	32.9#
66	40.1	29.9	44.9

