

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102125\
 Data File : BP025993.D
 Acq On : 21 Oct 2025 19:41
 Operator : RC/JU
 Sample : Q3381-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 VNJ-222

Quant Time: Oct 22 10:49:35 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 15 23:48:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.696	152	139889	20.00	ng	0.00
21) Naphthalene-d8	10.472	136	504712	20.00	ng	0.01
39) Acenaphthene-d10	14.319	164	286992	20.00	ng	0.00
64) Phenanthrene-d10	17.131	188	569206	20.00	ng	0.01
76) Chrysene-d12	21.595	240	628005	20.00	ng	0.01
86) Perylene-d12	24.942	264	281073	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.337	112	1061065	124.36	ng	0.00
7) Phenol-d6	6.908	99	1196830	100.51	ng	0.00
23) Nitrobenzene-d5	8.849	82	916419	81.61	ng	0.00
42) 2,4,6-Tribromophenol	15.842	330	496471	128.50	ng	0.00
45) 2-Fluorobiphenyl	12.919	172	1928779	86.14	ng	-0.01
79) Terphenyl-d14	19.860	244	2976249	75.74	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102125\
Data File : BP025993.D
Acq On : 21 Oct 2025 19:41
Operator : RC/JU
Sample : Q3381-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
VNJ-222

Quant Time: Oct 22 10:49:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP101525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 15 23:48:51 2025
Response via : Initial Calibration

