

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102323\
 Data File : BP017779.D
 Acq On : 24 Oct 2023 16:00
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Oct 25 01:55:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	154957	20.000	ng/ul	0.00
20) Naphthalene-d8	10.716	136	612462	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.581	164	374516	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.369	188	752293	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.516	240	636824	20.000	ng/ul	0.00
88) Perylene-d12	24.069	264	635843	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	32521	7.807	ng/uL	0.00
4) Pyridine-d5	3.687	84	215464	19.757	ng/ul	0.00
7) Phenol-d5	7.040	99	271371	21.045	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	177207	22.252	ng/ul	0.00
11) 2-Chlorophenol-d4	7.405	132	222414	21.148	ng/ul	0.00
15) 4-Methylphenol-d8	8.605	113	219133	21.640	ng/ul	0.00
21) Nitrobenzene-d5	9.093	128	107823	21.726	ng/ul	0.00
24) 2-Nitrophenol-d4	9.805	143	120901	22.048	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	224025	20.827	ng/ul	0.00
31) 4-Chloroaniline-d4	10.887	131	282939	20.100	ng/ul	0.00
46) Dimethylphthalate-d6	13.999	166	629395	20.019	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	728405	20.775	ng/ul	0.00
54) 4-Nitrophenol-d4	14.828	143	88329	19.718	ng/ul	0.00
60) Fluorene-d10	15.587	176	537122	19.764	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.734	200	91471	18.459	ng/ul	0.00
73) Anthracene-d10	17.475	188	798973	20.783	ng/ul	0.00
81) Pyrene-d10	19.733	212	880551	22.055	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.898	264	724602	20.461	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.287	88	34823	7.882	ng/uL	98
5) Pyridine	3.705	79	225863	20.214	ng/ul	94
6) Benzaldehyde	7.046	77	168184	24.245	ng/ul	97
8) Phenol	7.069	94	267848	21.067	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.322	93	230679	21.747	ng/ul	98
12) 2-Chlorophenol	7.434	128	223043	21.115	ng/ul	97
13) 2-Methylphenol	8.334	108	206674	21.539	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.399	45	192813	14.768	ng/ul	98
16) Acetophenone	8.740	105	348392	21.884	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.711	70	181734	24.067	ng/ul	98
18) 4-Methylphenol	8.669	108	222603	21.846	ng/ul	96
19) Hexachloroethane	8.946	117	102553	21.129	ng/ul	86
22) Nitrobenzene	9.134	77	280789	22.723	ng/ul	100
23) Isophorone	9.646	82	511725	23.158	ng/ul	97
25) 2-Nitrophenol	9.840	139	127080	21.712	ng/ul	96
26) 2,4-Dimethylphenol	9.881	107	248134	21.600	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.140	93	305989	22.389	ng/ul	97
29) 2,4-Dichlorophenol	10.363	162	214701	20.874	ng/ul	98
30) Naphthalene	10.763	128	720089	20.430	ng/ul	99
32) 4-Chloroaniline	10.916	127	269780	20.014	ng/ul	100
33) Hexachlorobutadiene	10.987	225	146434	17.534	ng/ul	99
34) Caprolactam	11.746	113	55272	20.869	ng/ul	89
35) 4-Chloro-3-methylphenol	12.028	107	218407	22.555	ng/ul	94
36) 2-Methylnaphthalene	12.387	142	488144	20.503	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102323\
 Data File : BP017779.D
 Acq On : 24 Oct 2023 16:00
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Oct 25 01:55:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.604	142	494265	20.300	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.740	216	260214	18.777	ng/ul#	97
40) Hexachlorocyclopentadiene	12.675	237	116930	14.686	ng/ul	99
41) 2,4,6-Trichlorophenol	13.004	196	167170	20.100	ng/ul	94
42) 2,4,5-Trichlorophenol	13.075	196	177082	20.594	ng/ul	99
43) 1,1'-Biphenyl	13.404	154	638619	20.698	ng/ul	99
44) 2-Chloronaphthalene	13.451	162	511901	20.523	ng/ul	99
45) 2-Nitroaniline	13.704	65	141523	24.969	ng/ul	98
47) Dimethylphthalate	14.046	163	619341	19.988	ng/ul	100
48) 2,6-Dinitrotoluene	14.198	165	120802	21.181	ng/ul	96
50) Acenaphthylene	14.310	152	818208	20.833	ng/ul	99
51) 3-Nitroaniline	14.540	138	115961	21.378	ng/ul	97
52) Acenaphthene	14.646	153	539720	20.329	ng/ul	98
53) 2,4-Dinitrophenol	14.751	184	54683	17.080	ng/ul	90
55) 4-Nitrophenol	14.840	109	104260	21.986	ng/ul	96
56) Dibenzofuran	14.987	168	750643	20.138	ng/ul	96
57) 2,4-Dinitrotoluene	14.987	165	174150	20.690	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	15.210	232	140085	18.358	ng/ul#	90
59) Diethylphthalate	15.410	149	619634	20.712	ng/ul	99
61) Fluorene	15.645	166	604791	19.752	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.634	204	300527	18.466	ng/ul	94
63) 4-Nitroaniline	15.722	138	111598	22.810	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.745	198	98575	18.573	ng/ul	92
67) N-Nitrosodiphenylamine	15.863	169	498762	21.884	ng/ul	100
68) 4-Bromophenyl-phenylether	16.545	248	161250	18.319	ng/ul#	86
69) Hexachlorobenzene	16.622	284	175015	17.319	ng/ul#	83
70) Atrazine	16.834	200	165752	19.788	ng/ul	96
71) Pentachlorophenol	16.998	266	107202	17.645	ng/ul	99
72) Phenanthrene	17.416	178	919035	20.854	ng/ul	100
74) Anthracene	17.510	178	950448	21.294	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.357	216	257936	20.434	ng/ul	98
76) Pentachlorobenzene	14.875	250	231145	18.894	ng/ul	100
77) Carbazole	17.804	167	803171	21.291	ng/ul	98
78) Di-n-butylphthalate	18.316	149	1021778	23.331	ng/ul	100
80) Fluoranthene	19.398	202	1056709	23.036	ng/ul#	95
82) Pyrene	19.763	202	1089456	22.417	ng/ul	96
83) Butylbenzylphthalate	20.633	149	430135	25.733	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.445	252	281700	19.579	ng/ul	98
85) Benzo(a)anthracene	21.498	228	1012047	20.905	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.375	149	631960	27.083	ng/ul#	98
87) Chrysene	21.557	228	932647	20.372	ng/ul	99
89) Di-n-octyl phthalate	22.357	149	1062301	26.376	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	903042	20.868	ng/ul	99
91) Benzo(k)fluoranthene	23.274	252	903042	20.590	ng/ul	99
93) Benzo(a)pyrene	23.957	252	847616	20.864	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.780	276	948254	19.585	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.810	278	791374	19.589	ng/ul#	97
96) Benzo(g,h,i)perylene	27.627	276	762112	19.536	ng/ul#	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102323\
 Data File : BP017779.D
 Acq On : 24 Oct 2023 16:00
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792

Quant Time: Oct 25 01:55:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.882 ng/uL
RT: 3.287 min Scan# 34
Delta R.T. -0.012 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 88	Resp:	34823
Ion	Ratio	Lower Upper
88	100	
43	33.8	24.4 36.6
58	66.1	53.0 79.6



#5
Pyridine
Concen: 20.214 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion: 79	Resp:	225863
Ion	Ratio	Lower Upper
79	100	
52	64.9	57.2 85.8
51	33.8	28.6 42.8









#6
Benzaldehyde
Concen: 24.245 ng/ul
RT: 7.046 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 77 Resp: 168184
Ion Ratio Lower Upper
77 100
105 90.2 75.4 113.2
106 88.1 72.3 108.5



#8
Phenol
Concen: 21.067 ng/ul
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion: 94 Resp: 267848
Ion Ratio Lower Upper
94 100
65 31.4 25.0 37.4
66 40.8 33.0 49.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.747 ng/ul
RT: 7.322 min Scan# 720
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 93 Resp: 230679
Ion Ratio Lower Upper
93 100
63 74.7 60.7 91.1
95 32.1 27.8 41.6



#12
2-Chlorophenol
Concen: 21.115 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:128 Resp: 223043
Ion Ratio Lower Upper
128 100
64 49.5 36.7 55.1
130 32.8 26.2 39.2









#13
2-Methylphenol
Concen: 21.539 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:108 Resp: 206674
Ion Ratio Lower Upper
108 100
107 88.3 73.4 110.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 14.768 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion: 45 Resp: 192813
Ion Ratio Lower Upper
45 100
77 16.3 14.2 21.4
79 12.8 10.3 15.5









#16
Acetophenone
Concen: 21.884 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:105 Resp: 348392
Ion Ratio Lower Upper
105 100
77 82.8 64.4 96.6
51 27.6 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 24.067 ng/ul
RT: 8.711 min Scan# 956
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion: 70 Resp: 181734
Ion Ratio Lower Upper
70 100
42 56.3 44.6 66.8
101 9.3 7.5 11.3
130 19.6 17.3 25.9









#18
4-Methylphenol
Concen: 21.846 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:108 Resp: 222603
Ion Ratio Lower Upper
108 100
107 113.6 87.5 131.3



#19
Hexachloroethane
Concen: 21.129 ng/ul
RT: 8.946 min Scan# 996
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:117 Resp: 102553
Ion Ratio Lower Upper
117 100
201 102.8 96.3 144.5
199 63.8 57.7 86.5









#22
Nitrobenzene
Concen: 22.723 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion: 77 Resp: 280789
Ion Ratio Lower Upper
77 100
123 42.1 33.5 50.3
65 13.6 10.8 16.2



#23
Isophorone
Concen: 23.158 ng/ul
RT: 9.646 min Scan# 1115
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion: 82 Resp: 511725
Ion Ratio Lower Upper
82 100
95 7.1 6.2 9.2
138 16.7 14.4 21.6









#25
2-Nitrophenol
Concen: 21.712 ng/ul
RT: 9.840 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:139 Resp: 127080
Ion Ratio Lower Upper
139 100
65 53.5 40.4 60.6
109 33.4 25.4 38.2



#26
2,4-Dimethylphenol
Concen: 21.600 ng/ul
RT: 9.881 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:107 Resp: 248134
Ion Ratio Lower Upper
107 100
121 47.5 40.6 61.0
122 81.7 66.9 100.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.389 ng/ul
 RT: 10.140 min Scan# 1199
 Delta R.T. -0.006 min
 Lab File: BP017779.D
 Acq: 24 Oct 2023 16:00




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792


Tgt Ion: 93 Resp: 305989
 Ion Ratio Lower Upper
 93 100
 95 33.3 24.7 37.1
 123 10.8 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 20.874 ng/ul
 RT: 10.363 min Scan# 1237
 Delta R.T. -0.006 min
 Lab File: BP017779.D
 Acq: 24 Oct 2023 16:00

Tgt Ion: 162 Resp: 214701
 Ion Ratio Lower Upper
 162 100
 164 66.5 53.8 80.8
 98 34.6 25.5 38.3









#30
Naphthalene
Concen: 20.430 ng/ul
RT: 10.763 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:128 Resp: 720089
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 20.014 ng/ul
RT: 10.916 min Scan# 1331
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:127 Resp: 269780
Ion Ratio Lower Upper
127 100
129 31.6 25.1 37.7





#33

Hexachlorobutadiene

Concen: 17.534 ng/ul

RT: 10.987 min Scan# 1343

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:225 Resp: 146434

Ion Ratio Lower Upper

225 100

223 62.9 50.6 75.8

227 61.8 50.1 75.1



#34

Caprolactam

Concen: 20.869 ng/ul

RT: 11.746 min Scan# 1472

Delta R.T. -0.023 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:113 Resp: 55272


Ion Ratio Lower Upper

113 100

55 159.1 112.1 168.1

56 113.3 87.0 130.6









#35
4-Chloro-3-methylphenol
Concen: 22.555 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:107 Resp: 218407
Ion Ratio Lower Upper
107 100
144 25.0 22.9 34.3
142 79.8 68.1 102.1



#36
2-Methylnaphthalene
Concen: 20.503 ng/ul
RT: 12.387 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:142 Resp: 488144
Ion Ratio Lower Upper
142 100
141 88.7 71.5 107.3





#37

1-Methylnaphthalene

Concen: 20.300 ng/ul

RT: 12.604 min Scan# 1618

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:142 Resp: 494265

Ion Ratio Lower Upper

142 100

141 93.8 72.5 108.7

116 4.1 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.777 ng/ul

RT: 12.740 min Scan# 1641

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Tgt Ion:216 Resp: 260214




Ion Ratio Lower Upper


216 100

214 79.2 62.2 93.2

179 22.2 15.8 23.8

108 19.0 12.6 18.8#









#40
Hexachlorocyclopentadiene
Concen: 14.686 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:237 Resp: 116930
Ion Ratio Lower Upper
237 100
235 63.1 50.3 75.5
272 11.1 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 20.100 ng/ul
RT: 13.004 min Scan# 1686
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:196 Resp: 167170
Ion Ratio Lower Upper
196 100
198 99.0 74.7 112.1
200 33.4 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 20.594 ng/ul
RT: 13.075 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:196 Resp: 177082
Ion Ratio Lower Upper
196 100
198 95.5 76.0 114.0
200 32.3 24.3 36.5



#43
1,1'-Biphenyl
Concen: 20.698 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:154 Resp: 638619
Ion Ratio Lower Upper
154 100
153 40.7 32.6 49.0
76 13.8 9.8 14.8





#44

2-Chloronaphthalene

Concen: 20.523 ng/ul

RT: 13.451 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:162 Resp: 511901

Ion Ratio Lower Upper

162 100

164 33.3 25.9 38.9

127 38.2 29.9 44.9



#45

2-Nitroaniline

Concen: 24.969 ng/ul

RT: 13.704 min Scan# 1805

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion: 65 Resp: 141523


Ion Ratio Lower Upper

65 100

92 68.4 56.2 84.4

138 105.2 85.9 128.9









#47
Dimethylphthalate
Concen: 19.988 ng/ul
RT: 14.046 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:163 Resp: 619341
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.181 ng/ul
RT: 14.198 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:165 Resp: 120802
Ion Ratio Lower Upper
165 100
89 63.0 48.0 72.0
121 22.2 16.6 24.8









#50
Acenaphthylene
Concen: 20.833 ng/ul
RT: 14.310 min Scan# 1908
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:152 Resp: 818208
Ion Ratio Lower Upper
152 100
151 20.3 15.9 23.9
153 13.1 10.8 16.2



#51
3-Nitroaniline
Concen: 21.378 ng/ul
RT: 14.540 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:138 Resp: 115961
Ion Ratio Lower Upper
138 100
108 11.1 9.3 13.9
92 122.8 95.2 142.8





#52

Acenaphthene

Concen: 20.329 ng/ul

RT: 14.646 min Scan# 1965

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:153 Resp: 539720

Ion Ratio Lower Upper

153 100

152 49.5 37.6 56.4

154 88.9 70.0 105.0



#53

2,4-Dinitrophenol

Concen: 17.080 ng/ul

RT: 14.751 min Scan# 1983

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:184 Resp: 54683


Ion Ratio Lower Upper

184 100

63 74.1 50.6 76.0

154 64.6 47.9 71.9









#55
4-Nitrophenol
Concen: 21.986 ng/ul
RT: 14.840 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:109 Resp: 104260
Ion Ratio Lower Upper
109 100
139 115.7 92.9 139.3
65 126.8 94.5 141.7



#56
Dibenzofuran
Concen: 20.138 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:168 Resp: 750643
Ion Ratio Lower Upper
168 100
139 38.0 28.6 43.0









#57
2,4-Dinitrotoluene
Concen: 20.690 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:165 Resp: 174150
Ion Ratio Lower Upper
165 100
63 73.5 51.1 76.7
182 2.9 2.7 4.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.358 ng/ul
RT: 15.210 min Scan# 2061
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:232 Resp: 140085
Ion Ratio Lower Upper
232 100
131 48.3 32.9 49.3
130 2.8 1.8 2.8#
166 33.9 23.3 34.9









#59
Diethylphthalate
Concen: 20.712 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:149 Resp: 619634
Ion Ratio Lower Upper
149 100
177 22.0 18.0 27.0
150 12.6 10.2 15.4



#61
Fluorene
Concen: 19.752 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:166 Resp: 604791
Ion Ratio Lower Upper
166 100
165 96.8 77.4 116.0
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.466 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:204 Resp: 300527
Ion Ratio Lower Upper
204 100
206 33.1 26.2 39.4
141 60.1 42.9 64.3



#63
4-Nitroaniline
Concen: 22.810 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:138 Resp: 111598
Ion Ratio Lower Upper
138 100
92 54.6 43.0 64.6
108 104.5 82.7 124.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.573 ng/ul
 RT: 15.745 min Scan# 2152
 Delta R.T. -0.006 min
 Lab File: BP017779.D
 Acq: 24 Oct 2023 16:00




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792


Tgt Ion:198 Resp: 98575
 Ion Ratio Lower Upper
 198 100
 182 4.8 3.9 5.9
 77 29.4 19.8 29.8



#67
 N-Nitrosodiphenylamine
 Concen: 21.884 ng/ul
 RT: 15.863 min Scan# 2172
 Delta R.T. -0.006 min
 Lab File: BP017779.D
 Acq: 24 Oct 2023 16:00

Tgt Ion:169 Resp: 498762
 Ion Ratio Lower Upper
 169 100
 168 67.0 53.7 80.5
 167 37.1 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 18.319 ng/ul
RT: 16.545 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:248 Resp: 161250
Ion Ratio Lower Upper
248 100
250 102.6 76.8 115.2
141 84.6 52.4 78.6#



#69
Hexachlorobenzene
Concen: 17.319 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:284 Resp: 175015
Ion Ratio Lower Upper
284 100
142 48.2 27.6 41.4#
249 35.8 24.5 36.7









#70
Atrazine
Concen: 19.788 ng/ul
RT: 16.834 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00




Instrument :
BNA_P
ClientSampleId :
SSTD020792


Tgt Ion:200 Resp: 165752
Ion Ratio Lower Upper
200 100
173 25.6 19.8 29.6
215 44.8 38.5 57.7



#71
Pentachlorophenol
Concen: 17.645 ng/ul
RT: 16.998 min Scan# 2365
Delta R.T. -0.006 min
Lab File: BP017779.D
Acq: 24 Oct 2023 16:00

Tgt Ion:266 Resp: 107202
Ion Ratio Lower Upper
266 100
264 63.1 49.4 74.2
268 62.7 50.3 75.5





#72

Phenanthrene

Concen: 20.854 ng/ul

RT: 17.416 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:178 Resp: 919035

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 19.1 15.3 22.9



#74

Anthracene

Concen: 21.294 ng/ul

RT: 17.510 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:178 Resp: 950448


Ion Ratio Lower Upper

178 100

179 15.3 12.0 18.0

176 18.1 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.434 ng/uL

RT: 13.357 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:216 Resp: 257936

Ion Ratio Lower Upper

216 100

214 78.9 60.7 91.1

218 48.1 38.2 57.4



#76

Pentachlorobenzene

Concen: 18.894 ng/uL

RT: 14.875 min Scan# 2004

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:250 Resp: 231145


Ion Ratio Lower Upper

250 100

248 62.6 50.2 75.4

252 62.3 49.6 74.4









#77
 Carbazole
 Concen: 21.291 ng/ul
 RT: 17.804 min Scan# 2502
 Delta R.T. -0.006 min
 Lab File: BP017779.D
 Acq: 24 Oct 2023 16:00




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020792


Tgt Ion:167 Resp: 803171
 Ion Ratio Lower Upper
 167 100
 166 21.7 16.7 25.1
 139 13.0 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 23.331 ng/ul
 RT: 18.316 min Scan# 2589
 Delta R.T. -0.006 min
 Lab File: BP017779.D
 Acq: 24 Oct 2023 16:00

Tgt Ion:149 Resp: 1021778
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.2 10.8
 104 5.7 4.3 6.5





#80

Fluoranthene

Concen: 23.036 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:202 Resp: 1056709

Ion Ratio Lower Upper

202 100

101 11.1 7.1 10.7#

100 8.4 5.7 8.5



#82

Pyrene

Concen: 22.417 ng/ul

RT: 19.763 min Scan# 2835

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:202 Resp: 1089456


Ion Ratio Lower Upper

202 100

101 12.0 8.3 12.5

100 10.3 7.4 11.0





#83

Butylbenzylphthalate

Concen: 25.733 ng/ul

RT: 20.633 min Scan# 2983

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:149 Resp: 430135

Ion Ratio Lower Upper

149 100

91 75.1 58.8 88.2

206 20.0 17.5 26.3



#84

3,3'-Dichlorobenzidine

Concen: 19.579 ng/ul

RT: 21.445 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:252 Resp: 281700


Ion Ratio Lower Upper

252 100

254 63.8 51.8 77.8

126 10.9 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.905 ng/ul

RT: 21.498 min Scan# 3130

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:228 Resp: 1012047

Ion Ratio Lower Upper

228 100

229 19.4 15.6 23.4

226 27.0 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 27.083 ng/ul

RT: 21.375 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:149 Resp: 631960


Ion Ratio Lower Upper

149 100

167 26.0 21.6 32.4

279 3.2 3.6 5.4#





#87

Chrysene

Concen: 20.372 ng/ul

RT: 21.557 min Scan# 3139

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:228 Resp: 932647

Ion Ratio Lower Upper

228 100

226 30.3 24.0 36.0

229 19.8 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 26.376 ng/ul




RT: 22.357 min Scan# 3276


Delta R.T. -0.006 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Tgt Ion:149 Resp: 1062301





#90

Benzo(b)fluoranthene

Concen: 20.868 ng/ul

RT: 23.274 min Scan# 3432

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:252 Resp: 903042

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 10.4 7.1 10.7



#91

Benzo(k)fluoranthene

Concen: 20.590 ng/ul

RT: 23.274 min Scan# 3432

Delta R.T. -0.053 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:252 Resp: 903042


Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 10.4 7.0 10.6





#93

Benzo(a)pyrene

Concen: 20.864 ng/ul

RT: 23.957 min Scan# 3547

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:252 Resp: 847616

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 11.3 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.585 ng/ul

RT: 26.780 min Scan# 4028

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00




Tgt Ion:276 Resp: 948254


Ion Ratio Lower Upper

276 100

138 25.2 15.6 23.4#

277 24.7 19.4 29.2





#95

Dibenzo(a,h)anthracene

Concen: 19.589 ng/ul

RT: 26.810 min Scan# 4033

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Instrument :

BNA_P

ClientSampleId :

SSTD020792





Tgt Ion:278 Resp: 791374

Ion Ratio Lower Upper

278 100

139 17.3 11.4 17.2#

279 23.4 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 19.536 ng/ul

RT: 27.627 min Scan# 4172

Delta R.T. 0.000 min

Lab File: BP017779.D

Acq: 24 Oct 2023 16:00

Tgt Ion:276 Resp: 762112

Ion Ratio Lower Upper

276 100

138 24.0 15.0 22.6#

277 24.4 18.5 27.7

