

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102323\
 Data File : BP017793.D
 Acq On : 25 Oct 2023 00:31
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793

Quant Time: Oct 25 01:59:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	155197	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	650868	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.581	164	407173	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.369	188	857277	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.510	240	724466	20.000	ng/u1	0.00
88) Perylene-d12	24.063	264	738977	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	34148	8.185	ng/uL	0.00
4) Pyridine-d5	3.687	84	231673	21.211	ng/u1	0.00
7) Phenol-d5	7.040	99	280745	21.738	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	182461	22.877	ng/u1	0.00
11) 2-Chlorophenol-d4	7.399	132	229376	21.776	ng/u1	-0.01
15) 4-Methylphenol-d8	8.605	113	227291	22.411	ng/u1	0.00
21) Nitrobenzene-d5	9.087	128	112001	21.236	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.805	143	122916	21.093	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	230402	20.156	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.887	131	301697	20.168	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	673132	19.693	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	767756	20.141	ng/u1	0.00
54) 4-Nitrophenol-d4	14.822	143	92907	19.076	ng/u1	0.00
60) Fluorene-d10	15.581	176	575374	19.473	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.728	200	101638	17.999	ng/u1	0.00
73) Anthracene-d10	17.469	188	872705	19.921	ng/u1	0.00
81) Pyrene-d10	19.728	212	968825	21.330	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.898	264	818621	19.889	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.287	88	35792	8.088	ng/uL	98
5) Pyridine	3.705	79	239224	21.377	ng/u1	99
6) Benzaldehyde	7.046	77	167595	24.123	ng/u1	96
8) Phenol	7.069	94	283882	22.294	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.322	93	235487	22.166	ng/u1	97
12) 2-Chlorophenol	7.434	128	227889	21.541	ng/u1	97
13) 2-Methylphenol	8.334	108	211849	22.044	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.399	45	169510	12.963	ng/u1	98
16) Acetophenone	8.740	105	354573	22.238	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.705	70	185669	24.550	ng/u1	98
18) 4-Methylphenol	8.669	108	226970	22.240	ng/u1	96
19) Hexachloroethane	8.940	117	103893	21.372	ng/u1	83
22) Nitrobenzene	9.134	77	287302	21.878	ng/u1	98
23) Isophorone	9.646	82	527267	22.453	ng/u1	98
25) 2-Nitrophenol	9.834	139	128885	20.721	ng/u1	89
26) 2,4-Dimethylphenol	9.881	107	256524	21.013	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.134	93	313251	21.568	ng/u1	99
29) 2,4-Dichlorophenol	10.357	162	221531	20.267	ng/u1	96
30) Naphthalene	10.763	128	736601	19.666	ng/u1	99
32) 4-Chloroaniline	10.910	127	288725	20.155	ng/u1	98
33) Hexachlorobutadiene	10.987	225	150088	16.911	ng/u1	97
35) 4-Chloro-3-methylphenol	12.028	107	234296	22.768	ng/u1	94
36) 2-Methylnaphthalene	12.381	142	501421	19.818	ng/u1	99
37) 1-Methylnaphthalene	12.604	142	504167	19.485	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102323\
 Data File : BP017793.D
 Acq On : 25 Oct 2023 00:31
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793

Quant Time: Oct 25 01:59:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) 1,2,4,5-Tetrachloroben...	12.734	216	266404	17.682	ng/ul#	96
40) Hexachlorocyclopentadiene	12.675	237	147023	16.985	ng/ul	99
41) 2,4,6-Trichlorophenol	12.999	196	178519	19.743	ng/ul	95
42) 2,4,5-Trichlorophenol	13.075	196	188729	20.188	ng/ul	99
43) 1,1'-Biphenyl	13.404	154	668042	19.915	ng/ul	99
44) 2-Chloronaphthalene	13.451	162	530834	19.575	ng/ul	98
45) 2-Nitroaniline	13.698	65	154445	25.063	ng/ul	93
47) Dimethylphthalate	14.046	163	665017	19.741	ng/ul	99
48) 2,6-Dinitrotoluene	14.198	165	133447	21.522	ng/ul	94
50) Acenaphthylene	14.304	152	862128	20.191	ng/ul	99
51) 3-Nitroaniline	14.540	138	125095	21.212	ng/ul	99
52) Acenaphthene	14.646	153	569620	19.734	ng/ul	100
53) 2,4-Dinitrophenol	14.751	184	61703	17.727	ng/ul#	81
55) 4-Nitrophenol	14.840	109	121782	23.621	ng/ul	96
56) Dibenzofuran	14.987	168	780144	19.250	ng/ul	94
57) 2,4-Dinitrotoluene	14.987	165	190819	20.852	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.204	232	149763	18.053	ng/ul#	84
59) Diethylphthalate	15.410	149	678713	20.867	ng/ul	99
61) Fluorene	15.640	166	650173	19.531	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.634	204	322584	18.231	ng/ul	94
63) 4-Nitroaniline	15.716	138	120271	22.611	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.745	198	107356	17.750	ng/ul#	91
67) N-Nitrosodiphenylamine	15.863	169	539775	20.783	ng/ul	99
68) 4-Bromophenyl-phenylether	16.540	248	177847	17.730	ng/ul#	89
69) Hexachlorobenzene	16.622	284	189509	16.457	ng/ul#	85
70) Atrazine	16.828	200	181870	19.053	ng/ul	99
71) Pentachlorophenol	16.998	266	116685	16.854	ng/ul	96
72) Phenanthrene	17.416	178	997583	19.865	ng/ul	99
74) Anthracene	17.504	178	1027308	20.197	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.357	216	268926	18.696	ng/uL	99
76) Pentachlorobenzene	14.875	250	244470	17.536	ng/uL	99
77) Carbazole	17.798	167	891281	20.733	ng/ul	98
78) Di-n-butylphthalate	18.310	149	1122305	22.488	ng/ul	99
80) Fluoranthene	19.398	202	1140100	21.847	ng/ul	96
82) Pyrene	19.757	202	1190744	21.537	ng/ul	96
83) Butylbenzylphthalate	20.628	149	500017	26.295	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.439	252	300509	18.360	ng/ul#	97
85) Benzo(a)anthracene	21.492	228	1111168	20.176	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.375	149	712286	26.833	ng/ul	99
87) Chrysene	21.551	228	1034963	19.872	ng/ul	99
89) Di-n-octyl phthalate	22.351	149	1189849	25.419	ng/ul	100
90) Benzo(b)fluoranthene	23.322	252	1034840	20.576	ng/ul	99
91) Benzo(k)fluoranthene	23.322	252	1034840	20.302	ng/ul	98
93) Benzo(a)pyrene	23.951	252	955290	20.233	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.774	276	1075762	19.117	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.798	278	890924	18.975	ng/ul#	96
96) Benzo(g,h,i)perylene	27.621	276	862406	19.022	ng/ul#	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102323\
 Data File : BP017793.D
 Acq On : 25 Oct 2023 00:31
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793

Quant Time: Oct 25 01:59:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.088 ng/uL
RT: 3.287 min Scan# 34
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 88 Resp: 35792
Ion Ratio Lower Upper
88 100
43 30.4 24.4 36.6
58 64.3 53.0 79.6



#5
Pyridine
Concen: 21.377 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion: 79 Resp: 239224
Ion Ratio Lower Upper
79 100
52 70.2 57.2 85.8
51 35.0 28.6 42.8









#6
Benzaldehyde
Concen: 24.123 ng/ul
RT: 7.046 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 77 Resp: 167595
Ion Ratio Lower Upper
77 100
105 90.5 75.4 113.2
106 86.3 72.3 108.5



#8
Phenol
Concen: 22.294 ng/ul
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion: 94 Resp: 283882
Ion Ratio Lower Upper
94 100
65 30.5 25.0 37.4
66 40.8 33.0 49.6









#10
Bis(2-Chloroethyl)ether
Concen: 22.166 ng/ul
RT: 7.322 min Scan# 720
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 93 Resp: 235487
Ion Ratio Lower Upper
93 100
63 74.0 60.7 91.1
95 32.5 27.8 41.6



#12
2-Chlorophenol
Concen: 21.541 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:128 Resp: 227889
Ion Ratio Lower Upper
128 100
64 48.3 36.7 55.1
130 31.9 26.2 39.2









#13
2-Methylphenol
Concen: 22.044 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:108 Resp: 211849
Ion Ratio Lower Upper
108 100
107 90.1 73.4 110.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 12.963 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion: 45 Resp: 169510
Ion Ratio Lower Upper
45 100
77 16.4 14.2 21.4
79 12.3 10.3 15.5









#16
Acetophenone
Concen: 22.238 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:105 Resp: 354573
Ion Ratio Lower Upper
105 100
77 84.9 64.4 96.6
51 27.8 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 24.550 ng/ul
RT: 8.705 min Scan# 955
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion: 70 Resp: 185669
Ion Ratio Lower Upper
70 100
42 57.1 44.6 66.8
101 9.2 7.5 11.3
130 19.5 17.3 25.9





#18

4-Methylphenol

Concen: 22.240 ng/ul

RT: 8.669 min Scan# 949

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :





SSTD020793

Tgt Ion:108 Resp: 226970

Ion Ratio Lower Upper

108 100

107 113.9 87.5 131.3



#19

Hexachloroethane

Concen: 21.372 ng/ul

RT: 8.940 min Scan# 995

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:117 Resp: 103893


Ion Ratio Lower Upper

117 100

201 100.6 96.3 144.5

199 59.8 57.7 86.5









#22
Nitrobenzene
Concen: 21.878 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 77 Resp: 287302
Ion Ratio Lower Upper
77 100
123 40.7 33.5 50.3
65 14.0 10.8 16.2



#23
Isophorone
Concen: 22.453 ng/ul
RT: 9.646 min Scan# 1115
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion: 82 Resp: 527267
Ion Ratio Lower Upper
82 100
95 7.2 6.2 9.2
138 16.8 14.4 21.6









#25
2-Nitrophenol
Concen: 20.721 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:139 Resp: 128885
Ion Ratio Lower Upper
139 100
65 59.8 40.4 60.6
109 35.8 25.4 38.2



#26
2,4-Dimethylphenol
Concen: 21.013 ng/ul
RT: 9.881 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:107 Resp: 256524
Ion Ratio Lower Upper
107 100
121 47.7 40.6 61.0
122 82.3 66.9 100.3





#27

Bis(2-Chloroethoxy)methane

Concen: 21.568 ng/ul

RT: 10.134 min Scan# 1199

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion: 93 Resp: 313251

Ion Ratio Lower Upper

93 100

95 31.9 24.7 37.1

123 11.0 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 20.267 ng/ul

RT: 10.357 min Scan# 1236

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:162 Resp: 221531


Ion Ratio Lower Upper

162 100

164 65.7 53.8 80.8

98 36.4 25.5 38.3









#30
Naphthalene
Concen: 19.666 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:128 Resp: 736601
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 12.7 10.5 15.7



#32
4-Chloroaniline
Concen: 20.155 ng/ul
RT: 10.910 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:127 Resp: 288725
Ion Ratio Lower Upper
127 100
129 32.4 25.1 37.7









#33
Hexachlorobutadiene
Concen: 16.911 ng/ul
RT: 10.987 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:225 Resp: 150088
Ion Ratio Lower Upper
225 100
223 60.4 50.6 75.8
227 61.3 50.1 75.1



#35
4-Chloro-3-methylphenol
Concen: 22.768 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:107 Resp: 234296
Ion Ratio Lower Upper
107 100
144 25.6 22.9 34.3
142 79.5 68.1 102.1









#36
2-Methylnaphthalene
Concen: 19.818 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:142 Resp: 501421
Ion Ratio Lower Upper
142 100
141 88.3 71.5 107.3



#37
1-Methylnaphthalene
Concen: 19.485 ng/ul
RT: 12.604 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:142 Resp: 504167
Ion Ratio Lower Upper
142 100
141 93.1 72.5 108.7
116 4.5 3.4 5.2









#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.682 ng/ul
RT: 12.734 min Scan# 1640
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:216 Resp: 266404
Ion Ratio Lower Upper
216 100
214 78.7 62.2 93.2
179 23.6 15.8 23.8
108 20.0 12.6 18.8#



#40
Hexachlorocyclopentadiene
Concen: 16.985 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:237 Resp: 147023
Ion Ratio Lower Upper
237 100
235 63.3 50.3 75.5
272 12.0 11.1 16.7





#41

2,4,6-Trichlorophenol

Concen: 19.743 ng/ul

RT: 12.999 min Scan# 1685

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:196 Resp: 178519

Ion Ratio Lower Upper

196 100

198 99.1 74.7 112.1

200 31.3 23.7 35.5



#42

2,4,5-Trichlorophenol

Concen: 20.188 ng/ul

RT: 13.075 min Scan# 1698

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:196 Resp: 188729


Ion Ratio Lower Upper

196 100

198 94.3 76.0 114.0

200 32.1 24.3 36.5









#43
1,1'-Biphenyl
Concen: 19.915 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:154 Resp: 668042
Ion Ratio Lower Upper
154 100
153 40.1 32.6 49.0
76 13.2 9.8 14.8



#44
2-Chloronaphthalene
Concen: 19.575 ng/ul
RT: 13.451 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:162 Resp: 530834
Ion Ratio Lower Upper
162 100
164 32.2 25.9 38.9
127 38.9 29.9 44.9









#45
2-Nitroaniline
Concen: 25.063 ng/ul
RT: 13.698 min Scan# 1804
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion: 65 Resp: 154445
Ion Ratio Lower Upper
65 100
92 66.6 56.2 84.4
138 98.4 85.9 128.9



#47
Dimethylphthalate
Concen: 19.741 ng/ul
RT: 14.046 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion: 163 Resp: 665017
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.0 8.2 12.4









#48
2,6-Dinitrotoluene
Concen: 21.522 ng/ul
RT: 14.198 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:165 Resp: 133447
Ion Ratio Lower Upper
165 100
89 65.4 48.0 72.0
121 21.3 16.6 24.8



#50
Acenaphthylene
Concen: 20.191 ng/ul
RT: 14.304 min Scan# 1907
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:152 Resp: 862128
Ion Ratio Lower Upper
152 100
151 19.6 15.9 23.9
153 13.2 10.8 16.2









#51
3-Nitroaniline
Concen: 21.212 ng/ul
RT: 14.540 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:138 Resp: 125095
Ion Ratio Lower Upper
138 100
108 10.2 9.3 13.9
92 119.4 95.2 142.8



#52
Acenaphthene
Concen: 19.734 ng/ul
RT: 14.646 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:153 Resp: 569620
Ion Ratio Lower Upper
153 100
152 46.6 37.6 56.4
154 87.6 70.0 105.0









#53
2,4-Dinitrophenol
Concen: 17.727 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:184 Resp: 61703
Ion Ratio Lower Upper
184 100
63 78.4 50.6 76.0#
154 74.5 47.9 71.9#



#55
4-Nitrophenol
Concen: 23.621 ng/ul
RT: 14.840 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:109 Resp: 121782
Ion Ratio Lower Upper
109 100
139 111.5 92.9 139.3
65 121.2 94.5 141.7









#56
Dibenzenofuran
Concen: 19.250 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:168 Resp: 780144
Ion Ratio Lower Upper
168 100
139 39.2 28.6 43.0



#57
2,4-Dinitrotoluene
Concen: 20.852 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:165 Resp: 190819
Ion Ratio Lower Upper
165 100
63 67.9 51.1 76.7
182 2.7 2.7 4.1









#58
2,3,4,6-Tetrachlorophenol
Concen: 18.053 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. -0.012 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:232 Resp: 149763
Ion Ratio Lower Upper
232 100
131 53.7 32.9 49.3#
130 3.1 1.8 2.8#
166 35.4 23.3 34.9#



#59
Diethylphthalate
Concen: 20.867 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:149 Resp: 678713
Ion Ratio Lower Upper
149 100
177 22.1 18.0 27.0
150 12.0 10.2 15.4





#61

Fluorene

Concen: 19.531 ng/ul

RT: 15.640 min Scan# 2134

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:166 Resp: 650173

Ion Ratio Lower Upper

166 100

165 95.8 77.4 116.0

167 13.5 10.9 16.3



#62

4-Chlorophenyl-phenylether

Concen: 18.231 ng/ul

RT: 15.634 min Scan# 2133

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:204 Resp: 322584


Ion Ratio Lower Upper

204 100

206 32.4 26.2 39.4

141 60.2 42.9 64.3









#63
4-Nitroaniline
Concen: 22.611 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:138 Resp: 120271
Ion Ratio Lower Upper
138 100
92 58.5 43.0 64.6
108 107.7 82.7 124.1



#66
4,6-Dinitro-2-methylphenol
Concen: 17.750 ng/ul
RT: 15.745 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:198 Resp: 107356
Ion Ratio Lower Upper
198 100
182 5.0 3.9 5.9
77 30.0 19.8 29.8#





#67

N-Nitrosodiphenylamine

Concen: 20.783 ng/ul

RT: 15.863 min Scan# 2172

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:169 Resp: 539775

Ion Ratio Lower Upper

169 100

168 65.7 53.7 80.5

167 36.2 29.4 44.0



#68

4-Bromophenyl-phenylether

Concen: 17.730 ng/ul

RT: 16.540 min Scan# 2287

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:248 Resp: 177847


Ion Ratio Lower Upper

248 100

250 95.8 76.8 115.2

141 86.0 52.4 78.6#









#69
Hexachlorobenzene
Concen: 16.457 ng/ul
RT: 16.622 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:284 Resp: 189509
Ion Ratio Lower Upper
284 100
142 45.7 27.6 41.4#
249 36.1 24.5 36.7



#70
Atrazine
Concen: 19.053 ng/ul
RT: 16.828 min Scan# 2336
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:200 Resp: 181870
Ion Ratio Lower Upper
200 100
173 25.5 19.8 29.6
215 47.2 38.5 57.7





#71

Pentachlorophenol

Concen: 16.854 ng/ul

RT: 16.998 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:266 Resp: 116685

Ion Ratio Lower Upper

266 100

264 57.9 49.4 74.2

268 64.6 50.3 75.5



#72

Phenanthrene

Concen: 19.865 ng/ul

RT: 17.416 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:178 Resp: 997583


Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 19.9 15.3 22.9





#74

Anthracene

Concen: 20.197 ng/uI

RT: 17.504 min Scan# 2451

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:178 Resp: 1027308

Ion Ratio Lower Upper

178 100

179 15.1 12.0 18.0

176 18.7 14.8 22.2



#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.696 ng/uI

RT: 13.357 min Scan# 1746

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:216 Resp: 268926


Ion Ratio Lower Upper

216 100

214 76.7 60.7 91.1

218 47.6 38.2 57.4





#76

Pentachlorobenzene

Concen: 17.536 ng/uL

RT: 14.875 min Scan# 2004

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:250 Resp: 244470

Ion Ratio Lower Upper

250 100

248 63.5 50.2 75.4

252 63.1 49.6 74.4



#77

Carbazole

Concen: 20.733 ng/uL

RT: 17.798 min Scan# 2501

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:167 Resp: 891281


Ion Ratio Lower Upper

167 100

166 21.6 16.7 25.1

139 13.4 9.8 14.6





#78

Di-n-butylphthalate

Concen: 22.488 ng/ul

RT: 18.310 min Scan# 2588

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:149 Resp: 1122305

Ion Ratio Lower Upper

149 100

150 9.4 7.2 10.8

104 5.8 4.3 6.5



#80

Fluoranthene

Concen: 21.847 ng/ul

RT: 19.398 min Scan# 2773

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:202 Resp: 1140100


Ion Ratio Lower Upper

202 100

101 10.5 7.1 10.7

100 8.0 5.7 8.5





#82

Pyrene

Concen: 21.537 ng/ul

RT: 19.757 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:202 Resp: 1190744

Ion Ratio Lower Upper

202 100

101 12.5 8.3 12.5

100 10.3 7.4 11.0



#83

Butylbenzylphthalate

Concen: 26.295 ng/ul

RT: 20.628 min Scan# 2982

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:149 Resp: 500017


Ion Ratio Lower Upper

149 100

91 76.4 58.8 88.2

206 19.4 17.5 26.3





#84

3,3'-Dichlorobenzidine

Concen: 18.360 ng/ul

RT: 21.439 min Scan# 3120

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:252 Resp: 300509

Ion Ratio Lower Upper

252 100

254 63.1 51.8 77.8

126 12.1 7.8 11.6#



#85

Benzo(a)anthracene

Concen: 20.176 ng/ul

RT: 21.492 min Scan# 3129

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:228 Resp: 1111168


Ion Ratio Lower Upper

228 100

229 19.8 15.6 23.4

226 26.6 21.1 31.7









#86
Bis(2-ethylhexyl)phthalate
Concen: 26.833 ng/ul
RT: 21.375 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31




Instrument :
BNA_P
ClientSampleId :
SSTD020793


Tgt Ion:149 Resp: 712286
Ion Ratio Lower Upper
149 100
167 26.9 21.6 32.4
279 3.6 3.6 5.4



#87
Chrysene
Concen: 19.872 ng/ul
RT: 21.551 min Scan# 3139
Delta R.T. -0.006 min
Lab File: BP017793.D
Acq: 25 Oct 2023 00:31

Tgt Ion:228 Resp: 1034963
Ion Ratio Lower Upper
228 100
226 30.3 24.0 36.0
229 19.2 15.5 23.3





#89

Di-n-octyl phthalate

Concen: 25.419 ng/ul

RT: 22.351 min Scan# 31

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31





Instrument :

BNA_P

ClientSampleId :

SSTD020793

Tgt Ion:149 Resp: 1189849



#90

Benzo(b)fluoranthene

Concen: 20.576 ng/ul

RT: 23.322 min Scan# 3440

Delta R.T. 0.047 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:252 Resp: 1034840


Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.0

125 10.0 7.1 10.7





#91

Benzo(k)fluoranthene

Concen: 20.302 ng/ul

RT: 23.322 min Scan# 3440

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:252 Resp: 1034840

Ion Ratio Lower Upper

252 100

253 21.5 17.5 26.3

125 10.0 7.0 10.6



#93

Benzo(a)pyrene

Concen: 20.233 ng/ul

RT: 23.951 min Scan# 3547

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:252 Resp: 955290


Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 11.5 7.8 11.6





#94

Indeno(1,2,3-cd)pyrene

Concen: 19.117 ng/ul

RT: 26.774 min Scan# 4027

Delta R.T. -0.006 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31

Instrument :

BNA_P

ClientSampleId :

SSTD020793





Tgt Ion:276 Resp: 1075762

Ion Ratio Lower Upper

276 100

138 24.1 15.6 23.4#

277 25.0 19.4 29.2



#95

Dibenzo(a,h)anthracene

Concen: 18.975 ng/ul

RT: 26.798 min Scan# 4031

Delta R.T. -0.012 min

Lab File: BP017793.D

Acq: 25 Oct 2023 00:31




Tgt Ion:278 Resp: 890924


Ion Ratio Lower Upper

278 100

139 18.5 11.4 17.2#

279 23.6 18.8 28.2





#96
 Benzo(g,h,i)perylene
 Concen: 19.022 ng/ul
 RT: 27.621 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: BP017793.D
 Acq: 25 Oct 2023 00:31

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020793

Tgt Ion:	276	Resp:	862406
Ion Ratio	Lower	Upper	
276	100		
138	22.8	15.0	22.6#
277	23.5	18.5	27.7

