

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103025\
 Data File : BP026045.D
 Acq On : 30 Oct 2025 16:23
 Operator : RC/JU
 Sample : Q3447-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 P001-TW2-251023-01

Quant Time: Oct 30 16:51:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	363836	20.000	ng	0.00
21) Naphthalene-d8	10.454	136	1417735	20.000	ng	0.00
39) Acenaphthene-d10	14.331	164	855859	20.000	ng	0.01
64) Phenanthrene-d10	17.136	188	1766660	20.000	ng	0.01
76) Chrysene-d12	21.589	240	1797152	20.000	ng	0.02
86) Perylene-d12	24.912	264	877217	20.000	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.302	112	2715620	123.161	ng	0.00
7) Phenol-d6	6.854	99	3104808	110.631	ng	0.00
23) Nitrobenzene-d5	8.825	82	2058853	90.531	ng	0.00
42) 2,4,6-Tribromophenol	15.848	330	1370195	148.086	ng	0.02
45) 2-Fluorobiphenyl	12.942	172	5046038	84.719	ng	0.01
79) Terphenyl-d14	19.883	244	8565276	96.831	ng	0.02
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.536	149	15458	2.799	ng	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103025\
Data File : BP026045.D
Acq On : 30 Oct 2025 16:23
Operator : RC/JU
Sample : Q3447-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
P001-TW2-251023-01

Quant Time: Oct 30 16:51:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 30 11:51:34 2025
Response via : Initial Calibration

