

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103123\
 Data File : BP017834.D
 Acq On : 31 Oct 2023 16:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020798

Quant Time: Nov 01 01:40:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	136533	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	530690	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.575	164	315130	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.369	188	644264	20.000	ng/u1	0.00
79) Chrysene-d12	21.516	240	550972	20.000	ng/u1	0.00
88) Perylene-d12	24.074	264	556969	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	33163	9.036	ng/uL	0.02
4) Pyridine-d5	3.705	84	195866	20.384	ng/u1	0.01
7) Phenol-d5	7.040	99	242505	21.344	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	157480	22.444	ng/u1	0.00
11) 2-Chlorophenol-d4	7.399	132	201131	21.705	ng/u1	-0.01
15) 4-Methylphenol-d8	8.604	113	189674	21.259	ng/u1	0.00
21) Nitrobenzene-d5	9.087	128	92803	21.581	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.804	143	101717	21.408	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	191052	20.498	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.887	131	244482	20.044	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	518957	19.617	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	599892	20.334	ng/u1	0.00
54) 4-Nitrophenol-d4	14.822	143	66271	17.581	ng/u1	0.00
60) Fluorene-d10	15.581	176	438438	19.173	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.728	200	72285	17.033	ng/u1	0.00
73) Anthracene-d10	17.475	188	648855	19.708	ng/u1	0.00
81) Pyrene-d10	19.733	212	733699	21.240	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	625678	20.169	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	34947	8.977	ng/uL	99
5) Pyridine	3.722	79	205340	20.857	ng/u1	94
6) Benzaldehyde	7.046	77	155279	25.406	ng/u1	97
8) Phenol	7.069	94	243501	21.736	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.322	93	201495	21.559	ng/u1	98
12) 2-Chlorophenol	7.434	128	202175	21.723	ng/u1	97
13) 2-Methylphenol	8.334	108	180014	21.292	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.393	45	145012	12.606	ng/u1	98
16) Acetophenone	8.740	105	300374	21.414	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.704	70	138412	20.803	ng/u1	97
18) 4-Methylphenol	8.669	108	191414	21.320	ng/u1	99
19) Hexachloroethane	8.940	117	95828	22.408	ng/u1	84
22) Nitrobenzene	9.134	77	248663	23.224	ng/u1	99
23) Isophorone	9.640	82	416263	21.740	ng/u1	97
25) 2-Nitrophenol	9.834	139	106770	21.052	ng/u1	90
26) 2,4-Dimethylphenol	9.875	107	212593	21.358	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.134	93	256215	21.636	ng/u1	99
29) 2,4-Dichlorophenol	10.357	162	181882	20.408	ng/u1	96
30) Naphthalene	10.763	128	623803	20.426	ng/u1	99
32) 4-Chloroaniline	10.910	127	235041	20.123	ng/u1	98
33) Hexachlorobutadiene	10.981	225	133152	18.400	ng/u1	97
34) Caprolactam	11.740	113	41858	18.240	ng/u1	98
35) 4-Chloro-3-methylphenol	12.022	107	183909	21.918	ng/u1	94
36) 2-Methylnaphthalene	12.381	142	411498	19.947	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103123\
 Data File : BP017834.D
 Acq On : 31 Oct 2023 16:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020798

Quant Time: Nov 01 01:40:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.604	142	414071	19.627	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.734	216	221121	18.963	ng/ul#	96
40) Hexachlorocyclopentadiene	12.675	237	109570	16.355	ng/ul	97
41) 2,4,6-Trichlorophenol	12.998	196	139717	19.965	ng/ul	92
42) 2,4,5-Trichlorophenol	13.069	196	144410	19.959	ng/ul	99
43) 1,1'-Biphenyl	13.404	154	530994	20.453	ng/ul	99
44) 2-Chloronaphthalene	13.451	162	422878	20.149	ng/ul	97
45) 2-Nitroaniline	13.698	65	116896	24.511	ng/ul	94
47) Dimethylphthalate	14.045	163	515243	19.762	ng/ul	99
48) 2,6-Dinitrotoluene	14.198	165	100743	20.993	ng/ul	98
50) Acenaphthylene	14.304	152	671585	20.322	ng/ul	99
51) 3-Nitroaniline	14.539	138	90559	19.841	ng/ul	95
52) Acenaphthene	14.645	153	443465	19.851	ng/ul	100
53) 2,4-Dinitrophenol	14.751	184	35774	13.280	ng/ul#	79
55) 4-Nitrophenol	14.839	109	82596	20.700	ng/ul	95
56) Dibenzofuran	14.981	168	607997	19.384	ng/ul	93
57) 2,4-Dinitrotoluene	14.986	165	143768	20.299	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.204	232	117884	18.360	ng/ul#	86
59) Diethylphthalate	15.404	149	515256	20.469	ng/ul	97
61) Fluorene	15.639	166	494422	19.190	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.633	204	250829	18.317	ng/ul	92
63) 4-Nitroaniline	15.722	138	89087	21.641	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.745	198	77024	16.946	ng/ul#	87
67) N-Nitrosodiphenylamine	15.863	169	409619	20.986	ng/ul	99
68) 4-Bromophenyl-phenylether	16.545	248	141996	18.837	ng/ul#	87
69) Hexachlorobenzene	16.622	284	154049	17.801	ng/ul#	87
70) Atrazine	16.828	200	138275	19.275	ng/ul	99
71) Pentachlorophenol	16.998	266	85842	16.499	ng/ul	98
72) Phenanthrene	17.416	178	756807	20.053	ng/ul	100
74) Anthracene	17.510	178	772095	20.199	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.351	216	221028	20.446	ng/ul	99
76) Pentachlorobenzene	14.875	250	200130	19.102	ng/ul	98
77) Carbazole	17.804	167	660989	20.460	ng/ul	99
78) Di-n-butylphthalate	18.316	149	836405	22.300	ng/ul	99
80) Fluoranthene	19.398	202	859911	21.667	ng/ul	96
82) Pyrene	19.763	202	903167	21.480	ng/ul	98
83) Butylbenzylphthalate	20.633	149	358706	24.803	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.445	252	226322	18.181	ng/ul#	99
85) Benzo(a)anthracene	21.498	228	846871	20.219	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.380	149	506570	25.092	ng/ul	99
87) Chrysene	21.557	228	791112	19.973	ng/ul	99
89) Di-n-octyl phthalate	22.363	149	852288	24.158	ng/ul	100
90) Benzo(b)fluoranthene	23.280	252	773342	20.402	ng/ul	99
91) Benzo(k)fluoranthene	23.333	252	780408	20.314	ng/ul	99
93) Benzo(a)pyrene	23.962	252	726819	20.424	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.792	276	822008	19.381	ng/ul	97
95) Dibenzo(a,h)anthracene	26.815	278	679232	19.194	ng/ul	98
96) Benzo(g,h,i)perylene	27.639	276	658693	19.276	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP103123\
 Data File : BP017834.D
 Acq On : 31 Oct 2023 16:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020798

Quant Time: Nov 01 01:40:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.977 ng/uL
 RT: 3.311 min Scan# 31
 Delta R.T. 0.012 min
 Lab File: BP017834.D
 Acq: 31 Oct 2023 16:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020798


Tgt Ion: 88 Resp: 34947
 Ion Ratio Lower Upper
 88 100
 43 31.5 24.4 36.6
 58 67.1 53.0 79.6



#5
 Pyridine
 Concen: 20.857 ng/uL
 RT: 3.722 min Scan# 108
 Delta R.T. 0.012 min
 Lab File: BP017834.D
 Acq: 31 Oct 2023 16:42

Tgt Ion: 79 Resp: 205340
 Ion Ratio Lower Upper
 79 100
 52 78.9 57.2 85.8
 51 35.4 28.6 42.8









#6
Benzaldehyde
Concen: 25.406 ng/ul
RT: 7.046 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion: 77 Resp: 155279
Ion Ratio Lower Upper
77 100
105 91.0 75.4 113.2
106 88.4 72.3 108.5



#8
Phenol
Concen: 21.736 ng/ul
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion: 94 Resp: 243501
Ion Ratio Lower Upper
94 100
65 33.2 25.0 37.4
66 44.2 33.0 49.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.559 ng/ul
RT: 7.322 min Scan# 719
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion: 93 Resp: 201495
Ion Ratio Lower Upper
93 100
63 76.1 60.7 91.1
95 31.4 27.8 41.6



#12
2-Chlorophenol
Concen: 21.723 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:128 Resp: 202175
Ion Ratio Lower Upper
128 100
64 48.3 36.7 55.1
130 33.3 26.2 39.2





#13

2-Methylphenol

Concen: 21.292 ng/ul

RT: 8.334 min Scan# 893

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :





SSTD020798

Tgt Ion:108 Resp: 180014

Ion Ratio Lower Upper

108 100

107 89.6 73.4 110.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.606 ng/ul

RT: 8.393 min Scan# 902

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion: 45 Resp: 145012


Ion Ratio Lower Upper

45 100

77 18.7 14.2 21.4

79 13.4 10.3 15.5









#16
Acetophenone
Concen: 21.414 ng/ul
RT: 8.740 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:105 Resp: 300374
Ion Ratio Lower Upper
105 100
77 84.0 64.4 96.6
51 27.0 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.803 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion: 70 Resp: 138412
Ion Ratio Lower Upper
70 100
42 57.5 44.6 66.8
101 9.3 7.5 11.3
130 19.6 17.3 25.9









#18
4-Methylphenol
Concen: 21.320 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:108 Resp: 191414
Ion Ratio Lower Upper
108 100
107 110.0 87.5 131.3



#19
Hexachloroethane
Concen: 22.408 ng/ul
RT: 8.940 min Scan# 995
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:117 Resp: 95828
Ion Ratio Lower Upper
117 100
201 98.0 96.3 144.5
199 64.1 57.7 86.5









#22
Nitrobenzene
Concen: 23.224 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion: 77 Resp: 248663
Ion Ratio Lower Upper
77 100
123 41.3 33.5 50.3
65 13.7 10.8 16.2



#23
Isophorone
Concen: 21.740 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion: 82 Resp: 416263
Ion Ratio Lower Upper
82 100
95 7.5 6.2 9.2
138 16.4 14.4 21.6









#25
2-Nitrophenol
Concen: 21.052 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:139 Resp: 106770
Ion Ratio Lower Upper
139 100
65 57.5 40.4 60.6
109 37.5 25.4 38.2



#26
2,4-Dimethylphenol
Concen: 21.358 ng/ul
RT: 9.875 min Scan# 1154
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:107 Resp: 212593
Ion Ratio Lower Upper
107 100
121 47.8 40.6 61.0
122 79.8 66.9 100.3





#27

Bis(2-Chloroethoxy)methane

Concen: 21.636 ng/ul

RT: 10.134 min Scan# 1199

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion: 93 Resp: 256215

Ion Ratio Lower Upper

93 100

95 31.6 24.7 37.1

123 10.9 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 20.408 ng/ul

RT: 10.357 min Scan# 1236

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:162 Resp: 181882


Ion Ratio Lower Upper

162 100

164 66.5 53.8 80.8

98 37.6 25.5 38.3









#30
Naphthalene
Concen: 20.426 ng/ul
RT: 10.763 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:128 Resp: 623803
Ion Ratio Lower Upper
128 100
129 10.4 8.7 13.1
127 12.6 10.5 15.7



#32
4-Chloroaniline
Concen: 20.123 ng/ul
RT: 10.910 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:127 Resp: 235041
Ion Ratio Lower Upper
127 100
129 32.3 25.1 37.7





#33

Hexachlorobutadiene

Concen: 18.400 ng/ul

RT: 10.981 min Scan# 11

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:225 Resp: 133152

Ion Ratio Lower Upper

225 100

223 65.0 50.6 75.8

227 64.9 50.1 75.1



#34

Caprolactam

Concen: 18.240 ng/ul

RT: 11.740 min Scan# 1471

Delta R.T. -0.030 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:113 Resp: 41858


Ion Ratio Lower Upper

113 100

55 141.2 112.1 168.1

56 113.4 87.0 130.6









#35
4-Chloro-3-methylphenol
Concen: 21.918 ng/ul
RT: 12.022 min Scan# 1519
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:107 Resp: 183909
Ion Ratio Lower Upper
107 100
144 24.8 22.9 34.3
142 79.4 68.1 102.1



#36
2-Methylnaphthalene
Concen: 19.947 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:142 Resp: 411498
Ion Ratio Lower Upper
142 100
141 88.8 71.5 107.3





#37

1-Methylnaphthalene

Concen: 19.627 ng/ul

RT: 12.604 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:142 Resp: 414071

Ion Ratio Lower Upper

142 100

141 93.7 72.5 108.7

116 4.4 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.963 ng/ul

RT: 12.734 min Scan# 1640

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Tgt Ion:216 Resp: 221121




Ion Ratio Lower Upper


216 100

214 79.0 62.2 93.2

179 23.6 15.8 23.8

108 19.3 12.6 18.8#









#40
Hexachlorocyclopentadiene
Concen: 16.355 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:237 Resp: 109570
Ion Ratio Lower Upper
237 100
235 64.6 50.3 75.5
272 11.8 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 19.965 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:196 Resp: 139717
Ion Ratio Lower Upper
196 100
198 101.9 74.7 112.1
200 31.9 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 19.959 ng/ul
RT: 13.069 min Scan# 1699
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:196 Resp: 144410
Ion Ratio Lower Upper
196 100
198 94.8 76.0 114.0
200 31.3 24.3 36.5



#43
1,1'-Biphenyl
Concen: 20.453 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:154 Resp: 530994
Ion Ratio Lower Upper
154 100
153 40.3 32.6 49.0
76 12.7 9.8 14.8









#44
2-Chloronaphthalene
Concen: 20.149 ng/ul
RT: 13.451 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:162 Resp: 422878
Ion Ratio Lower Upper
162 100
164 32.7 25.9 38.9
127 40.8 29.9 44.9



#45
2-Nitroaniline
Concen: 24.511 ng/ul
RT: 13.698 min Scan# 1804
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion: 65 Resp: 116896
Ion Ratio Lower Upper
65 100
92 67.6 56.2 84.4
138 99.6 85.9 128.9





#47

Dimethylphthalate

Concen: 19.762 ng/ul

RT: 14.045 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:163 Resp: 515243

Ion Ratio Lower Upper

163 100

194 3.9 3.4 5.0

164 9.8 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 20.993 ng/ul

RT: 14.198 min Scan# 1889

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:165 Resp: 100743


Ion Ratio Lower Upper

165 100

89 61.0 48.0 72.0

121 22.3 16.6 24.8









#50
Acenaphthylene
Concen: 20.322 ng/ul
RT: 14.304 min Scan# 1907
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:152 Resp: 671585
Ion Ratio Lower Upper
152 100
151 20.1 15.9 23.9
153 13.0 10.8 16.2



#51
3-Nitroaniline
Concen: 19.841 ng/ul
RT: 14.539 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:138 Resp: 90559
Ion Ratio Lower Upper
138 100
108 11.3 9.3 13.9
92 124.7 95.2 142.8





#52

Acenaphthene

Concen: 19.851 ng/ul

RT: 14.645 min Scan# 1965

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:153 Resp: 443465

Ion Ratio Lower Upper

153 100

152 46.9 37.6 56.4

154 88.1 70.0 105.0



#53

2,4-Dinitrophenol

Concen: 13.280 ng/ul

RT: 14.751 min Scan# 1983

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:184 Resp: 35774


Ion Ratio Lower Upper

184 100

63 83.3 50.6 76.0#

154 72.1 47.9 71.9#









#55
4-Nitrophenol
Concen: 20.700 ng/ul
RT: 14.839 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:109 Resp: 82596
Ion Ratio Lower Upper
109 100
139 107.4 92.9 139.3
65 120.6 94.5 141.7



#56
Dibenzofuran
Concen: 19.384 ng/ul
RT: 14.981 min Scan# 2022
Delta R.T. -0.012 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:168 Resp: 607997
Ion Ratio Lower Upper
168 100
139 40.1 28.6 43.0





#57

2,4-Dinitrotoluene

Concen: 20.299 ng/ul

RT: 14.986 min Scan# 20

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:165 Resp: 143768

Ion Ratio Lower Upper

165 100

63 71.2 51.1 76.7

182 3.0 2.7 4.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.360 ng/ul

RT: 15.204 min Scan# 2060

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Tgt Ion:232 Resp: 117884




Ion Ratio Lower Upper


232 100

131 52.9 32.9 49.3#

130 2.9 1.8 2.8#

166 33.5 23.3 34.9









#59
Diethylphthalate
Concen: 20.469 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:149 Resp: 515256
Ion Ratio Lower Upper
149 100
177 20.9 18.0 27.0
150 12.3 10.2 15.4



#61
Fluorene
Concen: 19.190 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:166 Resp: 494422
Ion Ratio Lower Upper
166 100
165 97.6 77.4 116.0
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.317 ng/ul
RT: 15.633 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:204 Resp: 250829
Ion Ratio Lower Upper
204 100
206 32.6 26.2 39.4
141 63.0 42.9 64.3



#63
4-Nitroaniline
Concen: 21.641 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:138 Resp: 89087
Ion Ratio Lower Upper
138 100
92 56.6 43.0 64.6
108 116.7 82.7 124.1









#66
4,6-Dinitro-2-methylphenol
Concen: 16.946 ng/ul
RT: 15.745 min Scan# 2153
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:198 Resp: 77024
Ion Ratio Lower Upper
198 100
182 4.7 3.9 5.9
77 32.3 19.8 29.8#



#67
N-Nitrosodiphenylamine
Concen: 20.986 ng/ul
RT: 15.863 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:169 Resp: 409619
Ion Ratio Lower Upper
169 100
168 66.7 53.7 80.5
167 36.2 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 18.837 ng/ul
RT: 16.545 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:248 Resp: 141996
Ion Ratio Lower Upper
248 100
250 99.7 76.8 115.2
141 85.6 52.4 78.6#



#69
Hexachlorobenzene
Concen: 17.801 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:284 Resp: 154049
Ion Ratio Lower Upper
284 100
142 44.8 27.6 41.4#
249 34.5 24.5 36.7









#70
Atrazine
Concen: 19.275 ng/ul
RT: 16.828 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020798


Tgt Ion:200 Resp: 138275
Ion Ratio Lower Upper
200 100
173 25.4 19.8 29.6
215 47.1 38.5 57.7



#71
Pentachlorophenol
Concen: 16.499 ng/ul
RT: 16.998 min Scan# 2365
Delta R.T. -0.006 min
Lab File: BP017834.D
Acq: 31 Oct 2023 16:42

Tgt Ion:266 Resp: 85842
Ion Ratio Lower Upper
266 100
264 64.2 49.4 74.2
268 63.7 50.3 75.5





#72

Phenanthrene

Concen: 20.053 ng/ul

RT: 17.416 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:178 Resp: 756807

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.9 15.3 22.9



#74

Anthracene

Concen: 20.199 ng/ul

RT: 17.510 min Scan# 2452

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:178 Resp: 772095


Ion Ratio Lower Upper

178 100

179 14.9 12.0 18.0

176 18.5 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.446 ng/uL

RT: 13.351 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:216 Resp: 221028

Ion Ratio Lower Upper

216 100

214 78.0 60.7 91.1

218 47.8 38.2 57.4



#76

Pentachlorobenzene

Concen: 19.102 ng/uL

RT: 14.875 min Scan# 2004

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:250 Resp: 200130


Ion Ratio Lower Upper

250 100

248 63.2 50.2 75.4

252 64.5 49.6 74.4





#77

Carbazole

Concen: 20.460 ng/ul

RT: 17.804 min Scan# 2502

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:167 Resp: 660989

Ion Ratio Lower Upper

167 100

166 20.9 16.7 25.1

139 13.2 9.8 14.6



#78

Di-n-butylphthalate

Concen: 22.300 ng/ul

RT: 18.316 min Scan# 2589

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:149 Resp: 836405


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 5.8 4.3 6.5





#80

Fluoranthene

Concen: 21.667 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:202 Resp: 859911

Ion Ratio Lower Upper

202 100

101 10.7 7.1 10.7

100 8.3 5.7 8.5



#82

Pyrene

Concen: 21.480 ng/ul

RT: 19.763 min Scan# 2835

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:202 Resp: 903167


Ion Ratio Lower Upper

202 100

101 11.2 8.3 12.5

100 9.9 7.4 11.0





#83

Butylbenzylphthalate

Concen: 24.803 ng/ul

RT: 20.633 min Scan# 2984

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:149 Resp: 358706

Ion Ratio Lower Upper

149 100

91 75.9 58.8 88.2

206 20.3 17.5 26.3



#84

3,3'-Dichlorobenzidine

Concen: 18.181 ng/ul

RT: 21.445 min Scan# 3121

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:252 Resp: 226322


Ion Ratio Lower Upper

252 100

254 64.6 51.8 77.8

126 12.2 7.8 11.6#





#85

Benzo(a)anthracene

Concen: 20.219 ng/ul

RT: 21.498 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:228 Resp: 846871

Ion Ratio Lower Upper

228 100

229 19.9 15.6 23.4

226 27.8 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 25.092 ng/ul

RT: 21.380 min Scan# 3110

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:149 Resp: 506570


Ion Ratio Lower Upper

149 100

167 26.7 21.6 32.4

279 3.7 3.6 5.4





#87

Chrysene

Concen: 19.973 ng/ul

RT: 21.557 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:228 Resp: 791112

Ion Ratio Lower Upper

228 100

226 30.8 24.0 36.0

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 24.158 ng/ul




RT: 22.363 min Scan# 3277


Delta R.T. -0.000 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Tgt Ion:149 Resp: 852288





#90

Benzo(b)fluoranthene

Concen: 20.402 ng/ul

RT: 23.280 min Scan# 3434

Delta R.T. 0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:252 Resp: 773342

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 9.8 7.1 10.7



#91

Benzo(k)fluoranthene

Concen: 20.314 ng/ul

RT: 23.333 min Scan# 3442

Delta R.T. 0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:252 Resp: 780408


Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 9.3 7.0 10.6





#93

Benzo(a)pyrene

Concen: 20.424 ng/ul

RT: 23.962 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:252 Resp: 726819

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 10.7 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.381 ng/ul

RT: 26.792 min Scan# 4030

Delta R.T. 0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42




Tgt Ion:276 Resp: 822008


Ion Ratio Lower Upper

276 100

138 21.8 15.6 23.4

277 23.7 19.4 29.2





#95

Dibenzo(a,h)anthracene

Concen: 19.194 ng/ul

RT: 26.815 min Scan# 4036

Delta R.T. 0.006 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020798





Tgt Ion:278 Resp: 679232

Ion Ratio Lower Upper

278 100

139 16.2 11.4 17.2

279 23.9 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 19.276 ng/ul

RT: 27.639 min Scan# 4174

Delta R.T. 0.012 min

Lab File: BP017834.D

Acq: 31 Oct 2023 16:42

Tgt Ion:276 Resp: 658693

Ion Ratio Lower Upper

276 100

138 22.2 15.0 22.6

277 23.5 18.5 27.7

