

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110123\
 Data File : BP017882.D
 Acq On : 02 Nov 2023 12:43
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020654

Quant Time: Nov 02 23:35:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	178304	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.704	136	676108	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.575	164	390914	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.369	188	789020	20.000	ng/u1	0.00
79) Chrysene-d12	21.516	240	668865	20.000	ng/u1	0.00
88) Perylene-d12	24.074	264	693954	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	35628	7.433	ng/uL	0.01
4) Pyridine-d5	3.699	84	223982	17.849	ng/u1	0.00
7) Phenol-d5	7.040	99	305020	20.557	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	184672	20.153	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.399	132	254047	20.993	ng/u1	-0.01
15) 4-Methylphenol-d8	8.599	113	238692	20.485	ng/u1	-0.01
21) Nitrobenzene-d5	9.087	128	119178	21.753	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.799	143	135553	22.393	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.328	165	241763	20.360	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.881	131	309556	19.920	ng/u1	-0.01
46) Dimethylphthalate-d6	13.992	166	657823	20.046	ng/u1	-0.01
49) Acenaphthylene-d8	14.275	160	756206	20.663	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	88020	18.824	ng/u1	0.00
60) Fluorene-d10	15.581	176	541168	19.077	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.733	200	98795	19.009	ng/u1	0.00
73) Anthracene-d10	17.469	188	818266	20.294	ng/u1	0.00
81) Pyrene-d10	19.727	212	909508	21.689	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	762568	19.729	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	36862	7.251	ng/uL#	93
5) Pyridine	3.722	79	230350	17.916	ng/u1	98
6) Benzaldehyde	7.040	77	194498	24.368	ng/u1	99
8) Phenol	7.069	94	299243	20.455	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.322	93	243105	19.918	ng/u1	96
12) 2-Chlorophenol	7.434	128	252875	20.805	ng/u1	99
13) 2-Methylphenol	8.328	108	227593	20.614	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.393	45	185635	12.357	ng/u1	96
16) Acetophenone	8.734	105	374831	20.462	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.704	70	175314	20.177	ng/u1	99
18) 4-Methylphenol	8.663	108	239548	20.431	ng/u1	92
19) Hexachloroethane	8.934	117	118333	21.188	ng/u1	88
22) Nitrobenzene	9.128	77	302286	22.160	ng/u1	97
23) Isophorone	9.640	82	532619	21.834	ng/u1	99
25) 2-Nitrophenol	9.834	139	137880	21.339	ng/u1#	90
26) 2,4-Dimethylphenol	9.875	107	267262	21.075	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.128	93	307264	20.366	ng/u1	96
29) 2,4-Dichlorophenol	10.357	162	234272	20.632	ng/u1	97
30) Naphthalene	10.757	128	778293	20.003	ng/u1	99
32) 4-Chloroaniline	10.910	127	296392	19.918	ng/u1	99
33) Hexachlorobutadiene	10.981	225	173330	18.801	ng/u1	97
34) Caprolactam	11.740	113	60284	20.619	ng/u1	94
35) 4-Chloro-3-methylphenol	12.022	107	237206	22.190	ng/u1	92
36) 2-Methylnaphthalene	12.381	142	512596	19.503	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110123\
 Data File : BP017882.D
 Acq On : 02 Nov 2023 12:43
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 37 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020654


Quant Time: Nov 02 23:35:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 19 01:32:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	517503	19.253	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.728	216	288353	19.934	ng/ul	97
40) Hexachlorocyclopentadiene	12.669	237	152294	18.325	ng/ul	99
41) 2,4,6-Trichlorophenol	12.992	196	180327	20.772	ng/ul	99
42) 2,4,5-Trichlorophenol	13.069	196	193794	21.592	ng/ul	99
43) 1,1'-Biphenyl	13.398	154	660264	20.501	ng/ul	99
44) 2-Chloronaphthalene	13.445	162	531334	20.409	ng/ul	98
45) 2-Nitroaniline	13.698	65	152812	25.830	ng/ul	93
47) Dimethylphthalate	14.039	163	645293	19.952	ng/ul	99
48) 2,6-Dinitrotoluene	14.192	165	129743	21.795	ng/ul	94
50) Acenaphthylene	14.304	152	849163	20.714	ng/ul	99
51) 3-Nitroaniline	14.539	138	123540	21.820	ng/ul	99
52) Acenaphthene	14.639	153	548512	19.793	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	65323	19.548	ng/ul	89
55) 4-Nitrophenol	14.845	109	105893	21.394	ng/ul	96
56) Dibenzofuran	14.981	168	752997	19.353	ng/ul	91
57) 2,4-Dinitrotoluene	14.986	165	186174	21.190	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.204	232	153775	19.307	ng/ul#	91
59) Diethylphthalate	15.404	149	641893	20.556	ng/ul	99
61) Fluorene	15.639	166	618037	19.338	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.628	204	317405	18.685	ng/ul#	89
63) 4-Nitroaniline	15.716	138	117583	23.025	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.745	198	104085	18.698	ng/ul	92
67) N-Nitrosodiphenylamine	15.863	169	507871	21.247	ng/ul	99
68) 4-Bromophenyl-phenylether	16.539	248	183753	19.904	ng/ul#	89
69) Hexachlorobenzene	16.622	284	202276	19.085	ng/ul	92
70) Atrazine	16.828	200	176430	20.082	ng/ul	99
71) Pentachlorophenol	16.998	266	116867	18.341	ng/ul	98
72) Phenanthrene	17.410	178	926908	20.054	ng/ul	100
74) Anthracene	17.504	178	943636	20.157	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.351	216	282856	21.365	ng/uL	99
76) Pentachlorobenzene	14.869	250	256798	20.014	ng/uL	99
77) Carbazole	17.804	167	803452	20.307	ng/ul	98
78) Di-n-butylphthalate	18.310	149	1048778	22.832	ng/ul	99
80) Fluoranthene	19.398	202	1055211	21.901	ng/ul	99
82) Pyrene	19.763	202	1086633	21.288	ng/ul	100
83) Butylbenzylphthalate	20.633	149	460971	26.257	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.445	252	300609	19.893	ng/ul	99
85) Benzo(a)anthracene	21.498	228	1033801	20.331	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.380	149	664510	27.114	ng/ul	100
87) Chrysene	21.557	228	954028	19.840	ng/ul	99
89) Di-n-octyl phthalate	22.357	149	1103488	25.104	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	950901	20.134	ng/ul	100
91) Benzo(k)fluoranthene	23.327	252	938557	19.608	ng/ul	99
93) Benzo(a)pyrene	23.957	252	880205	19.852	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.792	276	1027808	19.450	ng/ul	99
95) Dibenzo(a,h)anthracene	26.821	278	858872	19.480	ng/ul	99
96) Benzo(g,h,i)perylene	27.639	276	824018	19.354	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 02 23:35:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP101823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 19 01:32:15 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.251 ng/uL
RT: 3.305 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion: 88 Resp: 36862
Ion Ratio Lower Upper
88 100
43 37.3 24.4 36.6#
58 70.2 53.0 79.6



#5
Pyridine
Concen: 17.916 ng/uL
RT: 3.722 min Scan# 108
Delta R.T. 0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion: 79 Resp: 230350
Ion Ratio Lower Upper
79 100
52 73.3 57.2 85.8
51 35.5 28.6 42.8









#6
Benzaldehyde
Concen: 24.368 ng/ul
RT: 7.040 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion: 77 Resp: 194498
Ion Ratio Lower Upper
77 100
105 93.7 75.4 113.2
106 88.8 72.3 108.5



#8
Phenol
Concen: 20.455 ng/ul
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion: 94 Resp: 299243
Ion Ratio Lower Upper
94 100
65 30.5 25.0 37.4
66 44.7 33.0 49.6









#10
Bis(2-Chloroethyl)ether
Concen: 19.918 ng/ul
RT: 7.322 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion: 93 Resp: 243105
Ion Ratio Lower Upper
93 100
63 72.3 60.7 91.1
95 32.3 27.8 41.6



#12
2-Chlorophenol
Concen: 20.805 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion: 128 Resp: 252875
Ion Ratio Lower Upper
128 100
64 47.3 36.7 55.1
130 32.5 26.2 39.2









#13
2-Methylphenol
Concen: 20.614 ng/ul
RT: 8.328 min Scan# 891
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:108 Resp: 227593
Ion Ratio Lower Upper
108 100
107 89.9 73.4 110.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 12.357 ng/ul
RT: 8.393 min Scan# 902
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion: 45 Resp: 185635
Ion Ratio Lower Upper
45 100
77 19.9 14.2 21.4
79 13.7 10.3 15.5









#16
Acetophenone
Concen: 20.462 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:105 Resp: 374831
Ion Ratio Lower Upper
105 100
77 85.9 64.4 96.6
51 27.4 21.2 31.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.177 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion: 70 Resp: 175314
Ion Ratio Lower Upper
70 100
42 55.1 44.6 66.8
101 9.7 7.5 11.3
130 21.2 17.3 25.9









#18
4-Methylphenol
Concen: 20.431 ng/ul
RT: 8.663 min Scan# 94
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:108 Resp: 239548
Ion Ratio Lower Upper
108 100
107 117.4 87.5 131.3



#19
Hexachloroethane
Concen: 21.188 ng/ul
RT: 8.934 min Scan# 994
Delta R.T. -0.018 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:117 Resp: 118333
Ion Ratio Lower Upper
117 100
201 106.2 96.3 144.5
199 64.0 57.7 86.5









#22
Nitrobenzene
Concen: 22.160 ng/ul
RT: 9.128 min Scan# 1030
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion: 77 Resp: 302286
Ion Ratio Lower Upper
77 100
123 40.0 33.5 50.3
65 14.0 10.8 16.2



#23
Isophorone
Concen: 21.834 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion: 82 Resp: 532619
Ion Ratio Lower Upper
82 100
95 8.3 6.2 9.2
138 17.8 14.4 21.6









#25
2-Nitrophenol
Concen: 21.339 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:139 Resp: 137880
Ion Ratio Lower Upper
139 100
65 56.8 40.4 60.6
109 38.3 25.4 38.2#



#26
2,4-Dimethylphenol
Concen: 21.075 ng/ul
RT: 9.875 min Scan# 1154
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:107 Resp: 267262
Ion Ratio Lower Upper
107 100
121 47.5 40.6 61.0
122 77.5 66.9 100.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.366 ng/ul

RT: 10.128 min Scan# 1197

Delta R.T. -0.018 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion: 93 Resp: 307264

Ion Ratio Lower Upper

93 100

95 33.4 24.7 37.1

123 10.8 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 20.632 ng/ul

RT: 10.357 min Scan# 1236

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion: 162 Resp: 234272


Ion Ratio Lower Upper

162 100

164 66.3 53.8 80.8

98 36.1 25.5 38.3









#30
Naphthalene
Concen: 20.003 ng/ul
RT: 10.757 min Scan# 11
Delta R.T. -0.018 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:128 Resp: 778293
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.4 10.5 15.7



#32
4-Chloroaniline
Concen: 19.918 ng/ul
RT: 10.910 min Scan# 1330
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:127 Resp: 296392
Ion Ratio Lower Upper
127 100
129 31.7 25.1 37.7





#33

Hexachlorobutadiene

Concen: 18.801 ng/ul

RT: 10.981 min Scan# 11

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:225 Resp: 173330

Ion Ratio Lower Upper

225 100

223 63.3 50.6 75.8

227 66.7 50.1 75.1



#34

Caprolactam

Concen: 20.619 ng/ul

RT: 11.740 min Scan# 1471

Delta R.T. -0.030 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:113 Resp: 60284


Ion Ratio Lower Upper

113 100

55 146.6 112.1 168.1

56 116.3 87.0 130.6









#35
4-Chloro-3-methylphenol
Concen: 22.190 ng/ul
RT: 12.022 min Scan# 1519
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:107 Resp: 237206
Ion Ratio Lower Upper
107 100
144 25.4 22.9 34.3
142 77.2 68.1 102.1



#36
2-Methylnaphthalene
Concen: 19.503 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:142 Resp: 512596
Ion Ratio Lower Upper
142 100
141 89.5 71.5 107.3





#37

1-Methylnaphthalene

Concen: 19.253 ng/ul

RT: 12.598 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:142 Resp: 517503

Ion Ratio Lower Upper

142 100

141 91.8 72.5 108.7

116 4.4 3.4 5.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.934 ng/ul

RT: 12.728 min Scan# 1639

Delta R.T. -0.018 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Tgt Ion:216 Resp: 288353




Ion Ratio Lower Upper


216 100

214 78.2 62.2 93.2

179 22.9 15.8 23.8

108 18.6 12.6 18.8





#40

Hexachlorocyclopentadiene

Concen: 18.325 ng/ul

RT: 12.669 min Scan# 1

Delta R.T. -0.018 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:237 Resp: 152294

Ion Ratio Lower Upper

237 100

235 63.1 50.3 75.5

272 12.4 11.1 16.7



#41

2,4,6-Trichlorophenol

Concen: 20.772 ng/ul

RT: 12.992 min Scan# 1684

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:196 Resp: 180327


Ion Ratio Lower Upper

196 100

198 92.5 74.7 112.1

200 30.4 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 21.592 ng/ul
RT: 13.069 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:196 Resp: 193794
Ion Ratio Lower Upper
196 100
198 94.8 76.0 114.0
200 29.6 24.3 36.5



#43
1,1'-Biphenyl
Concen: 20.501 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:154 Resp: 660264
Ion Ratio Lower Upper
154 100
153 40.2 32.6 49.0
76 12.2 9.8 14.8





#44

2-Chloronaphthalene

Concen: 20.409 ng/ul

RT: 13.445 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:162 Resp: 531334

Ion Ratio Lower Upper

162 100

164 31.9 25.9 38.9

127 39.7 29.9 44.9



#45

2-Nitroaniline

Concen: 25.830 ng/ul

RT: 13.698 min Scan# 1804

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion: 65 Resp: 152812


Ion Ratio Lower Upper

65 100

92 66.8 56.2 84.4

138 98.4 85.9 128.9









#47
Dimethylphthalate
Concen: 19.952 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:163 Resp: 645293
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.5 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.795 ng/ul
RT: 14.192 min Scan# 1888
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:165 Resp: 129743
Ion Ratio Lower Upper
165 100
89 64.8 48.0 72.0
121 22.7 16.6 24.8





#50

Acenaphthylene

Concen: 20.714 ng/ul

RT: 14.304 min Scan# 1907

Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:152 Resp: 849163

Ion Ratio Lower Upper

152 100

151 20.1 15.9 23.9

153 13.0 10.8 16.2



#51

3-Nitroaniline

Concen: 21.820 ng/ul

RT: 14.539 min Scan# 1947

Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:138 Resp: 123540


Ion Ratio Lower Upper

138 100

108 11.1 9.3 13.9

92 118.0 95.2 142.8









#52
Acenaphthene
Concen: 19.793 ng/ul
RT: 14.639 min Scan# 1964
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:153 Resp: 548512
Ion Ratio Lower Upper
153 100
152 48.2 37.6 56.4
154 88.2 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 19.548 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:184 Resp: 65323
Ion Ratio Lower Upper
184 100
63 70.3 50.6 76.0
154 70.2 47.9 71.9









#55
4-Nitrophenol
Concen: 21.394 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:109 Resp: 105893
Ion Ratio Lower Upper
109 100
139 108.7 92.9 139.3
65 119.0 94.5 141.7



#56
Dibenzofuran
Concen: 19.353 ng/ul
RT: 14.981 min Scan# 2022
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:168 Resp: 752997
Ion Ratio Lower Upper
168 100
139 41.2 28.6 43.0





#57

2,4-Dinitrotoluene

Concen: 21.190 ng/ul

RT: 14.986 min Scan# 20

Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:165 Resp: 186174

Ion Ratio Lower Upper

165 100

63 66.0 51.1 76.7

182 3.0 2.7 4.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.307 ng/ul

RT: 15.204 min Scan# 2060

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Tgt Ion:232 Resp: 153775




Ion Ratio Lower Upper


232 100

131 48.8 32.9 49.3

130 3.1 1.8 2.8#

166 31.8 23.3 34.9









#59
Diethylphthalate
Concen: 20.556 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:149 Resp: 641893
Ion Ratio Lower Upper
149 100
177 21.9 18.0 27.0
150 12.5 10.2 15.4



#61
Fluorene
Concen: 19.338 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:166 Resp: 618037
Ion Ratio Lower Upper
166 100
165 101.1 77.4 116.0
167 13.8 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.685 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:204 Resp: 317405
Ion Ratio Lower Upper
204 100
206 33.8 26.2 39.4
141 65.9 42.9 64.3#



#63
4-Nitroaniline
Concen: 23.025 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:138 Resp: 117583
Ion Ratio Lower Upper
138 100
92 60.2 43.0 64.6
108 111.9 82.7 124.1









#66
4,6-Dinitro-2-methylphenol
Concen: 18.698 ng/ul
RT: 15.745 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:198 Resp: 104085
Ion Ratio Lower Upper
198 100
182 4.3 3.9 5.9
77 29.7 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 21.247 ng/ul
RT: 15.863 min Scan# 2172
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:169 Resp: 507871
Ion Ratio Lower Upper
169 100
168 65.6 53.7 80.5
167 36.7 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 19.904 ng/ul
RT: 16.539 min Scan# 210
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:248 Resp: 183753
Ion Ratio Lower Upper
248 100
250 97.2 76.8 115.2
141 85.3 52.4 78.6#



#69
Hexachlorobenzene
Concen: 19.085 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:284 Resp: 202276
Ion Ratio Lower Upper
284 100
142 38.5 27.6 41.4
249 35.2 24.5 36.7





#70

Atrazine

Concen: 20.082 ng/ul

RT: 16.828 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:200 Resp: 176430

Ion Ratio Lower Upper

200 100

173 25.4 19.8 29.6

215 48.1 38.5 57.7



#71

Pentachlorophenol

Concen: 18.341 ng/ul

RT: 16.998 min Scan# 2365

Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:266 Resp: 116867


Ion Ratio Lower Upper

266 100

264 63.5 49.4 74.2

268 61.2 50.3 75.5









#72
Phenanthrene
Concen: 20.054 ng/ul
RT: 17.410 min Scan# 2438
Delta R.T. -0.012 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:178 Resp: 926908
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 19.2 15.3 22.9



#74
Anthracene
Concen: 20.157 ng/ul
RT: 17.504 min Scan# 2451
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:178 Resp: 943636
Ion Ratio Lower Upper
178 100
179 15.7 12.0 18.0
176 18.6 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.365 ng/uL

RT: 13.351 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:216 Resp: 282856

Ion Ratio Lower Upper

216 100

214 77.9 60.7 91.1

218 47.8 38.2 57.4



#76

Pentachlorobenzene

Concen: 20.014 ng/uL

RT: 14.869 min Scan# 2003

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:250 Resp: 256798


Ion Ratio Lower Upper

250 100

248 62.5 50.2 75.4

252 62.9 49.6 74.4





#77

Carbazole

Concen: 20.307 ng/ul

RT: 17.804 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:167 Resp: 803452

Ion Ratio Lower Upper

167 100

166 21.4 16.7 25.1

139 13.3 9.8 14.6



#78

Di-n-butylphthalate

Concen: 22.832 ng/ul

RT: 18.310 min Scan# 2588

Delta R.T. -0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:149 Resp: 1048778


Ion Ratio Lower Upper

149 100

150 9.1 7.2 10.8

104 6.0 4.3 6.5









#80
Fluoranthene
Concen: 21.901 ng/ul
RT: 19.398 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43




Instrument :
BNA_P
ClientSampleId :
SSTD020654


Tgt Ion:202 Resp: 1055211
Ion Ratio Lower Upper
202 100
101 9.5 7.1 10.7
100 7.3 5.7 8.5



#82
Pyrene
Concen: 21.288 ng/ul
RT: 19.763 min Scan# 2835
Delta R.T. -0.000 min
Lab File: BP017882.D
Acq: 02 Nov 2023 12:43

Tgt Ion:202 Resp: 1086633
Ion Ratio Lower Upper
202 100
101 10.2 8.3 12.5
100 9.1 7.4 11.0





#83

Butylbenzylphthalate

Concen: 26.257 ng/ul

RT: 20.633 min Scan# 2986

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:149 Resp: 460971

Ion Ratio Lower Upper

149 100

91 73.2 58.8 88.2

206 21.3 17.5 26.3



#84

3,3'-Dichlorobenzidine

Concen: 19.893 ng/ul

RT: 21.445 min Scan# 3121

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:252 Resp: 300609


Ion Ratio Lower Upper

252 100

254 65.1 51.8 77.8

126 10.7 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.331 ng/ul

RT: 21.498 min Scan# 3110

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:228 Resp: 1033801

Ion Ratio Lower Upper

228 100

229 19.0 15.6 23.4

226 27.5 21.1 31.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 27.114 ng/ul

RT: 21.380 min Scan# 3110

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:149 Resp: 664510


Ion Ratio Lower Upper

149 100

167 26.8 21.6 32.4

279 4.6 3.6 5.4





#87

Chrysene

Concen: 19.840 ng/ul

RT: 21.557 min Scan# 3143

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:228 Resp: 954028

Ion Ratio Lower Upper

228 100

226 30.9 24.0 36.0

229 19.8 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 25.104 ng/ul




RT: 22.357 min Scan# 3276


Delta R.T. -0.006 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Tgt Ion:149 Resp: 1103488





#90

Benzo(b)fluoranthene

Concen: 20.134 ng/ul

RT: 23.274 min Scan# 3436

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:252 Resp: 950901

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 8.8 7.1 10.7



#91

Benzo(k)fluoranthene

Concen: 19.608 ng/ul

RT: 23.327 min Scan# 3441

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:252 Resp: 938557


Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3

125 9.2 7.0 10.6





#93

Benzo(a)pyrene

Concen: 19.852 ng/ul

RT: 23.957 min Scan# 3

Delta R.T. -0.000 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:252 Resp: 880205

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 9.9 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.450 ng/ul

RT: 26.792 min Scan# 4030

Delta R.T. 0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43




Tgt Ion:276 Resp: 1027808


Ion Ratio Lower Upper

276 100

138 20.5 15.6 23.4

277 24.1 19.4 29.2





#95

Dibenzo(a,h)anthracene

Concen: 19.480 ng/ul

RT: 26.821 min Scan# 4041

Delta R.T. 0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Instrument :

BNA_P

ClientSampleId :

SSTD020654





Tgt Ion:278 Resp: 858872

Ion Ratio Lower Upper

278 100

139 13.8 11.4 17.2

279 23.8 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 19.354 ng/ul

RT: 27.639 min Scan# 4174

Delta R.T. 0.012 min

Lab File: BP017882.D

Acq: 02 Nov 2023 12:43

Tgt Ion:276 Resp: 824018

Ion Ratio Lower Upper

276 100

138 18.8 15.0 22.6

277 23.9 18.5 27.7

