

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110623\
 Data File : BP017914.D
 Acq On : 06 Nov 2023 14:47
 Operator : MA/JU
 Sample : SST00540
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005603

Quant Time: Nov 07 03:06:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:03:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	55055	20.000	ng/ul	0.00
20) Naphthalene-d8	10.681	136	252815	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.528	164	173273	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.298	188	399954	20.000	ng/ul	0.00
79) Chrysene-d12	21.416	240	414625	20.000	ng/ul	0.00
88) Perylene-d12	23.898	264	463235	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	2856	1.957	ng/uL	0.00
4) Pyridine-d5	3.693	84	14387	3.604	ng/ul	0.02
7) Phenol-d5	7.034	99	19095	3.699	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	12618	4.052	ng/ul	0.00
11) 2-Chlorophenol-d4	7.393	132	17633	4.350	ng/ul	0.00
15) 4-Methylphenol-d8	8.587	113	15744	3.690	ng/ul	0.00
21) Nitrobenzene-d5	9.069	128	8144	3.754	ng/ul	0.00
24) 2-Nitrophenol-d4	9.781	143	8386	3.337	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	15793	3.446	ng/ul	0.02
31) 4-Chloroaniline-d4	10.863	131	21500	3.478	ng/ul	0.01
46) Dimethylphthalate-d6	13.945	166	70550	4.401	ng/ul	0.00
49) Acenaphthylene-d8	14.222	160	75953	4.335	ng/ul	0.00
54) 4-Nitrophenol-d4	14.751	143	20	0.009	ng/ul	-0.02
60) Fluorene-d10	15.522	176	60134	4.502	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	8511	2.910	ng/ul	0.00
73) Anthracene-d10	17.398	188	98121	4.509	ng/ul	0.00
81) Pyrene-d10	19.645	212	115192	4.170	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.733	264	119536	4.390	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.281	88	3145	1.945	ng/uL#	89
5) Pyridine	3.711	79	16847	4.059	ng/ul	91
6) Benzaldehyde	7.040	77	11664	4.062	ng/ul	98
8) Phenol	7.063	94	19654	3.870	ng/ul#	90
10) Bis(2-Chloroethyl)ether	7.311	93	17678	4.346	ng/ul	93
12) 2-Chlorophenol	7.422	128	16919	4.188	ng/ul	97
13) 2-Methylphenol	8.322	108	15013	3.845	ng/ul	95
14) 2,2'-oxybis(1-Chloropr...	8.375	45	9750	2.082	ng/ul#	92
16) Acetophenone	8.722	105	28265	4.127	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.681	70	15795	4.244	ng/ul	98
18) 4-Methylphenol	8.658	108	14626	3.482	ng/ul	85
19) Hexachloroethane	8.916	117	9319	4.531	ng/ul	96
22) Nitrobenzene	9.116	77	22835	3.925	ng/ul	98
23) Isophorone	9.616	82	42605	4.005	ng/ul	98
25) 2-Nitrophenol	9.810	139	9172	3.539	ng/ul#	87
26) 2,4-Dimethylphenol	9.857	107	21118	3.901	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.110	93	22248	3.786	ng/ul	97
29) 2,4-Dichlorophenol	10.346	162	14274	3.182	ng/ul#	86
30) Naphthalene	10.734	128	70288	4.597	ng/ul	98
32) 4-Chloroaniline	10.887	127	19761	3.328	ng/ul	100
33) Hexachlorobutadiene	10.946	225	17035	4.685	ng/ul	94
34) Caprolactam	11.722	113	4885	3.680	ng/ul	90
35) 4-Chloro-3-methylphenol	11.999	107	19430	3.868	ng/ul	97
36) 2-Methylnaphthalene	12.346	142	45549	4.324	ng/ul	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110623\
 Data File : BP017914.D
 Acq On : 06 Nov 2023 14:47
 Operator : MA/JU
 Sample : SST00540
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005603

Quant Time: Nov 07 03:06:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:03:13 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.563	142	49021	4.523	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.693	216	28760	4.601	ng/ul	97
40) Hexachlorocyclopentadiene	12.628	237	7236	2.264	ng/ul	95
41) 2,4,6-Trichlorophenol	12.963	196	16244	4.078	ng/ul	95
42) 2,4,5-Trichlorophenol	13.046	196	16074	3.858	ng/ul	95
43) 1,1'-Biphenyl	13.357	154	65363	4.411	ng/ul	99
44) 2-Chloronaphthalene	13.404	162	51999	4.383	ng/ul	98
45) 2-Nitroaniline	13.663	65	10742	3.040	ng/ul	95
47) Dimethylphthalate	13.993	163	69775	4.415	ng/ul	98
48) 2,6-Dinitrotoluene	14.151	165	12579	4.022	ng/ul	91
50) Acenaphthylene	14.257	152	85170	4.336	ng/ul	99
51) 3-Nitroaniline	14.504	138	7797	2.783	ng/ul#	88
52) Acenaphthene	14.592	153	57457	4.405	ng/ul	95
53) 2,4-Dinitrophenol	14.728	184	1800	1.028	ng/ul	95
55) 4-Nitrophenol	14.875	109	174	0.062	ng/ul	94
56) Dibenzofuran	14.928	168	79442	4.384	ng/ul	99
57) 2,4-Dinitrotoluene	14.945	165	18902	4.076	ng/ul#	90
58) 2,3,4,6-Tetrachlorophenol	15.157	232	16262	4.199	ng/ul#	96
59) Diethylphthalate	15.345	149	70410	4.269	ng/ul	99
61) Fluorene	15.581	166	67838	4.436	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.575	204	35906	4.547	ng/ul	94
63) 4-Nitroaniline	15.675	138	6800	2.597	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.698	198	9460	3.096	ng/ul#	93
67) N-Nitrosodiphenylamine	15.804	169	55494	4.316	ng/ul	98
68) 4-Bromophenyl-phenylether	16.475	248	22580	4.567	ng/ul	96
69) Hexachlorobenzene	16.557	284	27472	4.867	ng/ul	99
70) Atrazine	16.763	200	21095	4.487	ng/ul	97
71) Pentachlorophenol	16.934	266	12228	3.724	ng/ul	97
72) Phenanthrene	17.339	178	109108	4.422	ng/ul	100
74) Anthracene	17.434	178	109605	4.372	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.310	216	29962	4.590	ng/ul	96
76) Pentachlorobenzene	14.822	250	30840	4.743	ng/ul	97
77) Carbazole	17.733	167	82398	3.866	ng/ul	97
78) Di-n-butylphthalate	18.233	149	115192	4.127	ng/ul	99
80) Fluoranthene	19.322	202	133041	4.126	ng/ul#	96
82) Pyrene	19.675	202	142038	4.233	ng/ul	96
83) Butylbenzylphthalate	20.539	149	55836	3.786	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.345	252	42278	4.079	ng/ul	99
85) Benzo(a)anthracene	21.398	228	147210	4.390	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.222	149	177	0.008	ng/ul	94
87) Chrysene	21.457	228	139035	4.479	ng/ul	99
89) Di-n-octyl phthalate	22.227	149	136707	3.494	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	143088	4.247	ng/ul	99
91) Benzo(k)fluoranthene	23.180	252	149804	4.459	ng/ul	98
93) Benzo(a)pyrene	23.786	252	133679	4.286	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.527	276	160732	4.300	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.557	278	135578	4.368	ng/ul#	96
96) Benzo(g,h,i)perylene	27.351	276	131925	4.430	ng/ul#	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110623\
 Data File : BP017914.D
 Acq On : 06 Nov 2023 14:47
 Operator : MA/JU
 Sample : SSTD00540
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005603

Quant Time: Nov 07 03:06:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:03:13 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.945 ng/uL
RT: 3.281 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:	88	Resp:	3145
Ion	Ratio	Lower	Upper
88	100		
43	49.5	26.2	39.4#
58	67.8	55.6	83.4



#5
Pyridine
Concen: 4.059 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:	79	Resp:	16847
Ion	Ratio	Lower	Upper
79	100		
52	70.1	51.0	76.6
51	43.2	29.8	44.8









#6
Benzaldehyde
Concen: 4.062 ng/ul
RT: 7.040 min Scan# 61
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion: 77 Resp: 11664
Ion Ratio Lower Upper
77 100
105 85.1 69.0 103.6
106 80.4 66.2 99.4



#8
Phenol
Concen: 3.870 ng/ul
RT: 7.063 min Scan# 676
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion: 94 Resp: 19654
Ion Ratio Lower Upper
94 100
65 44.4 27.7 41.5#
66 55.7 41.7 62.5









#10
Bis(2-Chloroethyl)ether
Concen: 4.346 ng/ul
RT: 7.311 min Scan# 718
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion: 93 Resp: 17678
Ion Ratio Lower Upper
93 100
63 79.6 57.2 85.8
95 32.0 26.2 39.4



#12
2-Chlorophenol
Concen: 4.188 ng/ul
RT: 7.422 min Scan# 737
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion: 128 Resp: 16919
Ion Ratio Lower Upper
128 100
64 47.3 39.8 59.6
130 31.0 25.8 38.8









#13
2-Methylphenol
Concen: 3.845 ng/ul
RT: 8.322 min Scan# 890
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:108 Resp: 15013
Ion Ratio Lower Upper
108 100
107 93.3 70.7 106.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 2.082 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion: 45 Resp: 9750
Ion Ratio Lower Upper
45 100
77 22.4 14.8 22.2#
79 17.0 11.2 16.8#









#16
Acetophenone
Concen: 4.127 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:105 Resp: 28265
Ion Ratio Lower Upper
105 100
77 86.0 67.7 101.5
51 31.1 23.4 35.2



#17
N-Nitroso-di-n-propylamine
Concen: 4.244 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion: 70 Resp: 15795
Ion Ratio Lower Upper
70 100
42 58.1 45.7 68.5
101 10.1 7.9 11.9
130 19.9 14.6 21.8









#18
4-Methylphenol
Concen: 3.482 ng/ul
RT: 8.658 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:108 Resp: 14626
Ion Ratio Lower Upper
108 100
107 124.3 87.2 130.8



#19
Hexachloroethane
Concen: 4.531 ng/ul
RT: 8.916 min Scan# 991
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:117 Resp: 9319
Ion Ratio Lower Upper
117 100
201 105.5 79.6 119.4
199 62.7 51.4 77.0









#22
Nitrobenzene
Concen: 3.925 ng/ul
RT: 9.116 min Scan# 110
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion: 77 Resp: 22835
Ion Ratio Lower Upper
77 100
123 38.7 31.2 46.8
65 12.4 11.5 17.3



#23
Isophorone
Concen: 4.005 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion: 82 Resp: 42605
Ion Ratio Lower Upper
82 100
95 7.9 7.2 10.8
138 16.0 13.4 20.2









#25
2-Nitrophenol
Concen: 3.539 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:139 Resp: 9172
Ion Ratio Lower Upper
139 100
65 68.1 49.0 73.6
109 52.8 32.8 49.2#



#26
2,4-Dimethylphenol
Concen: 3.901 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:107 Resp: 21118
Ion Ratio Lower Upper
107 100
121 44.2 37.8 56.8
122 77.2 61.9 92.9









#27
Bis(2-Chloroethoxy)methane
Concen: 3.786 ng/ul
RT: 10.110 min Scan# 1193
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion: 93 Resp: 22248
Ion Ratio Lower Upper
93 100
95 30.7 26.2 39.4
123 11.3 9.2 13.8



#29
2,4-Dichlorophenol
Concen: 3.182 ng/ul
RT: 10.346 min Scan# 1234
Delta R.T. 0.018 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion: 162 Resp: 14274
Ion Ratio Lower Upper
162 100
164 72.5 49.4 74.2
98 42.6 27.6 41.4#









#30
Naphthalene
Concen: 4.597 ng/ul
RT: 10.734 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:128 Resp: 70288
Ion Ratio Lower Upper
128 100
129 9.6 9.1 13.7
127 13.3 10.6 15.8



#32
4-Chloroaniline
Concen: 3.328 ng/ul
RT: 10.887 min Scan# 1326
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:127 Resp: 19761
Ion Ratio Lower Upper
127 100
129 32.5 26.2 39.2





#33

Hexachlorobutadiene

Concen: 4.685 ng/ul

RT: 10.946 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:225 Resp: 17035

Ion Ratio Lower Upper

225 100

223 66.7 47.7 71.5

227 65.9 50.8 76.2



#34

Caprolactam

Concen: 3.680 ng/ul

RT: 11.722 min Scan# 1468

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:113 Resp: 4885


Ion Ratio Lower Upper

113 100

55 129.7 114.1 171.1

56 105.5 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 3.868 ng/ul
RT: 11.999 min Scan# 1515
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:107 Resp: 19430
Ion Ratio Lower Upper
107 100
144 24.8 20.8 31.2
142 74.9 62.2 93.2



#36
2-Methylnaphthalene
Concen: 4.324 ng/ul
RT: 12.346 min Scan# 1574
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:142 Resp: 45549
Ion Ratio Lower Upper
142 100
141 84.8 70.9 106.3





#37

1-Methylnaphthalene

Concen: 4.523 ng/ul

RT: 12.563 min Scan# 1611

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:142 Resp: 49021

Ion Ratio Lower Upper

142 100

141 92.2 74.3 111.5

116 5.0 3.6 5.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.601 ng/ul

RT: 12.693 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Tgt Ion:216 Resp: 28760




Ion Ratio Lower Upper


216 100

214 75.8 63.4 95.2

179 24.5 18.9 28.3

108 19.4 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 2.264 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:237 Resp: 7236
Ion Ratio Lower Upper
237 100
235 59.9 51.3 76.9
272 12.7 9.4 14.2



#41
2,4,6-Trichlorophenol
Concen: 4.078 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:196 Resp: 16244
Ion Ratio Lower Upper
196 100
198 92.3 78.5 117.7
200 29.6 24.1 36.1









#42
2,4,5-Trichlorophenol
Concen: 3.858 ng/ul
RT: 13.046 min Scan# 1691
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:196 Resp: 16074
Ion Ratio Lower Upper
196 100
198 92.0 78.5 117.7
200 33.2 25.3 37.9



#43
1,1'-Biphenyl
Concen: 4.411 ng/ul
RT: 13.357 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:154 Resp: 65363
Ion Ratio Lower Upper
154 100
153 40.5 31.8 47.8
76 10.7 8.5 12.7





#44

2-Chloronaphthalene

Concen: 4.383 ng/ul

RT: 13.404 min Scan# 1754

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:162 Resp: 51999

Ion Ratio Lower Upper

162 100

164 34.9 26.7 40.1

127 42.7 33.0 49.6



#45

2-Nitroaniline

Concen: 3.040 ng/ul

RT: 13.663 min Scan# 1798

Delta R.T. 0.012 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion: 65 Resp: 10742


Ion Ratio Lower Upper

65 100

92 58.9 53.9 80.9

138 97.6 79.1 118.7









#47
Dimethylphthalate
Concen: 4.415 ng/ul
RT: 13.993 min Scan# 1854
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:163 Resp: 69775
Ion Ratio Lower Upper
163 100
194 4.4 3.2 4.8
164 9.7 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 4.022 ng/ul
RT: 14.151 min Scan# 1881
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:165 Resp: 12579
Ion Ratio Lower Upper
165 100
89 61.2 54.9 82.3
121 20.2 19.0 28.6





#50

Acenaphthylene

Concen: 4.336 ng/ul

RT: 14.257 min Scan# 1898

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:152 Resp: 85170

Ion Ratio Lower Upper

152 100

151 20.3 15.9 23.9

153 13.1 10.4 15.6



#51

3-Nitroaniline

Concen: 2.783 ng/ul

RT: 14.504 min Scan# 1941

Delta R.T. 0.018 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:138 Resp: 7797


Ion Ratio Lower Upper

138 100

108 19.4 8.4 12.6#

92 129.9 94.2 141.4









#52
Acenaphthene
Concen: 4.405 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:153 Resp: 57457
Ion Ratio Lower Upper
153 100
152 49.3 37.7 56.5
154 94.1 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 1.028 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. 0.024 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:184 Resp: 1800
Ion Ratio Lower Upper
184 100
63 75.4 54.5 81.7
154 69.8 56.4 84.6









#55
4-Nitrophenol
Concen: 0.062 ng/ul
RT: 14.875 min Scan# 2004
Delta R.T. 0.082 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:109 Resp: 174
Ion Ratio Lower Upper
109 100
139 91.6 79.6 119.4
65 104.6 87.8 131.8



#56
Dibenzofuran
Concen: 4.384 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:168 Resp: 79442
Ion Ratio Lower Upper
168 100
139 41.4 33.8 50.6





#57

2,4-Dinitrotoluene

Concen: 4.076 ng/ul

RT: 14.945 min Scan# 2016

Delta R.T. 0.012 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:165 Resp: 18902

Ion Ratio Lower Upper

165 100

63 58.2 53.3 79.9

182 3.5 1.9 2.9#



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.199 ng/ul

RT: 15.157 min Scan# 2052

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Tgt Ion:232 Resp: 16262




Ion Ratio Lower Upper


232 100

131 43.9 38.3 57.5

130 4.1 2.2 3.4#

166 30.4 25.2 37.8









#59
Diethylphthalate
Concen: 4.269 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:149 Resp: 70410
Ion Ratio Lower Upper
149 100
177 20.9 16.4 24.6
150 12.2 9.7 14.5



#61
Fluorene
Concen: 4.436 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:166 Resp: 67838
Ion Ratio Lower Upper
166 100
165 94.7 79.8 119.6
167 13.8 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 4.547 ng/ul
RT: 15.575 min Scan# 2119
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:204 Resp: 35906
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 60.1 53.6 80.4



#63
4-Nitroaniline
Concen: 2.597 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:138 Resp: 6800
Ion Ratio Lower Upper
138 100
92 62.7 45.4 68.2
108 136.7 103.1 154.7









#66
4,6-Dinitro-2-methylphenol
Concen: 3.096 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.012 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:198 Resp: 9460
Ion Ratio Lower Upper
198 100
182 7.6 4.4 6.6#
77 27.1 24.6 37.0



#67
N-Nitrosodiphenylamine
Concen: 4.316 ng/ul
RT: 15.804 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:169 Resp: 55494
Ion Ratio Lower Upper
169 100
168 67.0 53.4 80.0
167 37.3 28.0 42.0









#68
4-Bromophenyl-phenylether
Concen: 4.567 ng/ul
RT: 16.475 min Scan# 2276
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:248 Resp: 22580
Ion Ratio Lower Upper
248 100
250 100.0 77.9 116.9
141 83.8 62.4 93.6



#69
Hexachlorobenzene
Concen: 4.867 ng/ul
RT: 16.557 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:284 Resp: 27472
Ion Ratio Lower Upper
284 100
142 37.2 29.0 43.6
249 32.1 25.8 38.6





#70

Atrazine

Concen: 4.487 ng/ul

RT: 16.763 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:200 Resp: 21095

Ion Ratio Lower Upper

200 100

173 25.8 20.2 30.4

215 44.5 37.7 56.5



#71

Pentachlorophenol

Concen: 3.724 ng/ul

RT: 16.934 min Scan# 2354

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:266 Resp: 12228


Ion Ratio Lower Upper

266 100

264 65.3 53.8 80.6

268 61.8 52.4 78.6





#72

Phenanthrene

Concen: 4.422 ng/ul

RT: 17.339 min Scan# 2423

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:178 Resp: 109108

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 19.1 15.2 22.8



#74

Anthracene

Concen: 4.372 ng/ul

RT: 17.434 min Scan# 2439

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:178 Resp: 109605


Ion Ratio Lower Upper

178 100

179 14.6 12.5 18.7

176 19.2 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.590 ng/uL

RT: 13.310 min Scan# 1738

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:216 Resp: 29962

Ion Ratio Lower Upper

216 100

214 77.0 62.3 93.5

218 43.3 39.8 59.6



#76

Pentachlorobenzene

Concen: 4.743 ng/uL

RT: 14.822 min Scan# 1995

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:250 Resp: 30840


Ion Ratio Lower Upper

250 100

248 64.3 48.5 72.7

252 63.6 50.2 75.4





#77

Carbazole

Concen: 3.866 ng/ul

RT: 17.733 min Scan# 2489

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:167 Resp: 82398

Ion Ratio Lower Upper

167 100

166 20.9 17.7 26.5

139 15.8 11.0 16.4



#78

Di-n-butylphthalate

Concen: 4.127 ng/ul

RT: 18.233 min Scan# 2575

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:149 Resp: 115192


Ion Ratio Lower Upper

149 100

150 9.5 7.4 11.2

104 6.5 5.0 7.6









#80
Fluoranthene
Concen: 4.126 ng/ul
RT: 19.322 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47




Instrument :
BNA_P
ClientSampleId :
SSTD005603


Tgt Ion:202 Resp: 133041
Ion Ratio Lower Upper
202 100
101 7.7 7.1 10.7
100 5.0 5.3 7.9#



#82
Pyrene
Concen: 4.233 ng/ul
RT: 19.675 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BP017914.D
Acq: 06 Nov 2023 14:47

Tgt Ion:202 Resp: 142038
Ion Ratio Lower Upper
202 100
101 9.3 8.6 12.8
100 7.6 7.3 10.9





#83

Butylbenzylphthalate

Concen: 3.786 ng/ul

RT: 20.539 min Scan# 2967

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:149 Resp: 55836

Ion Ratio Lower Upper

149 100

91 70.1 60.4 90.6

206 20.5 15.3 22.9



#84

3,3'-Dichlorobenzidine

Concen: 4.079 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:252 Resp: 42278


Ion Ratio Lower Upper

252 100

254 62.9 50.7 76.1

126 10.4 7.9 11.9





#85

Benzo(a)anthracene

Concen: 4.390 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:228 Resp: 147210

Ion Ratio Lower Upper

228 100

229 18.7 15.4 23.2

226 27.4 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 0.008 ng/ul

RT: 21.222 min Scan# 3083

Delta R.T. -0.053 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:149 Resp: 177


Ion Ratio Lower Upper

149 100

167 23.2 21.0 31.6

279 3.1 3.0 4.6





#87

Chrysene

Concen: 4.479 ng/ul

RT: 21.457 min Scan# 3122

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:228 Resp: 139035

Ion Ratio Lower Upper

228 100

226 30.5 23.9 35.9

229 19.8 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 3.494 ng/ul




RT: 22.227 min Scan# 3254


Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Tgt Ion:149 Resp: 136707





#90

Benzo(b)fluoranthene

Concen: 4.247 ng/ul

RT: 23.127 min Scan# 3407

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:252 Resp: 143088

Ion Ratio Lower Upper

252 100

253 22.7 17.9 26.9

125 7.2 6.6 10.0



#91

Benzo(k)fluoranthene

Concen: 4.459 ng/ul

RT: 23.180 min Scan# 3416

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:252 Resp: 149804


Ion Ratio Lower Upper

252 100

253 20.8 17.2 25.8

125 7.7 7.0 10.6





#93

Benzo(a)pyrene

Concen: 4.286 ng/ul

RT: 23.786 min Scan# 3518

Delta R.T. 0.006 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47

Instrument :

BNA_P

ClientSampleId :

SSTD005603





Tgt Ion:252 Resp: 133679

Ion Ratio Lower Upper

252 100

253 22.7 17.3 25.9

125 8.0 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.300 ng/ul

RT: 26.527 min Scan# 3985

Delta R.T. 0.000 min

Lab File: BP017914.D

Acq: 06 Nov 2023 14:47




Tgt Ion:276 Resp: 160732


Ion Ratio Lower Upper

276 100

138 14.9 15.9 23.9#

277 24.3 19.5 29.3









#95
 Dibenzo(a,h)anthracene
 Concen: 4.368 ng/ul
 RT: 26.557 min Scan# 3989
 Delta R.T. 0.006 min
 Lab File: BP017914.D
 Acq: 06 Nov 2023 14:47

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005603

Tgt Ion: 278 Resp: 135578
 Ion Ratio Lower Upper
 278 100
 139 9.8 11.0 16.4#
 279 23.3 19.1 28.7



#96
 Benzo(g,h,i)perylene
 Concen: 4.430 ng/ul
 RT: 27.351 min Scan# 4125
 Delta R.T. 0.000 min
 Lab File: BP017914.D
 Acq: 06 Nov 2023 14:47

Tgt Ion: 276 Resp: 131925
 Ion Ratio Lower Upper
 276 100
 138 13.8 14.7 22.1#
 277 23.3 19.2 28.8

