

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110623\
 Data File : BP017916.D
 Acq On : 06 Nov 2023 16:29
 Operator : MA/JU
 Sample : SSTD02042
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605

Quant Time: Nov 07 03:07:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:03:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	77846	20.000	ng/u1	0.00
20) Naphthalene-d8	10.681	136	379576	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.522	164	253972	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	584279	20.000	ng/u1	0.00
79) Chrysene-d12	21.416	240	427885	20.000	ng/u1	0.00
88) Perylene-d12	23.898	264	457246	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	15100	7.316	ng/uL	0.00
4) Pyridine-d5	3.675	84	106792	18.918	ng/u1	0.00
7) Phenol-d5	7.028	99	133802	18.329	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	84867	19.272	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	110819	19.334	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	113937	18.883	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	56236	17.266	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	63708	16.884	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	132651	19.280	ng/u1	0.00
31) 4-Chloroaniline-d4	10.851	131	173302	18.672	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	437959	18.640	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	483320	18.818	ng/u1	0.00
54) 4-Nitrophenol-d4	14.775	143	58213	17.001	ng/u1	0.00
60) Fluorene-d10	15.522	176	368243	18.807	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.675	200	76119	17.815	ng/u1	0.00
73) Anthracene-d10	17.398	188	584716	18.392	ng/u1	0.00
81) Pyrene-d10	19.645	212	552644	19.386	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	489950	18.230	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.281	88	16788	7.342	ng/uL	100
5) Pyridine	3.699	79	111398	18.981	ng/u1	100
6) Benzaldehyde	7.028	77	91922	22.639	ng/u1	100
8) Phenol	7.058	94	132403	18.437	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.305	93	111148	19.325	ng/u1	100
12) 2-Chlorophenol	7.416	128	111072	19.446	ng/u1	100
13) 2-Methylphenol	8.316	108	105493	19.108	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.375	45	83580	12.624	ng/u1	100
16) Acetophenone	8.716	105	186338	19.244	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.681	70	100346	19.068	ng/u1	100
18) 4-Methylphenol	8.652	108	113199	19.060	ng/u1	100
19) Hexachloroethane	8.916	117	55379	19.042	ng/u1	100
22) Nitrobenzene	9.110	77	156906	17.962	ng/u1	100
23) Isophorone	9.616	82	279804	17.519	ng/u1	100
25) 2-Nitrophenol	9.810	139	66202	17.015	ng/u1	100
26) 2,4-Dimethylphenol	9.852	107	140693	17.309	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.104	93	152948	17.335	ng/u1	100
29) 2,4-Dichlorophenol	10.328	162	127664	18.954	ng/u1	100
30) Naphthalene	10.728	128	427575	18.627	ng/u1	100
32) 4-Chloroaniline	10.875	127	166676	18.699	ng/u1	100
33) Hexachlorobutadiene	10.946	225	107196	19.637	ng/u1	100
34) Caprolactam	11.716	113	34927	17.524	ng/u1	100
35) 4-Chloro-3-methylphenol	11.987	107	140792	18.667	ng/u1	100
36) 2-Methylnaphthalene	12.340	142	298350	18.862	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110623\
 Data File : BP017916.D
 Acq On : 06 Nov 2023 16:29
 Operator : MA/JU
 Sample : SST02042
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Quant Time: Nov 07 03:07:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:03:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.563	142	305557	18.779	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.693	216	180199	19.666	ng/ul	100
40) Hexachlorocyclopentadiene	12.628	237	83333	17.787	ng/ul	100
41) 2,4,6-Trichlorophenol	12.957	196	111996	19.180	ng/ul	100
42) 2,4,5-Trichlorophenol	13.034	196	116533	19.082	ng/ul	100
43) 1,1'-Biphenyl	13.357	154	405476	18.667	ng/ul	100
44) 2-Chloronaphthalene	13.404	162	328463	18.890	ng/ul	100
45) 2-Nitroaniline	13.651	65	93862	18.123	ng/ul	100
47) Dimethylphthalate	13.992	163	433266	18.702	ng/ul	100
48) 2,6-Dinitrotoluene	14.145	165	87420	19.072	ng/ul	100
50) Acenaphthylene	14.251	152	538433	18.700	ng/ul	100
51) 3-Nitroaniline	14.487	138	80869	19.695	ng/ul	100
52) Acenaphthene	14.587	153	354958	18.567	ng/ul	100
53) 2,4-Dinitrophenol	14.704	184	42750	16.662	ng/ul	100
55) 4-Nitrophenol	14.792	109	76630	18.623	ng/ul	100
56) Dibenzofuran	14.928	168	495778	18.668	ng/ul	100
57) 2,4-Dinitrotoluene	14.934	165	129282	19.020	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.151	232	110107	19.398	ng/ul	100
59) Diethylphthalate	15.345	149	449355	18.588	ng/ul	100
61) Fluorene	15.581	166	415050	18.517	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.569	204	218139	18.845	ng/ul	100
63) 4-Nitroaniline	15.663	138	76935	20.044	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.687	198	80504	18.035	ng/ul	100
67) N-Nitrosodiphenylamine	15.804	169	346215	18.434	ng/ul	100
68) 4-Bromophenyl-phenylether	16.475	248	138054	19.116	ng/ul	100
69) Hexachlorobenzene	16.557	284	161683	19.606	ng/ul	100
70) Atrazine	16.757	200	132257	19.257	ng/ul	100
71) Pentachlorophenol	16.928	266	93133	19.415	ng/ul	100
72) Phenanthrene	17.339	178	667673	18.525	ng/ul	100
74) Anthracene	17.433	178	673757	18.399	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.310	216	180581	18.935	ng/uL	100
76) Pentachlorobenzene	14.822	250	183842	19.354	ng/uL	100
77) Carbazole	17.728	167	575105	18.468	ng/ul	100
78) Di-n-butylphthalate	18.233	149	660795	16.205	ng/ul	100
80) Fluoranthene	19.316	202	647584	19.460	ng/ul	100
82) Pyrene	19.675	202	669982	19.346	ng/ul	100
83) Butylbenzylphthalate	20.539	149	305624	20.081	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.345	252	206129	19.271	ng/ul	100
85) Benzo(a)anthracene	21.398	228	651702	18.833	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.274	149	428323	19.609	ng/ul	100
87) Chrysene	21.451	228	600664	18.751	ng/ul	100
89) Di-n-octyl phthalate	22.227	149	727712	18.844	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	611953	18.400	ng/ul	100
91) Benzo(k)fluoranthene	23.174	252	611564	18.444	ng/ul	100
93) Benzo(a)pyrene	23.780	252	568456	18.465	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.527	276	665275	18.032	ng/ul	100
95) Dibenzo(a,h)anthracene	26.551	278	547589	17.873	ng/ul	100
96) Benzo(g,h,i)perylene	27.351	276	528731	17.989	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 07 03:07:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 07 03:03:13 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.342 ng/uL
RT: 3.281 min Scan# 33
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion: 88 Resp: 16788
Ion Ratio Lower Upper
88 100
43 32.8 26.2 39.4
58 69.5 55.6 83.4



#5
Pyridine
Concen: 18.981 ng/uL
RT: 3.699 min Scan# 104
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion: 79 Resp: 111398
Ion Ratio Lower Upper
79 100
52 63.8 51.0 76.6
51 37.3 29.8 44.8









#6
Benzaldehyde
Concen: 22.639 ng/ul
RT: 7.028 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion: 77 Resp: 91922
Ion Ratio Lower Upper
77 100
105 86.3 69.0 103.6
106 82.8 66.2 99.4



#8
Phenol
Concen: 18.437 ng/ul
RT: 7.058 min Scan# 675
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion: 94 Resp: 132403
Ion Ratio Lower Upper
94 100
65 34.6 27.7 41.5
66 52.1 41.7 62.5









#10
Bis(2-Chloroethyl)ether
Concen: 19.325 ng/ul
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion: 93 Resp: 111148
Ion Ratio Lower Upper
93 100
63 71.5 57.2 85.8
95 32.8 26.2 39.4



#12
2-Chlorophenol
Concen: 19.446 ng/ul
RT: 7.416 min Scan# 736
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:128 Resp: 111072
Ion Ratio Lower Upper
128 100
64 49.7 39.8 59.6
130 32.3 25.8 38.8









#13
2-Methylphenol
Concen: 19.108 ng/ul
RT: 8.316 min Scan# 889
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:108 Resp: 105493
Ion Ratio Lower Upper
108 100
107 88.4 70.7 106.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 12.624 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion: 45 Resp: 83580
Ion Ratio Lower Upper
45 100
77 18.5 14.8 22.2
79 14.0 11.2 16.8









#16
Acetophenone
Concen: 19.244 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:105 Resp: 186338
Ion Ratio Lower Upper
105 100
77 84.6 67.7 101.5
51 29.3 23.4 35.2



#17
N-Nitroso-di-n-propylamine
Concen: 19.068 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion: 70 Resp: 100346
Ion Ratio Lower Upper
70 100
42 57.1 45.7 68.5
101 9.9 7.9 11.9
130 18.2 14.6 21.8









#18
4-Methylphenol
Concen: 19.060 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:108 Resp: 113199
Ion Ratio Lower Upper
108 100
107 109.0 87.2 130.8



#19
Hexachloroethane
Concen: 19.042 ng/ul
RT: 8.916 min Scan# 991
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:117 Resp: 55379
Ion Ratio Lower Upper
117 100
201 99.5 79.6 119.4
199 64.2 51.4 77.0









#22
Nitrobenzene
Concen: 17.962 ng/ul
RT: 9.110 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion: 77 Resp: 156906
Ion Ratio Lower Upper
77 100
123 39.0 31.2 46.8
65 14.4 11.5 17.3



#23
Isophorone
Concen: 17.519 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion: 82 Resp: 279804
Ion Ratio Lower Upper
82 100
95 9.0 7.2 10.8
138 16.8 13.4 20.2









#25
2-Nitrophenol
Concen: 17.015 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:139 Resp: 66202
Ion Ratio Lower Upper
139 100
65 61.3 49.0 73.6
109 41.0 32.8 49.2



#26
2,4-Dimethylphenol
Concen: 17.309 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:107 Resp: 140693
Ion Ratio Lower Upper
107 100
121 47.3 37.8 56.8
122 77.4 61.9 92.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.335 ng/ul
 RT: 10.104 min Scan# 1193
 Delta R.T. 0.000 min
 Lab File: BP017916.D
 Acq: 06 Nov 2023 16:29




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion: 93 Resp: 152948
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.2 39.4
 123 11.5 9.2 13.8



#29
 2,4-Dichlorophenol
 Concen: 18.954 ng/ul
 RT: 10.328 min Scan# 1231
 Delta R.T. 0.000 min
 Lab File: BP017916.D
 Acq: 06 Nov 2023 16:29

Tgt Ion: 162 Resp: 127664
 Ion Ratio Lower Upper
 162 100
 164 61.8 49.4 74.2
 98 34.5 27.6 41.4









#30
Naphthalene
Concen: 18.627 ng/ul
RT: 10.728 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:128 Resp: 427575
Ion Ratio Lower Upper
128 100
129 11.4 9.1 13.7
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 18.699 ng/ul
RT: 10.875 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:127 Resp: 166676
Ion Ratio Lower Upper
127 100
129 32.7 26.2 39.2









#33
Hexachlorobutadiene
Concen: 19.637 ng/ul
RT: 10.946 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:225 Resp: 107196
Ion Ratio Lower Upper
225 100
223 59.6 47.7 71.5
227 63.5 50.8 76.2



#34
Caprolactam
Concen: 17.524 ng/ul
RT: 11.716 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:113 Resp: 34927
Ion Ratio Lower Upper
113 100
55 142.6 114.1 171.1
56 114.6 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 18.667 ng/ul
RT: 11.987 min Scan# 1513
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:107 Resp: 140792
Ion Ratio Lower Upper
107 100
144 26.0 20.8 31.2
142 77.7 62.2 93.2



#36
2-Methylnaphthalene
Concen: 18.862 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:142 Resp: 298350
Ion Ratio Lower Upper
142 100
141 88.6 70.9 106.3









#37
1-Methylnaphthalene
Concen: 18.779 ng/ul
RT: 12.563 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:142 Resp: 305557
Ion Ratio Lower Upper
142 100
141 92.9 74.3 111.5
116 4.5 3.6 5.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.666 ng/ul
RT: 12.693 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:216 Resp: 180199
Ion Ratio Lower Upper
216 100
214 79.3 63.4 95.2
179 23.6 18.9 28.3
108 19.1 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 17.787 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:237 Resp: 83333
Ion Ratio Lower Upper
237 100
235 64.1 51.3 76.9
272 11.8 9.4 14.2



#41
2,4,6-Trichlorophenol
Concen: 19.180 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:196 Resp: 111996
Ion Ratio Lower Upper
196 100
198 98.1 78.5 117.7
200 30.1 24.1 36.1





#42

2,4,5-Trichlorophenol

Concen: 19.082 ng/ul

RT: 13.034 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:196 Resp: 116533

Ion Ratio Lower Upper

196 100

198 98.1 78.5 117.7

200 31.6 25.3 37.9



#43

1,1'-Biphenyl

Concen: 18.667 ng/ul

RT: 13.357 min Scan# 1746

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:154 Resp: 405476


Ion Ratio Lower Upper

154 100

153 39.8 31.8 47.8

76 10.6 8.5 12.7









#44
2-Chloronaphthalene
Concen: 18.890 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:162 Resp: 328463
Ion Ratio Lower Upper
162 100
164 33.4 26.7 40.1
127 41.3 33.0 49.6



#45
2-Nitroaniline
Concen: 18.123 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion: 65 Resp: 93862
Ion Ratio Lower Upper
65 100
92 67.4 53.9 80.9
138 98.9 79.1 118.7









#47
Dimethylphthalate
Concen: 18.702 ng/ul
RT: 13.992 min Scan# 1854
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:163 Resp: 433266
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 19.072 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:165 Resp: 87420
Ion Ratio Lower Upper
165 100
89 68.6 54.9 82.3
121 23.8 19.0 28.6









#50
Acenaphthylene
Concen: 18.700 ng/ul
RT: 14.251 min Scan# 1898
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:152 Resp: 538433
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 19.695 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:138 Resp: 80869
Ion Ratio Lower Upper
138 100
108 10.5 8.4 12.6
92 117.8 94.2 141.4





#52

Acenaphthene

Concen: 18.567 ng/ul

RT: 14.587 min Scan# 1955

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:153 Resp: 354958

Ion Ratio Lower Upper

153 100

152 47.1 37.7 56.5

154 88.0 70.4 105.6



#53

2,4-Dinitrophenol

Concen: 16.662 ng/ul

RT: 14.704 min Scan# 1975

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:184 Resp: 42750


Ion Ratio Lower Upper

184 100

63 68.1 54.5 81.7

154 70.5 56.4 84.6









#55
4-Nitrophenol
Concen: 18.623 ng/ul
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:109 Resp: 76630
Ion Ratio Lower Upper
109 100
139 99.5 79.6 119.4
65 109.8 87.8 131.8



#56
Dibenzofuran
Concen: 18.668 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:168 Resp: 495778
Ion Ratio Lower Upper
168 100
139 42.2 33.8 50.6









#57
2,4-Dinitrotoluene
Concen: 19.020 ng/ul
RT: 14.934 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:165 Resp: 129282
Ion Ratio Lower Upper
165 100
63 66.6 53.3 79.9
182 2.4 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.398 ng/ul
RT: 15.151 min Scan# 2051
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:232 Resp: 110107
Ion Ratio Lower Upper
232 100
131 47.9 38.3 57.5
130 2.8 2.2 3.4
166 31.5 25.2 37.8









#59
Diethylphthalate
Concen: 18.588 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:149 Resp: 449355
Ion Ratio Lower Upper
149 100
177 20.5 16.4 24.6
150 12.1 9.7 14.5



#61
Fluorene
Concen: 18.517 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:166 Resp: 415050
Ion Ratio Lower Upper
166 100
165 99.7 79.8 119.6
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.845 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:204 Resp: 218139
Ion Ratio Lower Upper
204 100
206 32.5 26.0 39.0
141 67.0 53.6 80.4



#63
4-Nitroaniline
Concen: 20.044 ng/ul
RT: 15.663 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:138 Resp: 76935
Ion Ratio Lower Upper
138 100
92 56.8 45.4 68.2
108 128.9 103.1 154.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.035 ng/ul
 RT: 15.687 min Scan# 2142
 Delta R.T. 0.000 min
 Lab File: BP017916.D
 Acq: 06 Nov 2023 16:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605


Tgt Ion:198 Resp: 80504
 Ion Ratio Lower Upper
 198 100
 182 5.5 4.4 6.6
 77 30.8 24.6 37.0



#67
 N-Nitrosodiphenylamine
 Concen: 18.434 ng/ul
 RT: 15.804 min Scan# 2162
 Delta R.T. 0.000 min
 Lab File: BP017916.D
 Acq: 06 Nov 2023 16:29

Tgt Ion:169 Resp: 346215
 Ion Ratio Lower Upper
 169 100
 168 66.7 53.4 80.0
 167 35.0 28.0 42.0









#68
4-Bromophenyl-phenylether
Concen: 19.116 ng/ul
RT: 16.475 min Scan# 2276
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:248 Resp: 138054
Ion Ratio Lower Upper
248 100
250 97.4 77.9 116.9
141 78.0 62.4 93.6



#69
Hexachlorobenzene
Concen: 19.606 ng/ul
RT: 16.557 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:284 Resp: 161683
Ion Ratio Lower Upper
284 100
142 36.3 29.0 43.6
249 32.2 25.8 38.6





#70

Atrazine

Concen: 19.257 ng/ul

RT: 16.757 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:200 Resp: 132257

Ion Ratio Lower Upper

200 100

173 25.3 20.2 30.4

215 47.1 37.7 56.5



#71

Pentachlorophenol

Concen: 19.415 ng/ul

RT: 16.928 min Scan# 2353

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:266 Resp: 93133


Ion Ratio Lower Upper

266 100

264 67.2 53.8 80.6

268 65.5 52.4 78.6





#72

Phenanthrene

Concen: 18.525 ng/ul

RT: 17.339 min Scan# 2423

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:178 Resp: 667673

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.0 15.2 22.8



#74

Anthracene

Concen: 18.399 ng/ul

RT: 17.433 min Scan# 2439

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:178 Resp: 673757


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.6 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.935 ng/uL

RT: 13.310 min Scan# 1738

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:216 Resp: 180581

Ion Ratio Lower Upper

216 100

214 77.9 62.3 93.5

218 49.7 39.8 59.6



#76

Pentachlorobenzene

Concen: 19.354 ng/uL

RT: 14.822 min Scan# 1995

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:250 Resp: 183842


Ion Ratio Lower Upper

250 100

248 60.6 48.5 72.7

252 62.8 50.2 75.4





#77

Carbazole

Concen: 18.468 ng/ul

RT: 17.728 min Scan# 2489

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:167 Resp: 575105

Ion Ratio Lower Upper

167 100

166 22.1 17.7 26.5

139 13.7 11.0 16.4



#78

Di-n-butylphthalate

Concen: 16.205 ng/ul

RT: 18.233 min Scan# 2575

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:149 Resp: 660795


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 6.3 5.0 7.6





#80

Fluoranthene

Concen: 19.460 ng/ul

RT: 19.316 min Scan# 2759

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:202 Resp: 647584

Ion Ratio Lower Upper

202 100

101 8.9 7.1 10.7

100 6.6 5.3 7.9



#82

Pyrene

Concen: 19.346 ng/ul

RT: 19.675 min Scan# 2820

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:202 Resp: 669982


Ion Ratio Lower Upper

202 100

101 10.7 8.6 12.8

100 9.1 7.3 10.9





#83

Butylbenzylphthalate

Concen: 20.081 ng/ul

RT: 20.539 min Scan# 2967

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:149 Resp: 305624

Ion Ratio Lower Upper

149 100

91 75.5 60.4 90.6

206 19.1 15.3 22.9



#84

3,3'-Dichlorobenzidine

Concen: 19.271 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:252 Resp: 206129


Ion Ratio Lower Upper

252 100

254 63.4 50.7 76.1

126 9.9 7.9 11.9









#85
Benzo(a)anthracene
Concen: 18.833 ng/ul
RT: 21.398 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:228 Resp: 651702
Ion Ratio Lower Upper
228 100
229 19.3 15.4 23.2
226 27.0 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.609 ng/ul
RT: 21.274 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BP017916.D
Acq: 06 Nov 2023 16:29

Tgt Ion:149 Resp: 428323
Ion Ratio Lower Upper
149 100
167 26.3 21.0 31.6
279 3.8 3.0 4.6





#87

Chrysene

Concen: 18.751 ng/ul

RT: 21.451 min Scan# 3122

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:228 Resp: 600664

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 20.1 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 18.844 ng/ul




RT: 22.227 min Scan# 3254


Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Tgt Ion:149 Resp: 727712





#90

Benzo(b)fluoranthene

Concen: 18.400 ng/ul

RT: 23.127 min Scan# 3407

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:252 Resp: 611953

Ion Ratio Lower Upper

252 100

253 22.4 17.9 26.9

125 8.3 6.6 10.0



#91

Benzo(k)fluoranthene

Concen: 18.444 ng/ul

RT: 23.174 min Scan# 3415

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:252 Resp: 611564


Ion Ratio Lower Upper

252 100

253 21.5 17.2 25.8

125 8.8 7.0 10.6





#93

Benzo(a)pyrene

Concen: 18.465 ng/ul

RT: 23.780 min Scan# 3518

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:252 Resp: 568456

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 9.4 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.032 ng/ul

RT: 26.527 min Scan# 3985

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29




Tgt Ion:276 Resp: 665275


Ion Ratio Lower Upper

276 100

138 19.9 15.9 23.9

277 24.4 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 17.873 ng/ul

RT: 26.551 min Scan# 3989

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:278 Resp: 547589

Ion Ratio Lower Upper

278 100

139 13.7 11.0 16.4

279 23.9 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 17.989 ng/ul

RT: 27.351 min Scan# 4125

Delta R.T. 0.000 min

Lab File: BP017916.D

Acq: 06 Nov 2023 16:29

Tgt Ion:276 Resp: 528731

Ion Ratio Lower Upper

276 100

138 18.4 14.7 22.1

277 24.0 19.2 28.8

