

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110723\
 Data File : BP017929.D
 Acq On : 07 Nov 2023 17:43
 Operator : MA/JU
 Sample : PB156694BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS694

Quant Time: Nov 07 21:49:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:40:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.858	152	76140	20.000	ng/ul	0.00
20) Naphthalene-d8	10.675	136	314610	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.522	164	203314	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.292	188	452290	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.410	240	428718	20.000	ng/ul	0.00
88) Perylene-d12	23.892	264	447532	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	10318	5.186	ng/uL	0.00
4) Pyridine-d5	3.676	84	131669	23.848	ng/ul	0.00
7) Phenol-d5	7.022	99	177816	24.904	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.211	67	107780	25.024	ng/ul	0.00
11) 2-Chlorophenol-d4	7.381	132	140898	25.681	ng/ul	0.00
15) 4-Methylphenol-d8	8.581	113	144794	24.535	ng/ul	0.00
21) Nitrobenzene-d5	9.063	128	69566	27.669	ng/ul	0.00
24) 2-Nitrophenol-d4	9.775	143	80243	27.710	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	142961	26.929	ng/ul	0.00
31) 4-Chloroaniline-d4	10.852	131	182673	23.745	ng/ul	0.00
46) Dimethylphthalate-d6	13.946	166	466248	26.153	ng/ul	0.00
49) Acenaphthylene-d8	14.222	160	509970	26.069	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	74347	27.124	ng/ul	0.00
60) Fluorene-d10	15.522	176	387118	25.944	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	84082	25.421	ng/ul	0.00
73) Anthracene-d10	17.398	188	609206	25.906	ng/ul	0.00
81) Pyrene-d10	19.645	212	725256	26.649	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.727	264	674612	26.978	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.281	88	22274	10.254	ng/uL	97
5) Pyridine	3.699	79	139488	24.299	ng/ul	98
6) Benzaldehyde	7.028	77	103448	26.049	ng/ul	96
8) Phenol	7.052	94	187505	26.695	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.305	93	152152	27.047	ng/ul	94
12) 2-Chlorophenol	7.417	128	154044	28.272	ng/ul	99
13) 2-Methylphenol	8.311	108	142288	26.350	ng/ul	94
14) 2,2'-oxybis(1-Chloropr...	8.369	45	103168	15.930	ng/ul	98
16) Acetophenone	8.716	105	249883	26.385	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.681	70	136907	27.622	ng/ul	99
18) 4-Methylphenol	8.646	108	156782	26.990	ng/ul	97
19) Hexachloroethane	8.911	117	72715	26.717	ng/ul	97
22) Nitrobenzene	9.105	77	209871	30.607	ng/ul	99
23) Isophorone	9.611	82	375976	30.151	ng/ul	98
25) 2-Nitrophenol	9.805	139	91393	30.497	ng/ul	97
26) 2,4-Dimethylphenol	9.846	107	162719	25.709	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.105	93	208796	30.440	ng/ul	98
29) 2,4-Dichlorophenol	10.328	162	151699	29.548	ng/ul	93
30) Naphthalene	10.728	128	511540	28.015	ng/ul	99
32) 4-Chloroaniline	10.875	127	176175	23.845	ng/ul	97
33) Hexachlorobutadiene	10.946	225	108731	25.454	ng/ul	97
34) Caprolactam	11.710	113	48235	29.480	ng/ul	97
35) 4-Chloro-3-methylphenol	11.981	107	178326	30.004	ng/ul	93
36) 2-Methylnaphthalene	12.340	142	349881	28.082	ng/ul	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110723\
 Data File : BP017929.D
 Acq On : 07 Nov 2023 17:43
 Operator : MA/JU
 Sample : PB156694BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS694


Quant Time: Nov 07 21:49:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:40:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	352959	27.422	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.687	216	186563	26.491	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	80486	21.460	ng/ul	98
41) 2,4,6-Trichlorophenol	12.951	196	123321	27.699	ng/ul	99
42) 2,4,5-Trichlorophenol	13.028	196	128921	28.120	ng/ul	98
43) 1,1'-Biphenyl	13.351	154	474439	28.615	ng/ul	98
44) 2-Chloronaphthalene	13.404	162	376570	28.510	ng/ul	98
45) 2-Nitroaniline	13.651	65	127079	32.132	ng/ul	95
47) Dimethylphthalate	13.993	163	504503	28.695	ng/ul	99
48) 2,6-Dinitrotoluene	14.146	165	103777	29.739	ng/ul	97
50) Acenaphthylene	14.251	152	631747	28.853	ng/ul	100
51) 3-Nitroaniline	14.487	138	97004	29.511	ng/ul	94
52) Acenaphthene	14.587	153	415080	28.604	ng/ul	99
53) 2,4-Dinitrophenol	14.698	184	53899	26.241	ng/ul#	82
55) 4-Nitrophenol	14.787	109	105266	29.600	ng/ul	94
56) Dibenzofuran	14.928	168	570765	28.475	ng/ul	98
57) 2,4-Dinitrotoluene	14.934	165	156482	30.008	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.146	232	114572	26.817	ng/ul#	88
59) Diethylphthalate	15.345	149	538833	29.415	ng/ul	98
61) Fluorene	15.575	166	488163	28.763	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.569	204	241552	27.502	ng/ul	99
63) 4-Nitroaniline	15.657	138	98291	31.608	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.687	198	96381	27.893	ng/ul	93
67) N-Nitrosodiphenylamine	15.798	169	404450	29.054	ng/ul	98
68) 4-Bromophenyl-phenylether	16.475	248	143648	26.971	ng/ul	95
69) Hexachlorobenzene	16.551	284	158795	25.976	ng/ul#	87
70) Atrazine	16.757	200	129053	24.274	ng/ul	96
71) Pentachlorophenol	16.928	266	94038	25.325	ng/ul	96
72) Phenanthrene	17.340	178	787637	29.617	ng/ul	99
74) Anthracene	17.434	178	774099	28.636	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.304	216	185124	26.117	ng/ul	98
76) Pentachlorobenzene	14.816	250	176669	24.977	ng/ul	98
77) Carbazole	17.728	167	709832	29.447	ng/ul	99
78) Di-n-butylphthalate	18.228	149	919978	30.672	ng/ul	100
80) Fluoranthene	19.316	202	937568	29.519	ng/ul	99
82) Pyrene	19.675	202	985984	29.763	ng/ul	99
83) Butylbenzylphthalate	20.539	149	438753	30.103	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.339	252	305154	28.474	ng/ul#	98
85) Benzo(a)anthracene	21.398	228	970725	29.401	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.275	149	631567	29.924	ng/ul	99
87) Chrysene	21.451	228	903269	29.503	ng/ul	99
89) Di-n-octyl phthalate	22.227	149	1087421	28.770	ng/ul	100
90) Benzo(b)fluoranthene	23.122	252	921595	29.836	ng/ul	98
91) Benzo(k)fluoranthene	23.174	252	940378	30.359	ng/ul	99
93) Benzo(a)pyrene	23.780	252	866294	30.169	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.527	276	1011642	29.580	ng/ul	98
95) Dibenzo(a,h)anthracene	26.545	278	842890	29.627	ng/ul	99
96) Benzo(g,h,i)perylene	27.351	276	807395	29.653	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 07 21:49:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 07 03:40:56 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.254 ng/uL
RT: 3.281 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion	88	Resp	22274
Ion Ratio	100		
Lower	26.2		
Upper	39.4		



#5
Pyridine
Concen: 24.299 ng/uL
RT: 3.699 min Scan# 104
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion	79	Resp	139488
Ion Ratio <td>100</td> <td></td> <td></td>	100		
Lower	51.0		
Upper	76.6		









#6
Benzaldehyde
Concen: 26.049 ng/ul
RT: 7.028 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion: 77 Resp: 103448
Ion Ratio Lower Upper
77 100
105 89.8 69.0 103.6
106 86.3 66.2 99.4



#8
Phenol
Concen: 26.695 ng/ul
RT: 7.052 min Scan# 674
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion: 94 Resp: 187505
Ion Ratio Lower Upper
94 100
65 34.8 27.7 41.5
66 51.1 41.7 62.5





#10

Bis(2-Chloroethyl)ether

Concen: 27.047 ng/ul

RT: 7.305 min Scan# 717

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion: 93 Resp: 152152

Ion Ratio Lower Upper

93 100

63 77.2 57.2 85.8

95 31.1 26.2 39.4



#12

2-Chlorophenol

Concen: 28.272 ng/ul

RT: 7.417 min Scan# 736

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:128 Resp: 154044


Ion Ratio Lower Upper

128 100

64 49.0 39.8 59.6

130 32.4 25.8 38.8





#13

2-Methylphenol

Concen: 26.350 ng/ul

RT: 8.311 min Scan# 888

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :





SLCS694

Tgt Ion:108 Resp: 142288

Ion Ratio Lower Upper

108 100

107 93.7 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 15.930 ng/ul

RT: 8.369 min Scan# 898

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion: 45 Resp: 103168


Ion Ratio Lower Upper

45 100

77 19.5 14.8 22.2

79 13.3 11.2 16.8









#16
Acetophenone
Concen: 26.385 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:105 Resp: 249883
Ion Ratio Lower Upper
105 100
77 87.1 67.7 101.5
51 28.8 23.4 35.2



#17
N-Nitroso-di-n-propylamine
Concen: 27.622 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion: 70 Resp: 136907
Ion Ratio Lower Upper
70 100
42 57.3 45.7 68.5
101 9.4 7.9 11.9
130 19.3 14.6 21.8









#18
4-Methylphenol
Concen: 26.990 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:108 Resp: 156782
Ion Ratio Lower Upper
108 100
107 112.3 87.2 130.8



#19
Hexachloroethane
Concen: 26.717 ng/ul
RT: 8.911 min Scan# 990
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:117 Resp: 72715
Ion Ratio Lower Upper
117 100
201 98.8 79.6 119.4
199 58.8 51.4 77.0









#22
Nitrobenzene
Concen: 30.607 ng/ul
RT: 9.105 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion: 77 Resp: 209871
Ion Ratio Lower Upper
77 100
123 38.2 31.2 46.8
65 14.1 11.5 17.3



#23
Isophorone
Concen: 30.151 ng/ul
RT: 9.611 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion: 82 Resp: 375976
Ion Ratio Lower Upper
82 100
95 8.2 7.2 10.8
138 15.8 13.4 20.2









#25
2-Nitrophenol
Concen: 30.497 ng/ul
RT: 9.805 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:139 Resp: 91393
Ion Ratio Lower Upper
139 100
65 65.1 49.0 73.6
109 41.5 32.8 49.2



#26
2,4-Dimethylphenol
Concen: 25.709 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:107 Resp: 162719
Ion Ratio Lower Upper
107 100
121 47.8 37.8 56.8
122 78.4 61.9 92.9





#27

Bis(2-Chloroethoxy)methane

Concen: 30.440 ng/ul

RT: 10.105 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion: 93 Resp: 208796

Ion Ratio Lower Upper

93 100

95 33.8 26.2 39.4

123 10.8 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 29.548 ng/ul

RT: 10.328 min Scan# 1231

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:162 Resp: 151699


Ion Ratio Lower Upper

162 100

164 67.6 49.4 74.2

98 37.9 27.6 41.4









#30
Naphthalene
Concen: 28.015 ng/ul
RT: 10.728 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:128 Resp: 511540
Ion Ratio Lower Upper
128 100
129 10.6 9.1 13.7
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 23.845 ng/ul
RT: 10.875 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:127 Resp: 176175
Ion Ratio Lower Upper
127 100
129 30.8 26.2 39.2









#33
Hexachlorobutadiene
Concen: 25.454 ng/ul
RT: 10.946 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:225 Resp: 108731
Ion Ratio Lower Upper
225 100
223 62.8 47.7 71.5
227 65.4 50.8 76.2



#34
Caprolactam
Concen: 29.480 ng/ul
RT: 11.710 min Scan# 1466
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:113 Resp: 48235
Ion Ratio Lower Upper
113 100
55 141.1 114.1 171.1
56 119.1 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 30.004 ng/ul
RT: 11.981 min Scan# 1512
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:107 Resp: 178326
Ion Ratio Lower Upper
107 100
144 22.4 20.8 31.2
142 72.0 62.2 93.2



#36
2-Methylnaphthalene
Concen: 28.082 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:142 Resp: 349881
Ion Ratio Lower Upper
142 100
141 92.4 70.9 106.3









#37
1-Methylnaphthalene
Concen: 27.422 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:142 Resp: 352959
Ion Ratio Lower Upper
142 100
141 91.3 74.3 111.5
116 4.9 3.6 5.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 26.491 ng/ul
RT: 12.687 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:216 Resp: 186563
Ion Ratio Lower Upper
216 100
214 80.0 63.4 95.2
179 24.3 18.9 28.3
108 21.8 15.3 22.9





#40

Hexachlorocyclopentadiene

Concen: 21.460 ng/ul

RT: 12.628 min Scan# 1622

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:237 Resp: 80486

Ion Ratio Lower Upper

237 100

235 65.6 51.3 76.9

272 11.5 9.4 14.2



#41

2,4,6-Trichlorophenol

Concen: 27.699 ng/ul

RT: 12.951 min Scan# 1677

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:196 Resp: 123321


Ion Ratio Lower Upper

196 100

198 98.7 78.5 117.7

200 32.5 24.1 36.1





#42

2,4,5-Trichlorophenol

Concen: 28.120 ng/ul

RT: 13.028 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:196 Resp: 128921

Ion Ratio Lower Upper

196 100

198 95.4 78.5 117.7

200 32.2 25.3 37.9



#43

1,1'-Biphenyl

Concen: 28.615 ng/ul

RT: 13.351 min Scan# 1745

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:154 Resp: 474439


Ion Ratio Lower Upper

154 100

153 40.4 31.8 47.8

76 12.0 8.5 12.7









#44
2-Chloronaphthalene
Concen: 28.510 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:162 Resp: 376570
Ion Ratio Lower Upper
162 100
164 33.4 26.7 40.1
127 43.0 33.0 49.6



#45
2-Nitroaniline
Concen: 32.132 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion: 65 Resp: 127079
Ion Ratio Lower Upper
65 100
92 64.9 53.9 80.9
138 93.1 79.1 118.7









#47
Dimethylphthalate
Concen: 28.695 ng/ul
RT: 13.993 min Scan# 1854
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:163 Resp: 504503
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 29.739 ng/ul
RT: 14.146 min Scan# 1880
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:165 Resp: 103777
Ion Ratio Lower Upper
165 100
89 71.7 54.9 82.3
121 22.9 19.0 28.6





#50

Acenaphthylene

Concen: 28.853 ng/ul

RT: 14.251 min Scan# 1898

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:152 Resp: 631747

Ion Ratio Lower Upper

152 100

151 20.0 15.9 23.9

153 12.9 10.4 15.6



#51

3-Nitroaniline

Concen: 29.511 ng/ul

RT: 14.487 min Scan# 1938

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:138 Resp: 97004


Ion Ratio Lower Upper

138 100

108 11.1 8.4 12.6

92 125.0 94.2 141.4









#52
Acenaphthene
Concen: 28.604 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:153 Resp: 415080
Ion Ratio Lower Upper
153 100
152 48.7 37.7 56.5
154 88.5 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 26.241 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:184 Resp: 53899
Ion Ratio Lower Upper
184 100
63 89.1 54.5 81.7#
154 79.6 56.4 84.6









#55
4-Nitrophenol
Concen: 29.600 ng/ul
RT: 14.787 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:109 Resp: 105266
Ion Ratio Lower Upper
109 100
139 96.2 79.6 119.4
65 118.3 87.8 131.8



#56
Dibenzofuran
Concen: 28.475 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:168 Resp: 570765
Ion Ratio Lower Upper
168 100
139 43.7 33.8 50.6





#57

2,4-Dinitrotoluene

Concen: 30.008 ng/ul

RT: 14.934 min Scan# 2014

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:165 Resp: 156482

Ion Ratio Lower Upper

165 100

63 66.3 53.3 79.9

182 2.5 1.9 2.9



#58

2,3,4,6-Tetrachlorophenol

Concen: 26.817 ng/ul

RT: 15.146 min Scan# 2050

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Tgt Ion:232 Resp: 114572




Ion Ratio Lower Upper


232 100

131 57.5 38.3 57.5#

130 3.5 2.2 3.4#

166 36.6 25.2 37.8









#59
Diethylphthalate
Concen: 29.415 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:149 Resp: 538833
Ion Ratio Lower Upper
149 100
177 21.2 16.4 24.6
150 12.7 9.7 14.5



#61
Fluorene
Concen: 28.763 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:166 Resp: 488163
Ion Ratio Lower Upper
166 100
165 98.2 79.8 119.6
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 27.502 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:204 Resp: 241552
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 67.6 53.6 80.4



#63
4-Nitroaniline
Concen: 31.608 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:138 Resp: 98291
Ion Ratio Lower Upper
138 100
92 56.4 45.4 68.2
108 120.5 103.1 154.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 27.893 ng/ul
 RT: 15.687 min Scan# 2142
 Delta R.T. 0.000 min
 Lab File: BP017929.D
 Acq: 07 Nov 2023 17:43




Instrument :
 BNA_P
 ClientSampleId :
 SLCS694


Tgt Ion:198 Resp: 96381
 Ion Ratio Lower Upper
 198 100
 182 4.4 4.4 6.6
 77 35.2 24.6 37.0



#67
 N-Nitrosodiphenylamine
 Concen: 29.054 ng/ul
 RT: 15.798 min Scan# 2161
 Delta R.T. -0.006 min
 Lab File: BP017929.D
 Acq: 07 Nov 2023 17:43

Tgt Ion:169 Resp: 404450
 Ion Ratio Lower Upper
 169 100
 168 67.6 53.4 80.0
 167 36.6 28.0 42.0









#68
4-Bromophenyl-phenylether
Concen: 26.971 ng/ul
RT: 16.475 min Scan# 2276
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:248 Resp: 143648
Ion Ratio Lower Upper
248 100
250 95.2 77.9 116.9
141 85.9 62.4 93.6



#69
Hexachlorobenzene
Concen: 25.976 ng/ul
RT: 16.551 min Scan# 2289
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:284 Resp: 158795
Ion Ratio Lower Upper
284 100
142 47.3 29.0 43.6#
249 35.6 25.8 38.6





#70

Atrazine

Concen: 24.274 ng/ul

RT: 16.757 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:200 Resp: 129053

Ion Ratio Lower Upper

200 100

173 24.0 20.2 30.4

215 43.9 37.7 56.5



#71

Pentachlorophenol

Concen: 25.325 ng/ul

RT: 16.928 min Scan# 2353

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:266 Resp: 94038


Ion Ratio Lower Upper

266 100

264 62.1 53.8 80.6

268 64.1 52.4 78.6





#72

Phenanthrene

Concen: 29.617 ng/ul

RT: 17.340 min Scan# 2423

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:178 Resp: 787637

Ion Ratio Lower Upper

178 100

179 14.7 12.3 18.5

176 19.2 15.2 22.8



#74

Anthracene

Concen: 28.636 ng/ul

RT: 17.434 min Scan# 2439

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:178 Resp: 774099


Ion Ratio Lower Upper

178 100

179 14.7 12.5 18.7

176 18.4 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 26.117 ng/uL

RT: 13.304 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:216 Resp: 185124

Ion Ratio Lower Upper

216 100

214 80.8 62.3 93.5

218 49.0 39.8 59.6



#76

Pentachlorobenzene

Concen: 24.977 ng/uL

RT: 14.816 min Scan# 1994

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:250 Resp: 176669


Ion Ratio Lower Upper

250 100

248 64.1 48.5 72.7

252 63.0 50.2 75.4





#77

Carbazole

Concen: 29.447 ng/ul

RT: 17.728 min Scan# 2489

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:167 Resp: 709832

Ion Ratio Lower Upper

167 100

166 21.4 17.7 26.5

139 14.4 11.0 16.4



#78

Di-n-butylphthalate

Concen: 30.672 ng/ul

RT: 18.228 min Scan# 2574

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:149 Resp: 919978


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 6.6 5.0 7.6





#80

Fluoranthene

Concen: 29.519 ng/uI

RT: 19.316 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:202 Resp: 937568

Ion Ratio Lower Upper

202 100

101 8.8 7.1 10.7

100 7.0 5.3 7.9



#82

Pyrene

Concen: 29.763 ng/uI

RT: 19.675 min Scan# 2820

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:202 Resp: 985984


Ion Ratio Lower Upper

202 100

101 10.3 8.6 12.8

100 8.7 7.3 10.9





#83

Butylbenzylphthalate

Concen: 30.103 ng/ul

RT: 20.539 min Scan# 2967

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:149 Resp: 438753

Ion Ratio Lower Upper

149 100

91 74.7 60.4 90.6

206 20.4 15.3 22.9



#84

3,3'-Dichlorobenzidine

Concen: 28.474 ng/ul

RT: 21.339 min Scan# 3103

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:252 Resp: 305154


Ion Ratio Lower Upper

252 100

254 64.3 50.7 76.1

126 12.3 7.9 11.9#





#85

Benzo(a)anthracene

Concen: 29.401 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:228 Resp: 970725

Ion Ratio Lower Upper

228 100

229 19.4 15.4 23.2

226 27.7 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 29.924 ng/ul

RT: 21.275 min Scan# 3092

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:149 Resp: 631567


Ion Ratio Lower Upper

149 100

167 25.6 21.0 31.6

279 3.9 3.0 4.6





#87

Chrysene

Concen: 29.503 ng/ul

RT: 21.451 min Scan# 3122

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:228 Resp: 903269

Ion Ratio Lower Upper

228 100

226 30.4 23.9 35.9

229 19.4 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 28.770 ng/ul




RT: 22.227 min Scan# 3254


Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Tgt Ion:149 Resp: 1087421









#90
Benzo(b)fluoranthene
Concen: 29.836 ng/ul
RT: 23.122 min Scan# 3406
Delta R.T. -0.006 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43




Instrument :
BNA_P
ClientSampleId :
SLCS694


Tgt Ion:252 Resp: 921595
Ion Ratio Lower Upper
252 100
253 21.1 17.9 26.9
125 8.7 6.6 10.0



#91
Benzo(k)fluoranthene
Concen: 30.359 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. 0.000 min
Lab File: BP017929.D
Acq: 07 Nov 2023 17:43

Tgt Ion:252 Resp: 940378
Ion Ratio Lower Upper
252 100
253 21.7 17.2 25.8
125 8.5 7.0 10.6





#93

Benzo(a)pyrene

Concen: 30.169 ng/ul

RT: 23.780 min Scan# 3519

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:252 Resp: 866294

Ion Ratio Lower Upper

252 100

253 21.1 17.3 25.9

125 9.8 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.580 ng/ul

RT: 26.527 min Scan# 3985

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43




Tgt Ion:276 Resp: 1011642


Ion Ratio Lower Upper

276 100

138 21.8 15.9 23.9

277 24.2 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 29.627 ng/ul

RT: 26.545 min Scan# 3988

Delta R.T. -0.006 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Instrument :

BNA_P

ClientSampleId :

SLCS694





Tgt Ion:278 Resp: 842890

Ion Ratio Lower Upper

278 100

139 14.1 11.0 16.4

279 23.7 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 29.653 ng/ul

RT: 27.351 min Scan# 4125

Delta R.T. 0.000 min

Lab File: BP017929.D

Acq: 07 Nov 2023 17:43

Tgt Ion:276 Resp: 807395

Ion Ratio Lower Upper

276 100

138 18.7 14.7 22.1

277 23.6 19.2 28.8

