

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110723\
 Data File : BP017936.D
 Acq On : 07 Nov 2023 21:41
 Operator : MA/JU
 Sample : 05026-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH335

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/09/2023
 Supervised By :mohammad ahmed 11/11/2023

Quant Time: Nov 07 22:48:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:40:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.858	152	64142	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	270617	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.522	164	172282	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	387283	20.000	ng/u1	0.00
79) Chrysene-d12	21.410	240	376622	20.000	ng/u1	0.00
88) Perylene-d12	23.892	264	407033	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	6752	4.029	ng/uL	0.00
4) Pyridine-d5	3.687	84	51687	11.113	ng/u1	0.01
7) Phenol-d5	7.034	99	37248	6.193	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	122483	33.757	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	114822	24.843	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	72120	14.507	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	75905	35.099	ng/u1	0.00
24) 2-Nitrophenol-d4	9.769	143	82790	33.237	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	136642	29.923	ng/u1	0.00
31) 4-Chloroaniline-d4	10.852	131	174990	26.445	ng/u1	0.00
46) Dimethylphthalate-d6	13.940	166	539234	35.695	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	546263	32.954	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	7905m	3.403	ng/u1	0.04
60) Fluorene-d10	15.516	176	433158	34.258	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	80685	28.489	ng/u1	0.00
73) Anthracene-d10	17.392	188	739175	36.709	ng/u1	0.00
81) Pyrene-d10	19.645	212	944796	39.518	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.727	264	917571	40.345	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110723\
Data File : BP017936.D
Acq On : 07 Nov 2023 21:41
Operator : MA/JU
Sample : 05026-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH335

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/09/2023
Supervised By :mohammad ahmed 11/11/2023

Quant Time: Nov 07 22:48:15 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 07 03:40:56 2023
Response via : Initial Calibration

