

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110723\
 Data File : BP017941.D
 Acq On : 08 Nov 2023 00:30
 Operator : MA/JU
 Sample : 05026-11
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH361

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/09/2023
 Supervised By :mohammad ahmed 11/11/2023

Quant Time: Nov 08 01:07:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 07 03:40:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.858	152	73983	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	304792	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.516	164	192035	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	424348	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.410	240	386152	20.000	ng/u1	0.00
88) Perylene-d12	23.886	264	417609	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	7131	3.689	ng/uL	0.00
4) Pyridine-d5	3.687	84	30984	5.776	ng/u1	0.01
7) Phenol-d5	7.028	99	40840	5.887	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	134568	32.154	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	128304	24.067	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	83152	14.501	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	83690	34.359	ng/u1	0.00
24) 2-Nitrophenol-d4	9.769	143	95712	34.117	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	150258	29.216	ng/u1	0.00
31) 4-Chloroaniline-d4	10.852	131	180525	24.222	ng/u1	0.00
46) Dimethylphthalate-d6	13.940	166	465196	27.627	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	638431	34.553	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	10899m	4.210	ng/u1	0.04
60) Fluorene-d10	15.516	176	513615	36.443	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	94090	30.320	ng/u1	0.00
73) Anthracene-d10	17.392	188	871981	39.521	ng/u1	0.00
81) Pyrene-d10	19.639	212	1014391	41.382	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.727	264	938039	40.201	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110723\
Data File : BP017941.D
Acq On : 08 Nov 2023 00:30
Operator : MA/JU
Sample : 05026-11
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH361

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/09/2023
Supervised By :mohammad ahmed 11/11/2023

Quant Time: Nov 08 01:07:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110623.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 07 03:40:56 2023
Response via : Initial Calibration

