




Data Path : Z:\svoasrv\HPCHEM1\BNA P\Data\BP110819\
Data File : BP001006.D
Acq On : 08 Nov 2019 08:17
Operator : JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 08 18:10:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP110619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 06 17:07:13 2019
Response via : Initial Calibration



(70) Pentachlorophenol (C)




15.845min (+0.476) 13.09ng

response 143

Ion	Exp%	Act%
265.70	100	100
268.00	63.60	35.08#
264.00	63.30	26.25#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA P\Data\BP110819\
 Data File : BP001006.D
 Acq On : 08 Nov 2019 08:17
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 08 18:10:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP110619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 06 17:07:13 2019
 Response via : Initial Calibration



(77) Benzidine

17.886min (+0.294) 593.23ng

response 284

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	13.27
183.00	12.20	13.27
0.00	0.00	0.00