

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP110819\
 Data File : BP001007.D
 Acq On : 08 Nov 2019 08:47
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 11:20:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP110619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 06 17:07:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.478	0.439	8.2	76	-0.01
3	Pyridine	1.263	1.196	5.3	76	-0.01
4	n-Nitrosodimethylamine	0.440	0.438	0.5	83	-0.01
5 S	2-Fluorophenol	1.099	1.052	4.3	78	0.00
6	Aniline	1.815	1.731	4.6	78	0.00
7 S	Phenol-d6	1.389	1.347	3.0	79	0.00
8	2-Chlorophenol	1.165	1.134	2.7	80	0.00
9	Benzaldehyde	0.969	0.945	2.5	79	0.00
10 C	Phenol	1.520	1.512	0.5	83	0.00
11	bis(2-Chloroethyl)ether	1.183	1.130	4.5	79	0.00
12	1,3-Dichlorobenzene	1.462	1.426	2.5	80	0.00
13 C	1,4-Dichlorobenzene	1.472	1.438	2.3	80	0.00
14	1,2-Dichlorobenzene	1.362	1.316	3.4	79	0.00
15	Benzyl Alcohol	1.059	1.062	-0.3	82	0.00
16	2,2'-oxybis(1-Chloropropane	1.236	1.163	5.9	78	0.00
17	2-Methylphenol	0.989	0.960	2.9	80	0.00
18	Hexachloroethane	0.541	0.532	1.7	81	0.00
19 P	n-Nitroso-di-n-propylamine	0.834	0.800	4.1	79	0.00
20	3+4-Methylphenols	1.227	1.201	2.1	80	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
22	Acetophenone	0.517	0.502	2.9	82	0.00
23 S	Nitrobenzene-d5	0.362	0.354	2.2	82	0.00
24	Nitrobenzene	0.363	0.351	3.3	81	0.00
25	Isophorone	0.665	0.635	4.5	81	0.00
26 C	2-Nitrophenol	0.136	0.133	2.2	83	0.00
27	2,4-Dimethylphenol	0.254	0.238	6.3	79	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.408	6.2	79	0.00
29 C	2,4-Dichlorophenol	0.306	0.295	3.6	80	0.00
30	1,2,4-Trichlorobenzene	0.376	0.364	3.2	81	0.00
31	Naphthalene	0.999	0.965	3.4	81	0.00
32	Benzoic acid	0.137	0.142	-3.6	90	0.00
33	4-Chloroaniline	0.428	0.413	3.5	80	0.00
34 C	Hexachlorobutadiene	0.249	0.248	0.4	84	0.00
35	Caprolactam	0.097	0.095	2.1	82	0.00
36 C	4-Chloro-3-methylphenol	0.318	0.314	1.3	83	0.00
37	2-Methylnaphthalene	0.699	0.682	2.4	82	0.00
38	1-Methylnaphthalene	0.661	0.639	3.3	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	0.666	0.645	3.2	83	0.00
41 P	Hexachlorocyclopentadiene	0.334	0.322	3.6	84	0.00
42 S	2,4,6-Tribromophenol	0.239	0.243	-1.7	88	0.00
43 C	2,4,6-Trichlorophenol	0.402	0.397	1.2	86	0.00
44	2,4,5-Trichlorophenol	0.429	0.424	1.2	85	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP110819\
 Data File : BP001007.D
 Acq On : 08 Nov 2019 08:47
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 11:20:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP110619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 06 17:07:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.328	1.303	1.9	84	0.00
46	1,1'-Biphenyl	1.495	1.448	3.1	83	0.00
47	2-Chloronaphthalene	1.204	1.149	4.6	82	0.00
48	2-Nitroaniline	0.315	0.319	-1.3	87	0.00
49	Acenaphthylene	1.830	1.787	2.3	84	0.00
50	Dimethylphthalate	1.479	1.457	1.5	86	0.00
51	2,6-Dinitrotoluene	0.308	0.313	-1.6	87	0.00
52 C	Acenaphthene	1.095	1.040	5.0	82	0.00
53	3-Nitroaniline	0.336	0.331	1.5	85	0.00
54 P	2,4-Dinitrophenol	0.076	0.081	-6.6	93	0.00
55	Dibenzofuran	1.710	1.661	2.9	83	0.00
56 P	4-Nitrophenol	0.234	0.234	0.0	87	0.00
57	2,4-Dinitrotoluene	0.387	0.399	-3.1	88	0.00
58	Fluorene	1.361	1.328	2.4	84	0.00
59	2,3,4,6-Tetrachlorophenol	0.365	0.373	-2.2	87	0.00
60	Diethylphthalate	1.485	1.461	1.6	88	0.00
61	4-Chlorophenyl-phenylether	0.719	0.691	3.9	84	0.00
62	4-Nitroaniline	0.361	0.362	-0.3	86	0.00
63	Azobenzene	1.300	1.228	5.5	83	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	0.056	0.056	0.0	93	0.00
66 c	n-Nitrosodiphenylamine	0.575	0.548	4.7	84	0.00
67	4-Bromophenyl-phenylether	0.235	0.225	4.3	84	0.00
68	Hexachlorobenzene	0.273	0.262	4.0	86	0.00
69	Atrazine	0.209	0.199	4.8	84	0.00
70 C	Pentachlorophenol	0.121	0.125	-3.3	93	0.00
71	Phenanthrene	1.024	0.982	4.1	84	0.00
72	Anthracene	0.999	0.960	3.9	84	0.00
73	Carbazole	0.915	0.874	4.5	84	0.00
74	Di-n-butylphthalate	1.105	1.068	3.3	85	0.00
75 C	Fluoranthene	1.157	1.115	3.6	85	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
77	Benzidine	0.737	0.629	14.7	73	0.00
78	Pyrene	1.413	1.393	1.4	85	0.00
79 S	Terphenyl-d14	1.005	1.014	-0.9	87	0.00
80	Butylbenzylphthalate	0.560	0.548	2.1	87	0.00
81	Benzo(a)anthracene	1.347	1.305	3.1	84	0.00
82	3,3'-Dichlorobenzidine	0.491	0.486	1.0	85	0.00
83	Chrysene	1.183	1.176	0.6	85	0.00
84	Bis(2-ethylhexyl)phthalate	0.712	0.697	2.1	85	0.00
85 c	Di-n-octyl phthalate	1.281	1.239	3.3	85	0.00
86	Indeno(1,2,3-cd)pyrene	1.618	1.571	2.9	86	0.00
87 I	Perylene-d12	1.000	1.000	0.0	85	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP110819\
Data File : BP001007.D
Acq On : 08 Nov 2019 08:47
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 08 11:20:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP110619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 06 17:07:13 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.192	1.139	4.4	85	0.00
89	Benzo(k)fluoranthene	1.122	1.112	0.9	85	0.00
90 C	Benzo(a)pyrene	1.096	1.065	2.8	85	0.00
91	Dibenzo(a,h)anthracene	1.119	1.088	2.8	85	0.00
92	Benzo(g,h,i)perylene	1.127	1.081	4.1	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0