




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110821\
Data File : BP007861.D
Acq On : 08 Nov 2021 15:15
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 08 16:32:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 08 16:32:09 2021
Response via : Initial Calibration



TIC: BP007861.D\data.ms

(70) Pentachlorophenol (C)




16.916min (0.000) 820221.20 ng

response 299773

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.60	64.33
264.00	60.90	64.45
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110821\
 Data File : BP007861.D
 Acq On : 08 Nov 2021 15:15
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 08 16:32:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 08 16:32:09 2021
 Response via : Initial Calibration



TIC: BP007861.D\data.ms

(77) Benzidine

19.451min (0.000) 231090.91 ng

response 1921974

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.09
183.00	11.40	12.02
0.00	0.00	0.00