

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110821\
 Data File : BP007862.D
 Acq On : 08 Nov 2021 15:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 16:37:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 08 16:36:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	79	0.00
2	1,4-Dioxane	40.000	41.126	-2.8	82	0.00
3	Pyridine	40.000	42.739	-6.8	91	0.00
4	n-Nitrosodimethylamine	40.000	38.533	3.7	80	0.00
5 S	2-Fluorophenol	80.000	79.844	0.2	82	0.00
6	Aniline	40.000	40.219	-0.5	81	0.00
7 S	Phenol-d6	80.000	79.896	0.1	80	0.00
8	2-Chlorophenol	40.000	39.892	0.3	81	0.00
9	Benzaldehyde	40.000	36.937	7.7	78	0.00
10 C	Phenol	40.000	40.210	-0.5	81	0.00
11	bis(2-Chloroethyl)ether	40.000	38.804	3.0	79	0.00
12	1,3-Dichlorobenzene	40.000	39.186	2.0	80	0.00
13 C	1,4-Dichlorobenzene	40.000	39.124	2.2	81	0.00
14	1,2-Dichlorobenzene	40.000	37.294	6.8	76	0.00
15	Benzyl Alcohol	40.000	39.563	1.1	77	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.793	8.0	75	0.00
17	2-Methylphenol	40.000	37.975	5.1	76	0.00
18	Hexachloroethane	40.000	38.148	4.6	77	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.122	7.2	73	0.00
20	3+4-Methylphenols	40.000	37.778	5.6	74	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	72	0.00
22	Acetophenone	40.000	40.477	-1.2	75	0.00
23 S	Nitrobenzene-d5	80.000	81.674	-2.1	74	0.00
24	Nitrobenzene	40.000	40.579	-1.4	75	0.00
25	Isophorone	40.000	40.538	-1.3	73	0.00
26 C	2-Nitrophenol	40.000	41.284	-3.2	73	0.00
27	2,4-Dimethylphenol	40.000	40.290	-0.7	73	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.686	3.3	71	0.00
29 C	2,4-Dichlorophenol	40.000	39.874	0.3	71	0.00
30	1,2,4-Trichlorobenzene	40.000	39.035	2.4	72	0.00
31	Naphthalene	40.000	39.830	0.4	74	0.00
32	Benzoic acid	40.000	42.407	-6.0	69	0.00
33	4-Chloroaniline	40.000	40.101	-0.3	73	0.00
34 C	Hexachlorobutadiene	40.000	39.331	1.7	72	0.00
35	Caprolactam	40.000	42.081	-5.2	75	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.975	0.1	71	0.00
37	2-Methylnaphthalene	40.000	39.426	1.4	72	0.00
38	1-Methylnaphthalene	40.000	39.351	1.6	72	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	72	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.809	3.0	71	0.00
41 P	Hexachlorocyclopentadiene	40.000	32.792	18.0	56	0.00
42 S	2,4,6-Tribromophenol	80.000	84.482	-5.6	78	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.871	0.3	71	0.00
44	2,4,5-Trichlorophenol	40.000	40.051	-0.1	72	0.00
45 S	2-Fluorobiphenyl	80.000	78.460	1.9	72	0.00
46	1,1'-Biphenyl	40.000	39.555	1.1	73	0.00
47	2-Chloronaphthalene	40.000	39.937	0.2	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110821\
 Data File : BP007862.D
 Acq On : 08 Nov 2021 15:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 16:37:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 08 16:36:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.637	-6.6	75	0.00
49	Acenaphthylene	40.000	40.335	-0.8	73	0.00
50	Dimethylphthalate	40.000	39.377	1.6	72	0.00
51	2,6-Dinitrotoluene	40.000	40.427	-1.1	72	0.00
52 C	Acenaphthene	40.000	39.948	0.1	73	0.00
53	3-Nitroaniline	40.000	41.584	-4.0	74	0.00
54 P	2,4-Dinitrophenol	40.000	40.126	-0.3	68	0.00
55	Dibenzofuran	40.000	39.866	0.3	73	0.00
56 P	4-Nitrophenol	40.000	42.397	-6.0	75	0.00
57	2,4-Dinitrotoluene	40.000	41.843	-4.6	73	0.00
58	Fluorene	40.000	41.200	-3.0	74	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.555	-1.4	72	0.00
60	Diethylphthalate	40.000	41.138	-2.8	74	0.00
61	4-Chlorophenyl-phenylether	40.000	40.234	-0.6	73	0.00
62	4-Nitroaniline	40.000	44.053	-10.1	77	0.00
63	Azobenzene	40.000	42.194	-5.5	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	78	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.407	-1.0	71	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.562	1.1	78	0.00
67	4-Bromophenyl-phenylether	40.000	39.368	1.6	77	0.00
68	Hexachlorobenzene	40.000	39.204	2.0	77	0.00
69	Atrazine	40.000	40.287	-0.7	78	0.00
70 C	Pentachlorophenol	40.000	39.670	0.8	73	0.00
71	Phenanthrene	40.000	39.651	0.9	78	0.00
72	Anthracene	40.000	40.396	-1.0	80	0.00
73	Carbazole	40.000	40.518	-1.3	81	0.00
74	Di-n-butylphthalate	40.000	42.103	-5.3	83	0.00
75 C	Fluoranthene	40.000	40.917	-2.3	82	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77	Benzydine	40.000	42.244	-5.6	83	0.00
78	Pyrene	40.000	37.518	6.2	82	0.00
79 S	Terphenyl-d14	80.000	75.097	6.1	83	0.00
80	Butylbenzylphthalate	40.000	40.713	-1.8	88	0.00
81	Benzo(a)anthracene	40.000	38.885	2.8	86	0.00
82	3,3'-Dichlorobenzidine	40.000	40.966	-2.4	89	0.00
83	Chrysene	40.000	39.132	2.2	87	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.711	-1.8	89	0.00
85 c	Di-n-octyl phthalate	40.000	42.651	-6.6	92	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.300	-0.7	95	0.00
88	Benzo(b)fluoranthene	40.000	38.059	4.9	89	0.00
89	Benzo(k)fluoranthene	40.000	39.047	2.4	92	0.00
90 C	Benzo(a)pyrene	40.000	39.256	1.9	92	0.00
91	Dibenzo(a,h)anthracene	40.000	40.065	-0.2	95	0.00
92	Benzo(g,h,i)perylene	40.000	40.736	-1.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110821\
Data File : BP007862.D
Acq On : 08 Nov 2021 15:49
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 08 16:37:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 08 16:36:40 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0