

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110923\
 Data File : BP017989.D
 Acq On : 09 Nov 2023 15:35
 Operator : MA/JU
 Sample : SSTD02048
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612

Quant Time: Nov 09 21:50:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 09 21:43:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	59456	20.000	ng/u1	0.00
20) Naphthalene-d8	10.687	136	243978	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.540	164	158927	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	366043	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	372835	20.000	ng/u1	0.00
88) Perylene-d12	23.968	264	414045	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	12340	7.105	ng/uL	0.00
4) Pyridine-d5	3.681	84	83159	18.422	ng/u1	0.00
7) Phenol-d5	7.034	99	98277	17.589	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	61890	18.286	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	79989	17.892	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	78747	17.572	ng/u1	0.00
21) Nitrobenzene-d5	9.075	128	38193	17.481	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	42947	17.662	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	79036	18.144	ng/u1	0.00
31) 4-Chloroaniline-d4	10.869	131	109786	18.391	ng/u1	0.00
46) Dimethylphthalate-d6	13.963	166	265999	18.207	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	287077	18.329	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	34317	16.038	ng/u1	0.00
60) Fluorene-d10	15.545	176	221034	18.253	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	44116	17.260	ng/u1	0.00
73) Anthracene-d10	17.422	188	357549	18.230	ng/u1	0.00
81) Pyrene-d10	19.675	212	442499	18.441	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.798	264	436939	18.156	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.281	88	13400	6.938	ng/uL	100
5) Pyridine	3.699	79	85195	18.385	ng/u1	100
6) Benzaldehyde	7.034	77	65221	21.207	ng/u1	100
8) Phenol	7.063	94	97204	17.667	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.311	93	78994	18.191	ng/u1	100
12) 2-Chlorophenol	7.422	128	81113	17.837	ng/u1	100
13) 2-Methylphenol	8.322	108	73442	17.810	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.381	45	46877	9.316	ng/u1	100
16) Acetophenone	8.722	105	130693	18.238	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.687	70	69654	17.747	ng/u1	100
18) 4-Methylphenol	8.652	108	79422	17.761	ng/u1	100
19) Hexachloroethane	8.922	117	38885	17.887	ng/u1	100
22) Nitrobenzene	9.116	77	108491	18.250	ng/u1	100
23) Isophorone	9.622	82	190032	18.055	ng/u1	100
25) 2-Nitrophenol	9.816	139	45923	18.366	ng/u1	100
26) 2,4-Dimethylphenol	9.857	107	95100	17.793	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.116	93	107382	18.071	ng/u1	100
29) 2,4-Dichlorophenol	10.340	162	74459	17.826	ng/u1	100
30) Naphthalene	10.740	128	265627	17.956	ng/u1	100
32) 4-Chloroaniline	10.893	127	104909	18.369	ng/u1	100
33) Hexachlorobutadiene	10.957	225	56760	18.290	ng/u1	100
34) Caprolactam	11.740	113	23333	17.505	ng/u1	100
35) 4-Chloro-3-methylphenol	12.004	107	91812	18.460	ng/u1	100
36) 2-Methylnaphthalene	12.351	142	181855	17.985	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110923\
 Data File : BP017989.D
 Acq On : 09 Nov 2023 15:35
 Operator : MA/JU
 Sample : SSTD02048
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612

Quant Time: Nov 09 21:50:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110923.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 09 21:43:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	185344	17.993	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.704	216	98494	18.277	ng/ul	100
40) Hexachlorocyclopentadiene	12.640	237	39722	15.417	ng/ul	100
41) 2,4,6-Trichlorophenol	12.969	196	64197	18.322	ng/ul	100
42) 2,4,5-Trichlorophenol	13.051	196	67916	18.363	ng/ul	100
43) 1,1'-Biphenyl	13.369	154	245898	18.197	ng/ul	100
44) 2-Chloronaphthalene	13.422	162	195265	18.266	ng/ul	100
45) 2-Nitroaniline	13.675	65	62994	18.628	ng/ul	100
47) Dimethylphthalate	14.010	163	267185	18.128	ng/ul	100
48) 2,6-Dinitrotoluene	14.163	165	50439	18.225	ng/ul	100
50) Acenaphthylene	14.269	152	328325	18.304	ng/ul	100
51) 3-Nitroaniline	14.510	138	47947	18.624	ng/ul	100
52) Acenaphthene	14.610	153	216095	18.130	ng/ul	100
53) 2,4-Dinitrophenol	14.728	184	23385	15.814	ng/ul	100
55) 4-Nitrophenol	14.810	109	52999	18.283	ng/ul	100
56) Dibenzofuran	14.945	168	299942	18.230	ng/ul	100
57) 2,4-Dinitrotoluene	14.957	165	77964	18.418	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.169	232	60065	18.314	ng/ul	100
59) Diethylphthalate	15.369	149	347015	14.630	ng/ul	100
61) Fluorene	15.598	166	250271	18.036	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.592	204	121032	17.769	ng/ul	100
63) 4-Nitroaniline	15.687	138	46154	18.939	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.710	198	48317	18.045	ng/ul	100
67) N-Nitrosodiphenylamine	15.822	169	212328	18.111	ng/ul	100
68) 4-Bromophenyl-phenylether	16.498	248	74060	18.531	ng/ul	100
69) Hexachlorobenzene	16.581	284	80717	17.816	ng/ul	100
70) Atrazine	16.781	200	76989	18.977	ng/ul	100
71) Pentachlorophenol	16.957	266	47303	16.999	ng/ul	100
72) Phenanthrene	17.369	178	411960	18.081	ng/ul	100
74) Anthracene	17.457	178	416756	18.092	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.322	216	99230	18.068	ng/uL	100
76) Pentachlorobenzene	14.840	250	96968	17.968	ng/uL	100
77) Carbazole	17.757	167	375125	18.859	ng/ul	100
78) Di-n-butylphthalate	18.263	149	477926	17.865	ng/ul	100
80) Fluoranthene	19.345	202	504007	18.148	ng/ul	100
82) Pyrene	19.704	202	529294	18.032	ng/ul	100
83) Butylbenzylphthalate	20.575	149	230174	18.108	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.380	252	168363	19.057	ng/ul	100
85) Benzo(a)anthracene	21.433	228	546499	18.233	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.310	149	319858	17.946	ng/ul	100
87) Chrysene	21.492	228	508495	18.011	ng/ul	100
89) Di-n-octyl phthalate	22.274	149	554604	17.647	ng/ul	100
90) Benzo(b)fluoranthene	23.180	252	547768	18.408	ng/ul	100
91) Benzo(k)fluoranthene	23.233	252	533810	17.973	ng/ul	100
93) Benzo(a)pyrene	23.851	252	509343	18.159	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.633	276	596279	17.936	ng/ul	100
95) Dibenzo(a,h)anthracene	26.657	278	490433	17.843	ng/ul	100
96) Benzo(g,h,i)perylene	27.468	276	479925	18.090	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 6.938 ng/uL
 RT: 3.281 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BP017989.D
 Acq: 09 Nov 2023 15:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612


Tgt Ion:	88	Resp:	13400
Ion	Ratio	Lower	Upper
88	100		
43	32.9	26.3	39.5
58	67.5	54.0	81.0



#5
 Pyridine
 Concen: 18.385 ng/uL
 RT: 3.699 min Scan# 104
 Delta R.T. 0.000 min
 Lab File: BP017989.D
 Acq: 09 Nov 2023 15:35

Tgt Ion:	79	Resp:	85195
Ion	Ratio	Lower	Upper
79	100		
52	71.0	56.8	85.2
51	38.6	30.9	46.3









#6
Benzaldehyde
Concen: 21.207 ng/ul
RT: 7.034 min Scan# 671
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion: 77 Resp: 65221
Ion Ratio Lower Upper
77 100
105 94.7 75.8 113.6
106 87.0 69.6 104.4



#8
Phenol
Concen: 17.667 ng/ul
RT: 7.063 min Scan# 676
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion: 94 Resp: 97204
Ion Ratio Lower Upper
94 100
65 33.1 26.5 39.7
66 46.3 37.0 55.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.191 ng/ul
RT: 7.311 min Scan# 718
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion: 93 Resp: 78994
Ion Ratio Lower Upper
93 100
63 72.9 58.3 87.5
95 32.4 25.9 38.9



#12
2-Chlorophenol
Concen: 17.837 ng/ul
RT: 7.422 min Scan# 737
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:128 Resp: 81113
Ion Ratio Lower Upper
128 100
64 48.6 38.9 58.3
130 32.4 25.9 38.9









#13
2-Methylphenol
Concen: 17.810 ng/ul
RT: 8.322 min Scan# 890
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:108 Resp: 73442
Ion Ratio Lower Upper
108 100
107 92.5 74.0 111.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 9.316 ng/ul
RT: 8.381 min Scan# 900
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion: 45 Resp: 46877
Ion Ratio Lower Upper
45 100
77 20.7 16.6 24.8
79 13.3 10.6 16.0









#16
Acetophenone
Concen: 18.238 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:105 Resp: 130693
Ion Ratio Lower Upper
105 100
77 85.3 68.2 102.4
51 30.1 24.1 36.1



#17
N-Nitroso-di-n-propylamine
Concen: 17.747 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion: 70 Resp: 69654
Ion Ratio Lower Upper
70 100
42 58.4 46.7 70.1
101 9.6 7.7 11.5
130 20.1 16.1 24.1









#18
4-Methylphenol
Concen: 17.761 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:108 Resp: 79422
Ion Ratio Lower Upper
108 100
107 116.4 93.1 139.7



#19
Hexachloroethane
Concen: 17.887 ng/ul
RT: 8.922 min Scan# 992
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:117 Resp: 38885
Ion Ratio Lower Upper
117 100
201 95.0 76.0 114.0
199 59.5 47.6 71.4









#22
Nitrobenzene
Concen: 18.250 ng/ul
RT: 9.116 min Scan# 1025
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion: 77 Resp: 108491
Ion Ratio Lower Upper
77 100
123 36.8 29.4 44.2
65 14.3 11.4 17.2



#23
Isophorone
Concen: 18.055 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion: 82 Resp: 190032
Ion Ratio Lower Upper
82 100
95 9.4 7.5 11.3
138 15.6 12.5 18.7









#25
2-Nitrophenol
Concen: 18.366 ng/ul
RT: 9.816 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:139 Resp: 45923
Ion Ratio Lower Upper
139 100
65 64.5 51.6 77.4
109 41.6 33.3 49.9



#26
2,4-Dimethylphenol
Concen: 17.793 ng/ul
RT: 9.857 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:107 Resp: 95100
Ion Ratio Lower Upper
107 100
121 47.2 37.8 56.6
122 77.7 62.2 93.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.071 ng/ul
 RT: 10.116 min Scan# 1195
 Delta R.T. 0.000 min
 Lab File: BP017989.D
 Acq: 09 Nov 2023 15:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612


Tgt Ion: 93 Resp: 107382
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.4 38.2
 123 10.6 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 17.826 ng/ul
 RT: 10.340 min Scan# 1233
 Delta R.T. 0.000 min
 Lab File: BP017989.D
 Acq: 09 Nov 2023 15:35

Tgt Ion: 162 Resp: 74459
 Ion Ratio Lower Upper
 162 100
 164 63.8 51.0 76.6
 98 40.7 32.6 48.8





#30

Naphthalene

Concen: 17.956 ng/ul

RT: 10.740 min Scan# 1301

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNP_P

ClientSampleId :

SSTD020612





Tgt Ion:128 Resp: 265627

Ion Ratio Lower Upper

128 100

129 10.7 8.6 12.8

127 13.1 10.5 15.7



#32

4-Chloroaniline

Concen: 18.369 ng/ul

RT: 10.893 min Scan# 1327

Delta R.T. 0.000 min

Lab File: BP017989.D




Acq: 09 Nov 2023 15:35


Tgt Ion:127 Resp: 104909

Ion Ratio Lower Upper

127 100

129 31.0 24.8 37.2









#33
Hexachlorobutadiene
Concen: 18.290 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:225 Resp: 56760
Ion Ratio Lower Upper
225 100
223 63.2 50.6 75.8
227 59.9 47.9 71.9



#34
Caprolactam
Concen: 17.505 ng/ul
RT: 11.740 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:113 Resp: 23333
Ion Ratio Lower Upper
113 100
55 131.6 105.3 157.9
56 117.7 94.2 141.2









#35
4-Chloro-3-methylphenol
Concen: 18.460 ng/ul
RT: 12.004 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:107 Resp: 91812
Ion Ratio Lower Upper
107 100
144 25.0 20.0 30.0
142 76.6 61.3 91.9



#36
2-Methylnaphthalene
Concen: 17.985 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:142 Resp: 181855
Ion Ratio Lower Upper
142 100
141 85.0 68.0 102.0





#37

1-Methylnaphthalene

Concen: 17.993 ng/ul

RT: 12.575 min Scan# 1613

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:142 Resp: 185344

Ion Ratio Lower Upper

142 100

141 97.4 77.9 116.9

116 4.9 3.9 5.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.277 ng/ul

RT: 12.704 min Scan# 1635

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Tgt Ion:216 Resp: 98494




Ion Ratio Lower Upper


216 100

214 76.5 61.2 91.8

179 23.3 18.6 28.0

108 21.1 16.9 25.3





#40

Hexachlorocyclopentadiene

Concen: 15.417 ng/ul

RT: 12.640 min Scan# 1624

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:237 Resp: 39722

Ion Ratio Lower Upper

237 100

235 61.8 49.4 74.2

272 10.5 8.4 12.6



#41

2,4,6-Trichlorophenol

Concen: 18.322 ng/ul

RT: 12.969 min Scan# 1680

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:196 Resp: 64197


Ion Ratio Lower Upper

196 100

198 91.9 73.5 110.3

200 31.6 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 18.363 ng/ul

RT: 13.051 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:196 Resp: 67916

Ion Ratio Lower Upper

196 100

198 93.0 74.4 111.6

200 29.8 23.8 35.8



#43

1,1'-Biphenyl

Concen: 18.197 ng/ul

RT: 13.369 min Scan# 1748

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:154 Resp: 245898


Ion Ratio Lower Upper

154 100

153 41.8 33.4 50.2

76 12.3 9.8 14.8









#44
2-Chloronaphthalene
Concen: 18.266 ng/ul
RT: 13.422 min Scan# 1757
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:162 Resp: 195265
Ion Ratio Lower Upper
162 100
164 32.8 26.2 39.4
127 43.4 34.7 52.1



#45
2-Nitroaniline
Concen: 18.628 ng/ul
RT: 13.675 min Scan# 1800
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion: 65 Resp: 62994
Ion Ratio Lower Upper
65 100
92 68.8 55.0 82.6
138 98.6 78.9 118.3









#47
Dimethylphthalate
Concen: 18.128 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:163 Resp: 267185
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.7 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 18.225 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:165 Resp: 50439
Ion Ratio Lower Upper
165 100
89 79.0 63.2 94.8
121 24.9 19.9 29.9





#50

Acenaphthylene

Concen: 18.304 ng/ul

RT: 14.269 min Scan# 1901

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:152 Resp: 328325

Ion Ratio Lower Upper

152 100

151 19.4 15.5 23.3

153 12.9 10.3 15.5



#51

3-Nitroaniline

Concen: 18.624 ng/ul

RT: 14.510 min Scan# 1942

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:138 Resp: 47947


Ion Ratio Lower Upper

138 100

108 10.8 8.6 13.0

92 129.4 103.5 155.3





#52

Acenaphthene

Concen: 18.130 ng/ul

RT: 14.610 min Scan# 1959

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNP_P

ClientSampleId :

SSTD020612





Tgt Ion:153 Resp: 216095

Ion Ratio Lower Upper

153 100

152 47.7 38.2 57.2

154 86.0 68.8 103.2



#53

2,4-Dinitrophenol

Concen: 15.814 ng/ul

RT: 14.728 min Scan# 1979

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:184 Resp: 23385


Ion Ratio Lower Upper

184 100

63 85.0 68.0 102.0

154 80.9 64.7 97.1









#55
4-Nitrophenol
Concen: 18.283 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:109 Resp: 52999
Ion Ratio Lower Upper
109 100
139 93.5 74.8 112.2
65 115.7 92.6 138.8



#56
Dibenzofuran
Concen: 18.230 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:168 Resp: 299942
Ion Ratio Lower Upper
168 100
139 43.9 35.1 52.7









#57
2,4-Dinitrotoluene
Concen: 18.418 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:165 Resp: 77964
Ion Ratio Lower Upper
165 100
63 66.5 53.2 79.8
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.314 ng/ul
RT: 15.169 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:232 Resp: 60065
Ion Ratio Lower Upper
232 100
131 55.1 44.1 66.1
130 3.6 2.9 4.3
166 36.1 28.9 43.3





#59

Diethylphthalate

Concen: 14.630 ng/ul

RT: 15.369 min Scan# 2088

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:149 Resp: 347015

Ion Ratio Lower Upper

149 100

177 20.9 16.7 25.1

150 12.4 9.9 14.9



#61

Fluorene

Concen: 18.036 ng/ul

RT: 15.598 min Scan# 2127

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:166 Resp: 250271


Ion Ratio Lower Upper

166 100

165 100.2 80.2 120.2

167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 17.769 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:204 Resp: 121032
Ion Ratio Lower Upper
204 100
206 34.8 27.8 41.8
141 73.0 58.4 87.6



#63
4-Nitroaniline
Concen: 18.939 ng/ul
RT: 15.687 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:138 Resp: 46154
Ion Ratio Lower Upper
138 100
92 61.7 49.4 74.0
108 140.3 112.2 168.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.045 ng/ul
 RT: 15.710 min Scan# 2146
 Delta R.T. 0.000 min
 Lab File: BP017989.D
 Acq: 09 Nov 2023 15:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612


Tgt Ion:198 Resp: 48317
 Ion Ratio Lower Upper
 198 100
 182 5.8 4.6 7.0
 77 34.7 27.8 41.6



#67
 N-Nitrosodiphenylamine
 Concen: 18.111 ng/ul
 RT: 15.822 min Scan# 2165
 Delta R.T. 0.000 min
 Lab File: BP017989.D
 Acq: 09 Nov 2023 15:35

Tgt Ion:169 Resp: 212328
 Ion Ratio Lower Upper
 169 100
 168 67.4 53.9 80.9
 167 36.7 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 18.531 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:	248	Resp:	74060
Ion Ratio	Lower	Upper	
248	100		
250	96.5	77.2	115.8
141	94.5	75.6	113.4



#69
Hexachlorobenzene
Concen: 17.816 ng/ul
RT: 16.581 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:	284	Resp:	80717
Ion Ratio	Lower	Upper	
284	100		
142	45.8	36.6	55.0
249	35.8	28.6	43.0





#70

Atrazine

Concen: 18.977 ng/ul

RT: 16.781 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:200 Resp: 76989

Ion Ratio Lower Upper

200 100

173 25.3 20.2 30.4

215 48.3 38.6 58.0



#71

Pentachlorophenol

Concen: 16.999 ng/ul

RT: 16.957 min Scan# 2358

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:266 Resp: 47303


Ion Ratio Lower Upper

266 100

264 60.5 48.4 72.6

268 62.9 50.3 75.5





#72

Phenanthrene

Concen: 18.081 ng/ul

RT: 17.369 min Scan# 2428

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:178 Resp: 411960

Ion Ratio Lower Upper

178 100

179 14.5 11.6 17.4

176 18.9 15.1 22.7



#74

Anthracene

Concen: 18.092 ng/ul

RT: 17.457 min Scan# 2443

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:178 Resp: 416756


Ion Ratio Lower Upper

178 100

179 16.1 12.9 19.3

176 18.9 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.068 ng/uL
RT: 13.322 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:216 Resp: 99230
Ion Ratio Lower Upper
216 100
214 80.0 64.0 96.0
218 45.9 36.7 55.1



#76
Pentachlorobenzene
Concen: 17.968 ng/uL
RT: 14.840 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BP017989.D
Acq: 09 Nov 2023 15:35

Tgt Ion:250 Resp: 96968
Ion Ratio Lower Upper
250 100
248 60.4 48.3 72.5
252 65.2 52.2 78.2





#77

Carbazole

Concen: 18.859 ng/ul

RT: 17.757 min Scan# 2494

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:167 Resp: 375125

Ion Ratio Lower Upper

167 100

166 21.5 17.2 25.8

139 14.6 11.7 17.5



#78

Di-n-butylphthalate

Concen: 17.865 ng/ul

RT: 18.263 min Scan# 2580

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:149 Resp: 477926


Ion Ratio Lower Upper

149 100

150 8.8 7.0 10.6

104 6.4 5.1 7.7





#80

Fluoranthene

Concen: 18.148 ng/ul

RT: 19.345 min Scan# 2764

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:202 Resp: 504007

Ion Ratio Lower Upper

202 100

101 9.2 7.4 11.0

100 7.2 5.8 8.6



#82

Pyrene

Concen: 18.032 ng/ul

RT: 19.704 min Scan# 2825

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:202 Resp: 529294


Ion Ratio Lower Upper

202 100

101 11.2 9.0 13.4

100 9.8 7.8 11.8





#83

Butylbenzylphthalate

Concen: 18.108 ng/ul

RT: 20.575 min Scan# 2973

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:149 Resp: 230174

Ion Ratio Lower Upper

149 100

91 74.4 59.5 89.3

206 19.4 15.5 23.3



#84

3,3'-Dichlorobenzidine

Concen: 19.057 ng/ul

RT: 21.380 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:252 Resp: 168363


Ion Ratio Lower Upper

252 100

254 63.7 51.0 76.4

126 11.5 9.2 13.8





#85

Benzo(a)anthracene

Concen: 18.233 ng/ul

RT: 21.433 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:228 Resp: 546499

Ion Ratio Lower Upper

228 100

229 19.5 15.6 23.4

226 27.9 22.3 33.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.946 ng/ul

RT: 21.310 min Scan# 3098

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:149 Resp: 319858


Ion Ratio Lower Upper

149 100

167 25.2 20.2 30.2

279 3.7 3.0 4.4





#87

Chrysene

Concen: 18.011 ng/ul

RT: 21.492 min Scan# 3129

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:228 Resp: 508495

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.2 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 17.647 ng/ul




RT: 22.274 min Scan# 3262


Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Tgt Ion:149 Resp: 554604





#90

Benzo(b)fluoranthene

Concen: 18.408 ng/ul

RT: 23.180 min Scan# 3416

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:252 Resp: 547768

Ion Ratio Lower Upper

252 100

253 21.2 17.0 25.4

125 9.3 7.4 11.2



#91

Benzo(k)fluoranthene

Concen: 17.973 ng/ul

RT: 23.233 min Scan# 3425

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:252 Resp: 533810


Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.6

125 9.3 7.4 11.2





#93

Benzo(a)pyrene

Concen: 18.159 ng/ul

RT: 23.851 min Scan# 3530

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:252 Resp: 509343

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 9.8 7.8 11.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.936 ng/ul

RT: 26.633 min Scan# 4003

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35




Tgt Ion:276 Resp: 596279


Ion Ratio Lower Upper

276 100

138 20.3 16.2 24.4

277 23.5 18.8 28.2





#95

Dibenzo(a,h)anthracene

Concen: 17.843 ng/ul

RT: 26.657 min Scan# 4007

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:278 Resp: 490433

Ion Ratio Lower Upper

278 100

139 14.9 11.9 17.9

279 24.6 19.7 29.5



#96

Benzo(g,h,i)perylene

Concen: 18.090 ng/ul

RT: 27.468 min Scan# 4145

Delta R.T. 0.000 min

Lab File: BP017989.D

Acq: 09 Nov 2023 15:35

Tgt Ion:276 Resp: 479925

Ion Ratio Lower Upper

276 100

138 18.8 15.0 22.6

277 23.8 19.0 28.6

