

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP022940.D
 Acq On : 08 Nov 2024 18:53
 Operator : RC/JU
 Sample : SST00549
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005628

Quant Time: Nov 08 23:16:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 21:50:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.592	152	42297	20.000	ng/ul	0.00
20) Naphthalene-d8	10.345	136	158731	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.216	164	137746	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.027	188	357500	20.000	ng/ul	0.00
79) Chrysene-d12	21.462	240	445026	20.000	ng/ul	0.00
88) Perylene-d12	24.692	264	586620	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	1635	1.502	ng/uL	0.00
4) Pyridine-d5	3.569	84	9452	3.163	ng/ul	0.01
7) Phenol-d5	6.804	99	12192	3.424	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	6.945	67	8296	3.964	ng/ul	0.00
11) 2-Chlorophenol-d4	7.145	132	9948	3.938	ng/ul	0.00
15) 4-Methylphenol-d8	8.310	113	10396	3.659	ng/ul	0.00
21) Nitrobenzene-d5	8.739	128	4811	3.942	ng/ul	0.00
24) 2-Nitrophenol-d4	9.451	143	5654	4.001	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.004	165	11584	3.857	ng/ul	0.01
31) 4-Chloroaniline-d4	10.504	131	13863	3.907	ng/ul	0.00
46) Dimethylphthalate-d6	13.616	166	45619	4.168	ng/ul	0.00
49) Acenaphthylene-d8	13.898	160	45648	3.927	ng/ul	0.00
54) 4-Nitrophenol-d4	14.457	143	6	0.003	ng/ul	-0.01
60) Fluorene-d10	15.233	176	40604	4.277	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.380	200	7328	3.117	ng/ul	0.01
73) Anthracene-d10	17.121	188	69122	4.110	ng/ul	0.00
81) Pyrene-d10	19.515	212	89715	3.812	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.468	264	123012	3.927	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	1921	1.641	ng/uL#	74
5) Pyridine	3.587	79	10845	3.540	ng/ul	90
6) Benzaldehyde	6.775	77	9653	5.017	ng/ul	92
8) Phenol	6.834	94	13096	3.535	ng/ul#	88
10) Bis(2-Chloroethyl)ether	7.034	93	10459	3.773	ng/ul	97
12) 2-Chlorophenol	7.181	128	10716	4.102	ng/ul	94
13) 2-Methylphenol	8.045	108	9674	3.599	ng/ul	84
14) 2,2'-oxybis(1-Chloropr...	8.104	45	12688	3.938	ng/ul	95
16) Acetophenone	8.410	105	18501	3.944	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.387	70	9234	3.885	ng/ul#	90
18) 4-Methylphenol	8.369	108	10773	3.612	ng/ul	100
19) Hexachloroethane	8.645	117	5446	4.476	ng/ul	86
22) Nitrobenzene	8.787	77	14672	4.027	ng/ul	94
23) Isophorone	9.304	82	25519	3.907	ng/ul	99
25) 2-Nitrophenol	9.481	139	6135	4.029	ng/ul#	86
26) 2,4-Dimethylphenol	9.557	107	14237	4.301	ng/ul	96
27) Bis(2-Chloroethoxy)met...	9.769	93	13920	3.753	ng/ul#	96
29) 2,4-Dichlorophenol	10.034	162	12352	4.177	ng/ul	98
30) Naphthalene	10.392	128	36731	4.345	ng/ul	98
32) 4-Chloroaniline	10.522	127	12956	3.699	ng/ul	91
33) Hexachlorobutadiene	10.669	225	12892	5.172	ng/ul	95
34) Caprolactam	11.316	113	2459	2.579	ng/ul#	73
35) 4-Chloro-3-methylphenol	11.739	107	326	0.101	ng/ul	89
36) 2-Methylnaphthalene	12.016	142	27601	4.447	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP022940.D
 Acq On : 08 Nov 2024 18:53
 Operator : RC/JU
 Sample : SST00549
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005628

Quant Time: Nov 08 23:16:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 21:50:22 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.227	142	27984	4.418	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.386	216	21898	4.299	ng/ul#	91
40) Hexachlorocyclopentadiene	12.422	237	9	0.003	ng/ul	92
41) 2,4,6-Trichlorophenol	12.639	196	12521	3.908	ng/ul	88
42) 2,4,5-Trichlorophenol	12.739	196	12825	3.760	ng/ul	87
43) 1,1'-Biphenyl	13.033	154	40663	4.096	ng/ul	99
44) 2-Chloronaphthalene	13.080	162	33037	4.065	ng/ul	98
45) 2-Nitroaniline	13.304	65	8017	3.303	ng/ul	99
47) Dimethylphthalate	13.663	163	44517	4.082	ng/ul#	99
48) 2,6-Dinitrotoluene	13.816	165	8034	3.895	ng/ul	98
50) Acenaphthylene	13.927	152	45872	3.802	ng/ul	97
51) 3-Nitroaniline	14.163	138	6824	3.522	ng/ul	88
52) Acenaphthene	14.274	153	34362	4.028	ng/ul	97
53) 2,4-Dinitrophenol	14.345	184	9	0.006	ng/ul	89
55) 4-Nitrophenol	14.469	109	6	0.003	ng/ul#	29
56) Dibenzofuran	14.621	168	54752	4.273	ng/ul	93
57) 2,4-Dinitrotoluene	14.616	165	12612	3.935	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	14.880	232	13378	4.202	ng/ul#	98
59) Diethylphthalate	15.057	149	44354	4.239	ng/ul	96
61) Fluorene	15.292	166	43992	4.213	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.274	204	26665	4.466	ng/ul	96
63) 4-Nitroaniline	15.327	138	4245	2.178	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.398	198	7773	3.069	ng/ul	97
67) N-Nitrosodiphenylamine	15.510	169	37383	3.905	ng/ul	97
68) 4-Bromophenyl-phenylether	16.198	248	19339	4.486	ng/ul	95
69) Hexachlorobenzene	16.316	284	24939	4.700	ng/ul	97
70) Atrazine	16.486	200	16880	4.242	ng/ul	94
71) Pentachlorophenol	16.686	266	12227	3.583	ng/ul	95
72) Phenanthrene	17.063	178	79290	4.146	ng/ul	96
74) Anthracene	17.157	178	78079	4.091	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.004	216	21985	4.039	ng/uL	98
76) Pentachlorobenzene	14.551	250	25770	4.283	ng/uL	96
77) Carbazole	17.451	167	65973	3.906	ng/ul	98
78) Di-n-butylphthalate	18.033	149	74230	3.972	ng/ul	98
80) Fluoranthene	19.168	202	105718	3.908	ng/ul	97
82) Pyrene	19.551	202	115724	3.991	ng/ul#	94
83) Butylbenzylphthalate	20.492	149	35429	3.509	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.380	252	43226	4.075	ng/ul	98
85) Benzo(a)anthracene	21.445	228	124664	3.996	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.380	149	47758	3.296	ng/ul	98
87) Chrysene	21.503	228	121443	4.198	ng/ul	99
89) Di-n-octyl phthalate	22.609	149	86256	3.091	ng/ul	100
90) Benzo(b)fluoranthene	23.680	252	144843	3.923	ng/ul	98
91) Benzo(k)fluoranthene	23.680	252	144843	4.006	ng/ul	97
93) Benzo(a)pyrene	24.539	252	68401	2.013	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	28.362	276	201211	4.437	ng/ul	99
95) Dibenzo(a,h)anthracene	28.438	278	159288	4.337	ng/ul	98
96) Benzo(g,h,i)perylene	29.491	276	54010	1.531	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
Data File : BP022940.D
Acq On : 08 Nov 2024 18:53
Operator : RC/JU
Sample : SSTD00549
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD005628

Quant Time: Nov 08 23:16:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 08 21:50:22 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.641 ng/uL
RT: 3.181 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion: 88 Resp: 1921
Ion Ratio Lower Upper
88 100
43 32.8 24.2 36.4
58 44.7 59.2 88.8#



#5
Pyridine
Concen: 3.540 ng/uL
RT: 3.587 min Scan# 102
Delta R.T. 0.012 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion: 79 Resp: 10845
Ion Ratio Lower Upper
79 100
52 82.9 59.5 89.3
51 43.1 30.4 45.6









#6
Benzaldehyde
Concen: 5.017 ng/ul
RT: 6.775 min Scan# 641
Delta R.T. 0.017 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion: 77	Resp: 9653
Ion Ratio	Lower Upper
77 100	
105 102.3	72.5 108.7
106 79.4	65.5 98.3



#8
Phenol
Concen: 3.535 ng/ul
RT: 6.834 min Scan# 654
Delta R.T. 0.011 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion: 94	Resp: 13096
Ion Ratio	Lower Upper
94 100	
65 36.1	26.2 39.4
66 61.2	40.2 60.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 3.773 ng/ul
 RT: 7.034 min Scan# 688
 Delta R.T. -0.000 min
 Lab File: BP022940.D
 Acq: 08 Nov 2024 18:53




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005628


Tgt Ion: 93	Resp: 10459
Ion Ratio	Lower Upper
93 100	
63 72.5	60.6 91.0
95 37.2	29.0 43.4



#12
 2-Chlorophenol
 Concen: 4.102 ng/ul
 RT: 7.181 min Scan# 713
 Delta R.T. 0.006 min
 Lab File: BP022940.D
 Acq: 08 Nov 2024 18:53

Tgt Ion: 128	Resp: 10716
Ion Ratio	Lower Upper
128 100	
64 53.0	39.0 58.6
130 29.4	26.4 39.6





#13

2-Methylphenol

Concen: 3.599 ng/ul

RT: 8.045 min Scan# 80

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :





SSTD005628

Tgt Ion:108 Resp: 9674

Ion Ratio Lower Upper

108 100

107 107.1 73.3 109.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 3.938 ng/ul

RT: 8.104 min Scan# 870

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion: 45 Resp: 12688


Ion Ratio Lower Upper

45 100

77 25.8 18.2 27.2

79 13.5 11.7 17.5









#16
Acetophenone
Concen: 3.944 ng/ul
RT: 8.410 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:	105	Resp:	18501
Ion	Ratio	Lower	Upper
105	100		
77	90.1	69.4	104.0
51	29.3	23.2	34.8



#17
N-Nitroso-di-n-propylamine
Concen: 3.885 ng/ul
RT: 8.387 min Scan# 918
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:	70	Resp:	9234
Ion	Ratio	Lower	Upper
70	100		
42	59.9	42.3	63.5
101	13.1	7.4	11.2#
130	22.8	15.4	23.0









#18
4-Methylphenol
Concen: 3.612 ng/ul
RT: 8.369 min Scan# 914
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:108 Resp: 10773
Ion Ratio Lower Upper
108 100
107 114.7 91.5 137.3



#19
Hexachloroethane
Concen: 4.476 ng/ul
RT: 8.645 min Scan# 962
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:117 Resp: 5446
Ion Ratio Lower Upper
117 100
201 150.2 105.0 157.6
199 92.9 65.9 98.9









#22
Nitrobenzene
Concen: 4.027 ng/ul
RT: 8.787 min Scan# 984
Delta R.T. 0.011 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion: 77	Resp: 14672
Ion Ratio	Lower Upper
77 100	
123 35.7	32.2 48.4
65 14.3	10.9 16.3



#23
Isophorone
Concen: 3.907 ng/ul
RT: 9.304 min Scan# 1074
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion: 82	Resp: 25519
Ion Ratio	Lower Upper
82 100	
95 10.2	7.6 11.4
138 16.5	13.4 20.2









#25
2-Nitrophenol
Concen: 4.029 ng/ul
RT: 9.481 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:139 Resp: 6135
Ion Ratio Lower Upper
139 100
65 52.2 52.8 79.2#
109 42.1 38.9 58.3



#26
2,4-Dimethylphenol
Concen: 4.301 ng/ul
RT: 9.557 min Scan# 1117
Delta R.T. 0.011 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:107 Resp: 14237
Ion Ratio Lower Upper
107 100
121 38.7 34.4 51.6
122 68.7 56.7 85.1





#27

Bis(2-Chloroethoxy)methane

Concen: 3.753 ng/ul

RT: 9.769 min Scan# 1153

Delta R.T. -0.000 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion: 93 Resp: 13920

Ion Ratio Lower Upper

93 100

95 31.7 27.0 40.6

123 12.7 8.2 12.4#



#29

2,4-Dichlorophenol

Concen: 4.177 ng/ul

RT: 10.034 min Scan# 1198

Delta R.T. 0.017 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:162 Resp: 12352


Ion Ratio Lower Upper

162 100

164 65.9 51.4 77.0

98 33.4 27.0 40.6









#30
Naphthalene
Concen: 4.345 ng/ul
RT: 10.392 min Scan# 1259
Delta R.T. -0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:128 Resp: 36731
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 14.2 10.2 15.2



#32
4-Chloroaniline
Concen: 3.699 ng/ul
RT: 10.522 min Scan# 1281
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:127 Resp: 12956
Ion Ratio Lower Upper
127 100
129 27.9 26.4 39.6









#33
Hexachlorobutadiene
Concen: 5.172 ng/ul
RT: 10.669 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:225 Resp: 12892
Ion Ratio Lower Upper
225 100
223 57.4 49.7 74.5
227 60.6 51.4 77.0



#34
Caprolactam
Concen: 2.579 ng/ul
RT: 11.316 min Scan# 1416
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:113 Resp: 2459
Ion Ratio Lower Upper
113 100
55 124.6 123.0 184.6
56 88.2 99.8 149.8#









#35
4-Chloro-3-methylphenol
Concen: 0.101 ng/ul
RT: 11.739 min Scan# 1474
Delta R.T. 0.082 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:	107	Resp:	326
Ion	Ratio	Lower	Upper
107	100		
144	22.8	18.8	28.2
142	62.5	59.8	89.8



#36
2-Methylnaphthalene
Concen: 4.447 ng/ul
RT: 12.016 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:	142	Resp:	27601
Ion	Ratio	Lower	Upper
142	100		
141	82.7	73.1	109.7









#37
1-Methylnaphthalene
Concen: 4.418 ng/ul
RT: 12.227 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:142 Resp: 27984
Ion Ratio Lower Upper
142 100
141 94.2 73.8 110.8
116 4.6 4.2 6.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.299 ng/ul
RT: 12.386 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:216 Resp: 21898
Ion Ratio Lower Upper
216 100
214 84.5 59.4 89.2
179 21.4 17.3 25.9
108 18.1 11.8 17.6#





#40

Hexachlorocyclopentadiene

Concen: 0.003 ng/ul

RT: 12.422 min Scan# 1

Delta R.T. 0.064 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:237 Resp: 9

Ion Ratio Lower Upper

237 100

235 68.2 48.6 73.0

272 13.6 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 3.908 ng/ul

RT: 12.639 min Scan# 1641

Delta R.T. -0.000 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:196 Resp: 12521


Ion Ratio Lower Upper

196 100

198 85.8 80.5 120.7

200 30.6 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 3.760 ng/ul

RT: 12.739 min Scan# 1658

Delta R.T. 0.011 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:196 Resp: 12825

Ion Ratio Lower Upper

196 100

198 85.9 79.6 119.4

200 26.1 25.0 37.6



#43

1,1'-Biphenyl

Concen: 4.096 ng/ul

RT: 13.033 min Scan# 1708

Delta R.T. -0.000 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:154 Resp: 40663


Ion Ratio Lower Upper

154 100

153 39.2 32.0 48.0

76 10.7 8.4 12.6









#44
2-Chloronaphthalene
Concen: 4.065 ng/ul
RT: 13.080 min Scan# 1716
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:	162	Resp:	33037
Ion	Ratio	Lower	Upper
162	100		
164	33.0	26.6	40.0
127	36.0	30.2	45.2



#45
2-Nitroaniline
Concen: 3.303 ng/ul
RT: 13.304 min Scan# 1754
Delta R.T. 0.017 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:	65	Resp:	8017
Ion	Ratio	Lower	Upper
65	100		
92	68.4	54.2	81.2
138	82.0	65.1	97.7









#47
Dimethylphthalate
Concen: 4.082 ng/ul
RT: 13.663 min Scan# 1815
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:163 Resp: 44517
Ion Ratio Lower Upper
163 100
194 3.5 3.8 5.6#
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 3.895 ng/ul
RT: 13.816 min Scan# 1841
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:165 Resp: 8034
Ion Ratio Lower Upper
165 100
89 62.5 50.6 75.8
121 23.3 15.8 23.8





#50

Acenaphthylene

Concen: 3.802 ng/ul

RT: 13.927 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:152 Resp: 45872

Ion Ratio Lower Upper

152 100

151 17.8 15.9 23.9

153 12.9 10.2 15.2



#51

3-Nitroaniline

Concen: 3.522 ng/ul

RT: 14.163 min Scan# 1900

Delta R.T. 0.011 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:138 Resp: 6824


Ion Ratio Lower Upper

138 100

108 14.4 12.7 19.1

92 119.6 107.8 161.8









#52
Acenaphthene
Concen: 4.028 ng/ul
RT: 14.274 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:153 Resp: 34362
Ion Ratio Lower Upper
153 100
152 49.1 39.6 59.4
154 93.2 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 0.006 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. -0.018 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:184 Resp: 9
Ion Ratio Lower Upper
184 100
63 52.4 50.1 75.1
154 81.0 58.2 87.4









#55
4-Nitrophenol
Concen: 0.003 ng/ul
RT: 14.469 min Scan# 1954
Delta R.T. -0.012 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:109 Resp: 6
Ion Ratio Lower Upper
109 100
139 26.7 64.6 96.8#
65 180.0 80.6 120.8#



#56
Dibenzofuran
Concen: 4.273 ng/ul
RT: 14.621 min Scan# 1978
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:168 Resp: 54752
Ion Ratio Lower Upper
168 100
139 41.3 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 3.935 ng/ul
RT: 14.616 min Scan# 1977
Delta R.T. 0.011 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:165 Resp: 12612
Ion Ratio Lower Upper
165 100
63 66.2 41.3 61.9#
182 2.6 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.202 ng/ul
RT: 14.880 min Scan# 2022
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:232 Resp: 13378
Ion Ratio Lower Upper
232 100
131 34.5 27.0 40.6
130 3.4 1.9 2.9#
166 26.3 21.8 32.6









#59
Diethylphthalate
Concen: 4.239 ng/ul
RT: 15.057 min Scan# 2052
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:149 Resp: 44354
Ion Ratio Lower Upper
149 100
177 27.6 19.7 29.5
150 12.5 9.8 14.6



#61
Fluorene
Concen: 4.213 ng/ul
RT: 15.292 min Scan# 2092
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:166 Resp: 43992
Ion Ratio Lower Upper
166 100
165 97.4 79.5 119.3
167 12.0 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 4.466 ng/ul
RT: 15.274 min Scan# 2089
Delta R.T. -0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:204 Resp: 26665
Ion Ratio Lower Upper
204 100
206 32.2 27.2 40.8
141 58.8 44.6 66.8



#63
4-Nitroaniline
Concen: 2.178 ng/ul
RT: 15.327 min Scan# 2098
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:138 Resp: 4245
Ion Ratio Lower Upper
138 100
92 62.7 53.9 80.9
108 132.9 106.6 159.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 3.069 ng/ul
 RT: 15.398 min Scan# 2110
 Delta R.T. 0.011 min
 Lab File: BP022940.D
 Acq: 08 Nov 2024 18:53




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005628


Tgt Ion:198 Resp: 7773
 Ion Ratio Lower Upper
 198 100
 182 4.9 3.8 5.8
 77 26.8 20.0 30.0



#67
 N-Nitrosodiphenylamine
 Concen: 3.905 ng/ul
 RT: 15.510 min Scan# 2129
 Delta R.T. 0.006 min
 Lab File: BP022940.D
 Acq: 08 Nov 2024 18:53

Tgt Ion:169 Resp: 37383
 Ion Ratio Lower Upper
 169 100
 168 69.6 53.4 80.0
 167 37.7 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 4.486 ng/ul
RT: 16.198 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:248 Resp: 19339
Ion Ratio Lower Upper
248 100
250 91.2 77.1 115.7
141 67.0 50.7 76.1



#69
Hexachlorobenzene
Concen: 4.700 ng/ul
RT: 16.316 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:284 Resp: 24939
Ion Ratio Lower Upper
284 100
142 27.3 21.5 32.3
249 28.9 25.6 38.4









#70
Atrazine
Concen: 4.242 ng/ul
RT: 16.486 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:200 Resp: 16880
Ion Ratio Lower Upper
200 100
173 26.7 18.6 27.8
215 51.9 38.3 57.5



#71
Pentachlorophenol
Concen: 3.583 ng/ul
RT: 16.686 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:266 Resp: 12227
Ion Ratio Lower Upper
266 100
264 67.6 50.7 76.1
268 61.7 51.9 77.9





#72

Phenanthrene

Concen: 4.146 ng/ul

RT: 17.063 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:178 Resp: 79290

Ion Ratio Lower Upper

178 100

179 14.4 12.7 19.1

176 17.6 15.4 23.2



#74

Anthracene

Concen: 4.091 ng/ul

RT: 17.157 min Scan# 2409

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:178 Resp: 78079


Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 16.2 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.039 ng/uL

RT: 13.004 min Scan# 1703

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:216 Resp: 21985

Ion Ratio Lower Upper

216 100

214 75.0 60.2 90.2

218 45.1 38.1 57.1



#76

Pentachlorobenzene

Concen: 4.283 ng/uL

RT: 14.551 min Scan# 1966

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:250 Resp: 25770


Ion Ratio Lower Upper

250 100

248 59.9 52.2 78.2

252 61.8 50.1 75.1









#77
Carbazole
Concen: 3.906 ng/ul
RT: 17.451 min Scan# 2458
Delta R.T. 0.006 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:167 Resp: 65973
Ion Ratio Lower Upper
167 100
166 23.2 17.5 26.3
139 12.5 9.6 14.4



#78
Di-n-butylphthalate
Concen: 3.972 ng/ul
RT: 18.033 min Scan# 2558
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:149 Resp: 74230
Ion Ratio Lower Upper
149 100
150 9.7 7.2 10.8
104 6.9 5.0 7.4





#80

Fluoranthene

Concen: 3.908 ng/ul

RT: 19.168 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:202 Resp: 105718

Ion Ratio Lower Upper

202 100

101 5.4 5.2 7.8

100 4.6 4.1 6.1



#82

Pyrene

Concen: 3.991 ng/ul

RT: 19.551 min Scan# 2816

Delta R.T. 0.011 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:202 Resp: 115724


Ion Ratio Lower Upper

202 100

101 5.1 6.2 9.2#

100 5.0 4.9 7.3





#83

Butylbenzylphthalate

Concen: 3.509 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.000 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:149 Resp: 35429

Ion Ratio Lower Upper

149 100

91 70.2 55.6 83.4

206 28.8 20.2 30.2



#84

3,3'-Dichlorobenzidine

Concen: 4.075 ng/ul

RT: 21.380 min Scan# 3127

Delta R.T. 0.017 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:252 Resp: 43226


Ion Ratio Lower Upper

252 100

254 63.7 52.0 78.0

126 5.7 5.0 7.6





#85

Benzo(a)anthracene

Concen: 3.996 ng/ul

RT: 21.445 min Scan# 3137

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:228 Resp: 124664

Ion Ratio Lower Upper

228 100

229 20.6 15.5 23.3

226 27.2 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.296 ng/ul

RT: 21.380 min Scan# 3127

Delta R.T. 0.011 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:149 Resp: 47758


Ion Ratio Lower Upper

149 100

167 29.6 24.5 36.7

279 7.5 5.5 8.3





#87

Chrysene

Concen: 4.198 ng/ul

RT: 21.503 min Scan# 3147

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:228 Resp: 121443

Ion Ratio Lower Upper

228 100

226 31.0 24.2 36.2

229 20.3 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 3.091 ng/ul




RT: 22.609 min Scan# 3336


Delta R.T. 0.017 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Tgt Ion:149 Resp: 86256





#90

Benzo(b)fluoranthene

Concen: 3.923 ng/ul

RT: 23.680 min Scan# 3518

Delta R.T. 0.006 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53

Instrument :

BNA_P

ClientSampleId :

SSTD005628





Tgt Ion:252 Resp: 144843

Ion Ratio Lower Upper

252 100

253 20.9 17.4 26.2

125 4.0 3.7 5.5



#91

Benzo(k)fluoranthene

Concen: 4.006 ng/ul

RT: 23.680 min Scan# 3518

Delta R.T. -0.059 min

Lab File: BP022940.D

Acq: 08 Nov 2024 18:53




Tgt Ion:252 Resp: 144843


Ion Ratio Lower Upper

252 100

253 20.9 17.8 26.8

125 4.0 3.0 4.4









#93
Benzo(a)pyrene
Concen: 2.013 ng/ul
RT: 24.539 min Scan# 30
Delta R.T. -0.000 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53




Instrument :
BNA_P
ClientSampleId :
SSTD005628


Tgt Ion:252 Resp: 68401
Ion Ratio Lower Upper
252 100
253 23.8 17.6 26.4
125 5.3 4.2 6.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.437 ng/ul
RT: 28.362 min Scan# 4314
Delta R.T. 0.017 min
Lab File: BP022940.D
Acq: 08 Nov 2024 18:53

Tgt Ion:276 Resp: 201211
Ion Ratio Lower Upper
276 100
138 7.5 7.0 10.4
277 24.5 19.8 29.8









#95
 Dibenzo(a,h)anthracene
 Concen: 4.337 ng/ul
 RT: 28.438 min Scan# 41
 Delta R.T. 0.035 min
 Lab File: BP022940.D
 Acq: 08 Nov 2024 18:53

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005628

Tgt Ion:278 Resp: 159288
 Ion Ratio Lower Upper
 278 100
 139 7.0 5.4 8.0
 279 23.1 19.4 29.2



#96
 Benzo(g,h,i)perylene
 Concen: 1.531 ng/ul
 RT: 29.491 min Scan# 4506
 Delta R.T. 0.011 min
 Lab File: BP022940.D
 Acq: 08 Nov 2024 18:53

Tgt Ion:276 Resp: 54010
 Ion Ratio Lower Upper
 276 100
 138 8.8 7.1 10.7
 277 23.3 19.4 29.2

